

**REPORT ON THE U.S. GEOLOGICAL SURVEY'S EVALUATION PROGRAM  
FOR STANDARD REFERENCE SAMPLES DISTRIBUTED IN SEPTEMBER 1996:  
T-143 (TRACE CONSTITUENTS), T-145 (TRACE CONSTITUENTS),  
M-140 (MAJOR CONSTITUENTS), N-51 (NUTRIENT CONSTITUENTS),  
N-52 (NUTRIENT CONSTITUENTS), P-27 (LOW IONIC STRENGTH  
CONSTITUENTS), AND Hg-23 (MERCURY)**

**by Jerry W. Farrar and H. Keith Long**

---

**U.S. GEOLOGICAL SURVEY**

**Open-File Report 97-20**



**Lakewood, Colorado  
1997**

**DEPARTMENT OF THE INTERIOR**

**BRUCE BABBITT, Secretary**

**U.S. GEOLOGICAL SURVEY**

**Gordon P. Eaton, Director**

---

For additional information  
write to:  
William J. Shampine  
U.S. Geological Survey  
Water Resources Division  
Box 25046, Mail Stop 414  
Denver Federal Center  
Denver, Colorado 80225-0046

Copies of this report can be  
purchased from:  
U.S. Geological Survey  
Branch of Information Services  
Box 25286  
Denver, Colorado 80225-0286

## CONTENTS

	Page
Abstract .....	1
Introduction .....	1
Purpose and scope .....	2
Preparation of standard reference samples .....	5
Laboratory Analyses .....	7
Laboratory performance ratings .....	9
Statistical presentation of data .....	9
Reference .....	10

## FIGURE

Figure 1. Statistical parameters shown on reported-data graphs .....	10
--	----

## TABLES

Table 1. Laboratory participants in the analyses of standard reference samples distributed in September 1996 .....	3
2. Analytes determined in standard reference samples distributed in September 1996.....	7
3. Analytical methods codes .....	8
4. Overall laboratory performance ratings for standard reference water samples distributed in September 1996.....	11
5. Laboratory performance ratings for standard reference water sample T-143 (trace constituents) .....	13
6. Laboratory performance ratings for standard reference water sample T-145 (trace constituents) .....	21
7. Laboratory performance ratings for standard reference water sample M-140 (major constituents) .....	29
8. Laboratory performance ratings for standard reference water sample N-51 (nutrient constituents) .....	35
9. Laboratory performance ratings for standard reference water sample N-52 (nutrient constituents) .....	37
10. Laboratory performance ratings for standard reference water sample P-27 (low ionic strength constituents) .....	39
11. Laboratory performance ratings for standard reference water sample Hg-23 (mercury) .....	43
12. Statistical summary of reported data for standard reference water sample T-143 (trace constituents) .....	44
13. Statistical summary of reported data for standard reference water sample T-145 (trace constituents) .....	73
14. Statistical summary of reported data for standard reference water sample M-140 (major constituents) .....	102
15. Statistical summary of reported data for standard reference water sample N-51 (nutrient constituents) .....	119
16. Statistical summary of reported data for standard reference water sample N-52 (nutrient constituents) .....	125
17. Statistical summary of reported data for standard reference water sample P-27 (low ionic strength constituents) .....	131
18. Statistical summary of reported data for standard reference water sample Hg-23 (mercury) .....	143
19. Most probable values for constituents and properties in standard reference samples distributed in September 1996 ...	145

REPORT ON THE U.S. GEOLOGICAL SURVEY'S EVALUATION PROGRAM FOR  
STANDARD REFERENCE SAMPLES DISTRIBUTED IN SEPTEMBER 1996:  
T-143 (TRACE CONSTITUENTS), T-145 (TRACE CONSTITUENTS),  
M-140 (MAJOR CONSTITUENTS), N-51 (NUTRIENT CONSTITUENTS),  
N-52 (NUTRIENT CONSTITUENTS), P-27 (LOW IONIC STRENGTH  
CONSTITUENTS), AND Hg-23 (MERCURY)

By Jerry W. Farrar and H. Keith Long

## ABSTRACT

This report presents the results of the U.S. Geological Survey's analytical evaluation program for 7 standard reference samples -- T-143 (trace constituents), T-145 (trace constituents), M-140 (major constituents), N-51 (nutrient constituents), N-52 (nutrient constituents), P-27 (low ionic strength constituents), and Hg-23 (mercury) -- that were distributed in September 1996 to 167 laboratories registered in the U.S. Geological Survey sponsored interlaboratory testing program. Analytical data that were received from 140 of the laboratories were evaluated with respect to: overall laboratory performance and relative laboratory performance for each analyte in the seven reference samples. Results of these evaluations are presented in tabular form. Also presented are tables and graphs summarizing the analytical data provided by each laboratory for each analyte in the seven standard reference samples. The most probable value for each analyte was determined using nonparametric statistics.

## INTRODUCTION

The U.S. Geological Survey (USGS) conducts an interlaboratory evaluation program semiannually. This program provides a variety of reference materials to accomplish quality assurance testing of laboratories and to provide an adequate supply of samples that contribute to quality control programs of participating laboratories. Natural-matrix reference materials are preferred for use in this interlaboratory evaluation program. A series of samples are prepared and distributed each spring and fall. Occasionally, sediment samples are provided.

The program began in 1962 with a single sample containing major constituents that was prepared from distilled water and reagent grade chemicals. Twenty-three USGS laboratories participated in the 1962 determinations of six analytes in the major standard reference sample (SRS). Since that time, objectives of the program have been to:

- (1) evaluate and improve the performance of USGS and other participating laboratories;
- (2) provide a library of carefully prepared, homogeneous, stable reference materials for use in the quality control programs of laboratories;
- (3) identify analytical problem areas;
- (4) identify quality assurance needs with respect to environmental analyses and develop new reference materials to meet these needs; and
- (5) ascertain the accuracy and precision of analytical methods.

Two hundred fifteen USGS and non-USGS laboratories are registered in the program, which can currently provide eight standard reference sample types:

1. Trace constituents.
2. Major constituents.
3. Nutrient constituents.
4. Low ionic strength constituents.
5. Mercury.
6. Whole water (water with suspended sediment).
7. Acid mine drainage constituents.
8. Sediment (bed material) for major and trace constituents.

When sufficient data are available, a most probable value is statistically determined for each analyte in the SRS.

Though this is not a laboratory certification program, participation in this continuing quality assurance program is mandatory for all laboratories providing water-analyses data for USGS data storage or use (publications). Federal, State, municipal, and university laboratories can participate even though they do not provide data to the USGS. Analyses of these SRS provides the means to alert participating laboratories of possible deficiencies in their analytical operations, and also provides reference materials for in-house quality control programs. Participating laboratories are identified only by a confidential code number.

A library of SRS, from previous evaluations, are available. Participating laboratories can purchase previous SRS for further testing, continuing quality assurance, and quality control programs by contacting:

U.S. Geological Survey  
Branch of Technical Development and Quality Systems  
Denver Federal Center  
Box 25046 MS 401  
Denver, Colorado 80225-0046  
(303) 236-1870

### Purpose and Scope

This report summarizes the analytical results submitted by 140 of the 167 laboratories that requested and were shipped SRS for the January 1997 evaluation (table 1). Not all SRS are requested, nor necessarily analyzed by all the laboratories; nor do all laboratories enrolled in the program participate in each evaluation. Analytical results for the following, which were mailed the week of September 9, 1996, are presented in this report:

T-143	Trace constituents
T-145	Trace constituents
M-140	Major constituents
N-51	Nutrient constituents
N-52	Nutrient constituents
P-27	Low ionic strength constituents (precipitation)
Hg-23	Mercury

The USGS requested that analytical results be returned by November 15, 1996 for evaluation and preparation of this report. Each participating laboratory is requested to perform those determinations routinely made on the respective SRS for USGS investigations and to indicate the information was provided, it has been included in the respective data table. The analytical data are represented in ways

that allow participants to evaluate data distribution, scatters, outliers, central tendency, bias, skewness, and method relationships.

**Table 1.-Laboratory participants in the analyses of standard reference samples distributed in September 1996**

<u>State</u>	<u>City</u>	<u>Participating Laboratory</u>
Alabama	Tuscaloosa	Geological Survey of Alabama
Alaska	Soldotna	Alaska Department of Fish and Game
Arizona	Yuma	Burns and Roe Services Corporation
Arkansas	Arkadelphia	Ouachita Baptist University
	Fayetteville	University of Arkansas
	Little Rock	Arkansas Department of Pollution Control and Ecology
California	Davis	University of California - Davis
	La Verne	Metropolitan Water District
	Martinez	Central Contra Costa Sanitary District
	Oakland	East Bay Municipal Utility District
	Perris	Eastern Municipal Water District
	Sacramento	Anlab
	Sacramento	US Bureau of Reclamation
	Sacramento	USGS WRD
	Santa Fe Springs	West Coast Analytical Service, Inc.
	Tahoe City	Tahoe Research Group
	West Sacramento	California Department of Water Resources
	West Sacramento	Quanterra
Colorado	Alamosa	Bureau of Reclamation
	Arvada	Quanterra
	Arvada	USGS-NWQL
	Aurora	Core Laboratories, Inc.
	Colorado Springs	City of Colorado Springs
	Denver	US Bureau of Reclamation
	Denver	USGS Colorado District Toxic Project
	Denver	Denver Water Department
	Fort Collins	City of Fort Collins - Water Quality
	Fort Collins	CSU - Soil Testing Laboratory
	Fort Collins	USDA Forest Service
	Golden	Kaiser - Hill Rocky Flats
	Loveland	Northern Colorado Water Conservation
	Northglenn	Northglenn Water Treatment Plant
	Westminster	City of Westminster
Florida	Bradenton	Manatee County Environmental Management
	Brooksville	SW Florida Water Management District
	Ocala	USGS WRD QWSU
	Orlando	Post, Buckley, Schuh, and Jernigan, Inc.
	Ormond Beach	Environmental Laboratory
	Palatka	St. John's River Management District
	Tallahassee	City of Tallahassee
	Tallahassee	Florida Department of Environmental Regulations
	Tallahassee	Savannah Laboratories
	Tampa	Hillsborough County Environmental Protection Commission
	West Palm Beach	South Florida Water Management District
Georgia	Athens	University of Georgia
	Atlanta	Georgia Department of Natural Resources
	Atlanta	USGS WRD
	Decatur	Dekalb County Water Quality Laboratory
	Tifton	USDA - ARS
Hawaii	Honolulu	University of Hawaii - SOEST Analytical Services

**Table 1.-Laboratory participants in the analyses of standard reference samples distributed in September 1996--Continued**

State	City	Participating Laboratory
Idaho	Boise	US Bureau of Reclamation
	Pocatello	Idaho State University
Illinois	Champaign	Hazardous Waste Research Center
	Champaign	Illinois Environmental Protection Agency
Indiana	Indianapolis	Indianapolis Department of Public Works
Iowa	Des Moines	University Hygienic Laboratory, Des Moines Branch
Kansas	Lawrence	Kansas Geological Survey
	Topeka	City of Topeka
	Topeka	Kansas Department of Health and Environment
	Wichita	City of Wichita
Kentucky	Frankfort	Division of Environmental Studies
	Lexington	Kentucky Geological Survey
	Louisville	Metropolitan Sewer District
Maryland	Baltimore	Maryland Department of Health and Mental Hygiene
Michigan	Ann Arbor	University of Michigan
	Detroit	Detroit Water and Sewerage Department
Minnesota	Minneapolis	University of Minnesota, Department of Geology and Geophysics
	St. Paul	Metro Waste Control Commission
	St. Paul	University of Minnesota
	Columbia	University of Missouri
Missouri	Jefferson City	Missouri Department of Health
	Butte	Montana Bureau of Mines & Geology
Montana	Missoula	University of Montana
	Boulder City	US Bureau of Reclamation
Nevada	Las Vegas	University of Nevada - Las Vegas
	Reno	Desert Research Institute
	Reno	Nevada State Health Laboratory
	Reno	Reno-Sparks Wastewater Treatment
	Sutcliffe	Pyramid Lake Fisheries
	Brewster	NYC DEP Brewster Lab
	Brockport	SUNY - Brockport
New York	Grahamsville	New York City Department of Environmental Protection
	Hauppauge	Suffolk County Water Authority
	Hempstead	Nassau County Department of Health
	Ithaca	Cornell Nutrient Analysis Lab
	Milbrook	Institute of Ecosystem Studies
	North Babylon	Ecotest Laboratories
	Rochester	Monroe County
	Shokan	New York City Department of Environmental Protection
	Syracuse	Onandaga County DDS
	Troy	USGS-WRD
North Carolina	Valhalla	Department of Environmental Protection
	Wantagh	Cedar Creeks Projects laboratory
	Yorktown	New York City Department of Environmental Protection
	Chapel Hill	City of Durham Water Resources
	Charlotte	Mecklenburg County
North Carolina	Greensboro	City of Greensboro
North Dakota	Bismarck	North Dakota State Health Department
	Bismarck	North Dakota State Water Commission
Ohio	Bismarck	US BOR
	Cincinnati	US EPA
	Cuyahoga Heights	Northeast Ohio Regional Sewer District
	Tiffin	Heidelberg College
	Wooster	The Ohio State University

**Table 1.-Laboratory participants in the analyses of standard reference samples distributed in September 1996--Continued**

<u>State</u>	<u>City</u>	<u>Participating Laboratory</u>
Oklahoma	Norman	Oklahoma Geological Survey
	Oklahoma City	Oklahoma Department of Environmental Quality
Oregon	Corvallis	USDA - CCAL
	Tigard	Unified Sewerage Agency
Pennsylvania	Harrisburg	Pennsylvania Department of Environmental Resources
	Somerset	Geochemical Testing
Puerto Rico	San Juan	Department of Natural Resources
South Dakota	Brookings	SDSU - Water Quality Laboratory
Tennessee	Chattanooga	TVA Environmental Chemistry
Texas	Austin	Lower Colorado River Authority
	College Station	Texas A & M
	Seguin	Guadalupe-Blanco River Authority
Vermont	Waterbury	Vermont Agency of Natural Resources
Virginia	Culpepper	ESS Labs
	Manassas	Occoquan Watershed Monitoring Laboratory
	Richmond	Consolidated Laboratory Services
Washington	Seattle	Frontier Geoscience
	Seattle	Brooks-Rand, Ltd.
Wisconsin	Madison	University of Wisconsin, Department of Hygiene
	Milwaukee	Milwaukee Metro Sewerage District

#### Middle East Laboratories

<u>Location</u>	<u>Participating Laboratory</u>
Gaza	Al-Azar University, Water Research Center Laboratory
	Islamic University, Environmental & Rural Research Center Laboratory
	Ministry of Agriculture Laboratory
	Ministry of Health, Public Health Laboratory
Israel	Geological Survey of Israel Laboratory
	Israeli Hydrological Service Laboratory
	Mekeroth Water Company, Central Laboratory
	Water Resources Research Center, Institute for Desert Research
Jordan	Royal Scientific Society of Jordan, Environmental Research Center Laboratory
	Water Authority of Jordan, Central Laboratory
West Bank	Al-Quds University, College of Science & Technology, Water Research Center
	Bethlehem University , Water and Soil Environmental Research Unit
	Birzeit University, Center for Environmental & Occupational Health Services
	Najah-WESC

#### Preparation of Standard Reference Samples

All of the SRS used in this evaluation were prepared by personnel of the USGS in Lakewood, Colorado and were analyzed for analyte concentrations and physical property values prior to mailing. A library of reference samples is maintained and can be requested by participating laboratories for use in their quality control programs.

Trace constituent sample T-143 was prepared using water collected from the North Platte River near Windover, Wyoming. The water was pumped through 0.45, 0.2- and 0.1- $\mu\text{m}$  filters, in series, into a 3500-L polypropylene drum. The water was continuously circulated and passed through a 0.1- $\mu\text{m}$  filter and ultraviolet sterilizer for 24 hours. Following this circulation, the water was acidified to pH 1.3 with nitric acid and chlorinated to 5 ppm free chlorine. The trace constituent concentrations were adjusted by adding reagent grade chemicals. The sample was circulated an additional 24 hours prior to bottling.

polypropylene bottles used were acid leached, deionized-water rinsed, and autoclave sterilized.

Trace constituent sample T-145 was prepared using water collected from the Yampa River near Steamboat Springs, Colorado. The water was pumped through 0.45, 0.2- and 0.1- $\mu\text{m}$  filters, in series, into a 3500-L polypropylene drum. The water was continuously circulated and passed through a 0.1- $\mu\text{m}$  filter and ultraviolet sterilizer for 24 hours. Following this circulation, the water was acidified to pH 1.3 with nitric acid and chlorinated to 5 ppm free chlorine. The trace constituent concentrations were adjusted by adding reagent grade chemicals. The sample was circulated an additional 24 hours prior to bottling. During bottling the sample was pumped through an ultraviolet sterilizer and a 0.1- $\mu\text{m}$  filter. The 500-mL polypropylene bottles used were acid leached, deionized-water rinsed, and autoclave sterilized.

Major constituent sample M-140 was prepared using water collected from the Arkansas River near Avondale, Colorado. The water was pumped through 0.45, 0.2- and 0.1- $\mu\text{m}$  filters, in series, into a 1300-L polypropylene drum. The water was chlorinated to 5-ppm free chlorine with sodium hypochlorite, continuously circulated, and passed through a 0.1- $\mu\text{m}$  filter and ultraviolet sterilizer for 24 hours prior to bottling. During bottling the sample was pumped through an ultraviolet sterilizer and a 0.1- $\mu\text{m}$  filter. The 500-mL polypropylene bottles used were acid leached, deionized-water rinsed, and autoclave sterilized.

Nutrient sample N-51 was prepared using water collected from the West Chicago Creek near Idaho Springs, Colorado. These samples were prepared the week prior to the mailing for this SRS evaluation. The water was pumped through 0.45, 0.2- and 0.1- $\mu\text{m}$  filters, in series, into a 600-L polypropylene drum and continuously circulated and passed through a 0.1- $\mu\text{m}$  filter for 24 hours. The desired nutrient concentrations were obtained by adding reagent-grade chemicals. The sample was continuously circulated for 24 hours prior to being bottled. The 250-mL polyethylene bottles used were new, amber, acid leached, and deionized-water rinsed.

Nutrient sample N-52 was prepared using deionized water. These samples were prepared the week prior to the mailing for this SRS evaluation. The water was pumped through 0.45, 0.2- and 0.1- $\mu\text{m}$  filters, in series, into a 600-L polypropylene drum and continuously circulated and passed through a 0.1- $\mu\text{m}$  filter for 24 hours. The desired nutrient concentrations were obtained by adding reagent-grade chemicals. The sample was continuously circulated for 24 hours prior to being bottled. The 30-mL glass vials used were new, amber, acid leached, and deionized-water rinsed.

Sample P-27 was prepared in a 400-L polypropylene drum using water collected from the West Chicago Creek near Idaho Springs, Colorado. The water was pumped into the drum through 0.45, 0.2- and 0.1- $\mu\text{m}$  filters in series. Desired phosphate and fluoride concentrations were obtained by adding reagent-grade chemicals. Prior to bottling, the sample was continuously mixed for 24 hours while being circulated through a 0.1- $\mu\text{m}$  filter and an ultraviolet sterilizer. During bottling the sample was pumped through an ultraviolet sterilizer and a 0.1- $\mu\text{m}$  filter. The 500-mL polypropylene bottles used were acid leached, deionized-water rinsed, and autoclave sterilized.

Sample Hg-23 was prepared using water collected from the Fall River near Idaho Springs, Colorado. The sample was prepared in a 190-L polypropylene drum. The river water was pumped into this drum through 0.45, 0.2- and 0.1- $\mu\text{m}$  filters in series. The water was continuously circulated and passed through a 0.1- $\mu\text{m}$  filter and ultraviolet sterilizer for 48 hours. Nitric acid (5-percent, v/v) and dichromate ion (0.05-percent, w/w) were added to stabilize the sample. The desired mercury concentration was obtained by adding a mercury standard solution. Following an additional 24 hours of circulation, the sample was bottled. The 250-mL glass bottles and tetrafluoroethylene fluorocarbon resin caps used were new, acid leached, and deionized-water rinsed.

## LABORATORY ANALYSES

The participating laboratories were asked to determine analytes which are summarized in table 2. The number of analytes varied from 28 in T-143 & T-145 (trace constituents) to 1 in Hg-23 (mercury).

**Table 2.-Analytes determined in standard reference samples distributed in September 1996**

<small>(mo/L, milligrams per liter, <math>\mu</math>g/L, micrograms per liter, <math>\mu</math>S/cm, microsiemens per centimeter at 25 degrees Celsius)</small>							
Analyte or property		Units	T-143, T-145	M-140	N-51, N 52	P-27	Hg-23
Acidity	Acidity as $\text{CaCO}_3$	mg/L				X	
Alk	Alkalinity as $\text{CaCO}_3$	mg/L		X			
Ag	Silver	$\mu$ g/L	X				
Al	Aluminum	$\mu$ g/L	X				
As	Arsenic	$\mu$ g/L	X				
B	Boron	$\mu$ g/L	X				
Ba	Barium	$\mu$ g/L	X				
Be	Beryllium	$\mu$ g/L	X				
Ca	Calcium	mg/L	X	X			X
Cd	Cadmium	$\mu$ g/L	X				
Cl	Chloride	mg/L		X			
Co	Cobalt	$\mu$ g/L	X				
Cr	Chromium, total	$\mu$ g/L	X				
Cu	Copper	$\mu$ g/L	X				
DSRD	Dissolved solids	mg/L		X			
F	Fluoride	mg/L		X			X
Fe	Iron	$\mu$ g/L	X				
Hg	Mercury	$\mu$ g/L					X
K	Potassium	mg/L	X	X			X
Li	Lithium	$\mu$ g/L	X				
Mg	Magnesium	mg/L	X	X			X
Mn	Manganese	$\mu$ g/L	X				
Mo	Molybdenum	$\mu$ g/L	X				
Na	Sodium	mg/L	X	X			X
$\text{NH}_3$ as N	Ammonia	mg/L			X		
$\text{NH}_3 + \text{Org N}$ as N	Ammonia + Organic N	mg/L			X		
Ni	Nickel	$\mu$ g/L	X				
$\text{NO}_3 + \text{NO}_2$ as N	Nitrate + Nitrite	mg/L			X		
Pb	Lead	$\mu$ g/L	X				
pH		unit		X		X	
$\text{PO}_4$ as P	Orthophosphate	mg/L			X		
total P as P	Phosphorus	mg/L		X	X	X	
Sb	Antimony	$\mu$ g/L	X				
Se	Selenium	$\mu$ g/L	X				
$\text{SiO}_2$	Silica	mg/L	X	X			
$\text{SO}_4$	Sulfate	mg/L		X			X
Sp Cond	Specific conductance	$\mu$ S/cm		X			X
Sr	Strontium	$\mu$ g/L	X	X			
Tl	Thallium	$\mu$ g/L	X				
U	Uranium	$\mu$ g/L	X				
V	Vanadium	$\mu$ g/L	X	X			
Zn	Zinc	$\mu$ g/L	X				

Laboratories were requested to identify the method used for each analyte according to table 3 analytical method codes.

**Table 3. Analytical methods codes**

Code	Method
0	Other
1	Atomic absorption: direct, air
2	Atomic absorption: direct, nitrous oxide
3	Atomic absorption: graphite furnace
4	Inductively coupled plasma
5	Direct current plasma
6	Inductively coupled plasma/Mass spectrometry
7	Ion chromatography
8	Atomic absorption: cold vapor
9	Atomic fluorescence
10	Atomic absorption: extraction [ <i>specify chelating agents</i> ]
11	Atomic absorption: hydride [ <i>specify reducing agent</i> ]
12	Flame emission
20	Titration: colorimetric [ <i>specify color reagent</i> ]
21	Titration: electrometric [ <i>specify reducing or oxidizing agent/color reagent</i> ]
22	Colorimetric: [ <i>specify reducing or oxidizing agent/color reagent</i> ]
40	Ion selective electrode
41	Electrometric [ <i>pH and Specific Conductance</i> ]
50	Gravimetric: [ <i>specify filtration, evaporation, and so forth</i> ]
51	Turbidimetric

Participating laboratories were also asked to use the references listed below to further define the methods.

1. American Public Health Association and others, 1992, Standard methods for the examination of water and wastewater 18th ed: Washington, D.C., American Public Health Association, 981p.
2. American Society for Testing and Materials, Annual book of ASTM standards: Philadelphia, v. 11.01, and v. 11.02.
3. Kopp, J.F., and McKee, G.F., 1979, Methods for chemical analysis of water and wastes: Cincinnati, U.S. Environmental Protection Agency, EPA 600/4-79-020, rev. 1983, 460 p.
4. Fishman, M.J., and Friedman, L.C., eds., 1989. Methods for determination of inorganic substances in water and fluvial sediments (3d ed.): U.S. Geological Survey Techniques of Water-Resources Investigations, Book 5, Chapter A1, 545 p.
5. Miscellaneous manufacturer's instrument manuals or references.

## LABORATORY PERFORMANCE RATINGS

To facilitate interlaboratory performance comparisons, laboratory performance ratings, based on the analyses reported for each SRS, are included in tables 4 through 11 in this report. Averages of the analyte ratings and the number of analyte values reported for each SRS are given for each participating laboratory. Laboratory performance for each analyte is rated on a scale 4 to 0, based on the absolute Z-value, as listed below:

<u>Rating</u>	<u>Absolute Z-value</u>
4 (Excellent)	0.00 to 0.50
3 (Good)	0.51 to 1.00
2 (Satisfactory)	1.01 to 1.50
1 (Questionable)	1.51 to 2.00
0 (Poor)	Greater than 2.00

Overall laboratory performance ratings greater than 2.4 are considered satisfactory. Overall laboratory performance ratings between 2.0 and 2.39 are considered marginal; those less than 2.0 are considered poor.

## STATISTICAL PRESENTATION OF DATA


Data in this report have been evaluated using nonparametric statistics as described by Hoaglin and others (1983). This statistical approach is a resistant statistic because the median is not influenced by outliers as is the mean in traditional statistics.

Analytical data for each analyte are presented in tabular and graphical forms in tables 12 through 18. Tabulated data for each analyte include the laboratory code number, reported values, analytical method, most probable value (MPV), number of reported values - excluding less than values (N), data range, Z-value, and the F-pseudosigma. (The Z-value is equivalent to the Z-score of traditional statistics, being the number of deviations the reported value is from the MPV. The F-pseudosigma is equivalent to the standard deviation ( $\sigma$ ) of traditional statistics when the data has a Gaussian distribution.) If an analyte has a sufficient number of determinations by a given method, usually 7, the F-pseudosigma for that analytical method is reported in the block of data listed for each analyte.

The median value is considered the MPV. Reported values of "less than" are used to establish the median, but are not considered in determining the data range. The median (midpoint) divides the ordered data into halves and is designated the MPV. The hinges include the middle 50-percent of the data and are the mid-values of the upper and lower halves of the data. (The hinges are similar to quartiles, but are not mathematically equivalent.) The range of data between the upper hinge (Hu) and the lower hinge (Hl), the hinge spread (H-spr), is used to calculate the F-pseudosigma, the laboratory performance rating, the upper warning level (UWL) and lower warning level (LWL), the upper control level (UCL) and the lower control level (LCL). The F-pseudosigma is calculated by comparison of the H-spr value to the Gaussian distribution relation; 67.45 percent of the data "hinges" between plus and minus  $1\sigma$ , resulting in a H-spr of  $2 \times 0.6745 = 1.349\sigma$ . This relation allows the calculation of the F-pseudosigma = (H-spr)/1.349. Laboratories reporting "less than" values are not performance rated unless their reported "less than" values are greater than two Z-values from the MPV.

The graphical plot of the reported data is shown in figure 1. The upper and lower boundaries of the graphical plots generally are +3 and -3 F-pseudosigma deviations from the median. (Computer-program

scaling constraints do not permit these boundaries to always be graphed at exactly these values.) The graphical plot is a box plot/control chart with reported values grouped by analytical method in ascending order of value. Lines designate the MPV, Hu, HI, and the (UWL) and (LWL) at +2 and -2 F-pseudosigma, respectively. "Less than" values are not plotted. The analytical data are presented in ways that allow participants to evaluate data distribution, scatter, outliers, central tendency, bias, skewness, and method relationships.



NOTE: vertical scale is the concentration value of the individual analyte in appropriate units (see table 2.) Methods shown are defined in Tables 3 and 12 through 18.

Figure 1.-Statistical parameters shown on reported-data graphs

#### REFERENCE

Hoaglin, D.C., Mosteller, F., and Tukey, J.W., eds., 1983, Understanding Robust and Exploratory Data Analysis: John Wiley and Sons, Inc., 447p.

**Table 4. Overall laboratory performance ratings for standard reference water samples distributed in September 1996**

[Lab, laboratory number; OWR, overall weighted rating for all sample types; OLR, overall laboratory rating for reported values of a sample type; V/94, number of reported values of 94 total possible values from all sample types; V/28, V/28, V/16, V/5, V/5, V/11, V/1 are number of reported values possible for T-143, T-145, M-140, N-51, N-52, P-27 and Hg-23 respectively]

Standard reference sample =		T-143		T-145		M-140		N-51		N-52		P-27		Hg-23		
Lab	OWR	V/94	OLR	V/28	OLR	V/28	OLR	V/16	OLR	V/5	OLR	V/5	OLR	V/11	OLR	V/1
1	3.5	90	3.6	28	3.4	28	3.5	16	3.0	3	3.4	5	3.8	9	3	1
2	1.7	18					1.5	10					2.0	8		
3	1.8	86	1.7	26	1.5	25	1.9	16	2.5	4	3.0	5	2.0	9	0	1
4	3.1	29	2.8	10	3.2	17	4.0	2								
5	3.0	49			3.4	23	3.4	13	0.5	2	2.0	3	2.3	8		
7	2.6	59	2.6	23	2.5	19	2.2	6	2.0	1	1.3	4	3.8	5	4	1
10	3.3	37	2.8	6	2.8	8	3.5	13	3.8	4	3.6	5			4	1
11	2.8	72	3.0	25	2.9	26	1.7	12	2.7	3	3.2	5			3	1
13	2.8	61	2.6	23	2.8	22	2.7	12			3.8	4				
15	1.9	70	2.2	21	1.6	20	1.8	12	0.0	2	2.6	5	1.9	9	3	1
16	2.8	80	3.2	27	2.7	27	2.6	15	2.2	5	1.6	5			4	1
18	3.3	66	3.5	22	3.6	20	2.9	14	2.8	4	2.6	5			3	1
19	3.2	33	3.2	13	3.3	6	3.0	10			3.5	4				
21	3.0	6	0.0	1					3.6	5						
22	3.3	3					4.0	1	2.0	1	4.0	1				
23	2.7	53	2.9	15	2.4	14	2.7	14	0.0	1	1.3	3	3.7	6		
24	2.9	47	2.6	20	2.8	14	3.7	13								
25	2.5	55	2.1	15	2.2	11	2.5	12	3.0	4	2.4	5	3.1	8		
26	3.4	73	3.3	24	3.5	24	3.4	13			2.0	2	3.4	9	4	1
28	1.5	46	1.5	16	1.5	16	2.3	7	0.0	2			1.5	4	0	1
30	3.1	49	3.2	21	2.9	21	3.2	5			4.0	2				
32	3.4	67	3.6	26	3.6	26	2.9	14							3	1
33	3.0	42	2.7	10	2.4	10	3.4	11	3.0	1	2.5	2	3.6	8		
34	3.6	7	3.3	3	3.7	3								4	1	
35	3.5	2	4.0	1	3.0	1										
36	1.4	59	1.3	18	0.8	20	1.8	11			2.0	5	2.5	4	3	1
38	3.4	27					3.3	10	3.0	5	3.6	5	3.6	7		
39	2.4	27	2.5	11			2.6	7	2.0	3			2.2	5	3	1
40	2.4	48	2.2	19	2.1	16	2.9	13								
42	2.3	72	3.0	25	2.0	26	1.8	12			0.3	3	2.4	5	4	1
43	3.6	25	3.4	7	3.4	7	3.7	11								
46	3.0	70	3.2	22	3.1	21	2.6	12	3.0	2	2.0	5	3.3	7	4	1
48	2.4	74	2.2	22	2.8	21	2.2	12	3.5	4	2.0	5	1.9	9	4	1
50	2.6	25	2.1	11			3.2	13						0	1	
51	2.7	15	2.7	3	2.7	3	2.8	8						3	1	
53	1.5	6							1.0	3	2.0	3				
55	2.8	32	2.7	10			2.6	12	2.8	4	3.6	5			3	1
56	2.2	18					2.3	9	3.3	4	1.0	5				
57	2.0	18					2.2	13			1.4	5				
58	1.7	34	2.4	9	1.0	8			1.0	4	1.5	4	1.9	9		
59	3.2	46			3.5	21	2.3	12	3.3	3	4.0	5	3.5	4	4	1
64	3.6	23	3.0	3			3.4	7	3.5	2	4.0	3	3.9	8		
68	2.2	71	2.2	25	1.9	24	2.5	13	2.3	4	2.8	4			2	1
69	2.9	48	2.9	18	2.9	18	2.6	10			4.0	1			4	1
70	3.4	55	3.3	19	3.3	16	3.5	13	4.0	1	3.4	5			3	1
73	2.6	22	2.7	11	2.5	11										
75	3.5	51	3.3	21	3.6	19	3.7	10			2.0	1				
76	2.8	26	2.7	10	2.8	9	2.6	5			4.0	1			3	1
80	2.3	30	2.8	8	2.6	7	1.8	12			2.0	3				
81	2.7	76	2.2	21	2.3	22	3.4	14	3.3	3	2.6	5	3.4	10	3	1
83	3.4	51	3.6	15	3.6	16	3.0	9	3.0	1	2.8	4	3.2	6		
85	3.1	52	3.2	17	2.8	13	3.1	12	3.4	5	3.2	5				
86	3.3	55	3.4	21	3.2	21	3.3	9			3.0	3			3	1
87	2.0	58	2.6	18	1.3	18	1.8	12	2.8	4	2.4	5			0	1
88	1.2	6							1.3	3	1.0	3				
89	3.0	76	2.8	22	2.9	21	3.0	13	3.0	4	3.8	5	3.4	10	3	1
90	1.8	6					1.0	4	4.0	1	3.0	1				
91	3.4	10	3.5	2	3.5	2			3.5	2	3.3	4				
92	3.6	14					3.8	6	3.7	3	3.7	3	3.0	2		
96	3.0	42	2.5	13	2.9	13	3.3	7	3.3	3	3.8	5			4	1
97	2.9	72	2.7	25	3.1	24	2.7	14	2.8	4	3.2	5				
102	1.8	63	1.8	22	1.8	23	1.5	10	2.7	3	2.2	5				
104	3.7	12	4.0	1	3.0	1			3.4	5	4.0	5				
105	3.1	81	3.0	25	3.4	25	2.9	14	3.3	3	3.0	5	2.9	8	4	1
107	3.1	44	2.6	11	3.1	11	3.5	11			3.0	4	3.6	7		
108	2.6	5									3.3	4			0	1
109	2.5	43	2.1	11	2.5	11	2.8	11					2.6	10		
110	2.9	11	2.0	4					4.0	1			3.3	6		
113	3.1	71	3.2	21	3.2	21	2.9	14	4.0	1	2.6	5	3.0	8	3	1
114	2.4	31	2.1	12	1.7	7	3.6	8	4.0	1	1.3	3				

**Table 4. Overall laboratory performance ratings for standard reference water samples distributed in September 1996—Continued**

[Lab, laboratory number; OWR, overall weighted rating for all sample types; OLR, overall laboratory rating for reported values of a sample type; V/94, number of reported values of 94 total possible values from all sample types; V/28, V/28, V/16, V/5, V/5, V/11, V/1 are number of reported values possible for T-143, T-145, M-140, N-51, N-52, P-27 and Hg-23 respectively]

Standard reference sample =	T-143		T-145		M-140		N-51		N-52		P-27		Hg-23			
Lab	OWR	V/94	OLR	V/28	OLR	V/28	OLR	V/16	OLR	V/5	OLR	V/5	OLR	V/11	OLR	V/1
118	1.9	35	0.6	10	2.7	10	2.7	6	2.0	4	2.2	5				
119	3.0	85	3.0	26	2.8	26	3.2	14	3.0	4	2.6	5	3.1	9	3	1
121	3.3	36	3.2	16	3.1	14	4.0	6							4	1
127	3.1	46	3.0	24			3.4	14	3.5	2	2.4	5				
128	3.0	68	2.8	25	2.8	25	3.3	12	3.0	2	3.5	4				
129	2.3	41	2.0	9	1.7	9	2.5	14	3.8	4	1.8	5				
132	2.5	56	2.8	16	2.1	16	3.1	9	2.5	4	2.3	4	2.1	7		
133	2.8	45	3.1	14	2.5	14	2.8	6	3.0	5	2.4	5		4	1	
134	3.7	89	3.8	27	3.7	27	3.3	15	3.6	5	3.6	5	3.9	9	3	1
138	3.4	86	3.3	26	3.3	26	3.8	15	3.0	4	3.2	5	3.9	9	4	1
140	2.4	54	2.9	13	1.9	13	2.5	11	3.7	3	1.8	5	2.2	9		
141	2.9	72	3.0	22	2.4	19	3.3	12	1.0	3	3.4	5	3.5	10	2	1
142	2.9	83	3.1	28	2.9	28	2.8	16	2.6	5	2.8	5			2	1
143	3.6	18					3.6	5	3.6	5	3.4	5	3.7	3		
144	2.7	9	2.3	4	2.8	4									4	1
145	1.8	78	1.0	23	1.2	23	2.9	15	3.5	2	3.2	5	2.8	9	2	
146	2.6	55	2.9	18	2.5	16	2.8	11			0.0	4	2.8	5	4	1
149	3.0	22			3.0	11	3.2	6	2.5	4				4		
151	3.3	54	3.3	19	3.5	19	3.0	13	3.3	3						
155	2.9	23	4.0	1	1.0	1	3.3	8	3.2	5	2.6	5	2.0	3		
158	2.4	56	2.5	16	1.9	17	3.5	8	3.0	3	3.8	4	1.3	8		
180	3.4	64	3.4	20	3.6	18	2.8	12	3.5	2	3.4	5	3.9	7		
183	1.5	13	0.0	1	0.0	1	1.7	3	2.0	2	2.0	3	1.7	3		
190	2.4	70	2.4	18	2.6	18	2.4	14	1.8	5	1.6	5	3.1	10		
191	3.2	70	3.3	25	3.4	25	3.0	11	4.0	1	3.5	2	2.2	6		
193	2.3	35	2.3	15	1.9	12	3.0	3	3.0	1	2.5	2	3.0	1	2	1
196	3.5	48	3.7	21	3.3	21	2.7	3					4.0	3		
197	3.5	6							3.0	2	4.0	2	3.5	2		
203	2.4	36	2.7	9	2.3	9	2.2	6	3.5	4	2.4	5	1.3	3		
204	2.5	48	2.3	14	2.2	13	3.0	8	3.0	3	2.8	4	2.7	6		
212	2.4	80	2.1	27	2.3	28	2.7	16	3.3	3	2.8	5		3		
213	2.3	34	1.8	12	2.9	12	3.0	4	4.0	1	1.5	4		0	1	
215	2.4	79	2.4	24	2.0	24	2.8	14	2.0	2	2.2	5	2.8	9	0	1
217	2.6	69	3.0	28	2.7	27	1.6	14								
218	2.2	18	2.8	5	2.6	5	1.6	8								
219	2.7	50	2.7	21	2.8	19	2.7	9							3	1
220	3.0	40	3.1	10	3.3	10	3.3	7	2.7	3	1.3	3	2.9	7		
221	3.0	62	3.2	18	3.3	18	3.1	8	2.8	5	2.4	5	2.4	7	3	1
224	1.9	71	1.6	19	1.8	19	1.5	13	2.2	5	2.0	5	3.2	10		
234	3.2	78	3.1	27	3.1	27	3.5	16	3.3	3	3.3	4		4	1	
235	2.0	60	2.4	23	1.8	22	1.1	9					2.0	5	2	
236	2.2	65	1.8	25	2.2	25	2.9	15								
240	2.0	64	1.7	18	1.2	18	3.0	12	0.7	3	2.6	5	3.1	8		
241	2.1	80	1.7	23	1.6	23	2.3	14	3.6	5	3.4	5	2.2	9	3	1
243	2.7	10					2.7	3	3.0	2	3.0	3	2.0	2		
244	2.9	7	0.0	1	0.0	1	4.0	3					4.0	2		
245	2.4	11			2.3	10								3	1	
246	0.0	5	0.0	5												
247	2.7	15					2.3	6	4.0	1	3.0	2	2.8	6		
248	2.0	4							2.0	4						
249	1.6	38	1.4	12	1.8	12	1.8	9			1.4	5				
253	1.4	19	0.5	4	0.8	4	2.0	5	4.0	1	1.4	5				
255	3.4	71	3.5	23	3.5	23	3.1	14			3.5	4	3.7	6	4	1
256	1.6	29	1.8	13			1.6	9					1.3	7		
257	1.4	62	1.3	19	0.7	19	2.0	13					2.2	10	0	1
258	1.0	20					1.4	11					0.4	9		
259	2.1	42	1.5	13	1.5	13	3.1	15					0.9	7	3	1
261	1.1	24	2.0	4	0.8	4	1.0	9					2.4	9		
262	2.2	20					1.9	11								
265	3.2	77	3.3	28	3.5	28	3.3	13					2.4	7	1	1
268	2.6	24	3.3	4	2.7	3	2.0	9					3.0	8		
270	0.3	12	0.0	3	0.7	3	0.0	3					0.3	3		
271	1.3	24	2.5	4	1.0	4	1.1	8					1.1	8		
272	0.6	26	0.0	4	0.8	4	0.4	9					1.0	9		
273	1.1	57	1.1	18	0.7	18	1.1	12					1.7	9		
274	0.3	34	0.5	11	0.1	11	0.4	12					1.8	6		
276	1.3	12					0.7	6								

Table 5. Laboratory performance ratings for standard reference water sample T-143 (trace constituents)

(MPV, most probable value; µg/L, micrograms per liter; mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/28, number of reported values of 28 possible values; RV, reported value; <, less than)

Rating	Absolute Z-value	Rating	Absolute Z-value													
4 (Excellent)	0.00 - 0.50	1 (Questionable)	1.51 - 2.00													
3 (Good)	0.51 - 1.00	0 (Poor)	greater than 2.00													
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)														
Analyte = Ag (Silver)				Al (Aluminium)	As (Arsenic)	B (Boron)	Ba (Barium)	Be (Beryllium)	Ca (Calcium)							
MPV = 19.8 µg/L				22.1 µg/L	15.2 µg/L	35.0 µg/L	81.9 µg/L	8.50 µg/L	53.7 mg/L							
F-pseudosigma = 1.4				8.3	1.2	5.2	4.5	0.66	2.2							
Lab	OLR	V/28	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
1	3.6	28	19.6	4	19.1	4	16.2	3	34.9	4	82.2	4	8.13	3	52.5	4
3	1.7	26	15.0	0	206.0	0	14.3	3	30.0	3	84.4	3	8.00	3	54.4	4
4	2.8	10	< 2000		NR	< 2000		< 2000		87.0	2	9.00	3	53.4	4	
7	2.6	23	21.6	2	37.6	1	< 120	NR	< 2000		83.6	4	8.10	3	55.8	3
10	2.8	6	< 2000		< 2000		16.0	3	< 2000		< 2000		< 2000		52.2	2
11	3.0	25	21.0	3	< 2000		15.0	4	37.0	4	81.0	4	8.40	4	56.3	2
13	2.6	23	20.1	4	22.2	4	15.5	4	< 2000		88.1	2	8.74	4	59.0	0
15	2.2	21	17.6	1	< 50	NR	< 100	NR	< 50	NR	77.9	3	8.60	4	50.6	2
16	3.2	27	19.9	4	18.5	4	15.4	4	57.1	0	82.2	4	7.70	2	54.0	4
18	3.5	22	20.9	3	< 100	NR	14.6	4	< 50	NR	79.0	3	8.50	4	53.5	4
19	3.2	13	< 2000		< 2000		< 2000		< 2000		82.2	4	< 2000		53.6	4
21	0.0	1	< 2000		< 2000		< 2000		< 2000		< 2000		< 2000		52.5	3
23	2.9	15	22.2	1	< 50	NR	20.5	0	< 2000		30.7	3	78.8	3	< 2000	
24	2.6	20	< 2000		< 2000		< 2000		< 2000		84.4	3	9.90	0	54.8	3
25	2.1	15	< 2000		< 2000		< 2000		< 2000		< 2000		< 2000		53.4	4
26	3.3	24	15.6	0	18.2	4	14.9	4	26.1	1	80.9	4	8.66	4	53.8	4
28	1.5	16	< 2000		32.3	2	< 2000		42.2	2	85.1	3	< 2000		55.4	3
30	3.2	21	19.0	3	< 2000		14.0	3	< 2000		80.0	4	7.30	1	54.0	4
32	3.6	26	19.9	4	22.6	4	14.9	4	< 2000		77.9	3	9.40	2	55.0	3
33	2.7	10	< 2000		100.0	0	< 2000		< 2000		102.0	0	< 2000		54.2	4
34	3.3	3	< 2000		< 2000		15.6	4	< 2000		< 2000		< 2000		56.7	2
35	4.0	1	< 2000		< 2000		< 2000		< 2000		< 2000		< 2000		54.0	4
36	1.3	18	13.0	0	< 200	NR	11.0	0	< 2000		90.0	1	10.00	0	52.0	3
39	2.5	11	20.2	4	< 2000		18.2	0	< 2000		79.1	3	9.50	1	< 2000	
40	2.2	19	19.3	4	< 2000		< 2000		< 2000		78.1	3	7.90	3	< 2000	
42	3.0	25	19.0	3	23.0	4	15.0	4	36.0	4	< 2000		< 2000		56.7	2
43	3.4	7	< 2000		< 2000		< 2000		< 2000		< 2000		< 2000		54.0	4
46	3.2	22	20.0	4	< 2000		13.5	2	22.7	0	85.0	3	8.76	4	56.0	2
48	2.2	22	20.0	4	18.5	4	14.2	3	< 100	NR	97.8	0	10.00	0	55.9	2
50	2.1	11	< 2000		< 2000		14.0	3	< 2000		90.0	1	< 2000		52.9	3
51	2.7	3	< 2000		< 2000		< 2000		< 2000		< 2000		< 2000		54.9	3
55	2.7	10	< 2000		< 2000		< 2000		< 2000		78.2	3	< 2000		54.9	3
58	2.4	9	21.0	3	17.0	3	19.0	0	< 2000		< 2000		< 2000		54.0	4
64	3.0	3	< 2000		< 2000		< 2000		< 2000		< 2000		< 2000		53.4	4
68	2.2	25	40.0	0	54.5	0	9.6	0	140.0	0	83.0	4	8.90	3	56.5	2
69	2.9	18	19.8	4	16.5	3	16.2	3	< 2000		< 2000		< 2000		7.32	1
70	3.3	19	18.7	3	< 100	NR	15.3	4	< 50	NR	82.3	4	8.67	4	55.4	3
73	2.7	11	17.0	1	< 2000		22.0	0	< 2000		< 2000		< 2000		53.3	4
75	3.3	21	20.1	4	< 30	NR	14.6	4	< 2000		81.2	4	8.50	4	< 2000	
76	2.7	10	19.4	4	20.2	4	16.5	2	< 2000		< 2000		< 2000		9.68	1
80	2.8	8	< 2000		< 2000		16.3	3	< 2000		< 2000		< 2000		51.3	2
81	2.2	21	< 2000		< 2000		< 6	NR	16.0	3	< 2000		78.0	3	7.00	0
83	3.6	15	< 2000		< 2000		< 25	NR	< 2000		77.8	3	8.20	4	53.1	4
85	3.2	17	20.0	4	< 100	NR	< 2000		35.6	4	83.2	4	9.17	2	55.4	3
86	3.4	21	< 2000		< 2000		15.3	4	36.9	4	81.5	4	8.25	4	55.5	3
87	2.6	18	26.0	0	< 2000		14.9	4	< 2000		90.7	1	< 2000		51.3	2
89	2.8	22	19.1	4	23.1	4	16.4	2	< 2000		98.1	0	7.40	1	52.6	4
91	3.5	2	< 2000		< 2000		< 2000		< 2000		< 2000		< 2000		52.0	3
96	2.5	13	20.4	4	< 2000		14.6	4	< 2000		< 100	NR	10.00	0	< 2000	
97	2.7	25	19.0	3	25.7	4	17.0	1	< 2000		69.6	0	8.99	3	51.8	3
102	1.8	22	143.0	0	20.0	4	15.8	3	< 2000		94.0	0	8.00	3	58.0	1
104	4.0	1	< 2000		< 2000		< 2000		< 2000		< 2000		< 2000		52.8	4
105	3.0	25	17.9	2	21.1	4	15.3	4	< 2000		76.2	2	7.30	1	52.0	3
107	2.6	11	20.0	4	23.0	4	< 2000		< 2000		77.0	2	< 2000		52.0	3
109	2.1	11	< 2000		< 2000		13.1	1	< 2000		< 2000		< 2000		52.6	4
110	2.0	4	< 2000		< 2000											

Table 5. Laboratory performance ratings for standard reference water sample T-143 (trace constituents)—Continued

(MPV, most probable value; µg/L, micrograms per liter; mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/28, number of reported values of 28 possible values; RV, reported value; <, less than)

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Questionable)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Poor)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	
Analyte = Ag (Silver)			
MPV = 19.8 µg/L			
F-pseudosigma = 1.4			
Lab	OLR	V/28	RV
			Rating
133	3.1	14	19.6
134	3.8	27	19.0
138	3.3	26	20.6
140	2.9	13	
141	3.0	22	29.0
142	3.1	28	19.3
144	2.3	4	
145	1.0	23	
146	2.9	18	< 10
151	3.3	19	17.9
155	4.0	1	
158	2.5	16	
180	3.4	20	19.9
183	0.0	1	
190	2.4	18	22.4
191	3.3	25	
193	2.3	15	19.0
196	3.7	21	19.5
203	2.7	9	
204	2.3	14	
212	2.1	27	14.0
213	1.8	12	19.2
215	2.4	24	28.0
217	3.0	28	20.7
218	2.8	5	
219	2.7	21	
220	3.1	10	
221	3.2	18	22.1
224	1.6	19	
234	3.1	27	20.6
235	2.4	23	20.0
236	1.8	25	14.5
240	1.7	18	
241	1.7	23	18.3
244	0.0	1	
246	0.0	5	
249	1.4	12	16.0
253	0.5	4	
255	3.5	23	19.9
256	1.8	13	25.4
257	1.3	19	17.0
259	1.5	13	21.0
261	2.0	4	
265	3.3	28	20.2
268	3.3	4	
270	0.0	3	
271	2.5	4	
272	0.0	4	
273	1.1	18	33.0
274	0.5	11	

Table 5. Laboratory performance ratings for standard reference water sample T-143 (trace constituents)—Continued

(MPV, most probable value; µg/L, micrograms per liter; mg/L, milligrams per liter; Lab, laboratory number; CLR, overall laboratory rating for all reported values; V/28, number of reported values of 28 possible values; RV, reported value; <, less than)

Rating	Absolute Z-value	Rating	Absolute Z-value									
4 (Excellent)	0.00 - 0.50	1 (Questionable)	1.51 - 2.00									
3 (Good)	0.51 - 1.00	0 (Poor)	greater than 2.00									
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)										
Analyte = Cd (Cadmium)	Co (Cobalt)	Cr (Chromium)	Cu (Copper)	Fe (Iron)	K (Potassium)	Li (Lithium)						
MPV = 19.1 µg/L	17.0 µg/L	37.0 µg/L	22.3 µg/L	222 µg/L	2.50 mg/L	18.0 µg/L						
F-pseudosigma = 1.5	1.2	2.6	1.9	14	0.21	2.1						
Lab	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
1	20.1	3	17.4	4	35.5	3	22.2	4	211	3	2.55	4
3	19.5	4	15.0	1	39.0	3	24.0	3	261	0	5.40	0
4	< 100	NR	< 100	NR	< 100	NR	23.0	4	235	3	< 100	NR
7	21.0	2	19.9	0	38.1	4	23.4	3	225	4	2.84	1
10							22.8	4	246	1		
11	19.0	4	18.0	3	36.0	4	24.0	3	202	2	2.50	4
13	17.4	2	12.6	0	38.5	3	10.4	0	207	2	2.38	3
15	15.0	0	< 20	NR	35.1	3	21.3	4	208	2	2.28	2
16	19.7	4	16.9	4	34.2	2	21.8	4	200	1	2.10	1
18	18.7	4	16.1	3	36.3	4	22.3	4	222	4	2.40	4
19	19.8	4			37.8	4	16.0	0	227	4	2.50	4
21									152	0		
23	18.8	4			35.0	3	23.9	3	218	4	2.44	4
24	17.7	3	12.3	0	31.2	0	16.2	0	222	4	2.34	3
25	21.0	2			37.0	4	21.0	3	202	2	2.64	3
26	20.4	3	16.4	3	37.5	4	23.4	3	222	4	2.49	4
28	13.6	0					7.8	0			2.32	3
30	19.5	4	17.0	4	37.0	4	22.5	4	470	0		
32	18.6	4	17.2	4	37.3	4	23.2	4	365	0	2.50	4
33									250	1	2.54	4
34												
35									216	4		
36	19.0	4			37.0	4	10.0	0	250	1	1.80	0
39	18.0	3			28.0	0						
40	18.6	4	15.2	1	32.5	1	21.3	4	21	0	2.32	3
42	18.0	3	20.0	0	38.0	4	22.0	4	224	4	2.50	4
43									230	3	2.30	3
46	18.6	4	17.4	4	35.8	4	22.4	4	222	4	2.62	3
48	18.0	3	< 50	NR	41.8	1	19.7	2	160	0	2.51	4
50			16.0	3	36.0	4	24.0	3	251	0		
51											2.48	4
55							20.8	3	222	4		
58	20.0	3			37.0	4	< 50	NR	210	3		
64											2.57	4
68	19.5	4	18.0	3	38.5	3	27.0	0	225	4	2.80	2
69	20.0	3			36.0	4	21.4	4	226	4	2.72	2
70	18.2	3	< 50	NR	37.1	4	22.8	4	214	3	2.43	4
73	19.0	4			33.0	1	22.0	4	221	4		
75	19.2	4	18.9	1	37.1	4	22.2	4	214	3		
76					36.3	4					19.7	3
80	13.9	0					22.0	4	223	4		
81	20.0	3			41.0	1	22.0	4	220	4	2.37	3
83	18.5	4			37.2	4	23.0	4	219	4	2.50	4
85	19.6	4			38.4	3	25.2	1	224	4	2.82	2
86	17.7	3	16.8	4	29.2	0	23.1	4	213	3	2.58	4
87	21.0	2			38.2	4	21.0	3	212	3	2.42	4
89	20.3	3	17.8	3	36.0	4	21.4	4	244	1	2.35	3
91									212	3		
96	17.5	2			38.9	3	23.4	3	248	1		
97	19.3	4	17.2	4	39.0	3	21.2	3	233	3	2.54	4
102	21.0	2	20.0	0	41.0	1	22.0	4	244	1	2.50	4
104												
105	19.7	4	15.5	2	32.9	1	22.4	4	213	3	2.47	4
107							24.8	2	200	1	2.63	3
109									256	0	2.51	4
110											19.4	3
113	19.1	4			36.1	4	23.5	3	223	4	2.40	4
114	19.0	4					21.0	3	204	2	3.00	0
118	6.4	0			44.8	0	7.9	0				
119	18.4	4	15.9	3	35.0	3	26.0	1	226	4	3.20	0
121	18.0	3	16.0	3			18.0	0	210	3	2.55	4
127	17.4	2	17.3	4	39.2	3	22.4	4	220	4	2.68	3
128	19.5	4	17.0	4	32.4	1	19.2	1	205	2	2.11	1
129							30.0	0	185	0	2.40	4
132	18.5	4	17.0	4	40.5	2	42.0	0	228	4	2.64	3

Table 5. Laboratory performance ratings for standard reference water sample T-143 (trace constituents)—Continued

(MPV, most probable value; µg/L, micrograms per liter; mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/28, number of reported values of 28 possible values; RV, reported value; <, less than)

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Questionable)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Poor)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Analyte = Cd (Cadmium) MPV = 19.1 µg/L F-pseudosigma = 1.5	Co (Cobalt) 17.0 µg/L	Cr (Chromium) 37.0 µg/L	Cu (Copper) 22.3 µg/L	Fe (Iron) 222 µg/L	K (Potassium) 2.50 mg/L	Li (Lithium) 18.0 µg/L						
Lab	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
133	19.9	3	34.3	2	19.5	2	223	4				
134	18.8	4	16.7	4	36.2	4	218	4				
138	19.0	4	15.1	1	38.7	3	20.1	2	224	4	2.48	4
140	20.0	3			35.0	3	26.0	1	237	2	2.48	4
141	19.8	4	16.2	3	36.7	4	23.3	3	200	1	2.51	4
142	20.0	3	16.3	3	35.4	3	20.0	2	221	4	2.55	4
144							20.2	2				
145	22.3	0	24.2	0	42.1	1	33.2	0	216	4	2.66	3
146	18.6	4	16.2	3	36.4	4	25.2	1	220	4	2.93	1
151	19.6	4			34.7	3	22.1	4	200	1	16.9	3
155							228	4				
158	8.8	0	17.1	4	25.6	0	25.8	1	222	4		
180	19.6	4	17.3	4	36.8	4	23.3	3	222	4	2.32	3
183												
190	20.7	2			39.4	3	22.2	4	192	0	2.50	4
191	18.6	4	17.6	3	38.4	3	22.3	4	221	4	2.47	4
193	20.0	3			36.0	4	26.0	1			2.32	3
196	18.2	4	17.3	4	34.5	3	23.7	3				17.9
203							18.7	1	220	4	2.88	1
204					47.0	0	11.7	0	208	2	2.40	4
212	23.0	0	17.0	4	33.0	1	26.0	1	260	0	2.40	4
213	21.4	1	18.7	2	44.5	0	23.2	4	302	0		
215	19.0	4	19.0	1	35.0	3	30.4	0	224	4	3.10	0
217	17.1	2	16.8	4	37.6	4	23.9	3	230	3	2.92	1
218										2.35	3	
219	18.0	3	16.0	3	34.0	2	19.0	1	208	2	2.50	4
220									222	4	2.60	4
221	19.6	4	16.0	3	37.7	4	21.3	4	220	4	2.61	3
224	13.9	0	15.9	3			21.5	4	201	1	2.31	3
234	19.7	4	17.4	4	41.6	1	23.6	3	224	4	2.45	4
235	20.0	3	19.0	1	40.0	2	20.0	2	235	3	3.60	0
236	17.0	2	13.9	0	35.1	3	18.7	1	215	3	2.14	1
240							21.0	3	184	0		13.1
241	20.3	3			34.0	2	19.7	2	264	0	2.54	4
244										2.70	3	
246										< 0.2	0	
249	16.0	0			42.4	1	21.7	4	188	0	2.68	3
253					40.2	2	30.0	0	330	0		
255	19.6	4	17.2	4	36.0	4	21.7	4	217	4	2.38	3
256	20.4	3			40.0	2	23.6	3	221	4	3.52	0
257	26.0	0	23.0	0	52.0	0	28.0	0	170	0	2.90	1
259	14.0	0	10.0	0	34.0	2	25.0	2			2.70	3
261										2.35	3	
265	20.0	3	17.8	3	37.0	4	23.0	4	230	3	2.46	4
268										2.70	3	
270										3.05	0	
271										2.80	2	
272										3.00	0	
273	16.0	0	8.0	0	38.0	4	22.0	4	230	3	2.97	0
274	7.4	0					4.2	0	232	3	2.78	2

Table 5. Laboratory performance ratings for standard reference water sample T-143 (trace constituents)—Continued

(MPV, most probable value;  $\mu\text{g/L}$ , micrograms per liter;  $\text{mg/L}$ , milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/28, number of reported values of 28 possible values; RV, reported value; <, less than)

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Questionable)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Poor)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Analyte = Mg (Magnesium)	Mn (Manganese)	Mo (Molybdenum)	Na (Sodium)	Ni (Nickel)	Pb (Lead)	Sb (Antimony)
MPV = 10.4 mg/L	18.2 $\mu\text{g/L}$	36.1 $\mu\text{g/L}$	34.0 mg/L	71.0 $\mu\text{g/L}$	83.4 $\mu\text{g/L}$	16.6 $\mu\text{g/L}$
F-pseudosigma = 0.5	1.9	4.3	1.6	5.0	7.1	1.5
Lab	RV	Rating	RV	Rating	RV	Rating
1	10.1	3	17.8	4	34.4	4
3	10.3	4	14.7	1	27.0	0
4	10.9	3	19.0	4	< 500	NR
7	10.8	3	18.6	4	29.4	1
10		22.0	1			
11	11.1	2	18.0	4	37.0	4
13	10.7	3	16.7	3		
15	9.8	2	28.3	0	38.3	3
16	10.5	4	18.2	4	37.7	4
18	10.2	4	16.5	3	37.0	4
19	10.5	4	18.8	4		
21					34.3	4
23	10.2	4	17.8	4	< 100	NR
24	10.3	4	17.4	4	34.0	4
25	10.0	3	22.0	1		
26	10.9	3	14.7	1	37.0	4
28	10.8	3	25.1	0	26.8	0
30	10.0	3	17.0	3	36.0	4
32	10.8	3	17.7	4	36.4	4
33	10.3	4	20.0	3		
34					33.7	4
35						
36	8.7	0	19.0	4		
39					38.0	0
40	10.3	4	14.4	1	37.9	4
42	11.1	2	21.0	2	36.0	4
43	10.5	4	20.0	3		
46	10.5	4	16.2	2	36.0	4
48	10.8	3	20.0	3	37.6	4
50			21.0	2	32.0	3
51	8.8	0			34.3	4
55	11.3	1	13.0	0	30.0	2
58			< 50	NR		
64					33.8	4
68	11.0	2	18.5	4	37.0	4
69	10.1	3	< 20	NR		
70	10.5	4	< 20	NR	< 50	NR
73					34.2	4
75	10.6	4	18.0	4	31.7	2
76					34.5	4
80			15.3	2		
81	10.3	4	17.0	3	28.0	4
83	10.1	3	17.4	4		
85	10.8	3			32.9	3
86	10.4	4	18.0	4	40.5	2
87	10.2	4	20.5	2	35.1	4
89	10.4	4	17.6	4		
91			17.6	4		
96			22.0	1		
97	10.4	4	22.0	1	35.4	4
102	12.5	0	20.0	3		
104					30.9	1
105	10.3	4	17.5	4	40.2	3
107	10.7	3	20.0	3		
109	10.4	4	19.3	3	12.5	0
110	8.3	0			34.1	4
113	10.7	3	18.6	4		
114	10.0	3	20.0	3		
118					33.0	3
119	10.6	4	20.0	3	35.1	4
121	10.0	3	18.0	4		
127	10.4	4	16.7	3	32.1	3
128	10.0	3	15.7	2	31.3	2
129	10.0	3	20.0	3		
132	10.5	4	17.0	3	26.0	0

Table 5. Laboratory performance ratings for standard reference water sample T-143 (trace constituents)—Continued

(MPV, most probable value;  $\mu\text{g/L}$ , micrograms per liter; mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/28, number of reported values of 28 possible values; RV, reported value; <, less than)

Rating	Absolute Z-value	Rating	Absolute Z-value										
4 (Excellent)	0.00 - 0.50	1 (Questionable)	1.51 - 2.00										
3 (Good)	0.51 - 1.00	0 (Poor)	greater than 2.00										
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)											
Analyte = Mg (Magnesium)	Mn (Manganese)	Mo (Molybdenum)	Na (Sodium)	Ni (Nickel)	Pb (Lead)	Sb (Antimony)							
MPV = 10.4 mg/L	18.2 $\mu\text{g/L}$	36.1 $\mu\text{g/L}$	34.0 mg/L	71.0 $\mu\text{g/L}$	83.4 $\mu\text{g/L}$	16.6 $\mu\text{g/L}$							
F-pseudosigma = 0.5													
Lab	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	
133	9.9	3					74.3	3	92.0	2			
134	10.2	4	18.1	4	37.0	4	33.7	4	69.4	4	83.4	4	
138	10.7	3	18.0	4	35.9	4	34.1	4	70.0	4	77.1	3	
140	11.0	2	17.0	3			42.5	0	71.0	4	86.0	4	
141	10.9	3	18.2	4	37.3	4	35.4	3	69.0	4	101.0	0	
142	10.0	3	19.0	4	38.3	3	34.2	4	68.3	3	87.0	4	
144													
145	11.4	1	20.7	2	30.6	2	37.1	1	86.3	0	498.9	0	
146	10.1	3	17.8	4	36.2	4	35.5	3	72.4	4	84.2	4	
151					16.5	3	38.6	3	69.0	4	82.8	4	
155													
158	11.1	2	19.5	3					76.0	3	30.1	0	
180	10.6	4	18.1	4	41.0	2	34.0	4	76.0	3	92.1	2	
183					22.7	0							
190	10.4	4	19.3	3			34.0	4	59.4	0	60.8	0	
191	10.7	3	19.0	4	39.7	3	33.2	4	72.2	4	82.0	4	
193	9.8	2					31.6	2	66.0	3	88.0	3	
196					19.0	4	36.4	4		70.5	4	85.4	4
203	10.6	4	20.0	3			32.5	3				15.8	
204	7.6	0	14.0	0			31.2	1			82.2	4	
212	11.2	1	19.0	4	38.0	4	36.7	1	72.0	4	120.0	0	
213									91.2	0	90.0	3	
215	10.6	4	18.1	4	37.0	4	34.5	4	71.0	4	82.0	4	
217	10.0	3	18.2	4	37.8	4	33.9	4	74.1	3	90.1	3	
218	10.4	4					31.0	1				17.9	
219	10.0	3	17.0	3	27.0	0	33.0	3	66.0	3	70.0	1	
220	10.0	3	19.2	3			31.0	1			78.4	3	
221	9.7	2	19.7	3	35.4	4	34.1	4	71.2	4	84.5	4	
224	10.0	3	19.0	4	48.0	0	31.2	1	59.9	0	140.3	0	
234	9.9	3	18.4	4	34.3	4	33.1	3	74.6	3	89.3	3	
235	10.5	4	18.0	4	39.0	3	39.0	0	74.0	3	87.0	4	
236	10.7	4	16.4	3	19.7	0	32.8	3	65.5	2	62.4	0	
240	10.2	4	18.0	4	11.0	0			75.0	3	131.0	0	
241	9.4	1	17.0	3	52.0	0	31.0	1	67.6	3	99.2	0	
				31.5	0							15.0	
246	8.7	0					41.0	0					
249							35.6	2	69.2	4	79.2	3	
253													
255	10.3	4	18.0	4	36.3	4	32.8	3	70.9	4	83.4	4	
256			19.0	4			31.2	1	76.6	2	155.0	0	
257	8.7	0	18.0	4	44.0	1	33.0	3	68.0	3	91.0	2	
259							34.0	4	125.0	0	54.0	0	
261	10.5	4					40.2	0					
265	10.1	3	19.0	4	35.0	4	33.0	3	73.0	4	79.6	3	
268	10.2	4					35.8	2				16.8	
270							42.2	0					
271	25.4	0					34.0	4					
272	9.0	0					50.0	0					
273	11.2	1	21.0	2			36.2	2	82.0	0	65.0	0	
274	17.7	0	7.4	0			18.4	0		1.2	0		

Table 5. Laboratory performance ratings for standard reference water sample T-143 (trace constituents)—Continued

(MPV, most probable value;  $\mu\text{g/L}$ , micrograms per liter;  $\text{mg/L}$ , milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/28, number of reported values of 28 possible values; RV, reported value; <, less than)

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Questionable)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Poor)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Analyte = Se (Selenium)			SiO <sub>2</sub> (Silica)		Sr (Strontium)		Tl (Thallium)		U (Uranium)		V (Vanadium)		Zn (Zinc)			
MPV =	9.63	$\mu\text{g/L}$	23.4	mg/L	306	$\mu\text{g/L}$	10.0	$\mu\text{g/L}$	12.0	$\mu\text{g/L}$	30.0	$\mu\text{g/L}$	20.0	$\mu\text{g/L}$		
F-pseudosigma =	1.64		1.7		15		1.0		0.9		3.0		2.2			
Lab	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating		
1	8.80	4	22.1	3	301	4	10.4	4	10.8	2	31.1	4	18.7	3		
3	< 5	0	24.4	3	309	4	< 10	NR			23.0	0	17.1	2		
4	27.0	0	323	2							< 50	NR	< 200	NR		
7	< 50	NR	23.8	4	303	4			< 120	NR	32.4	3	19.1	4		
10	9.40	4											21.0	4		
11	13.60	0	18.1	0	299	4	14.0	0			30.0	4	16.0	1		
13	7.20	2	22.9	4			9.8	4			39.7	0	20.1	4		
15	10.40	4	25.6	2			8.0	1			20.8	0	18.2	3		
16	11.20	3			282	1	10.2	4	12.0	4	30.3	4	19.5	4		
18	8.20	3			290	2	9.7	4			28.0	3	< 100	NR		
19													21.1	4		
21																
23	8.92	4					< 5	0					< 20	NR		
24			24.1	4	306	4					24.0	1	17.9	3		
25			25.5	2	324	2					23.0	0				
26	8.57	3	23.3	4							30.0	4	20.1	4		
28					319	3					30.2	4	16.5	1		
30	12.00	2							11.0	2	30.0	4	20.0	4		
32	9.67	4	24.0	4	311	4	9.6	4			30.5	4	19.9	4		
33			22.2	3	308	4										
34	9.25	4														
35													22.0	3		
36	5.00	0														
39	10.60	3					10.8	3								
40					249	0					27.7	3	13.8	0		
42	12.00	2	24.8	3	299	4	< 5	0			32.0	3	19.0	4		
43			23.0	4												
46	9.30	4					10.9	3			32.2	3				
48	7.80	2					7.9	1			27.7	3	< 5	0		
50	9.10	4									22.0	0				
51																
55			23.5	4	306	4					28.3	3				
58	10.00	4										< 50	NR			
64																
68	6.25	0			310	4					26.0	2	21.5	3		
69	9.55	4										< 50	NR			
70	< 10	NR	22.4	3	309	4	11.0	3				< 50	NR			
73	26.00	0					7.8	0				20.1	4			
75	9.67	4										20.0	4			
76					26.1	1						< 100	NR			
80	9.30	4										29.6	4	19.2	4	
81			23.6	4	262	0	53.0	0				19.0	0	15.5	1	
83			21.3	2									17.0	2		
85					318	3						30.4	4	18.3	3	
86	10.90	3			301	4						31.3	4	20.5	4	
87	19.40	0	23.4	4									24.0	1		
89	8.00	3	21.2	2			< 10	NR					32.3	3	40.7	0
91																
96	9.70	4											22.0	3		
97	5.75	0	23.0	4	255	0	10.9	3				34.9	1	14.6	0	
102	19.00	0			356	0						33.0	2	20.0	4	
104					22.7	4										
105	10.30	4	21.2	2	285	2							29.4	4	19.3	4
107			26.3	1												
109					265	0										
110			25.4	2												
113	8.65	3	22.8	4	295	3	9.6	4						19.4	4	
114														18.0	3	
118	8.10	3	25.5	2										24.0	1	
119	10.40	4	23.0	4										44.0	0	
121			22.7	4	300	4								32.0	3	
127	10.30	4	21.4	2	285	2								31.1	4	
128	10.50	3	24.0	4										27.7	3	
129			23.7	4										20.2	4	
132														20.0	4	

Table 5. Laboratory performance ratings for standard reference water sample T-143 (trace constituents)—Continued

(MPV, most probable value; µg/L, micrograms per liter; mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/28, number of reported values of 28 possible values; RV, reported value; <, less than)

Rating	Absolute Z-value	Rating	Absolute Z-value											
4 (Excellent)	0.00 - 0.50	1 (Questionable)	1.51 - 2.00											
3 (Good)	0.51 - 1.00	0 (Poor)	greater than 2.00											
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)												
Analyte = Se (Selenium)	SiO <sub>2</sub> (Silica)	Sr (Strontium)	Tl (Thallium)	U (Uranium)	V (Vanadium)	Zn (Zinc)								
MPV = 9.63 µg/L	23.4 mg/L	306 µg/L	10.0 µg/L	12.0 µg/L	30.0 µg/L	20.0 µg/L								
F-pseudosigma = 1.64		1.7	15	1.0	0.9	2.2								
Lab	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating		
133	8.20	3									20.5	4		
134	10.65	3	22.9	4	295	3	10.2	4		29.7	4	20.4	4	
138	9.58	4	21.5	2	297	3	9.4	3		29.5	4	17.3	2	
140			22.8	4			< 50	NR				20.0	4	
141	10.30	4							30.8	4	19.7	4		
142	10.50	3	26.7	1	315	3	10.4	4	12.2	4	29.6	4	15.7	1
144	9.30	4								37.3	0	25.1	0	
145			26.0	1	333	1				30.0	4	< 20	NR	
146	< 10	NR					< 10	NR						
151	10.80	3			316	3	10.3	4				20.2	4	
155														
158			24.6	3						30.6	4	18.6	3	
180	< 50.1	NR					< 32.1	NR		30.7	4	20.4	4	
183														
190	31.50	0	23.5	4	613	0						21.3	3	
191	11.40	2	23.3	4	305	4	9.9	4		32.2	3	18.8	3	
193	7.00	1					7.0	0				< 50	NR	
196	10.80	3			312	4	10.1	4	11.5	3	29.9	4	19.8	4
203			23.7	4										
204			23.4	4									18.0	3
212	12.00	2	24.3	3	330	1	10.0	4	14.0	0	32.0	3	22.0	3
213							9.0	3					29.3	0
215	9.10	4	45.9	0			< 7	0					20.0	4
217	9.77	4	20.7	1	295	3	10.9	3	13.0	2	28.6	4	17.9	3
218					318	3								
219			23.0	4	310	4					28.0	3	17.0	2
220	7.80	2											20.0	4
221	9.16	4											34.8	0
224	6.10	0									34.5	1	19.0	4
234	9.56	4	22.5	4	310	4	7.6	0			27.6	3	15.4	0
235			25.9	2	311	4	5.0	0			36.0	1	19.5	4
236	174.10	0	14.3	0	297	3					31.1	4	15.5	1
240	15.00	0	21.0	2	294	3	9.0	3					23.0	2
241	6.44	1	24.7	3			10.5	4			34.7	1	16.0	1
244														
246			5.8	0									184.0	0
249													50.0	0
253													29.7	4
255	5.33	0			20.6	1							18.8	3
256													20.4	4
257											93.0	0	21.0	4
259	8.60	3											50.0	0
261														
265	11.85	2	21.2	2	300	4	8.9	2	11.0	2	30.0	4	22.0	3
268														
270													15.0	0
271													5.8	0
272														
273							336	1						
274					31.3	0								

Table 6. Laboratory performance ratings for standard reference water sample T-145 (trace constituents)

(MPV, most probable value; µg/L, micrograms per liter; mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/28, number of reported values of 28 possible values; RV, reported value; <, less than)

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Questionable)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Poor)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

	Analyte = Ag (Silver)		Al (Aluminium)		As (Arsenic)		B (Boron)		Ba (Barium)		Be (Beryllium)		Ca (Calcium)			
Lab	OLR	V/28	MPV	µg/L	67.6	µg/L	9.88	µg/L	45.6	µg/L	37.1	µg/L	9.04	µg/L	30.7	mg/L
			F-pseudosigma	=	11.0		1.04		5.8		1.9		0.70		1.3	
1	3.4	28	7.55	4	60.1	3	9.90	4	45.7	4	36.8	4	8.94	4	29.0	2
3	1.5	25	6.00	1	98.0	0	10.00	4	40.0	3	36.5	4	9.20	4	31.1	4
4	3.2	17	8.00	4	68.0	4	13.00	0			37.0	4			30.5	4
5	3.4	23	6.55	2	74.2	3	9.86	4	49.2	3	36.0	3	8.84	4	28.3	1
7	2.5	19	9.90	0	83.8	2	< 120	NR			38.5	3	8.70	4	31.8	3
10	2.8	8			10.00		4									
11	2.9	26	8.00	4	60.0	3	10.00	4	50.0	3	36.0	3	9.00	4	31.7	3
13	2.8	22	8.13	3	79.8	2	10.50	3			39.8	2	9.17	4	33.2	1
15	1.6	20	< 10	NR	82.1	2	< 100	NR	< 50	NR	34.2	1	8.70	4	28.6	1
16	2.7	27	7.60	4	59.1	3	10.40	4	77.0	0	38.0	4	8.40	3	30.1	4
18	3.6	20	8.40	3	< 100	NR	9.20	3	< 50	NR	36.2	4	9.20	4	30.6	4
19	3.3	6													31.0	4
23	2.4	14	7.79	4	< 50	NR	12.80	0			36.9	2	34.8	2	12.4	0
24	2.8	14									38.3	3			30.0	3
25	2.2	11											8.10	2	31.1	4
26	3.5	24	7.76	4	65.8	4	9.62	4	37.6	2	37.6	4	9.17	4	30.6	4
28	1.5	16			37.5	0			45.5	4	38.6	3			31.5	3
30	2.9	21	7.60	4			9.00	3			36.0	3	7.60	0	32.0	2
32	3.6	26	7.40	4	69.1	4	9.77	4			35.4	3	9.65	3	31.5	3
33	2.4	10			100.0	0					41.0	0			30.3	4
34	3.7	3					9.60	4								
35	3.0	1														
36	0.8	20	15.00	0	300.0	0	8.00	1			< 0.05	0	11.00	0	30.0	3
40	2.1	16			58.8	3					35.3	3	2.30	0		
42	2.0	26	7.50	4	67.0	4	11.00	2	47.0	4			13.00	0	12.1	0
43	3.4	7													31.0	4
46	3.1	21	7.10	4	86.8	1	9.10	3	22.4	0	37.6	4	8.96	4	32.0	2
48	2.8	21	7.60	4	49.2	1	9.80	4	< 100	NR	43.2	0	10.00	2	31.6	3
51	2.7	3														
58	1.0	8	9.00	1	117.0	0	12.00	1								
59	3.5	21	7.60	4	74.4	3	10.00	4			37.0	4	9.80	2	30.0	3
68	1.9	24	10.10	0	98.5	0	6.90	0	125.0	0	38.5	3	9.30	4		
69	2.9	18	6.97	3	50.0	1	10.20	4					7.94	1	30.2	4
70	3.3	16	< 10	NR	< 100	NR	< 10	NR	< 50	NR	< 50	NR	9.17	4	31.6	3
73	2.5	11			62.0	3	13.00	0								
75	3.6	19	6.12	1	67.2	4	9.44	4			37.2	4	9.10	4	30.0	3
76	2.8	9			66.7	4	10.60	3					10.20	1		
80	2.6	7					10.60	3								
81	2.3	22			51.0	1	10.00	4			35.0	2	8.00	2	30.1	4
83	3.6	16			57.0	3					35.5	3	8.80	4	30.0	3
85	2.8	13	8.20	3	< 100	NR			42.4	3	37.7	4	9.64	3	31.0	4
86	3.2	21					9.95	4	45.7	4	36.8	4	8.79	4	31.8	3
87	1.3	18	13.00	0			9.40	4			40.0	1			29.2	2
89	2.9	21	7.60	4	64.6	4	9.00	3			< 50	NR	8.20	2	29.5	3
91	3.5	2														
96	2.9	13	8.20	3			9.70	4			< 100	NR	10.00	2		
97	3.1	24	6.64	3	72.7	4	10.40	4			29.2	0	9.43	3	29.1	2
102	1.8	23	52.00	0	69.0	4	6.10	0			42.0	0	8.80	4	33.0	1
104	3.0	1														
105	3.4	25	6.78	3	66.0	4	10.30	4			34.8	2	10.00	2	30.3	4
107	3.1	11	7.40	4	65.0	4					39.0	2			30.9	4
109	2.5	11					8.50	2							28.7	1
113	3.2	21	6.80	3	68.1	4	11.50	1			37.0	4	10.10	1	33.2	1
114	1.7	7	< 10	NR									< 10	NR	24.0	0
118	2.7	10	6.90	3	< 2000	NR	10.60	3								
119	2.8	26	6.90	3	62.0	3	9.00	3	42.0	3	39.0	2	8.92	4	31.7	3
121	3.1	14									36.0	3	10.00	2	30.0	3
128	2.8	25	7.50	4	71.0	4	9.95	4	40.9	3	31.9	0	8.40	3	32.0	2
129	1.7	9							125.0	0					34.0	0
132	2.1	16			73.0	4			28.0	0					30.8	4

Table 6. Laboratory performance ratings for standard reference water sample T-145 (trace constituents)—Continued

(MPV, most probable value; µg/L, micrograms per liter; mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/28, number of reported values of 28 possible values; RV, reported value; <, less than)

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Questionable)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Poor)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Lab	OLR	Analyte = Ag (Silver)		Al (Aluminium)		As (Arsenic)		B (Boron)		Ba (Barium)		Be (Beryllium)		Ca (Calcium)	
		MPV =	7.55 µg/L	67.6 µg/L	9.88 µg/L	45.6 µg/L	37.1 µg/L	9.04 µg/L	30.7 mg/L						
		F-pseudosigma =	0.92	11.0	1.04	5.8	1.9	0.70	1.3						
133	2.5	14	6.10	1	7.30	0	36.4	4	9.40	3	30.4	4			
134	3.7	27	6.89	3	71.7	4	9.07	3	47.1	4	35.6	3	9.13	4	31.5
138	3.3	26	7.07	3	64.6	4	9.51	4	41.8	3	36.4	4	9.22	4	31.5
140	1.9	13													35.0
141	2.4	19	12.30	0	< 100	NR	8.65	2	38.7	2	38.2	3	9.07	4	32.0
142	2.9	28	6.60	2	66.3	4	9.61	4	46.6	4	37.2	4	9.35	4	30.1
144	2.8	4					9.60	4				< 0.01	0		
145	1.2	23			107.2	0	19.30	0	44.7	4	40.8	1	9.90	2	32.8
146	2.5	16	< 10	NR	< 200	NR	13.50	0			38.5	3	8.19	2	29.6
149	3.0	11	7.60	4	60.0	3					40.0	1			
151	3.5	19	7.64	4	64.1	4	9.29	3			37.8	4	8.97	4	
155	1.0	1													
158	1.9	17			71.5	4			47.8	4	38.3	3	9.90	2	32.6
180	3.6	18	6.30	2	< 40.6	NR			48.8	3	36.2	4	8.70	4	31.0
183	0.0	1													
190	2.6	18	8.78	2	61.7	3	11.20	2							30.3
191	3.4	25			68.0	4	11.20	2	44.0	4	38.0	4	10.00	2	30.6
193	1.9	12	6.00	1			8.00	1					9.00	4	29.4
196	3.3	21	7.71	4	67.2	4	9.85	4			36.3	4	8.83	4	
203	2.3	9			60.9	3									29.1
204	2.2	13			79.4	2	11.40	2			39.4	2			31.7
212	2.3	28	6.80	3	74.0	3	9.60	4	52.0	2	42.0	0	11.00	0	31.9
213	2.9	12	7.45	4			8.60	2					9.12	4	
215	2.0	24	10.00	0	127.0	0	11.00	2	121.0	0	39.0	2	8.70	4	31.7
217	2.7	27	6.83	3	81.9	2	9.14	3	45.9	4	37.1	4	8.60	3	26.9
218	2.6	5													31.0
219	2.8	19			68.0	4									30.0
220	3.3	10					9.50	4							31.0
221	3.3	18	7.36	4	68.5	4	9.31	3							32.4
224	1.8	19			54.9	2	8.10	1			33.2	0	13.20	0	28.5
234	3.1	27	7.34	4	67.0	4	10.90	3	44.1	4	38.3	3	8.94	4	31.0
235	1.8	22	7.00	3	52.0	2			40.0	3	40.0	1	9.00	4	33.5
236	2.2	25	4.70	0	71.2	4	42.70	0	37.4	2	36.3	4	8.50	3	30.2
240	1.2	18			115.0	0			34.0	1	31.0	0			28.4
241	1.6	23	6.90	3	54.4	2	7.40	0			44.5	0	7.20	0	28.4
244	0.0	1													
245	2.3	10	5.82	1									8.93	4	
249	1.8	12	5.26	0	77.8	3	11.00	2							
253	0.8	4													
255	3.5	23	7.93	4	58.1	3	10.00	4	46.8	4	36.8	4	8.83	4	30.2
257	0.7	19	13.00	0	50.0	1									34.0
259	1.5	13	10.50	0							35.0	2			
261	0.8	4													30.1
265	3.5	28	7.65	4	66.9	4	11.00	2	44.0	4	38.0	4	9.15	4	30.7
268	2.7	3													29.5
270	0.7	3													32.5
271	1.0	4													32.6
272	0.8	4													14.0
273	0.7	18	18.00	0	157.0	0			150.0	0					32.4
274	0.1	11													0.0

Table 6. Laboratory performance ratings for standard reference water sample T-145 (trace constituents)—Continued

(MPV, most probable value; µg/L, micrograms per liter; mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/28, number of reported values of 28 possible values; RV, reported value; <, less than)

Rating	Absolute Z-value	Rating	Absolute Z-value									
4 (Excellent)	0.00 - 0.50	1 (Questionable)	1.51 - 2.00									
3 (Good)	0.51 - 1.00	0 (Poor)	greater than 2.00									
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)										
Analyte = Cd (Cadmium)	Co (Cobalt)	Cr (Chromium)	Cu (Copper)	Fe (Iron)	K (Potassium)	Li (Lithium)						
MPV = 9.33 µg/L	10.0 µg/L	15.3 µg/L	11.0 µg/L	101 µg/L	2.13 mg/L	27.3 µg/L						
F-pseudosigma = 0.82	0.9	1.4	1.4	8	0.16	2.5						
Lab	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
1	9.79	3	10.0	4	15.5	4	11.1	4	96	3	2.17	4
3	10.10	3	6.0	0	17.0	2	8.0	0	120	0	4.50	0
4	9.50	4	10.6	3	15.0	4			102	4		
5	8.90	3	10.1	4	15.5	4	11.8	3	100	4	2.12	4
7	13.00	0	11.4	1	18.0	1	13.9	1	104	4	2.46	1
10					16.0	4	10.8	4	105	3		
11	9.00	4	11.0	2	15.0	4	12.0	3	90	2	2.12	4
13	9.44	4	7.6	0	16.1	3	11.3	4	117	1	2.08	4
15	17.20	0	< 20	NR	16.1	3	11.1	4	92	2	38.30	0
16	9.10	4	10.4	4	13.8	2	11.4	4	77	0	1.60	0
18	9.20	4	9.9	4	14.3	3	12.3	3	100	4	2.10	4
19							8.0	0			2.10	4
23	9.16	4			12.3	0	11.5	4	< 500	NR	2.09	4
24	7.60	0	4.8	0					98	4	1.94	2
25									83	0	2.31	2
26	10.40	2	10.4	4	15.4	4	11.7	4	102	4	2.13	4
28	8.10	1					4.6	0			2.32	2
30	9.60	4	10.2	4	14.3	3	11.0	4	230	0		
32	9.18	4	10.2	4	14.5	3	11.3	4	183	0	2.10	4
33									120	0	2.18	4
34												
35									97	3		
36	8.60	3			10.0	0	19.0	0	140	0	1.60	0
40	8.50	2	7.1	0	9.5	0	10.0	3	95	3	1.99	3
42	9.00	4	14.0	0	18.0	1	13.0	2	107	3	2.30	2
43									110	2	2.20	4
46	8.84	3	10.7	3	16.4	3	11.3	4	101	4	2.26	3
48	9.20	4	< 50	NR	17.4	2	11.3	4	30	0	2.13	4
51											2.16	4
58	37.00	0			17.0	2	< 50	NR	100	4		
59	9.00	4			16.0	4	10.8	4	100	4	1.70	0
68	10.00	3	11.5	1	16.5	3	15.0	0	105	3	2.10	4
69	10.40	2			14.2	3	10.2	3	108	3	2.34	2
70	8.60	3	< 50	NR	15.2	4	11.8	3	102	4	1.96	2
73	9.00	4			16.0	4	11.0	4	101	4		
75	8.99	4	10.8	3	16.1	3	10.2	3	98	4		
76					14.8	4					27.3	4
80	< 10	NR					10.7	4	99	4		
81	10.00	3	8.0	0	16.0	4	11.0	4	85	1	2.03	3
83	8.80	3			15.2	4	11.8	3	100	4	2.20	4
85	7.60	0			< 10	NR	12.6	2	103	4	2.50	0
86	9.20	4	10.5	3	14.7	4	10.9	4	103	4	2.21	4
87	12.00	0			14.4	3	8.3	1	128	0	5.15	0
89	9.46	4	10.0	4	17.5	1	10.2	3	126	0	2.01	3
91									96	3		
96	8.90	3			16.2	3	12.9	2	112	2		
97	9.20	4	9.7	4	16.2	3	10.2	3	102	4	2.14	4
102	9.90	3	11.0	2	16.0	4	11.0	4	111	2	2.00	3
104												
105	9.86	3	9.6	4	15.4	4	11.4	4	103	4	2.08	4
107							10.2	3	90	2	2.28	3
109									97	4	2.20	4
113	9.71	4			15.6	4	12.2	3	102	4	2.12	4
114	< 10	NR			20.0	0	< 10	NR	100	4	2.00	3
118	10.60	1			16.5	3	10.9	4				
119	9.20	4	9.4	3	13.0	1	10.0	3	104	4	1.60	0
121	8.00	1	10.0	4			11.0	4	101	4	2.15	4
128	10.10	3	11.0	2	13.0	1	9.3	2	91	2	1.99	3
129							30.0	0	78	0	2.10	4
132	8.50	2	9.5	3	15.5	4	29.0	0	129	0	2.29	3

Table 6. Laboratory performance ratings for standard reference water sample T-145 (trace constituents)—Continued<sup>a</sup>

(MPV, most probable value; µg/L, micrograms per liter; mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/28, number of reported values of 28 possible values; RV, reported value; <, less than)

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Questionable)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Poor)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Analyte = Cd (Cadmium)	Co (Cobalt)	Cr (Chromium)	Cu (Copper)	Fe (Iron)	K (Potassium)	Li (Lithium)
MPV = 9.33 µg/L	10.0 µg/L	15.3 µg/L	11.0 µg/L	101 µg/L	2.13 mg/L	27.3 µg/L
F-pseudosigma = 0.82	0.9	1.4	1.4	8	0.16	2.5
Lab	RV	Rating	RV	Rating	RV	Rating
133	10.60	1		15.3	4	8.9
134	9.26	4	10.2	4	14.6	4
138	9.08	4	8.6	1	10.9	0
140	9.00	4			16.0	4
141	8.98	4	< 10	NR	12.0	0
142	9.82	3	9.7	4	13.9	3
144					10.1	3
145	11.00	1	14.6	0	19.3	0
146	9.03	4	9.2	3	14.5	3
149	10.00	3			17.0	2
151	9.39	4			14.5	3
155					11.8	3
158	4.60	0	10.6	3	13.1	1
180	10.00	3	9.7	4	14.3	3
183					10.6	4
190	10.30	2			19.1	0
191	9.31	4	10.8	3	15.4	4
193	11.00	1			15.0	4
196	9.39	4	9.4	3	< 25	NR
203					11.2	4
204					7.7	0
212	11.00	1	10.0	4	19.8	0
213	8.19	2	10.6	3	12.0	0
215	10.00	3	9.0	2	15.4	4
217	7.10	0	9.5	3	14.0	0
218					12.7	1
219	8.80	3	9.3	3	13.0	1
220					9.4	2
221	9.63	4	10.0	4	15.0	4
224	5.90	0	10.1	4	13.2	1
234	9.60	4	10.6	3	16.9	2
235	9.80	3	13.0	0	12.4	3
236	7.40	0	5.6	0	8.5	1
240	12.00	0	5.8	0	8.8	1
241	9.80	3			14.7	4
244					12.0	3
245	9.35	4			14.9	0
249	8.00	1			14.4	3
253					13.0	1
255	9.87	3	10.5	3	20.1	0
257	16.00	0	19.0	0	10.3	4
259	6.00	0	9.0	2	30.0	0
261					11.0	4
265	10.00	3	9.5	3	13.0	2
268					16.0	4
270					11.8	3
271					97	3
272					2.09	4
273	17.20	0	11.0	2	2.55	0
274	6.90	0			5.0	0
					2.6	0
					96	3
					147	0
					2.98	0
					2.68	0

Table 6. Laboratory performance ratings for standard reference water sample T-145 (trace constituents)—Continued

(MPV, most probable value;  $\mu\text{g/L}$ , micrograms per liter;  $\text{mg/L}$ , milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/28, number of reported values of 28 possible values; RV, reported value; <, less than)

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Questionable)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Poor)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Analyte = Mg (Magnesium)	Mn (Manganese)	Mo (Molybdenum)	Na (Sodium)	Ni (Nickel)	Pb (Lead)	Sb (Antimony)
MPV = 8.68 mg/L	20.9 $\mu\text{g/L}$	9.23 $\mu\text{g/L}$	41.2 mg/L	11.0 $\mu\text{g/L}$	12.7 $\mu\text{g/L}$	8.80 $\mu\text{g/L}$
$\sigma$ -pseudosigma = 0.45	1.5	1.29	1.9	1.3	1.2	0.96
Lab	RV	Rating	RV	Rating	RV	Rating
1	8.17	2	20.3	4	8.67	4
3	8.52	4	17.0	0	6.00	0
4	8.70	4	21.0	4		
5	8.10	2	21.0	4	< 10	NR
7	9.06	3	21.4	4	< 15	NR
10			25.0	0		
11	9.20	2	21.0	4	10.00	3
13	9.07	3	19.5	3		
15	8.10	2	24.0	0	20.10	0
16	8.60	4	21.5	4	8.90	4
18	8.60	4	20.3	4	< 20	NR
19	8.80	4			41.9	4
23	8.72	4	21.8	3	< 100	NR
24	8.44	3	19.8	3		
25	8.40	3	21.0	4		
26	9.10	3	21.5	4	9.90	3
28	8.90	4	23.3	1	6.40	0
30	9.00	3	21.0	4	9.40	4
32	9.11	3	20.9	4	8.18	3
33	8.95	3	20.0	3		
34						
35						
36	7.60	0	22.0	3		
40	8.48	4	18.1	1	9.90	3
42	9.30	2	23.0	2	10.00	3
43	8.80	4	20.0	3		
46	8.67	4	20.4	4		
48	8.89	4	20.0	3	9.30	4
51	16.47	0				
58		< 50	NR			
59	8.50	4	21.0	4		
68	9.05	3	21.5	4	7.30	2
69	8.37	3	< 20	NR		
70	8.73	4	20.6	4	< 50	NR
73			21.0	4		
75	8.90	4	20.8	4	< 10	NR
76						
80			18.5	1		
81	8.58	4	20.0	3		
83	8.45	4	20.3	4		
85	8.92	3				
86	8.67	4	20.3	4	17.70	0
87	8.36	3	27.0	0	11.00	2
89	8.29	3	20.4	4		
91			20.3	4		
96			27.0	0		
97	8.55	4	24.2	0	9.58	4
102	10.30	0	23.0	2		
104						
105	8.68	4	19.6	3	9.38	4
107	8.91	3	20.0	3		
109	8.50	4	19.3	2	8.20	3
113	8.84	4	21.5	4		
114	8.00	1	21.0	4		
118						
119	9.10	3	21.0	4	8.33	3
121	8.30	3	22.0	3		
128	8.28	3	18.7	2	5.50	0
129	9.00	3	10.0	0		
132	8.79	4	19.0	2	12.00	0

Table 6. Laboratory performance ratings for standard reference water sample T-145 (trace constituents)—Continued

(MPV, most probable value; µg/L, micrograms per liter; mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/28, number of reported values of 28 possible values; RV, reported value; <, less than)

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Questionable)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Poor)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Analyte = Mg (Magnesium)	MPV = 8.68 mg/L	Mn (Manganese)	MPV = 20.9 µg/L	Mo (Molybdenum)	MPV = 9.23 µg/L	Na (Sodium)	MPV = 41.2 mg/L	Ni (Nickel)	MPV = 11.0 µg/L	Pb (Lead)	MPV = 12.7 µg/L	Sb (Antimony)	MPV = 8.80 µg/L
F-pseudosigma = 0.45													
133	8.39	3						9.9	3	21.7	0		
134	8.57	4	21.1	4	8.68	4	41.2	4	10.5	4	12.6	4	8.58
138	8.88	4	20.5	4	8.39	3	41.6	4	9.4	2	12.0	3	8.80
140	9.50	1	19.0	2			48.5	0	10.0	3	10.0	0	
141	9.13	2	20.9	4	< 10	NR	43.0	3	< 20	NR	13.6	3	5.13
142	8.18	2	22.0	3	9.31	4	41.8	4	11.0	4	12.7	4	11.30
144													
145	9.28	2	23.3	1	7.60	2	44.2	1	16.4	0	115.0	0	
146	8.20	2	20.6	4	< 10	NR	42.4	3	< 40	NR	15.4	0	< 20
149			22.0	3	10.00	3			10.0	3	12.0	3	
151								11.6	4	12.8	4	9.04	4
155													
158	9.50	1	22.7	2				12.5	2	6.6	0		
180	8.85	4	20.5	4	10.50	3	41.7	4	11.1	4	< 27.2	NR	< 31.4
183			25.2	0									
190	8.63	4	22.4	2			41.4	4	11.2	4	11.1	2	
191	8.99	3	22.0	3	9.30	4	40.1	3	11.9	3	12.8	4	
193	8.20	2					39.2	2	< 50	NR	14.0	2	< 10
196			19.1	2	8.54	3			10.6	4	12.6	4	8.21
203	8.54	4	20.0	3			39.4	3					
204			16.0	0			39.4	3			12.3	4	
212	9.40	1	22.0	3	8.40	3	44.4	1	14.0	0	18.0	0	8.60
213									10.3	3	13.6	3	
215	8.90	4	19.7	3	8.40	3	42.4	3	12.0	3	13.5	3	14.00
217	8.09	2	19.8	3	7.20	1	40.9	4			12.5	4	8.69
218	8.45	4					38.0	1					
219	8.60	4			5.00	0	40.0	3	9.7	3			
220	9.00	3	20.4	4			38.0	1			12.1	3	
221	8.35	3	22.4	2	8.19	3	41.2	4	10.2	3	12.8	4	
224	8.26	3	22.0	3	13.00	0	36.1	0	10.9	4	11.0	2	
234	8.51	4	20.9	4	10.70	2	39.2	2	13.3	1	14.4	2	10.50
235	9.60	0	20.0	3	8.00	3	48.0	0	12.0	3			
236	9.00	3	20.5	4	< 11	NR	40.3	4	10.2	3	5.6	0	26.10
240	8.50	4	19.0	2					23.0	0	27.0	0	14.00
241	7.40	0	23.0	2	11.40	1	37.0	0	10.5	4	16.4	0	8.30
244			30.7	0									
245			25.0	0	12.20	0			10.7	4	14.6	1	
249							42.1	4	8.7	1	13.7	3	
253													
255	8.62	4	20.8	4	9.23	4	40.2	3	11.7	3	13.1	4	11.93
257	7.40	0	17.0	0	72.00	0	40.0	3	14.0	0	49.0	0	2.00
259							42.0	4	16.0	0	10.0	0	
261	10.80	0					45.9	0					
265	8.70	4	22.6	2	9.00	4	40.6	4	10.0	3	12.4	4	9.00
268	8.65	4					41.0	4					
270							49.4	0					
271	25.15	0					40.0	3					
272	6.06	0					50.0	0					
273	9.60	0	25.1	0			44.9	1	12.0	3	7.0	0	
274	0.00	0	8.0	0			22.4	0			1.3	0	

Table 6. Laboratory performance ratings for standard reference water sample T-145 (trace constituents)—Continued

(MPV, most probable value; µg/L, micrograms per liter; mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/28, number of reported values of 28 possible values; RV, reported value; <, less than)

		Absolute Z-value	Rating	Absolute Z-value						V (Vanadium)	Zn (Zinc)
		4 (Excellent)	0.00 - 0.50	1 (Questionable)	1.51 - 2.00						
		3 (Good)	0.51 - 1.00	0 (Poor)	greater than 2.00						
		2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)							
Analyte = Se (Selenium)		SiO <sub>2</sub> (Silica)	Sr (Strontium)	Tl (Thallium)	U (Uranium)	V (Vanadium)	Zn (Zinc)				
MPV = 10.1 µg/L	F-pseudosigma = 1.3	11.3 mg/L	203 µg/L	15.3 µg/L	1.10 µg/L	11.7 µg/L	10.0 µg/L				
Lab	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV
1	10.0	4	10.6	2	198	3	16.7	3	12.0	4	11.4
3	< 10	NR	5.7	0	207	4	< 10	NR	6.0	0	7.0
4			11.4	4	205	4					14.0
5	10.0	4	11.1	4	205	4			12.4	4	9.7
7	< 50	NR	11.6	4	203	4	< 120	NR	12.6	3	11.0
10	9.0	3									16.0
11	12.6	1	8.9	0	202	4	20.0	1	11.0	4	
13	8.2	2	11.2	4			15.8	4	< 50	NR	9.9
15	12.1	1	12.3	2			11.6	2	< 10	NR	< 5
16	10.0	4			184	0	16.0	4	15.0	1	9.4
18	10.1	4			200	4	13.8	3	11.2	4	< 100
19											NR
23	9.3	3				< 5	0				< 20
24			11.6	4	204	4					10.2
25			12.7	1	215	2					
26	8.8	3	11.5	4					12.5	4	9.2
28					207	4			14.6	1	6.0
30	11.0	3					1.20	0	11.4	4	9.6
32	10.3	4	11.4	4	203	4	15.6	4	11.8	4	10.0
33			10.9	3	200	4					
34	10.2	4									
35											
36	8.0	1				8.0	0				16.0
40					169	0			10.4	3	
42	13.0	0	12.1	2	201	4	9.0	0	15.0	1	10.0
43			11.0	4							
46	9.5	4				16.2	4		10.4	3	
48	8.6	2				12.1	2		9.5	2	< 5
51											
58	< 10	NR									< 50
59	11.0	3			200	4	15.5	4			10.0
68	6.9	0			210	3			9.9	2	10.5
69	10.2	4				16.4	4				< 50
70	11.0	3	10.9	3	206	4	13.3	3	< 50	NR	< 20
73	33.0	0									13.0
75	10.1	4					< 100	NR	11.5	4	11.0
76			12.2	2		11.2	1				
80	9.3	3									4.8
81			11.5	4	175	0	84.0	0	9.0	1	8.0
83			10.4	2							9.8
85					214	2					
86	11.6	2			196	3			13.1	3	11.8
87	19.2	0	11.8	3							16.3
89	9.3	3	10.7	3		< 10	NR		14.2	2	13.8
91											
96	10.0	4									11.0
97	8.8	3	11.3	4	169	0	16.7	3	14.0	2	< 4.6
102	13.0	0			239	0	19.0	2	10.0	3	8.8
104			10.7	3							
105	11.3	3	10.9	3	194	2			11.1	4	9.0
107			11.9	3							
109					183	0					
113	9.5	4	11.0	4	203	4	13.1	3			8.5
114											< 10
118	9.6	4	11.4	4							15.0
119	10.0	4	12.0	2		15.9	4	1.07	3	10.3	3
121			11.3	4	195	3					12.0
128	11.2	3	11.8	3		15.3	4		10.2	3	10.1
129			11.5	4							
132											9.0

Table 6. Laboratory performance ratings for standard reference water sample T-145 (trace constituents)—Continued

(MPV, most probable value; µg/L, micrograms per liter; mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/28, number of reported values of 28 possible values; RV, reported value; <, less than)

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Questionable)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Poor)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Analyte =	Se (Selenium)	SiO <sub>2</sub> (Silica)	Sr (Strontium)	Tl (Thallium)	U (Uranium)	V (Vanadium)	Zn (Zinc)
MPV =	10.1 µg/L	11.3 mg/L	203 µg/L	15.3 µg/L	1.10 µg/L	11.7 µg/L	10.0 µg/L
F-pseudosigma =	1.3	0.7	9	2.7	0.04	1.7	2.4
Lab	RV	Rating	RV	Rating	RV	Rating	RV
133	10.8	3					10.1
134	10.7	4	11.2	4	196	3	11.7
138	9.6	4	10.8	3	198	3	8.0
140			11.2	4			14.0
141	9.3	3			< 50	NR	9.9
142	10.8	3	12.8	0	212	2	8.4
144	10.0	4			16.7	3	13.0
145			12.5	1	216	1	17.0
146	< 10	NR			14.0	4	< 20
149	10.0	4					
151	11.2	3			208	3	8.9
155					16.0	4	
158			12.4	1			7.2
180	< 50.1	NR			< 32.1	NR	9.5
183							
190	11.0	3	11.9	3	390	0	13.3
191	12.0	2	11.1	4	203	4	8.5
193	8.0	1			15.8	4	< 50
196	11.1	3			9.0	0	NF
203			10.8	3	194	2	14.5
204			11.4	4			12.0
212	11.0	3	11.8	3	210	3	7.2
213					16.0	4	12.0
215	5.0	0	11.3	4	12.4	2	14.0
217	10.2	4	9.9	1	< 7	0	0.0
					15.2	4	
218					1.10	3	10.2
219			10.8	3	229	0	11.0
220	8.6	2					10.0
221	9.7	4					11.3
224	10.5	4					10.2
234	9.6	4	11.0	4	206	4	6.7
235			12.6	1	208	3	8.3
236	128.3	0	7.0	0	7.9	0	8.6
240	18.0	0	10.5	2	199	4	12.1
241	7.0	0	11.8	3	10.0	1	11.0
					15.9	4	5.0
244							0
245							9.3
249							0.6
253							20.0
255	8.1	2					0
257						11.6	4
259	8.7	2				39.0	0
261						9.0	4
265	12.4	1	10.1	1	200	4	66.0
268					15.0	4	0
270					1.10	4	10.5
271					12.0	4	4
272							20.0
273			227.0	0			5.8
274			19.2	0			1

Table 7. Laboratory performance ratings for standard reference water sample M-140 (major constituents)

(MPV, most probable value; mg/L, micrograms per liter; mg/L, milligrams per liter;  $\mu\text{S}/\text{cm}$ , microsiemens per centimeter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/16, number of reported values of 16 possible values; RV, reported value; <, less than)

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Questionable)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Poor)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Analyte = Alkalinity MPV = 114 mg/L				B (Boron) MPV = 41.6 $\mu\text{g}/\text{L}$				Ca (Calcium) MPV = 60.7 mg/L				Cl (Chloride) MPV = 25.8 mg/L				DSRD	
Lab	OLR	V/16	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	16	382 mg/L	
1	3.5	16	115	4	43.6	4	56.3	2	24.7	3	370	3					
2	1.5	10	106	0					27.7	2							
3	1.9	16	116	3	43.0	4	67.5	0	26.7	3	370	3					
4	4.0	2							26.0	4							
5	3.4	13	113	4	44.5	3	59.8	4	24.8	3	386	4					
7	2.2	6	117	3					24.1	2							
10	3.5	13	116	3	60.0	0	59.2	4	26.3	4	386	4					
11	1.7	12	116	3			2.9	0	26.7	3	358	1					
13	2.7	12	108	0			65.4	1	24.8	3	388	4					
15	1.8	12	96	0	< 50	NR	53.2	0	26.4	4	367	3					
16	2.5	15	113	3	61.6	0	58.6	3	28.9	0	372	3					
18	2.9	14	108	0	< 50	NR	60.1	4	27.0	3	397	3					
19	3.0	10	115	4			60.3	4	25.0	3	374	4					
22	4.0	1															
23	2.7	14	111	2			65.3	2	26.2	4	371	3					
24	3.7	13	115	4	38.2	3	59.8	4	25.6	4							
25	2.5	12	116	3			64.4	2	24.6	3	368	3					
26	3.4	13	114	4	35.3	2	61.1	4	26.2	4	378	4					
28	2.3	7			35.0	2	61.5	4									
30	3.2	5					63.0	3	25.2	4							
32	2.9	14	116	3			61.3	4	26.9	3	372	3					
33	3.4	11	113	3			59.2	4	24.3	2							
36	1.8	11	115	4			56.0	2	26.0	4	410	1					
38	3.3	10	116	4			58.1	3			378	4					
39	2.6	7	100	0					30.0	0							
40	2.9	13	113	4			53.6	0	26.9	3	378	4					
42	1.8	12	140	0	21.0	0	63.3	3	35.8	0							
43	3.7	11	115	4			61.0	4	26.0	4	386	4					
46	2.6	12	113	4	21.5	0	58.8	3	26.6	3							
48	2.2	12	112	3	20.0	0	62.2	4	24.0	2							
50	3.2	13	118	2	51.0	1	63.0	3	25.0	3	382	4					
51	2.8	8	116	3					27.8	2							
55	2.6	12	108	0			59.2	4	25.8	4	380	4					
56	2.3	9	119	2			61.8	4	23.5	1							
57	2.2	13	120	1	< 100	NR	63.0	3	25.0	3	390	4					
59	2.3	12	117	3			68.0	0	24.0	2	371	3					
64	3.4	7							26.3	4							
68	2.5	13	116	3	145.0	0	63.0	3	26.5	4							
69	2.6	10	118	2			58.8	3	26.2	4	401	2					
70	3.5	13	115	4	< 50	NR	61.4	4	25.8	4	385	4					
75	3.7	10	114	4	40.9	4	58.2	3	26.1	4	398	3					
76	2.6	5	108	0					27.9	2	398	3					
80	1.8	12	112	3			66.0	1	25.0	3	411	1					
81	3.4	14	115	4			59.9	4	23.9	2	424	0					
83	3.0	9	113	3			58.4	3	25.2	4							
85	3.1	12	111	2	40.6	4	62.1	4	25.0	3	378	4					
86	3.3	9			44.1	4	62.6	3									
87	1.8	12	120	1			56.0	2	29.0	0	348	0					
89	3.0	13	114	4			58.3	3	25.0	3	394	3					
90	1.0	4	103	0			58.4	3			410	1					
92	3.8	6	113	4							374	4					
96	3.3	7	115	4					24.9	3	402	2					
97	2.7	14	117	3			57.9	3	25.7	4	406	1					
102	1.5	10					60.4	4	25.5	4							
105	2.9	14	111	2			61.0	4	25.2	4	365	2					
107	3.5	11	114	4			58.0	3	27.2	3							
109	2.8	11	122	0			58.7	3	23.3	1	387	4					
113	2.9	14	115	4			66.0	1	24.9	3	396	3					
114	3.6	8	115	4					25.0	3	375	4					
118	2.7	6	115	4							398	3					
119	3.2	14	110	2	43.0	4	61.1	4	24.7	3	370	3					
121	4.0	6					60.0	4									
127	3.4	14	116	3	39.4	4			26.4	4	387	4					
128	3.3	12	116	3	33.1	1	62.0	4	25.4	4							
129	2.6	14	118	2	55.0	0	62.0	4	25.0	3	379	4					

Table 7. Laboratory performance ratings for standard reference water sample M-140 (major constituents)—Continued

(MPV, most probable value; mg/L, micrograms per liter; mg/L, milligrams per liter;  $\mu\text{S}/\text{cm}$ , microsiemens per centimeter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/16, number of reported values of 16 possible values; RV, reported value; <, less than)

Rating	Absolute Z-value	Rating	Absolute Z-value	B (Boron)	Ca (Calcium)	Cl (Chloride)	DSRD
4 (Excellent)	0.00 - 0.50	1 (Questionable)	1.51 - 2.00	41.6 $\mu\text{g}/\text{L}$	60.7 mg/L	25.8 mg/L	382 mg/L
3 (Good)	0.51 - 1.00	0 (Poor)	greater than 2.00				
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)					
	Analyte = Alkalinity			B (Boron)	Ca (Calcium)	Cl (Chloride)	DSRD
	MPV = 114 mg/L			41.6 $\mu\text{g}/\text{L}$	60.7 mg/L	25.8 mg/L	382 mg/L
	F-pseudosigma = 3			5.2	3.1	1.4	16
Lab	OLR	V/16	RV	Rating	RV	Rating	RV
132	3.1	9	114	4	42.5	4	58.8
133	2.8	6	115	4			59.7
134	3.3	15	136	0	42.3	4	61.1
138	3.8	15	114	4	37.6	3	62.0
140	2.5	11					62.5
141	3.3	12	116	3	37.5	3	63.3
142	2.8	16	116	3	41.0	4	59.0
143	3.6	5	113	4			25.2
145	2.9	15	99	0	48.3	2	62.8
146	2.8	11	114	4			24.9
149	3.2	6	114	4			56.9
151	3.0	13	114	4			27.0
155	3.3	8	117	3			61.2
158	3.5	8	111	2			60.7
180	2.8	12	111	2	44.5	3	60.4
183	1.7	3					28.7
190	2.4	14	115	4			58.6
191	3.0	11	118	2			59.1
193	3.0	3	114	4			26.1
196	2.7	3					24.6
203	2.2	6	105	0			23.2
204	3.0	8	111	2			24.1
212	2.7	16	112	3	45.0	3	59.8
213	3.0	4	112	3			62.9
215	2.8	14	112	3	40.0	4	26.2
217	1.6	14	117	3	40.6	4	62.0
218	1.6	8	106	0			24.0
219	2.7	9			44.0	4	55.7
220	3.3	7	112	3			58.0
221	3.1	8					60.0
224	1.6	13	104	0			66.4
234	3.5	16	112	3	39.0	3	65.9
235	1.1	9			38.0	3	60.6
236	2.9	15	114	4	40.4	4	66.5
240	3.0	12	112	3	58.7	0	59.6
241	2.3	14	116	4			59.4
243	2.7	3					27.5
244	4.0	3	115	4			22.0
247	2.3	6	104	0			25.7
249	1.8	9	107	0			20.0
253	2.0	5			65.2	2	30.6
255	3.1	14	113	4	22.4	0	24.2
256	1.6	9	107	0	< 20	NR	26.3
257	2.0	13	115	4			63.2
258	1.4	11	124	0			28.2
259	3.1	15	114	4	85.0	0	68.0
261	1.0	9	120	1			67.0
262	1.9	11	108	0	22.0	0	59.6
265	3.3	13	136	0	43.0	4	64.6
268	2.0	9	118	2			63.0
270	0.0	3					23.5
271	1.1	8					61.0
272	0.4	9	310	0			25.7
273	1.1	12	124	0	70.0	0	40.4
274	0.4	12	229	0			42.5
276	0.7	6	117	2			27.0
					53.9	0	41.3
					63.9	2	35.5
							0

Table 7. Laboratory performance ratings for standard reference water sample M-140 (major constituents)—Continued

(MPV, most probable value; mg/L, micrograms per liter; mg/L, milligrams per liter;  $\mu\text{S}/\text{cm}$ , microsiemens per centimeter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/16, number of reported values of 16 possible values; RV, reported value; <, less than)

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Questionable)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Poor)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Analyte = F (Fluoride)		K (Potassium)		Mg (Magnesium)		Na (Sodium)		(total Phosphorus) as P	
MPV =	0.530 mg/L	2.58 mg/L	18.0 mg/L	39.0 mg/L	0.032 mg/L	0.011			
F-pseudosigma =	0.037	0.14							
Lab	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV
1	0.560	3	2.57	4	17.4	3	38.2	4	0.030
2	0.590	1	2.10	0	17.0	2	35.0	0	0.052
3	0.599	1	2.43	2	20.6	0	34.5	0	0.030
4			2.37	2	17.1	3	39.6	4	
5									
7	0.680	0							
10	0.520	4	2.58	4	18.4	4	38.8	4	
11	0.500	3			0.5	0	1.5	0	0.045
13	0.510	3	2.30	1	19.0	2	41.2	2	
15	0.521	4	2.36	1	16.0	0	34.5	0	< 0.05 NR
16	0.544	4	2.50	3	17.8	4	38.5	4	0.051
18	0.510	3	2.20	0	18.0	4	39.0	4	0.027
19			2.60	4	18.4	4	41.2	2	
22									0.032
23	0.556	3	2.68	3	18.0	4	34.8	0	0.034
24	0.560	3	2.45	3	18.1	4	38.8	4	
25	0.510	3			19.2	2	42.7	1	
26	0.570	2	2.68	3	19.2	2	40.9	3	
28			2.32	1	18.2	4	42.7	1	
30					17.0	2			
32	0.431	0	2.53	4	20.0	0	39.5	4	0.050
33			2.62	4	17.8	4	38.2	4	
36	0.467	1	2.10	0	17.0	2	35.0	0	
38			2.62	4	18.1	4	34.4	0	0.032
39	0.540	4							0.034
40	0.506	3	2.64	4	16.8	2	37.1	3	
42			2.70	3	19.3	2	39.2	4	
43			2.60	4	19.0	2	40.0	3	
46	0.523	4	2.25	0	17.2	3	39.6	4	
48	0.690	0	2.58	4	18.3	4	39.8	4	0.070
50	0.530	4	2.50	3	18.0	4	40.0	3	
51			2.56	4	11.0	0	38.8	4	
55	0.520	4			19.4	2			0.024
56			3.02	0	19.3	2	50.8	0	
57	0.520	4	3.20	0	20.0	0	36.0	1	0.040
59	0.550	3	< 5	NR	20.5	0	44.5	0	0.040
64			2.68	3			37.9	3	0.030
68			2.90	0	18.0	4	40.5	3	0.052
69	0.580	2	2.82	1	17.5	3	38.6	4	
70	0.580	2	2.48	3	18.1	4	39.4	4	< 0.1 NR
75					18.4	4	38.4	4	
76	0.542	4							
80	0.510	3	2.00	0	17.0	2	37.0	2	
81	0.538	4	2.53	4	18.2	4	39.1	4	0.028
83	0.650	0	2.60	4	17.2	3	38.1	4	< 100 NR
85			2.95	0	18.3	4	39.8	4	
86			2.64	4	18.1	4	41.2	2	
87			2.47	3	17.2	3	37.3	3	0.028
89	0.084	0	2.38	2	18.2	4	38.2	4	0.029
90									
92								0.027	4
96	0.525	4							
97	0.495	3	2.60	4	18.0	4	40.1	3	< 0.002 0
102			2.30	1	19.9	1	33.2	0	0.007 0
105	0.550	3	2.61	4	18.8	3	40.8	3	0.019 2
107	0.527	4	2.75	2	17.7	4	39.6	4	0.028 4
109	0.590	1	2.56	4	17.6	4	38.6	4	
113	0.525	4	2.77	2	19.9	1	35.9	1	0.028 4
114	0.530	4						0.031 4	
118								0.020 2	
119	0.530	4	2.50	3	18.6	3	39.9	4	0.030 4
121			2.56	4	17.7	4	38.9	4	
127	0.476	2	2.53	4	18.6	3	39.0	4	< 0.01 NR
128	0.520	4	2.58	4	17.2	3	38.1	4	
129	0.645	0	2.50	3	18.0	4	37.0	2	0.027 4

Table 7. Laboratory performance ratings for standard reference water sample M-140 (major constituents)—Continued

(MPV, most probable value; mg/L, micrograms per liter; mg/L, milligrams per liter;  $\mu\text{S}/\text{cm}$ , microsiemens per centimeter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/16, number of reported values of 16 possible values; RV, reported value; <, less than)

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Questionable)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Poor)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Analyte = F (Fluoride)	K (Potassium)	Mg (Magnesium)	Na (Sodium)	(total Phosphorus) as P
MPV = 0.530 mg/L	2.58 mg/L	18.0 mg/L	39.0 mg/L	0.032 mg/L
F-pseudosigma = 0.037	0.14	1.0	1.9	0.011
Lab	RV	Rating	RV	Rating
132			2.66	3
133				16.9
134	0.590	1	2.54	4
138	0.531	4	2.59	4
140	0.441	0	2.58	4
141	0.570	2	2.57	4
142	0.603	1	2.30	1
143			17.8	4
145	0.460	1	2.54	4
146	0.557	3	3.22	0
149	0.530	4		
151	0.510	3	2.68	3
155				18.3
158	0.540	4		
180			2.42	2
183				18.4
190	0.521	4	2.58	4
191			2.56	4
193				18.4
196	0.525	4		
203				
204				1.1
212	0.510	3	2.50	3
213				19.7
215	0.500	3	2.80	1
217	0.420	0	4.20	0
218			2.41	2
219			2.80	1
220			2.60	4
221			2.61	4
224	0.920	0	3.10	0
234	0.485	2	2.64	4
235			4.15	0
236			2.31	1
240	0.535	4		
241	0.590	1	2.56	4
243				
244				
247	0.510	3		
249			2.68	3
253				
255	0.556	3	2.62	4
256			3.91	0
257	0.560	3	3.30	0
258	0.510	3	2.75	2
259	0.500	3	2.50	3
261			2.35	1
262	0.502	3	2.83	1
265	0.520	4	2.52	4
268			2.85	1
270			3.21	0
271			2.50	3
272	2.000	0	3.00	0
273	0.430	0	2.80	1
274	0.760	0	3.08	0
276				16.0

Table 7. Laboratory performance ratings for standard reference water sample M-140 (major constituents)—Continued

(MPV, most probable value; mg/L, micrograms per liter; mg/L, milligrams per liter;  $\mu\text{S}/\text{cm}$ , microsiemens per centimeter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/16, number of reported values of 16 possible values; RV, reported value; <, less than)

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Questionable)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Poor)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Analyte = pH	SiO <sub>2</sub> (Silica)	SO <sub>4</sub> (Sulfate)	Sp Cond	Sr (Strontium)	V (Vanadium)
MPV =	7.35 mg/L	150 mg/L	600 $\mu\text{S}/\text{cm}$	671 $\mu\text{g}/\text{L}$	3.42 $\mu\text{g}/\text{L}$
F-pseudosigma =	0.17	0.46	7	19	3.01
Lab	RV	Rating	RV	Rating	RV
1	8.21	4	6.89	3	150
2	8.04	2			152
3	7.83	0	8.43	0	142
4					152
5	8.17	3	7.08	3	151
7	8.20	4			155
10	8.42	3	7.30	4	151
11	8.26	4			5
13	8.36	4	7.37	4	150
15	8.11	2	7.97	2	140
16	8.10	2			143
18	8.25	4	7.44	4	148
19	8.55	1			149
22					
23	8.28	4	7.75	3	158
24	8.20	4	7.76	3	150
25	8.36	4	11.04	0	148
26	8.33	4	7.53	4	150
28					
30	8.40	3			150
32	8.40	3	7.50	4	146
33	8.28	4	7.10	3	147
36	8.33	4			160
38	8.40	3	7.29	4	
39	8.40	3			150
40	8.35	4	7.18	4	119
42	7.90	0	8.00	2	197
43	8.21	4	7.50	4	150
46	8.38	3	7.04	3	185
48	8.00	1			104
50	8.31	4	7.30	4	149
51	8.29	4			156
55	8.00	1	7.36	4	133
56	8.35	4			149
57	8.20	4	7.40	4	140
59	8.40	3	7.70	3	148
64	8.38	3			151
68	8.38	3	7.05	3	
69	8.40	3			158
70	8.28	4	7.10	3	146
75	8.27	4			153
76	8.26	4			
80	8.17	3	5.50	0	130
81	8.37	3	7.24	4	145
83			6.76	2	151
85	8.37	3	7.32	4	148
86	8.35	4			
87	7.95	1	7.18	4	127
89	8.39	3	6.80	2	148
90					
92	8.14	3			147
96	8.34	4			145
97	8.37	3	7.37	4	153
102			6.66	2	143
105	8.34	4	7.45	4	140
107	8.49	2	7.21	4	
109	8.42	3			155
113	8.32	4	7.34	4	154
114	8.07	2			152
118	7.90	0	7.47	4	
119	8.50	2	7.00	3	143
121			7.12	4	
127	8.31	4	7.21	4	159
128	8.43	3	7.79	3	155
129	7.99	1	7.95	2	150

Table 7. Laboratory performance ratings for standard reference water sample M-140 (major constituents)—Continued

(MPV, most probable value; mg/L, micrograms per liter; mg/L, milligrams per liter;  $\mu\text{S}/\text{cm}$ , microsiemens per centimeter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/16, number of reported values of 16 possible values; RV, reported value; <, less than)

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Questionable)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Poor)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Analyte = pH	SiO <sub>2</sub> (Silica)		SO <sub>4</sub> (Sulfate)		Sp Cond		Sr (Strontium)		V (Vanadium)	
MPV =	8.28	7.35 mg/L	150 mg/L	600 $\mu\text{S}/\text{cm}$	19	671 $\mu\text{g}/\text{L}$	31	3.42 $\mu\text{g}/\text{L}$	3.01	
F-pseudosigma =	0.17	0.46	7							
Lab	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
132	8.05	2			457	0				
133	8.15	3								
134	8.44	3	7.47	4	152	4	606	4	654	3
138	8.29	4	7.14	4	149	4	597	4	658	4
140	8.33	4	7.40	4	155	3	708	0		
141	8.35	4			156	3	607	4		< 10
142	8.26	4	8.41	0	163	1	604	4	694	3
143	8.38	3								3.42
145	8.30	4	7.95	2	149	4	615	3	699	3
146	8.15	3			155	3	595	4		< 10
149	8.40	3			160	2	612	3		NR
151	8.39	3	5.88	0	151	4	602	4	710	2
155	8.24	4	7.08	3			631	1		
158	8.26	4			152	4	609	4		
180	8.40	3			140	2	620	2		4.95
183	8.33	4					535	0		3
190	8.10	2	7.70	3	177	0	604	4	958	0
191	8.39	3	8.71	0	149	4			1	0
193					149	4	568	1		
196					145	3				
203	8.31	4	6.90	3	168	0	598	4		
204	8.16	3	7.42	4	155	3	594	4		
212	8.30	4	7.90	2	156	3	637	1	660	4
213	8.35	4							2.10	4
215	8.25	4	8.08	1	155	3	574	2		
217	8.40	3	6.08	0	158	2	625	2	578	0
218	8.02	1					556	0	673	4
219			7.50	4	40	0			700	3
220					148	4				
221	8.08	2			146	3				
224	8.17	3			152	4	560	0		8.30
234	8.13	3	7.20	4	153	4	611	3	668	4
235			8.26	1	174	0			709	2
236	8.24	4	4.66	0	152	4	589	3	649	3
240	8.24	4	6.88	2	147	4	576	2		3.80
241	8.30	4	7.68	3	152	4	414	0		4
243	7.91	0					595	4		
244	8.34	4					600	4		
247	8.56	1			143	2	595	4		
249	7.98	1			141	2	617	3		
253	8.20	4			117	0	625	2		
255	8.49	2	7.38	4	120	0	595	4		1.40
256	8.05	2	6.65	1	154	3				39.00
257	8.21	4			195	0	600	4		0
258	8.20	4			200	0	539	0		
259	8.50	2	7.00	3	148	4	609	4	680	4
261	7.95	1			16	0	522	0		
262	7.42	0			150	4	582	3		
265	8.12	3	6.85	2	155	3			658	4
268	8.10	2			154	3	560	0		1.20
270							550	0		3
271	7.83	0			17	0	545	0		
272	8.28	4					645	0		
273	8.20	4					620	2	770	0
274	7.90	0	14.93	0	149	4	634	1		
276	7.70	0								

Table 8. Laboratory performance ratings for standard reference water sample N-51 (nutrient constituents)

(MPV, most probable value; ug/L, micrograms per liter; mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/5, number of reported values of 5 possible values; RV, reported value; <, less than)

Rating	Absolute Z-value	Rating	Absolute Z-value									
4 (Excellent)	0.00 - 0.50	1 (Questionable)	1.51 - 2.00									
3 (Good)	0.51 - 1.00	0 (Poor)	greater than 2.00									
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)										
Analyte = NH <sub>3</sub> as N (Ammonia)		NH <sub>3</sub> + Org N as N (Ammonia+Organic N)										
MPV = 0.07 mg/L		NO <sub>3</sub> + NO <sub>2</sub> as N (Nitrate + Nitrite)										
F-pseudosigma = 0.05		total P as P (total Phosphorus)										
Lab OLR V/5		PO <sub>4</sub> as P (Orthophosphate as P)										
1	3.0	3	0.03	3	0.22	3	< 0.005	NR	0.03	3	< 0.001	NR
2	2.5	4	0.05	4	0.50	0	< 0.02	NR	0.05	4	0.00	2
5	0.5	2	0.15	1			< 0.03	NR			0.04	0
7	2.0	1	0.13	2			< 0.05	NR				
10	3.8	4	0.09	4	0.26	4	< 0.01	NR	0.05	3	0.02	4
11	2.7	3	0.17	1	0.36	3			0.04	4		
13	NR	0	< 0.02	NR			< 0.06	NR	< 0.05	NR	< 0.05	NR
15	0.0	2	< 0.05	NR	0.53	0	< 0.02	NR	0.12	0	< 0.02	NR
16	2.2	5	0.08	4	0.11	1	0.07	1	0.07	1	0.02	4
18	2.8	4	0.14	2	0.17	2	< 0.01	NR	0.03	3	0.02	4
21	3.6	5	0.01	2	0.29	4	0.01	4	0.04	4	0.02	4
22	2.0	1	< 0.1	NR	0.50	0	< 0.05	NR	< 0.1	NR	< 0.1	NR
23	0.0	1	< 0.1	NR							0.02	4
25	3.0	4	0.08	4	0.08	0	0.01	4			0.04	0
28	0.0	2	0.34	0								
33	3.0	1	0.04	3							< 0.01	NR
36	NR	0					< 0.05	NR	< 0.05	NR		
38	3.0	5	0.01	2	0.25	4	0.00	4	0.06	2	0.01	3
39	2.0	3	0.10	3					0.00	0	0.01	3
46	3.0	2							0.03	3	0.01	3
48	3.5	4	0.11	3	0.39	3	< 0.08	NR	0.04	4	0.02	4
53	1.0	3	0.11	3			0.40	0			0.12	0
55	2.8	4	0.12	3	0.30	4			0.04	4	0.08	0
56	3.3	4	0.04	3	0.23	3	< 0.02	NR	0.03	3	0.02	4
58	1.0	4	0.03	3	0.48	1			0.13	0	0.10	0
59	3.3	3	0.05	4	0.20	3	< 0.04	NR	0.03	3	< 0.01	NR
64	3.5	2	0.06	4			< 0.02	NR			0.01	3
68	2.3	4	0.15	1	0.33	4	0.01	4	0.08	0		
69	NR	0					< 0.05	NR				
70	4.0	1	< 0.1	NR	0.29	4	< 0.1	NR	< 0.1	NR	< 0.1	NR
75	NR	0					< 0.1	NR				
80	NR	0	< 0.02	NR			< 0.01	NR			< 0.05	NR
81	3.3	3	< 0.05	NR	0.40	2	0.01	4	0.03	4	< 0.005	NR
83	3.0	1	< 0.01	NR			< 0.02	NR	< 0.075	NR	0.01	3
85	3.4	5	0.03	3	0.20	3	0.04	3	0.04	4	0.02	4
87	2.8	4	0.07	4	0.12	1	< 0.01	NR	0.05	3	0.01	3
88	1.3	3	0.04	4			1.16	0			0.05	0
89	3.0	4	0.02	3	0.15	2	< 0.05	NR	0.04	4	0.01	3
90	4.0	1	0.08	4								
91	3.5	2	< 0.03	NR	0.32	4	< 0.02	NR	0.05	3		
92	3.7	3					0.01	4	0.03	4	0.02	3
96	3.3	3	0.09	4	0.23	3	< 0.05	NR	0.03	3	< 0.01	NR
97	2.8	4	0.10	3	0.33	4	< 0.003	NR	0.01	0	0.01	4
102	2.7	3			0.30	4			0.03	2	0.00	2
104	3.4	5	0.01	2	0.27	4	0.00	4	0.05	4	0.01	3
105	3.3	3	0.08	4	< 0.2	NR	< 0.04	NR	0.02	2	0.02	4
110	4.0	1	0.10	4								
113	4.0	1	< 0.1	NR			< 0.015	NR	0.03	4	< 0.004	NR
114	4.0	1	< 0.1	NR			< 0.04	NR	0.04	4		
118	2.0	4	0.04	3	0.24	4	0.08	1	0.01	0	< 0.01	NR
119	3.0	4	0.03	3	0.36	3	0.01	4	0.06	2	0.00	NR
127	3.5	2	0.02	3	0.31	4	< 0.01	NR	< 0.01	NR	< 0.05	NR
128	3.0	2	0.11	3	0.20	3	< 0.01	NR			< 0.01	NR
129	3.8	4	0.00	NR	0.30	4	0.00	4	0.04	4	0.03	3
132	2.5	4	0.08	4			0.11	0	0.03	3	0.01	3
133	3.0	5	0.08	4	0.24	4	0.18	0	0.04	4	0.01	3
134	3.6	5	0.04	4	0.30	4	0.00	4	0.04	4	0.00	2
138	3.0	4	0.03	3	0.28	4	< 0.005	NR	0.05	3	0.00	2
140	3.7	3	0.07	4	0.21	3	0.00	4	< 0.02	NR	< 0.01	NR
141	1.0	3	0.17	1	< 1	NR	< 0.05	NR	0.06	2	0.05	0

Table 8. Laboratory performance ratings for standard reference water sample N-51 (nutrient constituents)—Continued

(MPV, most probable value; ug/L, micrograms per liter; mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/5, number of reported values of 5 possible values; RV, reported value; <, less than)

Rating	Absolute Z-value	Rating	Absolute Z-value									
4 (Excellent)	0.00 - 0.50	1 (Questionable)	1.51 - 2.00									
3 (Good)	0.51 - 1.00	0 (Poor)	greater than 2.00									
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)										
Analyte = NH <sub>3</sub> as N (Ammonia)		NH <sub>3</sub> + Org N as N (Ammonia+Organic N)										
MPV =	0.07 mg/L	0.29 mg/L	0.01 mg/L									
F-pseudosigma =	0.05	0.10	0.04									
Lab	OLR	V/5	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
142	2.6	5	0.02	3	0.20	3	0.02	4	0.05	3	0.08	0
143	3.6	5	0.01	2	0.29	4	0.03	4	0.05	4	0.01	4
145	3.5	2	< 0.02	NR	0.21	3	< 0.02	NR	0.04	4	< 0.01	NR
146	NR	0	< 0.03	NR			< 0.05	NR	< 0.1	NR	< 0.05	NR
149	2.5	4	0.01	2			0.12	0	0.04	4	0.02	4
151	3.3	3	0.06	4			< 0.05	NR	0.05	3	0.03	3
155	3.2	5	0.10	3	0.21	3	0.00	4	0.03	3	0.01	3
158	3.0	3	0.04	3			< 0.02	NR	0.03	3	0.02	3
180	3.5	2	< 0.02	NR	0.21	3	< 0.01	NR	0.05	4	< 0.01	NR
183	2.0	2							0.05	3	0.03	1
190	1.8	5	0.10	3	0.40	2	0.01	4	0.14	0	0.14	0
191	4.0	1					< 0.01	NR			0.02	4
193	3.0	1					< 0.02	NR	0.03	3		
196	NR	0					< 0.05	NR			< 0.05	NR
197	3.0	2	0.13	2			0.02	4				
203	3.5	4	0.08	4	0.35	3	< 0.02	NR	0.05	4	0.02	3
204	3.0	3	< 0.005	NR	0.19	3	< 0.02	NR	0.03	2	0.01	4
212	3.3	3	< 0.1	NR	0.33	4	< 0.1	NR	0.05	3	0.01	3
213	4.0	1	< 1	NR	< 1	NR			0.04	4	< 0.02	NR
215	2.0	2	< 0.01	NR	0.51	0	< 0.01	NR	0.04	4	< 0.01	NR
220	2.7	3	0.00	2			0.01	4			0.00	2
221	2.8	5	0.04	3	0.48	1	0.02	4	0.03	3	0.01	3
224	2.2	5	0.12	2	0.45	1	0.06	2	0.03	3	0.02	3
234	3.3	3	0.07	4			< 0.01	NR	0.03	3	0.01	3
240	0.7	3	0.19	0	0.47	1	< 0.1	NR	0.07	1	< 0.1	NR
241	3.6	5	0.01	2	0.26	4	0.01	4	0.03	4	0.02	4
243	3.0	2	0.04	3			< 0.01	NR	0.05	3		
247	4.0	1					< 0.01	NR			0.02	4
253	4.0	1			0.29	4						

Table 9. Laboratory performance ratings for standard reference water sample N-52 (nutrient constituents)

(MPV, most probable value; ug/L, micrograms per liter; mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/5, number of reported values of 5 possible values; RV, reported value; <, less than)

Rating	Absolute Z-value	Rating	Absolute Z-value									
4 (Excellent)	0.00 - 0.50	1 (Questionable)	1.51 - 2.00									
3 (Good)	0.51 - 1.00	0 (Poor)	greater than 2.00									
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)										
Analyte = NH <sub>3</sub> as N Ammonia) 0.09												
MPV = 1.33 mg/L												
F-pseudosigma = 0.09												
Lab	OLR	V/5	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
1	3.4	5	1.36	4	2.16	3	1.70	4	1.55	3	1.20	3
3	3.0	5	1.27	3	2.20	3	1.80	3	1.55	3	1.12	3
5	2.0	3	1.38	3			1.30	0			1.12	3
7	1.3	4	1.10	0			1.54	1	1.64	3	1.05	1
10	3.6	5	1.32	4	2.48	3	1.74	4	1.62	4	1.20	3
11	3.2	5	1.42	2	2.58	3	1.71	4	1.58	4	1.13	3
13	3.8	4	1.30	4			1.72	4	1.61	4	1.20	3
15	2.6	5	1.11	0	2.58	3	1.72	4	1.68	2	1.15	4
16	1.6	5	1.25	3	1.94	1	1.73	4	1.97	0	1.45	0
18	2.6	5	1.37	4	1.89	0	1.83	2	1.59	4	1.19	3
19	3.5	4	1.34	4			1.78	3	1.64	3	1.14	4
22	4.0	1							1.59	4		
23	1.3	3	0.93	0					1.43	0	1.14	4
25	2.4	5	1.14	0	2.64	2	1.77	4	1.68	2	1.17	4
26	2.0	2	1.42	2							1.24	2
30	4.0	2					1.74	4			1.16	4
33	2.5	2	1.22	2							1.12	3
36	2.0	5	1.44	2	2.15	3	1.90	1	1.40	0	1.15	4
38	3.6	5	1.33	4	2.23	3	1.70	4	1.57	4	1.13	3
42	0.3	3					28.10	0	1.50	1	0.58	0
46	2.0	5	1.32	4	2.24	3	1.48	0	1.55	3	0.12	0
48	2.0	5	1.20	2	2.30	4	1.61	2	1.50	1	1.05	1
53	2.0	3	1.30	4			1.84	2			3.49	0
55	3.6	5	1.37	4	2.40	4	1.73	4	1.56	3	1.21	3
56	1.0	5	1.50	1	2.74	1	1.79	3	1.10	0	0.66	0
57	1.4	5	1.22	2	3.60	0	1.90	1	1.60	4	1.30	0
58	1.5	4	1.18	1	2.49	3			1.54	2	0.92	0
59	4.0	5	1.35	4	2.30	4	1.74	4	1.60	4	1.16	4
64	4.0	3	1.34	4			1.72	4			1.17	4
68	2.8	4	1.59	0	2.42	4	1.68	4	1.64	3		
69	4.0	1					1.76	4				
70	3.4	5	1.24	3	2.34	4	1.58	2	1.62	4	1.16	4
75	2.0	1					1.61	2				
76	4.0	1	1.29	4								
80	2.0	3	1.30	4			0.86	0			1.24	2
81	2.6	5	1.50	1	2.15	3	1.72	4	1.59	4	1.06	1
83	2.8	4	1.20	2			1.59	2	1.60	4	1.20	3
85	3.2	5	1.25	3	2.43	4	1.55	1	1.61	4	1.15	4
86	3.0	3	1.37	4			1.72	4	1.69	1		
87	2.4	5	1.00	0	2.10	2	1.74	4	1.63	3	1.19	3
88	1.0	3	1.09	0			2.49	0			1.11	3
89	3.8	5	1.32	4	2.22	3	1.72	4	1.62	4	1.16	4
90	3.0	1	1.38	3								
91	3.3	4	1.34	4	2.43	4	1.62	2	1.63	3		
92	3.7	3					1.79	3	1.60	4	1.18	4
96	3.8	5	1.37	4	2.39	4	1.64	3	1.58	4	1.14	4
97	3.2	5	1.34	4	2.23	3	1.65	3	1.52	2	1.16	4
102	2.2	5	1.00	0	2.23	3	1.83	2	1.57	3	1.11	3
104	4.0	5	1.33	4	2.43	4	1.69	4	1.61	4	1.17	4
105	3.0	5	1.36	4	1.50	0	1.76	4	1.65	3	1.17	4
107	3.0	4	1.39	3			1.74	4	1.66	2	1.12	3
108	3.3	4	1.29	4	2.34	4	1.75	4			1.28	1
113	2.6	5	1.38	3	2.14	2	1.50	0	1.60	4	1.15	4
114	1.3	3	1.10	0			2.29	0	1.60	4		
118	2.2	5	1.26	3	2.47	4	1.94	0	1.90	0	1.18	4
119	2.6	5	1.42	2	1.87	0	1.64	3	1.62	4	1.18	4
127	2.4	5	1.68	0	2.66	2	1.71	4	1.54	2	1.16	4
128	3.5	4	1.30	4	2.50	3	1.72	4			1.19	3
129	1.8	5	1.18	1	2.18	3	1.75	4	1.50	1	1.04	0
132	2.3	4	1.32	4			1.53	1	1.63	3	1.25	1

Table 9. Laboratory performance ratings for standard reference water sample N-52 (nutrient constituents)—Continued

(MPV, most probable value; ug/L, micrograms per liter; mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/S, number of reported values of 5 possible values; RV, reported value; <, less than)

Rating	Absolute Z-value	Rating	Absolute Z-value									
4 (Excellent)	0.00 - 0.50	1 (Questionable)	1.51 - 2.00									
3 (Good)	0.51 - 1.00	0 (Poor)	greater than 2.00									
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)										
Analyte = NH <sub>3</sub> as N Ammonia	1.33 0.09	NH <sub>3</sub> + Org N as N (Ammonia+Organic N)	NO <sub>3</sub> + NO <sub>2</sub> as N (Nitrate + Nitrite)	total P as P (total Phosphorus)	PO <sub>4</sub> as P (Orthophosphate as P)							
MPV = 1.33 mg/L	2.37 mg/L	1.72 mg/L	1.60 mg/L	1.16 mg/L								
F-pseudosigma = 0.09	0.22	0.10	0.06	0.06								
Lab	OLR	V/S	RV	Rating	RV	Rating	RV	Rating	RV	Rating		
133	2.4	5	1.35	4	2.05	2	2.16	0	1.63	3	1.12	3
134	3.6	5	1.37	4	2.30	4	1.73	4	1.65	3	1.19	3
138	3.2	5	1.28	3	2.38	4	1.66	3	1.52	2	1.15	4
140	1.8	5	1.52	0	2.35	4	1.66	3	1.66	2	1.01	0
141	3.4	5	1.26	3	2.56	3	1.72	4	1.59	4	1.21	3
142	2.8	5	1.26	3	2.33	4	1.98	0	1.65	3	1.15	4
143	3.4	5	1.34	4	2.53	3	1.85	2	1.60	4	1.15	4
145	3.2	5	1.45	2	2.34	4	1.68	4	1.66	2	1.17	4
146	0.0	4	3.57	0			4.54	0	4.03	0	3.48	0
155	2.6	5	1.29	4	2.12	2	1.66	3	1.52	2	1.08	2
158	3.8	4	1.36	4			1.76	4	1.60	4	1.10	3
180	3.4	5	1.32	4	2.47	4	1.74	4	1.68	2	1.19	3
183	2.0	3					0.15	0	1.58	4	1.10	2
190	1.6	5	1.55	0	2.70	1	1.82	3	1.22	0	1.14	4
191	3.5	2					1.73	4			1.21	3
193	2.5	2					1.63	3	1.67	2		
197	4.0	2	1.34	4			1.74	4				
203	2.4	5	1.41	3	1.44	0	1.83	3	1.63	4	1.22	2
204	2.8	4	1.35	4	2.48	3	1.35	0	1.59	4		
212	2.8	5	1.30	4	2.50	3	16.40	0	1.60	4	1.20	3
213	1.5	4	1.50	1	4.00	0			1.58	4	1.06	1
215	2.2	5	1.35	4	2.51	3	1.72	4	0.82	0	0.54	0
220	1.3	3	1.36	4			2.04	0			1.32	0
221	2.4	5	1.31	4	2.30	4	1.87	2	1.76	0	1.09	2
224	2.0	5	1.40	3	3.70	0	1.70	4	1.55	3	1.31	0
234	3.3	4	1.40	3			1.65	3	1.62	4	1.11	3
240	2.6	5	1.32	4	2.42	4	1.60	2	2.03	0	1.21	3
241	3.4	5	1.35	4	2.15	3	1.81	3	1.56	3	1.16	4
243	3.0	3	1.24	3			1.76	4	1.67	2		
247	3.0	2					1.78	3			1.19	3
248	2.0	4	1.40	3			0.76	0	1.56	3	1.23	2
249	1.4	5	0.74	0	2.57	3	6.71	0	1.75	0	1.18	4
253	1.4	5	1.29	4	3.62	0	0.92	0	1.43	0	1.20	3
255	3.5	4	1.25	< 5	NR		1.72	4	1.58	4	1.13	3

Table 10. Laboratory performance ratings for standard reference water sample P-27 (low ionic strength)

(MPV, most probable value; ug/L, micrograms per liter; mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/11, number of reported values of 11 possible values; RV, reported value; <, less than)

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Questionable)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Poor)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Analyte = Acidity as CaCO <sub>3</sub>				Ca (Calcium)				Cl (Chloride)				F (Fluoride)				K (Potassium)				Mg (Magnesium)			
MPV = 4.74 mg/L				2.53 mg/L				1.20 mg/L				0.100 mg/L				0.336 mg/L				0.461 mg/L			
F-pseudosigma = 3.19				0.24				0.49				0.033				0.038				0.050			
Lab	OLR	V/11	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	
1	3.8	9	2.50	3	2.50	4	1.20	4	< 0.1	NR	0.330	4	0.450	4									
2	2.0	8			2.80	2	1.13	4			0.330	4	0.545	1									
3	2.0	9	< 10	NR	2.40	3	1.78	2	0.063	2	0.330	4	0.420	3									
5	2.3	8	7.42	3	2.41	4	2.39	0			< 1	NR	0.434	3									
7	3.8	5	5.20	4			1.13	4	< 0.5	NR													
15	1.9	9	< 2	NR	2.82	2	1.63	3	0.108	4	0.511	0	0.639	0									
23	3.7	6			2.60	4	1.27	4	< 0.1	NR	0.350	4	< 0.5	NR									
25	3.1	8	27.00	0	2.50	4	1.20	4	0.100	4			0.420	3									
26	3.4	9			2.96	1	1.13	4	0.120	3	0.350	4	0.470	4									
28	1.5	4			2.85	2					0.190	0	0.450	4									
33	3.8	8			2.37	3	1.24	4			0.300	3	0.440	4									
36	2.5	4			2.40	3	< 5	NR	< 0.1	NR	< 0.5	NR	< 0.5	NR									
38	3.6	7	6.85	3	2.47	4					0.320	4	0.437	4									
39	2.2	5					2.00	1	0.006	0													
42	2.4	5			3.30	0	1.10	4			0.400	1	0.500	3									
46	3.3	7			2.44	4	1.65	3			0.330	4	0.442	4									
48	1.9	9			2.53	4	2.00	1	0.240	0	0.380	2	0.500	3									
58	1.9	9			1.70	0	1.49	3	0.092	4	0.350	4	0.320	0									
59	3.5	4			< 5	NR	1.10	4	< 0.2	NR	< 5	NR	< 5	NR									
64	3.9	8			2.50	4	1.20	4			0.320	4	0.450	4									
81	3.4	10	0.50	2	2.42	4	1.15	4	0.069	3	0.336	4	0.433	3									
83	3.2	6	5.10	4	2.50	4	< 5	NR	0.170	0	< 2	NR	0.434	3									
89	3.4	10	2.26	3	2.29	3	1.29	4	0.537	0	0.333	4	0.453	4									
92	3.0	2																					
105	2.9	8	0.42	2	2.58	4	1.14	4	< 0.2	NR	< 0.5	NR	0.387	2									
107	3.6	7			2.39	3	< 0.6	NR	0.077	3	0.340	4	0.450	4									
109	2.6	10	6.57	3	2.58	4	1.01	4	0.090	4	0.270	1	0.490	3									
110	3.3	6			1.45	0	1.13	4							0.450	4							
113	3.0	8			2.86	2	1.98	1	0.090	4	0.288	2	0.504	3									
119	3.1	9			2.57	4	1.18	4	0.100	4	0.000	0	0.410	2									
132	2.1	7	5.08	4	2.71	3					0.550	0	0.450	4									
134	3.9	9			2.60	4	1.17	4	0.120	3	0.343	4	0.447	4									
138	3.9	9			2.60	4	1.07	4	0.101	4	0.330	4	0.460	4									
140	2.2	9			2.39	3	0.87	3	0.053	2	0.289	2	0.467	4									
141	3.5	10	2.40	3	2.74	3	1.40	4	0.130	3	0.351	4	0.486	3									
143	3.7	3					1.20	4															
145	2.8	9			2.67	3	1.04	4	0.080	3	0.290	2	0.490	3									
146	2.8	5	< 10	NR	2.43	4	1.53	3	< 0.2	NR	< 1	NR	< 0.5	NR									
155	2.0	3			2.72	3																	
158	1.3	8			3.16	0	0.27	1			0.350	4	0.550	1									
180	3.9	7			2.62	4	1.00	4			< 1.26	NR	0.469	4									
183	1.7	3					1.01	4															
190	3.1	10	6.00	4	2.19	2	1.09	4	0.073	3	0.319	4	0.427	3									
191	2.2	6			2.32	3	1.83	2			1.110	0	0.470	4									
193	3.0	1																					
196	4.0	3					1.19	4	0.106	4													
197	3.5	2					0.94	3															
203	1.3	3					< 2	NR															
204	2.7	6			2.75	3	1.40	4															
215	2.8	9	3.60	4	2.60	4	2.00	1	0.100	4	< 1	NR	0.460	4									
220	2.9	7	7.21	3	2.50	4	1.18	4			0.300	3	0.400	2									
221	2.4	7			2.52	4	1.47	3			0.329	4	0.461	4									
224	3.2	10	4.01	4	2.59	4	1.36	4	0.450	0	0.332	4	0.519	2									
235	2.0	5			2.75	3					0.460	0	0.490	3									
240	3.1	8	1.88	3	2.65	4	1.13	4	0.076	3			0.430	3									
241	2.2	9			2.30	3	3.50	0	0.116	4	0.370	3	0.300	0									
243	2.0	2																					
244	4.0	2																					
247	2.8	6	0.20	2			1.80	2	0.100	4													
255	3.7	6			2.57	4	< 5	NR	< 0.2	NR	0.367	3	0.470	4									

**Table 10. Laboratory performance ratings for standard reference water sample P-27 (low ionic strength)—Continued**

(MPV, most probable value; ug/L, micrograms per liter; mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/11, number of reported values of 11 possible values; RV, reported value; <, less than)

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Questionable)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Poor)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Analyte = Acidity as CaCO <sub>3</sub>				Ca (Calcium)		Cl (Chloride)		F (Fluoride)		K (Potassium)		Mg (Magnesium)		
	MPV =	4.74 mg/L			2.53 mg/L		1.20 mg/L		0.100 mg/L		0.336 mg/L		0.461 mg/L	
	F-pseudosigma =	3.19			0.24		0.49		0.033		0.038		0.050	
Lab	OLR	V/11	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	
256	1.3	7			2.50	4	1.95	1	0.000	0	1.500	0		
257	2.2	10	1.25	2	2.80	2	1.63	3	0.085	4	0.200	0	0.630	0
258	0.4	9			1.60	0	8.40	0	0.170	0	0.500	0	1.940	0
261	0.9	7			3.26	0	0.00	0	< 0.39	NR	1.950	0		
262	2.4	9			2.50	4	1.30	4	0.082	3	0.350	4	0.440	4
265	2.4	7			2.90	1	0.95	3	< 0.1	NR	0.370	3	0.720	0
268	3.0	8			1.42	0	1.11	4			0.365	3	0.480	4
270	0.3	3			1.06	0					0.410	1		
271	1.1	8			4.28	0	2.00	1			0.900	0	10.370	0
272	1.0	9	15.00	0	2.41	3	7.09	0	0.405	0	0.000	0	0.000	0
273	1.7	9	5.20	4	2.80	2	13.00	0	0.400	0	0.412	1	0.521	2
276	1.8	6	4.40	4	1.80	0	3.50	0					1.400	0

Table 10. Laboratory performance ratings for standard reference water sample P-27 (low ionic strength)—Continued

(MPV, most probable value; ug/L, micrograms per liter; mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/11, number of reported values of 11 possible values; RV, reported value; <, less than)

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Questionable)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Poor)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Analyte = Na (Sodium)		pH		PO <sub>4</sub> as P		SO <sub>4</sub> (Sulfate)		Specific Conductance	
MPV =	1.34 mg/L	6.92		0.002 mg/L	2.42 mg/L	26.1	μS/cm		
F-pseudosigma =	0.10	0.32		0.014	0.37	1.1			
Lab	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV
1	1.34	4	6.99	4	< 0.001	NR	2.43	4	26.8
2	2.02	0	6.33	1			2.32	4	21.3
3	0.78	0	6.22	0	< 0.01	NR	3.40	0	26.6
5	1.34	4	6.21	0			4.37	0	26.3
7			6.98	4	< 0.16	NR	2.40	4	25.3
15	1.76	0	6.23	0	< 0.02	NR	2.50	4	26.1
23	1.43	3	6.93	4	< 0.01	NR	< 2.5	NR	27.0
25	1.36	4	7.13	3					27.0
26	1.32	4	6.95	4	< 0.5	NR	2.46	4	27.2
28	0.94	0			0.002	NR			
33	1.30	4	6.74	3	< 0.01	NR	2.34	4	25.8
36	1.20	2	6.96	4	< 0.025	NR	< 5	NR	23.9
38	1.24	3	7.10	3	0.001	NR			26.1
39			6.90	4	< 0.005	NR	2.30	4	27.5
42	1.30	4			0.002	NR	< 2.5	NR	
46	1.31	4	7.58	0			2.40	4	
48	1.38	4	6.40	1	< 0.005	NR	60.00	0	24.5
58	0.96	0	8.07	0	0.010	NR	2.00	2	26.1
59	< 5	NR	6.69	3			2.30	4	25.4
64	1.29	4	7.01	4	0.001	NR	2.44	4	25.5
81	1.24	3	6.88	4	< 0.005	NR	2.65	3	26.3
83	1.33	4			0.009	NR	2.52	4	
89	1.31	4	7.01	4	< 0.002	NR	2.48	4	25.8
92			7.03	4	< 0.005	NR	1.88	2	
105	1.18	1	6.65	3	0.002	NR	2.28	4	27.0
107	1.34	4	7.15	3	< 0.002	NR			26.0
109	1.36	4	6.17	0			3.70	0	27.0
110	1.32	4	7.03	4			2.34	4	
113	1.30	4	7.05	4	< 0.004	NR	< 1	NR	26.0
119	1.35	4	6.49	2	0.000	NR	2.38	4	26.0
132	1.50	1	7.15	3	< 0.01	NR			20.5
134	1.34	4	6.93	4	0.000	NR	2.35	4	26.6
138	1.35	4	6.78	4	< 0.004	NR	2.31	4	26.7
140	1.43	3	6.42	1	< 0.01	NR	2.00	2	32.5
141	1.39	4	6.96	4	< 0.05	NR	2.57	4	26.8
143			6.69	3	0.002	NR			25.9
145	1.34	4	7.30	2	< 0.01	NR	2.19	3	28.0
146	1.26	3	6.90	4	< 0.05	NR	< 5	NR	30.6
155			6.53	2	0.001	NR			27.8
158	0.96	0	4.69	0			30.50	0	26.1
180	1.34	4	7.10	3	< 0.01	NR	2.60	4	26.0
183			7.46	1					23.2
190	1.39	4	6.60	3	0.021	NR	2.80	2	27.6
191	2.07	0			0.400	NR	2.38	4	
193									26.8
196					< 0.05	NR	2.40	4	
197							2.36	4	
203			6.50	2			2.86	2	263.0
204			6.74	3	< 0.002	NR	2.03	2	26.3
215	1.60	0	7.05	4	0.030	NR	3.16	1	26.7
220	1.30	4					8.90	0	
221	1.45	2	6.09	0			0.41	0	
224	1.34	4	6.76	4	0.001	NR	2.62	3	25.0
235	1.19	2			< 0.5	NR	2.89	2	
240			6.74	3	< 0.1	NR	2.68	3	24.7
241	1.30	4	6.83	4	0.003	NR	2.00	2	20.0
243			8.23	0					25.7
244			6.98	4					25.9
247			7.01	4	< 0.001	NR	2.80	2	27.1
255	1.33	4	7.20	3					26.2

Table 10. Laboratory performance ratings for standard reference water sample P-27 (low ionic strength)—Continued

(MPV, most probable value; ug/L, micrograms per liter; mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/11, number of reported values of 11 possible values; RV, reported value; <, less than)

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent) 0.00 - 0.50	1 (Questionable) 1.51 - 2.00		
3 (Good) 0.51 - 1.00	0 (Poor) greater than 2.00		
2 (Satisfactory) 1.01 - 1.50	NR (Not Rated)		

Analyte = Na (Sodium)	pH	PO <sub>4</sub> as P	SO <sub>4</sub> (Sulfate)	Specific Conductance
MPV = 1.34 mg/L	6.92	0.002 mg/L	2.42 mg/L	26.1 μS/cm
F-pseudosigma = 0.10	0.32	0.014	0.37	1.1
Lab	RV	Rating	RV	Rating
256	0.92	0	7.04	4
257	1.20	2	7.03	4
258	1.50	1	6.30	1
261	2.30	0	6.44	2
262	1.46	2	5.18	0
265	1.19	2	6.76	4
268	1.38	4	6.76	4
270	1.68	0		
271	4.30	0	6.98	4
272	2.00	0	7.21	3
273	1.46	2	7.03	4
276			6.89	4
				0.000
				NR
				17.04
				0
				2.97
				2
				25.5
				3
				1.99
				2
				29.4
				0
				< 0.5
				0
				26.1
				4
				6.38
				0
				24.0
				1
				50.7
				0
				25.1
				3

Table 11. Laboratory performance ratings for standard reference water sample Hg-23 (mercury)

(MPV, most probable value; ug/L, micrograms per liter; Lab, laboratory number;  
V/1 number of reported values of 1 value; RV, reported value; <, less than)

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Questionable)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Poor)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Analyte = Hg (Mercury)

MPV = 0.34  $\mu$  g/L


F-pseudosigma = 0.07

Lab	V/1	RV	Rating
1	1	0.30	3
3	1	4.40	0
7	1	0.34	4
10	1	0.33	4
11	1	0.28	3
13	1	< 0.4	NR
15	1	0.29	3
16	1	0.35	4
18	1	0.38	3
26	1	0.37	4
28	1	2.40	0
32	1	0.29	3
34	1	0.32	4
36	1	0.30	3
39	1	0.30	3
42	1	0.32	4
46	1	0.31	4
48	1	0.37	4
50	1	0.52	0
51	1	0.39	3
55	1	0.30	3
58	1	< 0.5	NR
59	1	0.32	4
68	1	0.25	2
69	1	0.33	4
70	1	0.40	3
76	1	0.29	3
81	1	0.30	3
86	1	0.31	3
87	1	0.50	0
89	1	0.39	3
96	1	0.37	4
97	1	< 0.42	NR
105	1	0.37	4
108	1	0.53	0
113	1	0.39	3
119	1	0.39	3
127	1	0.32	4
133	1	0.33	4
134	1	0.28	3
138	1	0.32	4
141	1	0.43	2
142	1	0.44	2
144	1	0.35	4
145	1	0.43	2
146	1	0.36	4
149	1	0.35	4
193	1	0.43	2
212	1	0.28	3
213	1	0.75	0
215	1	0.92	0
219	1	0.30	3
221	1	0.30	3
234	1	0.34	4
235	1	0.43	2
241	1	0.39	3
245	1	0.28	3
255	1	0.32	4
257	1	1.00	0
259	1	0.40	3
265	1	0.20	1

Table 12. Statistical summary of reported data for standard reference water sample T-143 (trace constituents)


Definition of analytical methods, abbreviations, and symbols					
<u>Analytical methods</u>					
0. Other/Not reported			= atomic absorption: direct,air		
1. AA: direct, air			= atomic absorption: direct,nitrous oxide		
2. AA: direct, N <sub>2</sub> O			= atomic absorption: graphite furnace		
3. AA: graphite furnace			= inductively coupled plasma		
4. ICP			= direct current plasma		
5. DCP			= inductively coupled plasma/mass spectrometry		
6. ICP/MS			= ion chromatography		
7. IC			= atomic absorption: extraction [chelating agent(s) specified]		
10. AA: extraction			= atomic absorption: hydride [reducing agent specified]		
11. AA: hydride			= atomic absorption: flame emission		
12. AA: flame emission			= colorimetric [color reagent specified]		
<u>Abbreviations and symbols</u>					
N = number of samples					
MPV = most probable value					
F-pseudosigma = nonparametric statistic deviation					
Hu = upper hinge value					
Hi = lower hinge value					
µg/L = micrograms per liter					
mg/L = milligrams per liter					
Lab = laboratory code number					
NR = not rated, less than value reported					
< = less than					
<u>Constituent</u>		<u>page</u>	<u>Constituent</u>		
Ag	Silver	45	Mg	Magnesium	59
Al	Aluminium	46	Mn	Manganese	60
As	Arsenic	47	Mo	Molybdenum	61
B	Boron	48	Na	Sodium	62
Ba	Barium	49	Ni	Nickel	63
Be	Beryllium	50	Pb	Lead	64
Ca	Calcium	51	Sb	Antimony	65
Cd	Cadmium	52	Se	Selenium	66
Co	Cobalt	53	SiO <sub>2</sub>	Silica	67
Cr	Chromium	54	Sr	Strontium	68
Cu	Copper	55	Tl	Thallium	69
Fe	Iron	56	U	Uranium	70
K	Potassium	57	V	Vanadium	71
Li	Lithium	58	Zn	Zinc	72

Table 12. Statistical summary of reported data for standard reference water sample T-143 (trace constituents)–Continued  
 Ag (Silver)  $\mu\text{g/L}$



1. AA: direct air				6. ICP/MS							
3. AA: graphite furnace											
4. ICP											
N =	6	24	21	14				MPV =	19.8		
Minimum =	13.0	6.7	14.5	14.0				F-pseudosigma =	1.4		
Maximum =	26.0	40.0	143.0	20.7				N =	65		
Median =	19.5	20.0	19.5					Hu =	20.6		
F-pseudosigma =	1.7	1.5	0.7					HI =	18.7		
Lab	Rating	Z-value	1	3	4	6		Lab	Rating	Z-value	1
1	4	-0.15				19.6		212	0	-4.12	
3	0	-3.41				15.0		213	4	-0.43	14.0
7	2	1.28				21.6		215	0	5.82	
11	3	0.85				21.0		217	3	0.64	28.0
13	4	0.21				20.1		221	1	1.63	20.7
15	1	-1.56				17.6		234	3	0.57	
16	4	0.07				19.9		235	4	0.14	20.6
18	3	0.78				20.9		236	0	-3.76	
23	1	1.70				22.2		241	2	-1.07	14.5
26	0	-2.98				15.6		249	0	-2.70	18.3
30	3	-0.57				19.0		255	4	0.09	
32	4	0.07				19.9		256	0	3.98	19.9
36	0	-4.83	13.0					257	1	-1.99	25.4
39	4	0.28				20.2		259	3	0.85	17.0
40	4	-0.36				19.3		265	4	0.28	
42	3	-0.57				19.0		273	0	9.37	21.0
46	4	0.14				20.0					20.2
48	4	0.14				20.0					
58	3	0.85				21.0					
68	0	14.34				40.0					
69	4	0.00				19.8					
70	3	-0.78				18.7					
73	1	-1.99				17.0					
75	4	0.21				20.1					
76	4	-0.28				19.4					
85	4	0.14	20.0								
87	0	4.40	26.0								
89	4	-0.50				19.1					
96	4	0.43				20.4					
97	3	-0.57				19.0					
102	0	87.47				143.0					
105	2	-1.35				17.9					
107	4	0.14				20.0					
113	2	-1.21				18.1					
114	2	-1.28	18.0								
118	0	-9.30				6.7					
119	0	-2.34				16.5					
127	3	-0.71				18.8					
128	4	-0.21				19.5					
133	4	-0.14				19.6					
134	3	-0.59				19.0					
138	3	0.57				20.6					
141	0	6.53				29.0					
142	4	-0.36				19.3					
146	0	-6.40				< 10					
151	2	-1.35				17.9					
180	4	0.07				19.9					
190	1	1.85				22.4					
193	3	-0.57				19.0					
196	4	-0.21				19.5					

Table 12. Statistical summary of reported data for standard reference water sample T-143 (trace constituents)—Continued  
Al (Aluminum)




1. AA: direct air		4. ICP		5. DCP		6. ICP/MS		
2. AA: direct nitrous oxide		17.0		110.0		15.2		
3. AA: graphite furnace		45.5		206.0		0.0		
N		1	1	13	20	1	14	
Minimum		17.0	22.15	15.2	0.0	100.0	18.1	
Maximum				45.5	206.0		26.9	
Median				22.2	29.4		20.6	
F-pseudosigma				4.7	25.8		2.1	
Lab	Rating	Z-value	1	2	3	4	5	6
1	4	-0.36				19.1		
3	0	22.15			206.0			
4	NR				< 2000			
7	1	1.87			37.6			
13	4			22.2				
15	NR				< 50			
16	4	-0.43				18.5		
18	NR				< 100			
23	NR				< 50			
26	4			18.2				
28	2	1.23			32.3			
32	4	0.06				22.6		
33	0	9.38				100.0		
36	NR			< 200				
42	4	0.11				23.0		
48	4				18.5			
58	3	-0.61	17.0					
68	0	3.90			54.5			
69	3				16.5			
70	NR				< 100			
75	NR				< 30			
76	4	-0.23				20.2		
81	NR					< 6		
83	NR					< 25		
85	NR					< 100		
89	4				23.1			
97	4				25.7			
102	4	-0.25				20.0		
105	4	-0.12					21.1	
107	4				23.0			
113	4	-0.23				20.2		
118	NR			< 2000				
119	4	-0.31				19.5		
127	NR					< 30		
128	4	-0.41				18.7		
132	0	3.24				49.0		
134	4	0.10				22.9		
138	4	-0.25				20.0		
141	NR					< 100		
142	3	0.62				27.3		
145	0	4.36				58.3		
146	NR					< 200		
151	4	-0.48				18.1		
180	NR					< 40.6		
190	3				16.7			
191	4	-0.01				22.0		
196	4	-0.01				22.0		
203	4				18.6			
204	4				24.6			
212	4	-0.13				21.0		

MPV =	22.1
F-pseudosigma =	8.3
N =	50
Hu =	29.7
HI =	18.5


Lab	Rating	Z-value	1	2	3	4	5	6
215	0	13.84				137.0		
217	0	3.73				53.1		
219	3	0.83				29.0		
221	3					27.8		
224	0	-2.66				0.0		
234	3	-0.51				17.9		
235	3	-0.86				15.0		
236	3	0.92				29.7		
240	0	19.14				181.0		
241	3					15.2		
249	0					45.5		
255	3	-0.86				15.0		
257	0	10.59				110.0		
265	3	0.57					26.9	
273	1	-1.94					6.0	

Table 12. Statistical summary of reported data for standard reference water sample T-143 (trace constituents)—Continued



3. AA: graphite furnace		11. AA: hydride					MPV = 15.2		
4. ICP		11na. AA: hydride NaBH <sub>4</sub>					F-pseudosigma = 1.2		
6. ICP/MS		N =	33	8	14	3	8	N =	66
		Minimum =	9.6	14.0	14.0	15.0	14.0	Hu =	16.2
		Maximum =	20.5	27.4	17.0	18.2	16.4	Hi =	14.6
		Median =	15.0	15.3	15.4		14.9		
		F-pseudosigma =	1.4	3.2	0.8		1.0		
Lab	Rating	Z-value	3	4	6	11	11na	Lab	Rating
1	3	0.89	16.2					151	3
3	3	-0.72		14.3				180	NR
7	NR			< 120				190	4
10	3	0.72				16.0		191	1
11	4	-0.13		15.0				193	1
13	4	0.30	15.5					196	4
15	NR			< 100				204	4
16	4	0.21			15.4			212	3
18	4	-0.46	14.6					213	2
23	0	4.51	20.5					215	4
26	4	-0.21				14.9		217	4
30	3	-0.97		14.0				220	4
32	4	-0.21			14.9			221	4
34	4	0.38				15.6		224	4
36	0	-3.50	11.0					234	1
39	0	2.57			18.2			236	NR
42	4	-0.13			15.0			241	3
46	2	-1.39	13.5					249	0
48	3	-0.80	14.2					255	4
50	3	-0.97				14.0		265	2
58	0	3.25	19.0						
68	0	-4.72	9.6						
69	3	0.89	16.2						
70	4	0.13	15.3						
73	0	5.78		22.0					
75	4	-0.46			14.6				
76	2	1.14			16.5				
80	3	0.97	16.3						
81	3	0.72	16.0						
86	4	0.13				15.3			
87	4	-0.21				14.9			
89	2	1.05				16.4			
96	4	-0.46	14.6						
97	1	1.56		17.0					
102	3	0.55			15.8				
105	4	0.13			15.3				
109	1	-1.73	13.1						
113	1	1.98	17.5						
118	0	2.32	17.9						
119	4	-0.13			15.0				
127	3	-0.63	14.4						
128	4	-0.30			14.8				
133	2	1.14		16.5					
134	3	-1.00				14.0			
138	3	-0.97			14.0				
141	4	-0.13	15.0						
142	4	0.04			15.2				
144	3	-0.55	14.5						
145	0	10.33		27.4					
146	4	-0.04			15.1				


Table 12. Statistical summary of reported data for standard reference water sample T-143 (trace constituents)—Continued  
 B (Boron)  $\mu\text{g/L}$



4. ICP	MPV =	35.0
6. ICP/MS	F-pseudosigma =	5.2
22az. Color: azomethine	N =	35
	Hu =	37.0
	HI =	30.0
N =	30	4
Minimum =	22.7	28.8
Maximum =	140.0	95.0
Median =	35.2	
F-pseudosigma =	5.2	

Lab	Rating	Z-value	4	6	22az
1	4	-0.02	34.9		
3	3	-0.97	30.0		
11	4	0.39	37.0		
15	NR		< 50		
16	0	4.29	57.1		
18	NR		< 50		
24	3	-0.83	30.7		
26	1	-1.73	26.1		
28	2	1.40	42.2		
42	4	0.19		36.0	
46	0	-2.39	22.7		
48	NR		< 100		
68	0	20.38	140.0		
70	NR		< 50		
85	4	0.12	35.6		
86	4	0.37	36.9		
119	3	-0.58	32.0		
127	4	0.14	35.7		
128	3	-0.97	30.0		
129	0	11.65		95.0	
132	1	-1.65	26.5		
134	4	0.15	35.8		
138	2	-1.20		28.8	
141	0	-2.12	24.1		
142	4	0.14	35.7		
145	4	0.08	35.4		
158	3	0.58	38.0		
180	4	-0.50	32.4		
191	4	-0.19		34.0	
212	NR		< 50		
215	0	13.78	106.0		
217	2	1.01	40.2		
219	4	0.00	35.0		
234	4	-0.35	33.2		
235	3	-0.97	30.0		
236	2	-1.13	29.2		
240	2	-1.36	28.0		
255	4	0.19	36.0		
265	4	-0.07		34.7	
273	0	4.46	58.0		

Table 12. Statistical summary of reported data for standard reference water sample T-143 (trace constituents)—Continued  
 Ba (Barium)  $\mu\text{g/L}$



2. AA: direct nitrous oxide

5. DCP

3. AA: graphite furnace

6. ICP/MS


4. ICP

	N =	1	8	45	1	13
Minimum =	90.0	69.6	64.5	102.0	41.0	
Maximum =		98.1	94.0		92.0	
Median =		90.4	81.5		79.1	
F-pseudosigma =		14.5	3.7		3.9	

	MPV =	81.9
F-pseudosigma =		4.5
N =		68
Hu =		84.6
Hi =		78.5

Lab	Rating	Z-value	2	3	4	5	6
1	4	0.08				82.2	
3	3	0.55			84.4		
4	2	1.13			87.0		
7	4	0.38			83.6		
11	4	-0.20			81.0		
13	2	1.37			88.1		
15	3	-0.88			77.9		
16	4	0.07				82.2	
18	3	-0.64			79.0		
19	4	0.07			82.2		
24	3	-0.69			78.8		
25	3	0.55			84.4		
26	4	-0.22			80.9		
28	3	0.71			85.1		
30	4	-0.42			80.0		
32	3	-0.88				77.9	
33	0	4.45				102.0	
36	1	1.79	90.0				
39	3	-0.62				79.1	
40	3	-0.84			78.1		
46	3	0.69			85.0		
48	0	3.52			97.8		
50	1	1.79			90.0		
55	3	-0.82			78.2		
68	4	0.24			83.0		
70	4	0.09			82.3		
75	4	-0.15			81.2		
81	3	-0.86				78.0	
83	3	-0.91			77.8		
85	4	0.29			83.2		
86	4	-0.09			81.5		
87	1	1.95			90.7		
89	0	3.58			98.1		
96	NR	< 100					
97	0	-2.72			69.6		
102	0	2.68			94.0		
105	2	-1.26				76.2	
107	2	-1.08	77.0				
113	4	-0.38			80.2		
119	4	0.46			84.0		
121	4	-0.20			81.0		
127	2	-1.35			75.8		
128	0	-2.61				70.1	
133	3	-0.93			77.7		
134	3	-0.67			78.9		
138	4	-0.31			80.5		
141	3	0.64			84.8		
142	4	-0.02				81.8	
145	1	1.99			90.9		
146	4	0.07			82.2		

Table 12. Statistical summary of reported data for standard reference water sample T-143 (trace constituents)—Continued  
Be (Beryllium)  $\mu\text{g/L}$



1. AA: direct air  
3. AA: graphite furnace  
4. ICP


6. ICP/MS

	N =	1	10	37	13
Minimum =	10.00	6.70	6.53	7.00	
Maximum =		10.00	12.60	9.80	
Median =		8.44	8.50	8.70	
F-pseudosigma =		1.56	0.47	0.88	

MPV = 8.50  
F-pseudosigma = 0.66  
N = 61  
Hu = 8.89  
HI = 8.00

Lab	Rating	Z-value	1	3	4	6
1	3	-0.56			8.13	
3	3	-0.76		8.00		
4	3	0.76		9.00		
7	3	-0.61		8.10		
11	4	-0.15		8.40		
13	4	0.36		8.74		
15	4	0.15		8.60		
16	2	-1.21			7.70	
18	4	0.00		8.50		
25	0	2.12		9.90		
26	4	0.24		8.66		
30	1	-1.82		7.30		
32	2	1.36			9.40	
36	0	2.27		10.00		
39	1	1.52		9.50		
40	3	-0.91			7.90	
46	4	0.39			8.76	
48	0	2.27		10.00		
68	3	0.61			8.90	
69	1	-1.79		7.32		
70	4	0.26			8.67	
75	4	0.00			8.50	
76	1	1.79			9.68	
81	0	-2.27			7.00	
83	4	-0.45			8.20	
85	2	1.02			9.17	
86	4	-0.38			8.25	
89	1	-1.67		7.40		
96	0	2.27		10.00		
97	3	0.74		8.99		
102	3	-0.76			8.00	
105	1	-1.82			7.30	
113	4	0.42			8.78	
119	2	-1.18		7.72		
121	3	-0.76			8.00	
127	0	-2.99			6.53	
128	2	-1.46			7.54	
133	4	0.30			8.70	
134	4	0.02			8.51	
138	4	0.23			8.65	
141	3	-0.53			8.15	
142	3	0.56			8.87	
144	0	-12.86	< 0.01			
145	1	1.52			9.50	
146	1	-1.92			7.23	
151	3	-0.52			8.16	
158	2	1.21			9.30	
180	4	0.15			8.60	
191	3	0.59			8.89	
193	3	-0.76		8.00		

Table 12. Statistical summary of reported data for standard reference water sample T-143 (trace constituents)—Continued  
 Ca (Calcium) mg/L




0. Other			4. ICP					
1. AA: direct air			6. ICP/MS					
2. AA: direct nitrous oxide			20. Titrate: colorimetric					
	N =	3	16	7	53	3	4	
	Minimum =	48.3	44.8	44.0	32.7	53.0	0.0	
	Maximum =	54.2	58.0	54.0	59.1	55.0	57.1	
	Median =	52.9	51.8	54.4				
	F-pseudosigma =	2.1	0.6	2.0				
Lab	Rating	Z-value	0	1	2	4	6	20
1	4	-0.50				52.5		
3	4	0.34				54.4		
4	4	-0.11				53.4		
7	3	0.97				55.8		
11	2	1.19				56.3		
13	0	2.41				59.0		
15	2	-1.37				50.6		
16	4	0.16				54.0		
18	4	-0.09				53.5		
19	4	-0.02				53.6		
23	3	-0.52	52.5					
24	4	-0.11				53.4		
25	3	0.52				54.8		
26	4	0.07				53.8		
28	3	0.79				55.4		
30	4	0.16		54.0				
32	3	0.61			55.0			
33	4	0.25	54.2					
36	3	-0.74		52.0				
42	2	1.37			56.7			
43	4	0.16			54.0			
46	2	1.06			56.0			
48	2	1.01			55.9			
55	3	0.56			54.9			
68	2	1.28			56.5			
69	3	-0.56	52.4					
70	3	0.79			55.4			
75	4	-0.16	53.3					
81	4	-0.29			53.0			
83	4	-0.25			53.1			
85	3	0.79	55.4					
86	3	0.83			55.5			
87	2	-1.06		51.3				
89	4	-0.47	52.6					
97	3	-0.83	51.8					
102	1	1.96			58.0			
105	4	-0.38			52.8			
107	3	-0.74	52.0					
109	4	-0.47			52.6			
110	3	0.61	55.0					
113	1	1.91			57.9			
114	0	-4.34		44.0				
119	4	0.02			53.7			
121	3	-0.65			52.2			
128	3	0.92			55.7			
129	1	1.96	58.0					
132	4	-0.12			53.4			
133	4	-0.29			53.0			
134	4	0.38			54.5			
138	3	0.56			54.9			

MPV = 53.7  
 F-pseudosigma = 2.2  
 N = 86  
 Hu = 55.4  
 HI = 52.4

Lab	Rating	Z-value	0	1	2	4	6	20
140	4	0.38	54.5					
141	2	1.15				56.2		
142	4	-0.17				53.3		
145	0	2.45				59.1		
146	1	-1.64				50.0		
158	4	0.47				54.7		
180	4	0.34				54.4		
190	4	-0.16	53.3					
191	4	-0.25				53.1		
193	3	-0.83				51.8		
203	0	-2.64	47.8					
204	4	-0.20				53.2		
212	2	1.28				56.5		
215	3	0.74				55.3		
217	1	-1.55				50.2		
218	3	-0.74				52.0		
219	1	-1.64				50.0		
220	3	0.61	55.0					
221	3	0.65	55.1					
224	2	-1.06				51.3		
234	3	0.83				55.5		
235	0	2.41				59.0		
236	4	-0.50				52.5		
240	0	-2.72				47.6		
241	0	-3.98	44.8					
246	0	-9.44				32.7		
255	4	-0.38				52.8		
257	2	-1.19				51.0		
261	1	1.55				57.1		
265	4	0.20				54.1		
268	4	-0.18	53.3					
270	0	-2.39	48.3					
271	4	0.12				53.9		
272	0	-13.49				23.7		
273	2	1.48				56.9		
274	0	-24.12				0.0		

Table 12. Statistical summary of reported data for standard reference water sample T-143 (trace constituents)—Continued




1. AA: direct air	6. ICP/MS					
3. AA: graphite furnace						
4. ICP						
N =	6      28      35      12					
Minimum =	13.9      6.4      13.6      18.0					
Maximum =	26.0      21.4      22.3      23.0					
Median =	19.2      19.0      19.6					
F-pseudosigma =	1.7      1.3      0.8					
Lab	Rating	Z-value	1	3	4	6
1	3	0.65		20.1		
3	4	0.27			19.5	
4	NR				< 100	
7	2	1.28			21.0	
11	4	-0.07			19.0	
13	2	-1.15			17.4	
15	0	-2.77		15.0		
16	4	0.40				19.7
18	4	-0.27			18.7	
19	4	0.47			19.8	
23	4	-0.20		18.8		
24	3	-0.94			17.7	
25	2	1.28			21.0	
26	3	0.88		20.4		
28	0	-3.71			13.6	
30	4	0.27			19.5	
32	4	-0.34				18.6
36	4	-0.07		19.0		
39	3	-0.74		18.0		
40	4	-0.34			18.6	
42	3	-0.74				18.0
46	4	-0.34		18.6		
48	3	-0.74		18.0		
58	3	0.61		20.0		
68	4	0.27			19.5	
69	3	0.61		20.0		
70	3	-0.61		18.2		
73	4	-0.07			19.0	
75	4	0.07			19.2	
80	0	-3.51	13.9			
81	3	0.61		20.0		
83	4	-0.40			18.5	
85	4	0.34	19.6			
86	3	-0.94			17.7	
87	2	1.28	21.0			
89	3	0.81		20.3		
96	2	-1.08		17.5		
97	4	0.13		19.3		
102	2	1.28			21.0	
105	4	0.40				19.7
113	4	0.00			19.1	
114	4	-0.07	19.0			
118	0	-8.57		6.4		
119	4	-0.47		18.4		
121	3	-0.74			18.0	
127	2	-1.15			17.4	
128	4	0.27				19.5
132	4	-0.40			18.5	
133	3	0.54			19.9	
134	4	-0.20			18.8	

MPV =	19.1
F-pseudosigma =	1.5
N =	81
Hu =	20.0
HI =	18.0

Lab	Rating	Z-value	1	3	4	6
138	4	-0.07			19.0	
140	3	0.61	20.0			
141	4	0.47		19.8		
142	3	0.61			20.0	
145	0	2.16		22.3		
146	4	-0.34		18.6		
151	4	0.34			19.6	
158	0	-6.95		8.8		
180	4	0.34		19.6		
190	2	1.08	20.7			
191	4	-0.34			18.6	
193	3	0.61	20.0			
196	4	0.07			19.2	
212	0	2.63			23.0	
213	1	1.55	21.4			
215	4	-0.07		19.0		
217	2	-1.35			17.1	
219	3	-0.74		18.0		
221	4	0.34	19.6			
224	0	-3.51		13.9		
234	4	0.40	19.7			
235	3	0.61		20.0		
236	2	-1.42			17.0	
241	3	0.81	20.3			
249	0	-2.09	16.0			
255	4	0.32			19.6	
256	3	0.88		20.4		
257	0	4.65	26.0			
259	0	-3.44		14.0		
265	3	0.61			20.0	
273	0	-2.09		16.0		
274	0	-7.88	7.4			


Table 12. Statistical summary of reported data for standard reference water sample T-143 (trace constituents)—Continued



1. AA: direct air	6. ICP/MS					
3. AA: graphite furnace						
4. ICP						
N =	1      9      32      11					
Minimum =	23.0      10.0      8.0      15.1					
Maximum =	18.7      24.2      20.0					
Median =	17.3      16.8      17.0					
F-pseudosigma =	1.1      1.3      1.0					
Lab	Rating	Z-value	1	3	4	6
1	4	0.36		17.4		
3	1	-1.69			15.0	
4	NR				< 100	
7	0	2.45			19.9	
11	3	0.84			18.0	
13	0	-3.71			12.6	
15	NR				< 20	
16	4	-0.08				16.9
18	3	-0.76			16.1	
24	0	-3.96			12.3	
26	3	-0.51			16.4	
30	4	0.00			17.0	
32	4	0.17				17.2
40	1	-1.52			15.2	
42	0	2.53				20.0
46	4	0.34			17.4	
48	NR				< 50	
50	3	-0.84		16.0		
68	3	0.84			18.0	
70	NR				< 50	
75	1	1.60			18.9	
86	4	-0.17			16.8	
89	3	0.67		17.8		
97	4	0.17		17.2		
102	0	2.53			20.0	
105	2	-1.26				15.5
119	3	-0.93				15.9
121	3	-0.84			16.0	
127	4	0.25		17.3		
128	4	0.00			17.0	
132	4	0.00			17.0	
134	4	-0.25			16.7	
138	1	-1.60				15.1
141	3	-0.67			16.2	
142	3	-0.59				16.3
145	0	6.07			24.2	
146	3	-0.67			16.2	
158	4	0.08			17.1	
180	4	0.25			17.3	
191	3	0.51				17.6
196	4	0.25				17.3
212	4	0.00				17.0
213	2	1.43		18.7		
215	1	1.69			19.0	
217	4	-0.17			16.8	
219	3	-0.84			16.0	
221	3	-0.84		16.0		
224	3	-0.93			15.9	
234	4	0.34		17.4		
235	1	1.69			19.0	

MPV =	17.0
F-pseudosigma =	1.2
N =	53
Hu =	17.6
HI =	16.0

Table 12. Statistical summary of reported data for standard reference water sample T-143 (trace constituents)—Continued  
 Cr (Chromium)  $\mu\text{g/L}$




1. AA: direct air	4. ICP
2. AA: direct nitrous oxide	6. ICP/MS
3. AA: graphite furnace	
N =	1      2      28      35      13
Minimum =	35.0      38.2      25.6      29.2      32.4
Maximum =	52.0      47.0      42.1      38.4
Median =	37.4      37.1      35.4
F-pseudosigma =	3.5      2.0      2.1

Lab	Rating	Z-value	1	2	3	4	6
1	3	-0.57				35.5	
3	3	0.76			39.0		
4	NR				< 100		
7	4	0.42			38.1		
11	4	-0.38			36.0		
13	3	0.57			38.5		
15	3	-0.72			35.1		
16	2	-1.06				34.2	
18	4	-0.27			36.3		
19	4	0.30			37.8		
23	3	-0.76		35.0			
24	0	-2.20			31.2		
25	4	0.00			37.0		
26	4	0.19			37.5		
30	4	0.00			37.0		
32	4	0.11				37.3	
36	4	0.00		37.0			
39	0	-3.42			28.0		
40	1	-1.71				32.5	
42	4	0.38				38.0	
46	4	-0.46			35.8		
48	1	1.82			41.8		
50	4	-0.38			36.0		
58	4	0.00			37.0		
68	3	0.57				38.5	
69	4	-0.38			36.0		
70	4	0.04			37.1		
73	1	-1.52			33.0		
75	4	0.04			37.1		
76	4	-0.27				36.3	
81	1	1.52		41.0			
83	4	0.08			37.2		
85	3	0.53			38.4		
86	0	-2.96			29.2		
87	4	0.46		38.2			
89	4	-0.38			36.0		
96	3	0.72			38.9		
97	3	0.76			39.0		
102	1	1.52				41.0	
105	1	-1.56				32.9	
113	4	0.42			38.1		
118	0	2.96			44.8		
119	3	-0.76			35.0		
127	3	0.84			39.2		
128	1	-1.75				32.4	
132	2	1.33				40.5	
133	2	-1.03				34.3	
134	4	-0.30		36.2			
138	3	0.65			38.7		
140	3	-0.76	35.0				

MPV = 37.0  
 F-pseudosigma = 2.6  
 N = 79  
 Hu = 38.6  
 HI = 35.1

Lab	Rating	Z-value	1	2	3	4	6
141	4	-0.11				36.7	
142	3	-0.61					35.4
145	1	1.94				42.1	
146	4	-0.23				36.4	
151	3	-0.87					34.7
158	0	-4.33			25.6		
180	4	-0.08				36.8	
190	3	0.91			39.4		
191	3	0.53					35.4
193	4	-0.38			36.0		
196	3	-0.95				34.5	
204	0	3.80			47.0		
212	1	-1.52				33.0	
213	0	2.85			44.5		
215	3	-0.76				35.0	
217	4	0.23				37.6	
219	2	-1.14				34.0	
221	4	0.27			37.7		
234	1	1.75			41.6		
235	2	1.14			40.0		
236	3	-0.72				35.1	
241	2	-1.14				34.0	
249	1	2.03			42.4		
253	2	1.23			40.2		
255	4	-0.39				36.0	
256	2	1.14			40.0		
257	0	5.70			52.0		
259	2	-1.14				34.0	
265	4	0.00				37.0	
273	4	0.38				38.0	

Table 12. Statistical summary of reported data for standard reference water sample T-143 (trace constituents)—Continued  
 Cu (Copper)  $\mu\text{g/L}$




1. AA: direct air				6. ICP/MS							
3. AA: graphite furnace											
4. ICP											
N =	13	23	42	12				MPV =	22.3		
Minimum =	18.7	4.2	7.8	19.2				F-pseudosigma =	1.9		
Maximum =	30.0	24.8	42.0	26.0				N =	90		
Median =	25.0	21.4	22.5	22.2				Hu =	23.6		
F-pseudosigma =	3.4	1.9	1.7	1.6				HI =	21.0		
Lab	Rating	Z-value		1	3	4	6	Lab	Rating	Z-value	
1	4	-0.02		22.2				129	0	4.01	30.0
3	3	0.90			24.0			132	0	10.24	42.0
4	4	0.38			23.0			133	2	-1.43	19.5
7	3	0.59			23.4			134	4	0.07	22.4
10	4	0.28	22.8					138	2	-1.12	20.1
11	3	0.90			24.0			140	1	1.94	26.0
13	0	-6.16			10.4			141	3	0.54	23.3
15	4	-0.50			21.3			142	2	-1.18	20.0
16	4	-0.24				21.8		144	2	-1.07	20.2
18	4	0.02			22.3			145	0	5.67	33.2
19	0	-3.25			16.0			146	1	1.52	25.2
23	3	0.85	23.9					151	4	-0.09	22.1
24	0	-3.15			16.2			158	1	1.83	25.8
25	3	-0.66			21.0			180	3	0.54	23.3
26	3	0.59			23.4			190	4	-0.03	22.2
28	0	-7.51			7.8			191	4	0.02	22.3
30	4	0.12			22.5			193	1	1.94	26.0
32	4	0.49				23.2		196	3	0.74	23.7
36	0	-6.36		10.0				203	1	-1.85	18.7
40	4	-0.50			21.3			204	0	-5.48	11.7
42	4	-0.14			22.0			212	1	1.94	26.0
46	4	0.07			22.4			213	4	0.49	23.2
48	2	-1.33			19.7			215	0	4.22	30.4
50	3	0.90			24.0			217	3	0.85	23.9
55	3	-0.76			20.8			219	1	-1.69	19.0
58	NR		< 50					221	4	-0.50	21.3
68	0	2.46			27.0			224	4	-0.40	21.5
69	4	-0.45			21.4			234	3	0.69	23.6
70	4	0.28			22.8			235	2	-1.18	20.0
73	4	-0.14			22.0			236	1	-1.85	18.7
75	4	-0.03			22.2			240	3	-0.66	21.0
80	4	-0.14			22.0			241	2	-1.33	19.7
81	4	-0.14			22.0			249	4	-0.32	21.7
83	4	0.38				23.0		253	0	4.01	30.0
85	1	1.52	25.2					255	4	-0.31	21.7
86	4	0.43			23.1			256	3	0.69	23.6
87	3	-0.66		21.0				257	0	2.98	28.0
89	4	-0.45		21.4				259	2	1.42	25.0
96	3	0.59			23.4			265	4	0.38	
97	3	-0.55			21.2			273	4	-0.14	22.0
102	4	-0.14			22.0			274	0	-9.37	4.2
105	4	0.07			22.4						
107	2	1.32			24.8						
113	3	0.64				23.5					
114	3	-0.66		21.0							
118	0	-7.45			7.9						
119	1	1.94			26.0						
121	0	-2.21			18.0						
127	4	0.07			22.4						
128	1	-1.57			19.2						

Table 12. Statistical summary of reported data for standard reference water sample T-143 (trace constituents)—Continued

Fe (Iron)

µg/L




1. AA: direct air			5. DCP		
3. AA: graphite furnace			6. ICP/MS		
4. ICP			22. Colorimetric		
N =	19	9	56	1	4
Minimum =	170	188	21	250	200
Maximum =	470	302	261		152
Median =	222	233	221		228
F-pseudosigma =	27	22	10		
Lab	Rating	Z-value	1	3	4
1	3	-0.77		211	
3	0	2.81		261	
4	3	0.94		235	
7	4	0.20		225	
10	1	1.73	246		
11	2	-1.44		202	
13	2	-1.08		207	
15	2	-1.01		208	
16	1	-1.59		200	
18	4	0.00		222	
19	4	0.36		227	
21	0	-5.05			152
23	4	-0.29	218		
24	4	0.00		222	
25	2	-1.44		202	
26	4	0.00		222	
30	0	17.89	470		
32	0	10.32			365
33	1	2.02		250	
35	4	-0.43			216
36	1	2.02	250		
40	0	-14.52		21	
42	4	0.14		224	
43	3	0.58		230	
46	4	0.00		222	
48	0	-4.47		160	
50	0	2.09	251		
55	4	0.00		222	
58	3	-0.87	210		
68	4	0.22		225	
69	4	0.29	226		
70	3	-0.58		214	
73	4	-0.07		221	
75	3	-0.58		214	
80	4	0.07	223		
81	4	-0.14		220	
83	4	-0.22		219	
85	4	0.14		224	
86	3	-0.65		213	
87	3	-0.72	212		
89	1	1.59		244	
91	3	-0.72		212	
96	1	1.88	248		
97	3	0.79		233	
102	1	1.59		244	
105	3	-0.65		213	
107	1	-1.59	200		
109	0	2.48	256		
113	4	0.07		223	
114	2	-1.30	204		

MPV = 222  
F-pseudosigma = 14  
N = 93  
Hu = 230  
HI = 211

Lab	Rating	Z-value	1	3	4	5	6	22
119	4	0.29		226				
121	3	-0.87		210				
127	4	-0.14		220				
128	2	-1.23		205				
129	0	-2.67						185
132	4	0.40		228				
133	4	0.07		223				
134	4	-0.28		218				
138	4	0.14		224				
140	2	1.08	237					
141	1	-1.59		200				
142	4	-0.07		221				
145	4	-0.42		216				
146	4	-0.14		220				
151	1	-1.59						200
155	4	0.46						228
158	4	0.00		222				
180	4	0.00		222				
190	0	-2.16	192					
191	4	-0.07						221
203	4	-0.14	220					
204	2	-1.01		208				
212	0	2.74		260				
213	0	5.77		302				
215	4	0.14		224				
217	3	0.58		230				
219	2	-1.01		208				
220	4	0.02	222					
221	4	-0.14		220				
224	1	-1.51		201				
234	4	0.14		224				
235	3	0.94		235				
236	3	-0.53		215				
240	0	-2.74		184				
241	0	3.03	264					
249	0	-2.45		188				
253	0	7.79	330					
255	4	-0.35						217
256	4	-0.07	221					
257	0	-3.75	170					
265	3	0.58		230				
273	3	0.58		230				
274	3	0.73		232				

Table 12. Statistical summary of reported data for standard reference water sample T-143 (trace constituents)—Continued  
 K (Potassium) mg/L




1. AA: direct air	6. ICP/MS
4. ICP	7. Ion chromatography
5. DCP	12. Flame emission
N = 26	41
Minimum = 1.80	2.11
Maximum = 3.00	5.40
Median = 2.54	2.50
F-pseudosigma = 0.16	0.19
	0.27

MPV = 2.50  
 F-pseudosigma = 0.21  
 N = 84  
 Hu = 2.69  
 HI = 2.40

Lab	Rating	Z-value	1	4	5	6	7	12	Lab	Rating	Z-value	1	4	5	6	7	12
1	4	0.23	2.55						142	2	-1.30	2.22					
3	0	13.49		5.40					145	3	0.74	2.66					
7	1	1.60		2.84					146	1	2.00	2.93					
11	4	0.00		2.50					180	3	-0.84	2.32					
13	3	-0.56		2.38					190	4	0.00		2.50				
15	2	-1.02		2.28					191	4	-0.14		2.47				
16	1	-1.86	2.10						193	3	-0.84	2.32					
18	4	-0.47		2.40					203	1	1.77	2.88					
19	4	0.00		2.50					204	4	-0.47		2.40				2.40
23	4	-0.28	2.44						212	4	-0.47	2.40					
24	3	-0.74		2.34					215	0	2.79	3.10					
25	3	0.65		2.64					217	1	1.95	2.92					
26	4	-0.05			2.49				218	3	-0.70	2.35					
28	3	-0.84		2.32					219	4	0.00	2.50					
32	4	0.00			2.50				220	4	0.47	2.60					
33	4	0.19		2.54					221	3	0.51	2.61					
36	0	-3.26	1.80						224	3	-0.88	2.31					
40	3	-0.84		2.32					234	4	-0.23	2.45					
42	4	0.00		2.50					235	0	5.12	3.60					
43	3	-0.93		2.30					236	1	-1.67	2.14					
46	3	0.56		2.62					241	4	0.19	2.54					
48	4	0.05		2.51					246	0	-10.95	< 0.2					
51	4	-0.09			2.48				249	3	0.84		2.68				
64	4	0.33	2.57						255	3	-0.56	2.38					
68	2	1.40		2.80					256	0	4.74		3.52				
69	2	1.02			2.72				257	1	1.86		2.90				
70	4	-0.33		2.43					259	3	0.93	2.70					
81	3	-0.60			2.37				261	3	-0.70		2.35				
83	4	0.00		2.50					265	4	-0.19	2.46					
85	2	1.49	2.82						268	3	0.93	2.70					
86	4	0.37		2.58					270	0	2.56		3.05				
87	4	-0.37		2.42					271	2	1.40		2.80				
89	3	-0.70		2.35					272	0	2.33		3.00				
97	4	0.19		2.54					273	0	2.16	2.97					
102	4	0.00		2.50					274	2	1.30		2.78				
105	4	-0.14		2.47													
107	3	0.60		2.63													
109	4	0.05		2.51													
113	4	-0.47		2.40													
114	0	2.33	3.00														
119	0	3.26		3.20													
121	4	0.23		2.55													
127	3	0.84		2.68													
128	1	-1.81		2.11													
129	4	-0.47		2.40													
132	3	0.65		2.64													
134	4	-0.09		2.48													
138	4	-0.09		2.48													
140	4	0.05	2.51														
141	4	0.23		2.55													


Table 12. Statistical summary of reported data for standard reference water sample T-143 (trace constituents)—Continued



1. AA: direct air	6. ICP/MS						
3. AA: graphite furnace	12. Flame emission						
4. ICP							
N =	3      2      20      8      1						
Minimum =	16.0      14.0      13.1      14.5      22.2						
Maximum =	19.4      19.2      26.0      21.0						
Median =	18.0      18.1						
F-pseudosigma =	2.4      1.2						
Lab	Rating	Z-value	1	3	4	6	12
1	3	-0.67			16.6		
3	0	3.77			26.0		
4	NR				< 100		
7	3	0.66			19.4		
16	1	-1.65				14.5	
24	0	3.02			24.4		
25	4	0.00			18.0		
26	4	-0.14			17.7		
30	3	-0.94			16.0		
32	4	-0.38				17.2	
40	3	-0.61			16.7		
42	0	-5.71			< 6		
64	1	1.98					22.2
68	4	0.00			18.0		
69	3	0.57		19.2			
75	3	0.80			19.7		
76	4	0.14				18.3	
85	4	-0.24	17.5				
105	2	-1.08			15.7		
109	3	0.66	19.4				
127	3	-0.71			16.5		
134	3	0.74			19.6		
142	4	0.47			19.0		
145	2	1.46			21.1		
151	3	-0.52				16.9	
191	2	1.41				21.0	
196	4	-0.05				17.9	
212	4	0.47				19.0	
217	4	-0.24			17.5		
219	4	-0.47			17.0		
234	3	0.85			19.8		
236	0	-2.31			13.1		
256	0	-8.52					< 0.1
257	3	-0.94	16.0				
259	1	-1.89			14.0		
265	4	0.16				18.4	
273	0	2.83			24.0		

MPV =	18.0
F-pseudosigma =	2.1
N =	34
Hu =	19.6
HI =	16.7


Table 12. Statistical summary of reported data for standard reference water sample T-143 (trace constituents)—Continued  
 Mg (Magnesium) mg/L



0. Other		4. ICP					
1. AA: direct air		6. ICP/MS					
2. AA: direct nitrous oxide		20. Titrate: colorimetric					
		N =	2	20	5	54	3
		Minimum =	10.3	8.3	8.7	7.6	10.3
		Maximum =	10.4	11.0	10.4	12.5	10.8
		Median =			10.2	10.5	25.4
		F-pseudosigma =			0.6	0.5	
Lab	Rating	Z-value	0	1	2	4	6
1	3	-0.58			10.1		
3	4	-0.19			10.3		
4	3	0.96			10.9		
7	3	0.85			10.8		
11	2	1.35			11.1		
13	3	0.58			10.7		
15	2	-1.18			9.8		
16	4	0.19			10.5		
18	4	-0.31			10.2		
19	4	0.19			10.5		
23	4	-0.33	10.2				
24	4	-0.19			10.3		
25	3	-0.77			10.0		
26	3	0.96			10.9		
28	3	0.77			10.8		
30	3	-0.77		10.0			
32	3	0.77			10.8		
33	4	-0.19	10.3				
36	0	-3.28		8.7			
40	4	-0.19			10.3		
42	2	1.35			11.1		
43	4	0.19			10.5		
46	4	0.19			10.5		
48	3	0.77			10.8		
51	0	-3.04	8.8				
55	1	1.73			11.3		
68	2	1.16			11.0		
69	3	-0.58	10.1				
70	4	0.19			10.5		
75	4	0.39	10.6				
81	4	-0.19			10.3		
83	3	-0.58			10.1		
85	3	0.77	10.8				
86	4	0.00			10.4		
87	4	-0.39	10.2				
89	4	0.00	10.4				
97	4	0.00	10.4				
102	0	4.05			12.5		
105	4	-0.19			10.3		
107	3	0.58	10.7				
109	4	0.00	10.4				
110	0	-4.09	8.3				
113	3	0.58			10.7		
114	3	-0.77	10.0				
119	4	0.39			10.6		
121	3	-0.77			10.0		
127	4	0.00			10.4		
128	3	-0.87			10.0		
129	3	-0.77	10.0				
132	4	0.27			10.5		

Lab	Rating	Z-value	0	1	2	4	6	20
133	3	-0.89					9.9	
134	4	-0.45					10.2	
138	3	0.58					10.7	
140	2	1.16				11.0		
141	3	0.96					10.9	
142	3	-0.77					10.0	
145	1	2.00					11.4	
146	3	-0.58					10.1	
158	2	1.35					11.1	
180	4	0.39					10.6	
190	4	0.00	10.4					
191	3	0.62					10.7	
193	2	-1.25				9.8		
203	4	0.42				10.6		
204	0	-5.40					7.6	
212	1	1.54					11.2	
215	4	0.39					10.6	
217	3	-0.77					10.0	
218	4	-0.02				10.4		
219	3	-0.77					10.0	
220	3	-0.77				10.0		
221	2	-1.27				9.7		
224	3	-0.83					10.0	
234	3	-0.89					9.9	
235	4	0.19					10.5	
236	4	0.50					10.7	
240	4	-0.39					10.2	
241	1	-1.93				9.4		
246	0	-3.33					8.7	
255	4	-0.19					10.3	
257	0	-3.28					8.7	
261	4	0.19						10.5
265	3	-0.58					10.1	
268	4	-0.39				10.2		
271	0	28.83						25.4
272	0	-2.75						9.0
273	1	1.60					11.2	
274	0	14.13						17.7

Table 12. Statistical summary of reported data for standard reference water sample T-143 (trace constituents)—Continued  
 Mn (Manganese)  $\mu\text{g/L}$




1. AA: direct air	5. DCP
3. AA: graphite furnace	6. ICP/MS
4. ICP	22. Colorimetric
N =	14      10      48      1      12      1
Minimum =	17.0      7.4      13.0      20.0      15.7      31.5
Maximum =	22.0      22.7      28.3      —      21.0      —
Median =	19.7      19.5      18.0      —      18.0      —
F-pseudosigma =	1.5      2.4      1.4      —      1.3      —

Lab	Rating	Z-value	1	3	4	5	6	22
1	4	-0.19						17.8
3	1	-1.82			14.7			
4	4	0.42			19.0			
7	4	0.21			18.6			
10	1	1.97	22.0					
11	4	-0.10			18.0			
13	3	-0.78			16.7			
15	0	5.24			28.3			
16	4	0.00				18.2		
18	3	-0.88			16.5			
19	4	0.31			18.8			
23	4	-0.21	17.8					
24	4	-0.42			17.4			
25	1	1.97			22.0			
26	1	-1.82			14.7			
28	0	3.58			25.1			
30	3	-0.62			17.0			
32	4	-0.26				17.7		
33	3	0.93				20.0		
36	4	0.42			19.0			
40	1	-1.97			14.4			
42	2	1.45				21.0		
43	3	0.93			20.0			
46	2	-1.04			16.2			
48	3	0.93			20.0			
50	2	1.45	21.0					
55	0	-2.70			13.0			
58	NR		< 50					
68	4	0.16			18.5			
69	NR		< 20					
70	NR				< 20			
73	4	-0.10			18.0			
75	4	-0.05			18.1			
80	2	-1.50			15.3			
81	3	-0.62				17.0		
83	4	-0.42			17.4			
86	4	-0.10			18.0			
87	2	1.19	20.5					
89	4	-0.31	17.6					
91	4	-0.31			17.6			
96	1	1.97	22.0					
97	1	1.97			22.0			
102	3	0.93			20.0			
105	4	-0.36				17.5		
107	3	0.93	20.0					
109	3	0.57	19.3					
113	4	0.21			18.6			
114	3	0.93	20.0					
119	3	0.93			20.0			
121	4	-0.10			18.0			

Lab	Rating	Z-value	1	3	4	5	6	22
127	3	-0.78						16.7
128	2	-1.30						15.7
129	3	0.93	20.0					
132	3	-0.62				17.0		
134	4	-0.08				18.1		
138	4	-0.10				18.0		
140	3	-0.62	17.0					
141	4	0.00				18.2		
142	4	0.42				19.0		
145	2	1.30				20.7		
146	4	-0.21				17.8		
151	3	-0.88					16.5	
158	3	0.67				19.5		
180	4	-0.05				18.1		
183	0	2.33	22.7					
190	3	0.57	19.3					
191	4	0.42				19.0		
196	4	0.42				19.0		
203	3	0.93	20.0					
204	0	-2.18				14.0		
212	4	0.42					19.0	
215	4	-0.05				18.1		
217	4	0.00				18.2		
219	3	-0.62				17.0		
220	3	0.52	19.2					
221	3	0.78				19.7		
224	4	0.42				19.0		
234	4	0.10				18.4		
235	4	-0.10				18.0		
236	3	-0.93				16.4		
240	4	-0.10				18.0		
241	3	-0.62	17.0					31.5
244	0	6.90						
255	4	-0.12				18.0		
256	4	0.42	19.0					
257	4	-0.10	18.0					
265	4	0.42				19.0		
273	2	1.45				21.0		
274	0	-5.60				7.4		


Table 12. Statistical summary of reported data for standard reference water sample T-143 (trace constituents)--Continued



2. AA: direct nitrous oxide		6. ICP/MS			
3. AA: graphite furnace					
4. ICP		N =	1	10	30
		Minimum =	44.0	12.5	11.0
		Maximum =		52.0	48.0
		Median =		34.8	36.3
		F-pseudosigma =		2.5	1.8
Lab	Rating	Z-value	2	3	4
1	4	-0.40		34.4	
3	0	-2.12			27.0
4	NR			< 500	
7	1	-1.56			29.4
11	4	0.21			37.0
15	3	0.51		38.3	
16	4	0.37			37.7
18	4	0.21		37.0	
23	NR		< 100		
24	4	-0.49		34.0	
26	4	0.21			37.0
28	0	-2.16		26.8	
30	4	-0.02			36.0
32	4	0.07			36.4
40	4	0.42		37.9	
42	4	-0.02			36.0
46	4	-0.02			36.0
48	4	0.35		37.6	
50	3	-0.95		32.0	
55	2	-1.42			30.0
68	4	0.21			37.0
70	NR			< 50	
75	2	-1.02			31.7
81	1	-1.88			28.0
86	2	1.02			40.5
87	4	-0.23		35.1	
97	4	-0.16		35.4	
105	3	0.95			40.2
109	0	-5.49		12.5	
119	4	-0.23			35.1
127	3	-0.93		32.1	
128	2	-1.12			31.3
132	0	-2.35			26.0
134	4	0.21			37.0
138	4	-0.05			35.9
141	4	0.28			37.3
142	3	0.51			38.3
145	2	-1.28			30.6
146	4	0.02			36.2
151	3	0.58			38.6
180	2	1.14		41.0	
191	3	0.84			39.7
196	4	0.07			36.4
212	4	0.44			38.0
215	4	0.21		37.0	
217	4	0.40			37.8
219	0	-2.12			27.0
221	4	-0.16		35.4	
224	0	2.77			48.0
234	4	-0.42		34.3	

MPV =	36.1
F-pseudosigma =	4.3
N =	54
Hu =	37.8
HI =	32.0

Table 12. Statistical summary of reported data for standard reference water sample T-143 (trace constituents)--Continued  
Na (Sodium) mg/L



1. AA: direct air

4. ICP

5. DCP

6. ICP/MS

7. Ion chromatography

12. Flame emission

MPV = 34.0

F-pseudosigma = 1.6

N = 88


Hu = 35.0

HI = 32.9

	N =	23	49	1	3	1	11
Minimum =		28.0	29.7	33.7	32.7	34.0	18.4
Maximum =		42.5	55.0		33.2		50.0
Median =		33.6	34.2				34.0
F-pseudosigma =		1.2	1.6				4.5

Lab	Rating	Z-value	1	4	5	6	7	12
1	4	-0.22		33.6				
3	0	13.24		55.0				
4	3	0.60		34.9				
7	2	1.01		35.6				
11	4	-0.09		33.8				
13	3	0.66		35.0				
15	2	-1.35		31.8				
16	3	-0.72		32.8				
18	4	-0.50		33.2				
19	4	0.22		34.3				
23	4	-0.35	33.4					
24	4	-0.41		33.3				
25	0	2.42		37.8				
26	4	0.41		34.6				
28	0	2.11		37.3				
32	3	-0.53			33.1			
33	4	-0.16			33.7			
36	0	2.55	38.0					
40	3	-0.66		32.9				
42	4	-0.03		33.9				
43	3	0.66		35.0				
46	2	1.04		35.6				
48	3	0.79		35.2				
51	4	0.22				34.3		
64	4	-0.09	33.8					
68	4	0.35		34.5				
69	3	-0.85				32.6		
70	4	0.16		34.2				
75	4	0.35		34.5				
81	3	-0.79			32.7			
83	3	-0.66		32.9				
85	4	0.35	34.5					
86	3	0.66		35.0				
87	3	-0.72		32.8				
89	4	-0.22		33.6				
97	4	0.35	34.5					
102	1	-1.92		30.9				
105	1	-1.98		30.8				
107	3	-0.72		32.8				
109	4	0.09		34.1				
110	3	-0.60		33.0				
113	0	-2.67			29.7			
114	0	-3.74	28.0					
119	4	0.16		34.2				
121	2	-1.23		32.0				
127	4	-0.47		33.2				
128	3	-0.66		32.9				
129	3	0.66	35.0					
132	4	0.28		34.4				
134	4	-0.14		33.7				

Table 12. Statistical summary of reported data for standard reference water sample T-143 (trace constituents)—Continued  
Ni (Nickel) µg/L




1. AA: direct air				6. ICP/MS			
3. AA: graphite furnace							
4. ICP							
N =	9	19	39	14			
Minimum =	58.5	30.0	31.8	50.0			
Maximum =	110.0	125.0	86.3	74.8			
Median =	71.0	69.2	71.1	70.8			
F-pseudosigma =	7.4	5.6	3.7	2.9			
Lab	Rating	Z-value	1	3	4	6	
1	4	0.02			71.1		
3	0	-2.78		57.0			
4	NR			< 200			
7	1	1.69		79.5			
11	4	0.40		73.0			
13	2	-1.31		64.4			
15	3	0.85	75.3				
16	4	0.38		72.9			
18	4	0.02		71.1			
19	0	2.72		84.7			
23	3	0.60	74.0				
24	3	-0.60		68.0			
26	3	0.71		74.6			
28	0	-7.78		31.8			
30	4	-0.20		70.0			
32	4	0.00		71.0			
36	1	-1.79		62.0			
39	3	-0.99		66.0			
40	0	-3.91		51.3			
42	4	-0.20		70.0			
46	3	0.63		74.2			
48	2	1.07		76.4			
50	0	-2.98		56.0			
58	0	7.74	110.0				
68	3	0.60		74.0			
69	0	-2.48	58.5				
70	4	-0.02		70.9			
73	4	0.40		73.0			
75	4	0.02		71.1			
76	3	0.75		74.6			
81	0	-4.17		50.0			
83	4	-0.08		70.6			
85	3	0.67		74.4			
86	2	-1.29		64.5			
87	2	1.39	78.0				
89	3	-0.69		67.5			
96	3	-0.67		67.6			
97	4	-0.40		69.0			
102	1	1.79		80.0			
105	3	-0.85		66.7			
113	4	-0.08		70.6			
114	4	-0.20	70.0				
118	0	-8.13		30.0			
119	0	6.55		104.0			
121	4	-0.20		70.0			
127	4	0.00		71.0			
128	2	-1.27		64.6			
132	3	0.79		75.0			
133	3	0.65		74.3			
134	4	-0.32		69.4			

Table 12. Statistical summary of reported data for standard reference water sample T-143 (trace constituents)—Continued

Pb (Lead)

µg/L



1. AA: direct air  
3. AA: graphite furnace  
4. ICP

5. ICP/MS


	N =	5	40	24	15
Minimum =	72.0	1.2	17.9	77.1	
Maximum =	155.0	183.0	498.9	120.0	
Median =		82.2	85.3	83.3	
F-pseudosigma =		8.3	7.2	3.3	

MPV = 83.4  
F-pseudosigma = 7.1  
N = 84  
Hu = 89.0  
HI = 79.4

Lab	Rating	Z-value	1	3	4	6
1	4	0.02			83.6	
3	3	0.73		88.6		
4	NR			< 400		
7	4	0.43		86.5		
11	4	0.42		86.4		
13	3	0.85	89.5			
15	1	1.71	95.6			
16	4	0.46		86.7		
18	4	0.01	83.5			
19	2	1.47	93.9			
23	3	-0.72	78.3			
24	0	-3.33		59.8		
26	4	-0.39	80.7			
28	0	-9.23		17.9		
30	4	-0.20	82.0			
32	4	-0.30		81.3		
34	2	-1.32	74.1			
36	3	-0.77	78.0			
39	4	-0.34	81.0			
42	4	-0.06		83.0		
46	3	-0.91	77.0			
48	4	0.14	84.4			
58	2	1.21	92.0			
68	0	14.03	183.0			
69	4	-0.43	80.4			
70	4	-0.17	82.2			
73	4	0.50		87.0		
75	1	1.99		97.6		
76	4	-0.02		83.3		
80	4	0.01	83.5			
81	4	0.50	87.0			
83	4	0.02	83.6			
85	3	-0.98	76.5			
86	4	-0.34	81.0			
87	3	-0.77	78.0			
89	4	0.25	85.2			
96	2	1.32	92.8			
97	4	-0.22	81.9			
102	1	1.63		95.0		
105	3	0.63		87.9		
109	0	-7.25		32.0		
113	4	-0.33		81.1		
114	1	-1.61	72.0			
118	0	-7.82		27.9		
119	4	-0.20		82.0		
127	3	0.53		87.2		
128	4	-0.09		82.8		
132	4	-0.41		80.5		
133	2	1.21		92.0		
134	4	-0.01		83.4		

Lab	Rating	Z-value	1	3	4	6
138	3	-0.89				77.1
140	4	0.36	86.0			
141	0	2.47		101.0		
142	4	0.50			87.0	
145	0	58.53			498.9	
146	4	0.11			84.2	
151	4	-0.09				82.8
158	0	-7.51		30.1		
180	2	1.22			92.1	
190	0	-3.19		60.8		
191	4	-0.20				82.0
193	3	0.64		88.0		
196	4	0.28			85.4	
204	4	-0.17		82.2		
212	0	5.15				120.0
213	3	0.92	90.0			
215	4	-0.20			82.0	
217	3	0.94				90.1
219	1	-1.89			70.0	
220	3	-0.71		78.4		
221	4	0.15	84.5			
224	0	8.00		140.3		
234	3	0.83		89.3		
235	4	0.50			87.0	
236	0	-2.96			62.4	
240	0	6.70				131.0
241	0	2.22		99.2		
249	3	-0.60		79.2		
255	4	-0.01			83.4	
256	0	10.08	155.0			
257	2	1.07	91.0			
259	0	-4.15			54.0	
265	3	-0.55				79.6
273	0	-2.60				65.0
274	0	-11.59			1.2	


Table 12. Statistical summary of reported data for standard reference water sample T-143 (trace constituents)—Continued  
 Sb (Antimony)  $\mu\text{g/L}$



3. AA: graphite furnace			11na. AA: hydride $\text{NaBH}_4$		
4. ICP			6. ICP/MS		
N =	19	11	13	3	
Minimum =	8.0	5.7	15.8	2.0	
Maximum =	21.2	49.7	19.9	16.4	
Median =	15.4	19.0	16.7		
F-pseudosigma =	1.5	7.5	0.7		
Lab	Rating	Z-value	3	4	6
1	4	-0.47	15.9		
3	1	1.62		19.0	
7	0	9.92		31.3	
11	4	0.13		16.8	
13	3	-0.81	15.4		
15	4	0.40	17.2		
16	4	0.00		16.6	
18	2	-1.35	14.6		
26	NR		< 20		
30	4	0.27	17.0		
32	4	-0.40		16.0	
36	0	-4.45	10.0		
39	4	0.07		16.7	
40	0	-7.35		5.7	
42	4	-0.40		16.0	
46	4	0.40	17.2		
48	0	-2.56	12.8		
68	0	-4.79	9.5		
69	2	-1.08	15.0		
70	0	3.10	21.2		
75	0	4.25		22.9	
81	0	-7.07		< 6	
89	2	1.42	18.7		
96	2	-1.01	15.1		
97	4	0.00	16.6		
102	4	0.27		17.0	
105	4	0.27			17.0
119	3	-0.94			15.2
127	3	-0.67	15.6		
128	4	-0.20			16.3
134	4	-0.15			16.4
138	4	0.13			16.8
141	4	-0.13	16.4		
142	0	2.23			19.9
146	NR		< 20		
151	4	0.40			17.2
180	0	10.12		31.6	
193	2	-1.08	15.0		
196	3	-0.54			15.8
212	4	-0.40			16.0
215	0	-5.80	8.0		
217	3	0.88			17.9
234	4	0.27	17.0		
235	NR		< 50		
236	0	22.33			49.7
240	3	0.94			18.0
241	2	-1.08	15.0		
255	0	2.54			20.4
257	0	-9.85			2.0
265	4	0.13			16.8

MPV = 16.6  
 F-pseudosigma = 1.5  
 N = 46  
 Hu = 17.2  
 HI = 15.2

Table 12. Statistical summary of reported data for standard reference water sample T-143 (trace constituents)—Continued  
Se (Selenium)  $\mu\text{g/L}$



3. AA: graphite furnace	11na. AA: hydride $\text{NaBH}_4$
4. ICP	
6. ICP/MS	


N = 27      7      13      13  
 Minimum = 5.0      6.1      9.6      8.0  
 Maximum = 31.5      174.1      12.0      19.4  
 Median = 8.8      15.0      10.8      9.4  
 F-pseudosigma = 1.5      7.2      0.8      1.3

Lab	Rating	Z-value	3	4	6	11na
1	4	-0.50	8.80			
3	0	-2.82		< 5		
7	NR			< 50		
10	4	-0.14			9.40	
11	0	2.42		13.60		
13	2	-1.48	7.20			
15	4	0.47	10.40			
16	3	0.96			11.20	
18	3	-0.87	8.20			
23	4	-0.43			8.92	
26	3	-0.64			8.57	
30	2	1.45		12.00		
32	4	0.03			9.67	
34	4	-0.23			9.25	
36	0	-2.82	5.00			
39	3	0.59			10.60	
42	2	1.45		12.00		
46	4	-0.20	9.30			
48	2	-1.11	7.80			
50	4	-0.32			9.10	
58	4	0.23	10.00			
68	0	-2.06	6.25			
69	4	-0.05	9.55			
70	NR		< 10			
73	0	9.97		26.00		
75	4	0.03			9.67	
80	4	-0.20	9.30			
86	3	0.78			10.90	
87	0	5.95			19.40	
89	3	-0.99			8.00	
96	4	0.05	9.70			
97	0	-2.36	5.75			
102	0	5.71		19.00		
105	4	0.41			10.30	
113	3	-0.59	8.65			
118	3	-0.93	8.10			
119	4	0.47			10.40	
127	4	0.41	10.30			
128	3	0.53			10.50	
133	3	-0.87	8.20			
134	3	0.62			10.65	
138	4	-0.03			9.58	
141	4	0.41	10.30			
142	3	0.53			10.50	
144	4	-0.20	9.30			
146	NR		< 10			
151	3	0.72			10.80	
180	NR		< 50.1			
190	0	13.32	31.50			
191	2	1.08			11.40	

MPV = 9.63  
 F-pseudosigma = 1.64  
 N = 60  
 Hu = 10.80  
 Hi = 8.59

Lab	Rating	Z-value	3	4	6	11na
193	1	-1.60	7.00			
196	3	0.72			10.80	
212	2	1.45			12.00	
215	4	-0.32	9.10			
217	4	0.09			9.77	
220	2	-1.11	7.80			
221	4	-0.28	9.16			
224	0	-2.15			6.10	
234	4	-0.04	9.56			
236	0	100.17			174.10	
240	0	3.27			15.00	
241	1	-1.94	6.44			
255	0	-2.62	5.33			
259	3	-0.62				8.60
265	2	1.36				11.85

Table 12. Statistical summary of reported data for standard reference water sample T-143 (trace constituents)—Continued  
 $\text{SiO}_2$  (Silica) mg/L



2. AA: direct nitrous oxide

4. ICP

5. DCP

6. ICP/MS  
 22mb. Color: molybdate blue

MPV = 23.4

F-pseudosigma = 1.7

N = 56


Hu = 24.7

Hi = 22.3

	N =	1	32	1	3	19
Minimum =	24.7	5.8	22.2	23.3	20.6	
Maximum =		45.9		26.1	31.3	
Median =		23.0		23.4		
F-pseudosigma =		2.3		1.4		

Lab	Rating	Z-value	2	4	5	6	22mb
1	3	-0.72	22.1				
3	3	0.60	24.4				
4	0	2.10	27.0				
7	4	0.26	23.8				
11	0	-3.01	18.1				
13	4	-0.26	22.9				
15	2	1.29					25.6
24	4	0.43	24.1				
25	2	1.23	25.5				
26	4	-0.03	23.3				
32	4	0.37			24.0		
33	3	-0.66			22.2		
42	3	0.83	24.8				
43	4	-0.20	23.0				
55	4	0.09	23.5				
70	3	-0.55			22.4		
76	1	1.58			26.1		
81	4	0.14			23.6		
83	2	-1.18	21.3				
87	4	0.03			23.4		
89	2	-1.23			21.2		
97	4	-0.20			23.0		
104	4	-0.39			22.7		
105	2	-1.23	21.2				
107	1	1.69			26.3		
110	2	1.19			25.4		
113	4	-0.32			22.8		
118	2	1.22			25.5		
119	4	-0.20	23.0				
121	4	-0.37	22.7				
127	2	-1.12	21.4				
128	4	0.37	24.0				
129	4	0.20			23.7		
134	4	-0.28	22.9				
138	2	-1.06			21.5		
140	4	-0.32			22.8		
142	1	1.92	26.7				
145	1	1.52	26.0				
158	3	0.72	24.6				
190	4	0.09			23.5		
191	4	-0.03			23.3		
203	4	0.21			23.7		
204	4	0.03			23.4		
212	3	0.55	24.3				
215	0	12.94	45.9				
217	1	-1.52	20.7				
219	4	-0.20	23.0				
234	4	-0.49	22.5				
235	2	1.46	25.9				
236	0	-5.19	14.3				


Table 12. Statistical summary of reported data for standard reference water sample T-143 (trace constituents)—Continued  
 Sr (Strontium)       $\mu\text{g/L}$



0. Other			4. ICP					
2. AA: direct nitrous oxide			5. DCP					
3. AA: graphite furnace			6. ICP/MS					
	N =	2	1	1	34	1	6	
	Minimum =	265	318	255	249	308	262	
	Maximum =	613			356		330	
	Median =				302			
	F-pseudosigma =				12			
Lab	Rating	Z-value	0	2	3	4	5	6
1	4	-0.32				301		
3	4	0.20				309		
4	2	1.14				323		
7	4	-0.20				303		
11	4	-0.47				299		
16	1	-1.62				282		
18	2	-1.07				290		
24	4	0.00				306		
25	2	1.21				324		
28	3	0.87				319		
32	4	0.34				311		
33	4	0.13				308		
40	0	-3.83				249		
42	4	-0.47				299		
55	4	0.00				306		
68	4	0.27				310		
70	4	0.20				309		
81	0	-2.95				262		
85	3	0.81				318		
86	4	-0.34				301		
97	0	-3.42				255		
102	0	3.36				356		
105	2	-1.41				285		
109	0	-2.77	265					
113	3	-0.74				295		
121	4	-0.40				300		
127	2	-1.41				285		
134	3	-0.74				295		
138	3	-0.60				297		
142	3	0.61				315		
145	1	1.79				333		
151	3	0.67				316		
190	0	20.60	613					
191	4	-0.07				305		
196	4	0.40				312		
212	1	1.61				330		
217	3	-0.74				295		
218	3	0.81	318					
219	4	0.27				310		
234	4	0.27				310		
235	4	0.34				311		
236	3	-0.62				297		
240	3	-0.81				294		
265	4	-0.40				300		
273	1	2.01				336		

MPV = 306  
 F-pseudosigma = 15  
 N = 45  
 Hu = 315  
 Hi = 295

Table 12. Statistical summary of reported data for standard reference water sample T-143 (trace constituents)—Continued  
 Tl (Thallium)  $\mu\text{g/L}$



3. AA: graphite furnace

4. ICP


6. ICP/MS

	N =	16	2	15
Minimum =	5.0	9.0	7.4	
Maximum =	11.0	14.0	53.0	
Median =	9.6		10.1	
F-pseudosigma =	1.9		0.5	

MPV =	10.0
F-pseudosigma =	1.0
N =	33
Hu =	10.4
HI =	9.0

Lab	Rating	Z-value	3	4	6
1	4	0.39	10.4		
3	NR		< 10		
11	0	3.85		14.0	
13	4	-0.19	9.8		
15	1	-1.95	8.0		
16	4	0.16		10.2	
18	4	-0.29	9.7		
23	0	-5.00	< 5		
32	4	-0.40		9.6	
39	3	0.77		10.8	
42	0	-5.00		< 5	
46	3	0.87	10.9		
48	1	-2.02	7.9		
69	3	0.96	11.0		
70	0	-2.12	7.8		
76	0	-2.54		7.4	
81	0	41.43		53.0	
89	NR		< 10		
97	3	0.87	10.9		
113	4	-0.40	9.6		
119	4	0.10		10.1	
128	4	0.00		10.0	
134	4	0.16	10.2		
138	3	-0.59		9.4	
141	NR		< 50		
142	4	0.39		10.4	
146	NR		< 10		
151	4	0.29		10.3	
180	NR		< 32.1		
191	4	-0.10		9.9	
193	0	-2.89	7.0		
196	4	0.10		10.1	
212	4	0.00		10.0	
213	3	-0.93	9.0		
215	0	-3.00	< 7		
217	3	0.87		10.9	
234	0	-2.35	7.6		
235	0	-4.87	5.0		
240	3	-0.96		9.0	
241	4	0.48	10.5		
265	2	-1.06		8.9	

Table 12. Statistical summary of reported data for standard reference water sample T-143 (trace constituents)—Continued




- 4. ICP
- 6. ICP/MS
- 0. Other

N =	1	8
Minimum =	11.0	10.8
Maximum =		14.0
Median =		12.0
eudosigma =		1.0

MPV =	12.0
F-pseudosigma =	0.9
N =	9
Hu =	12.2
HI =	11.0

F-pseudosigma =			1.0
Lab	Rating	Z-value	
			4
			6
1	2	-1.34	10.8
7	NR		< 120
16	4	0.00	12.0
30	2	-1.12	11.0
75	NR		< 100
119	4	0.00	12.0
142	4	0.22	12.2
196	3	-0.56	11.5
212	0	2.25	14.0
217	2	1.12	13.0
265	2	-1.12	11.0

Table 12. Statistical summary of reported data for standard reference water sample T-143 (trace constituents)—Continued  
 V (Vanadium)  $\mu\text{g/L}$



2. AA: direct nitrous oxide

6. ICP/MS

3. AA: graphite furnace

4. ICP

	N =	1	6	36	11
Minimum =	93.0	22.0	20.8	19.0	
Maximum =		34.9	39.7	32.2	
Median =			30.0	30.0	
F-pseudosigma =			2.5	1.3	

MPV =	30.0
F-pseudosigma =	3.0
N =	54
Hu =	32.0
HI =	28.0


Lab	Rating	Z-value	2	3	4	6
1	4	0.37		31.1		
3	0	-2.36		23.0		
4	NR			< 50		
7	3	0.81		32.4		
11	4	0.00		30.0		
13	0	3.27		39.7		
15	0	-3.10		20.8		
16	4	0.10			30.3	
18	3	-0.67		28.0		
24	1	-2.02		24.0		
25	0	-2.36		23.0		
26	4	0.00		30.0		
28	4	0.07		30.2		
30	4	0.00		30.0		
32	4	0.17			30.5	
40	3	-0.78		27.7		
42	3	0.67			32.0	
46	3	0.74				32.2
48	3	-0.78		27.7		
50	0	-2.70		22.0		
55	3	-0.57		28.3		
68	2	-1.35		26.0		
70	NR			< 50		
75	4	-0.13		29.6		
81	0	-3.71			19.0	
85	4	0.13		30.4		
86	4	0.44			31.3	
89	3	0.78		32.3		
97	1	1.65		34.9		
102	2	1.01		33.0		
105	4	-0.20			29.4	
119	3	-0.67			28.0	
121	3	0.67		32.0		
127	4	0.37		31.1		
128	3	-0.78		27.7		
134	4	-0.11			29.7	
138	4	-0.17			29.5	
141	4	0.27		30.8		
142	4	-0.13			29.6	
145	0	2.46		37.3		
146	4	0.00		30.0		
158	4	0.20		30.6		
180	4	0.24		30.7		
191	3	0.74			32.2	
196	4	-0.03			29.9	
212	3	0.67			32.0	
217	4	-0.47		28.6		
219	3	-0.67		28.0		
224	1	1.52		34.5		
234	3	-0.81		27.6		

Lab	Rating	Z-value	2	3	4	6
235	1	2.02		36.0		
236	4	0.37			31.1	
241	1	1.59			34.7	
255	4	-0.10				29.7
257	0	21.25	93.0			
265	4	0.00				30.0

Table 12. Statistical summary of reported data for standard reference water sample T-143 (trace constituents)—Continued

Zn (Zinc)

µg/L



1. AA: direct air		6. ICP/MS			
3. AA: graphite furnace					
4. ICP					
N =		15	7	40	13
Minimum =		15.5	5.8	13.8	15.7
Maximum =		50.0	184.0	44.0	22.0
Median =		21.3	19.5	19.6	19.3
F-pseudosigma =		2.8	16.6	1.7	0.9

Lab	Rating	Z-value	1	3	4	6
1	3	-0.57			18.7	
3	2	-1.30		17.1		
4	NR		< 200			
7	4	-0.40		19.1		
10	4	0.45	21.0			
11	1	-1.80		16.0		
13	4	0.04		20.1		
15	3	-0.81		18.2		
16	4	-0.22			19.5	
18	NR		< 100			
19	4	0.49		21.1		
23	NR		< 20			
24	3	-0.94		17.9		
26	4	0.04		20.1		
28	1	-1.57		16.5		
30	4	0.00		20.0		
32	4	-0.04			19.9	
36	3	0.90	22.0			
40	0	-2.79			13.8	
42	4	-0.45			19.0	
48	0	-6.82			< 5	
58	NR		< 50			
68	3	0.67		21.5		
69	NR		< 50			
70	4	0.04		20.1		
73	4	0.00		20.0		
75	4	-0.36		19.2		
80	1	-2.02	15.5			
81	2	-1.35			17.0	
83	3	-0.76			18.3	
86	4	0.22		20.5		
87	1	1.80	24.0			
89	0	9.31			40.7	
96	3	0.90	22.0			
97	0	-2.43			14.6	
102	4	0.00		20.0		
105	4	-0.31			19.3	
113	4	-0.27			19.4	
114	3	-0.90	18.0			
118	1	1.80	24.0			
119	0	10.79			44.0	
121	4	0.00			20.0	
127	1	-1.66			16.3	
128	4	0.09			20.2	
132	4	0.00			20.0	
133	4	0.22			20.5	
134	4	0.19			20.4	
138	2	-1.21				17.3
140	4	0.00	20.0			
141	4	-0.13				19.7


MPV = 20.0  
F-pseudosigma = 2.2  
N = 75  
Hu = 21.0  
Hi = 18.0

Lab	Rating	Z-value	1	3	4	6
142	1	-1.93			15.7	
145	0	2.29			25.1	
146	NR		< 20			
151	4	0.09			20.2	
158	3	-0.63			18.6	
180	4	0.18			20.4	
190	3	0.58	21.3			
191	3	-0.54			18.8	
193	NR		< 50			
196	4	-0.09			19.8	
204	3	-0.90			18.0	
212	3	0.90			22.0	
213	0	4.18	29.3			
215	4	0.00			20.0	
217	3	-0.94			17.9	
219	2	-1.35			17.0	
220	4	0.00	20.0			
221	0	6.66		34.8		
224	4	-0.45			19.0	
234	0	-2.07			15.4	
235	4	-0.22			19.5	
236	1	-2.02			15.5	
240	2	1.35			23.0	
241	1	-1.80			16.0	
249	0	73.75			184.0	
253	0	13.49	50.0			
255	3	-0.54			18.8	
256	4	0.18	20.4			
257	4	0.45	21.0			
259	0	13.49	50.0			
265	3	0.90			22.0	
273	0	-2.25			15.0	
274	0	-6.37			5.8	

Table 13. Statistical summary of reported data for standard reference water sample T-145 (trace constituents)

Definition of analytical methods, abbreviations, and symbols			
<u>Analytical methods</u>			
0. Other/Not reported			
1. AA: direct, air	= atomic absorption: direct,air		
2. AA: direct, N <sub>2</sub> O	= atomic absorption: direct,nitrous oxide		
3. AA: graphite furnace	= atomic absorption: graphite furnace		
4. ICP	= inductively coupled plasma		
5. DCP	= direct current plasma		
6. ICP/MS	= inductively coupled plasma/mass spectrometry		
10. AA: extraction	= atomic absorption: extraction [chelating agent(s) specified]		
11. AA: hydride	= atomic absorption: hydride [reducing agent specified]		
12. Flame emission	= flame emission		
22. Color:	= colorimetric [color reagent specified]		
<u>Abbreviations and symbols</u>			
N =	number of samples		
MPV =	most probable value		
F-pseudosigma =	nonparametric statistic deviation		
Hu =	upper hinge value		
Hi =	lower hinge value		
µg/L =	micrograms per liter		
mg/L =	milligrams per liter		
Lab =	laboratory code number		
NR =	not rated, less than value reported		
< =	less than		
<u>Constituent</u>			
Ag	Silver	<u>page</u>	74
Al	Aluminium		75
As	Arsenic		76
B	Boron		77
Ba	Barium		78
Be	Beryllium		79
Ca	Calcium		80
Cd	Cadmium		81
Co	Cobalt		82
Cr	Chromium		83
Cu	Copper		84
Fe	Iron		85
K	Potassium		86
Li	Lithium		87
<u>Constituent</u>			
Mg	Magnesium	<u>page</u>	88
Mn	Manganese		89
Mo	Molybdenum		90
Na	Sodium		91
Ni	Nickel		92
Pb	Lead		93
Sb	Antimony		94
Se	Selenium		95
SiO <sub>2</sub>	Silica		96
Sr	Strontium		97
Tl	Thallium		98
U	Uranium		99
V	Vanadium		100
Zn	Zinc		101

Table 13. Statistical summary of reported data for standard reference water sample T-145 (trace constituents)—Continued




1. AA: direct air	6. ICP/MS
3. AA: graphite furnace	
4. ICP	
	N = 4 26 17 14
Minimum =	8.20 5.26 4.70 6.60
Maximum =	15.00 10.50 52.00 7.71
Median =	7.38 7.93 7.50
F-pseudosigma =	0.66 1.56 0.57

MPV = 7.55  
 F-pseudosigma = 0.92  
 N = 61  
 Hu = 8.13  
 HI = 6.89

Lab	Rating	Z-value	1	3	4	6
1	4	0.00			7.55	
3	1	-1.89			6.00	
4	4	0.49			8.00	
5	2	-1.09		6.55		
7	0	2.56			9.90	
11	4	0.49			8.00	
13	3	0.63			8.13	
15	NR				< 10	
16	4	0.05			7.60	
16	3	0.92			8.40	
23	4	0.26		7.79		
26	4	0.23		7.76		
30	4	0.05			7.60	
32	4	-0.16			7.40	
36	0	8.10	15.00			
42	4	-0.05			7.50	
46	4	-0.49		7.10		
48	4	0.05		7.60		
58	1	1.58		9.00		
59	4	0.05			7.60	
68	0	2.77		10.10		
69	3	-0.63		6.97		
70	NR			< 10		
75	1	-1.56			6.12	
85	3	0.71	8.20			
87	0	5.93	13.00			
89	4	0.05		7.60		
96	3	0.71		8.20		
97	3	-0.99		6.64		
102	0	48.36			52.00	
105	3	-0.84				6.78
107	4	-0.16		7.40		
113	3	-0.82		6.80		
114	NR		< 10			
118	3	-0.71		6.90		
119	3	-0.71		6.90		
128	4	-0.05			7.50	
133	1	-1.58			6.10	
134	3	-0.72			6.89	
138	3	-0.52			7.07	
141	0	5.17		12.30		
142	2	-1.03			6.60	
146	NR			< 10		
149	4	0.05		7.60		
151	4	0.10			7.64	
180	2	-1.36			6.30	
190	2	1.34		8.78		
193	1	-1.69		6.00		
196	4	0.17			7.71	
212	3	-0.82			6.80	

Lab	Rating	Z-value	1	3	4	6
213	4	-0.11		7.45		
215	0	2.67		10.00		
217	3	-0.78				6.83
221	4	-0.21		7.36		
234	4	-0.23		7.34		
235	3	-0.60			7.00	
236	0	-3.10			4.70	
241	3	-0.71		6.90		
245	1	-1.88		5.82		
249	0	-2.49		5.26		
255	4	0.41			7.93	
257	0	5.93	13.00			
259	0	3.21		10.50		
265	4	0.11				7.65
273	0	11.37			18.00	

Table 13. Statistical summary of reported data for standard reference water sample T-145 (trace constituents)—Continued  
 Al (Aluminum)  $\mu\text{g/L}$



1. AA: direct air			4. ICP		
2. AA: direct nitrous oxide			5. DCP		
3. AA: graphite furnace			6. ICP/MS		
N =	1	2	13	31	1
Minimum =	117.0	50.0	49.2	37.5	100.0
Maximum =			300.0	79.8	74.4
Median =				64.8	70.1
F-pseudosigma =				9.4	16.2
					4.1
Lab	Rating	Z-value	1	2	3
1	3	-0.59			60.1
3	0	2.77		98.0	
4	4	0.04		68.0	
5	3	0.60		74.2	
7	2	1.48		83.8	
11	3	-0.69		60.0	
13	2	1.11	79.8		
15	2	1.32		82.1	
16	3	-0.77			59.1
18	NR		< 100		
23	NR		< 50		
26	4	-0.16		65.8	
28	0	-2.74		37.5	
32	4	0.14			69.1
33	0	2.95		100.0	
36	0	21.18	300.0		
40	3	-0.80		58.8	
42	4	-0.05			67.0
46	1	1.75		86.8	
48	1	-1.68	49.2		
58	0	4.50	117.0		
59	3	0.62			74.4
68	0	2.82		98.5	
69	1	-1.60	50.0		
70	NR		< 100		
73	3	-0.51		62.0	
75	4	-0.04		67.2	
76	4	-0.08			66.7
81	1	-1.51			51.0
83	3	-0.97		57.0	
85	NR		< 100		
89	4	-0.27		64.6	
97	4	0.46		72.7	
102	4	0.13		69.0	
105	4	-0.15			66.0
107	4	-0.24		65.0	
113	4	0.05		68.1	
118	NR		< 2000		
119	3	-0.51			62.0
128	4	0.31			71.0
132	4	0.49		73.0	
134	4	0.37		71.7	
138	4	-0.27			64.6
141	NR		< 100		
142	4	-0.12		66.3	
145	0	3.61		107.2	
146	NR		< 200		
149	3	-0.69	60.0		
151	4	-0.32			64.1
158	4	0.36		71.5	


MPV = 67.6  
 F-pseudosigma = 11.0  
 N = 64  
 Hu = 76.1  
 HI = 61.3

Lab	Rating	Z-value	1	2	3	4	5	6
180	NR					< 40.6		
190	3	-0.54			61.7			
191	4	0.04					68.0	
196	4	-0.04					67.2	
203	3	-0.61				60.9		
204	2	1.08				79.4		
212	3	0.58					74.0	
215	0	5.41				127.0		
217	2	1.30				81.9		
219	4	0.04				68.0		
221	4	0.08				68.5		
224	2	-1.16				54.9		
234	4	-0.05				67.0		
235	2	-1.42				52.0		
236	4	0.33				71.2		
240	0	4.32				115.0		
241	2	-1.20				54.4		
249	3	0.93				77.8		
255	3	-0.87				58.1		
257	1	-1.60			50.0			
265	4	-0.06					66.9	
273	0	8.15					157.0	

Table 13. Statistical summary of reported data for standard reference water sample T-145 (trace constituents)—Continued

As (Arsenic)

µg/L



3. AA: graphite furnace	11na. AA: hydride $\text{NaBH}_4$
4. ICP	
6. ICP/MS	


N = 32    10    15    9  
 Minimum = 6.90    6.10    9.14    9.00  
 Maximum = 12.80    42.70    11.20    10.00  
 Median = 9.88    11.50    9.95    9.44  
 F-pseudosigma = 1.22    3.34    0.66    0.41

Lab	Rating	Z-value	3	4	6	11na
1	4	0.02	9.90			
3	4	0.12		10.00		
4	0	3.01		13.00		
5	4	-0.02	9.86			
7	NR		< 120			
10	4	0.12			10.00	
11	4	0.12		10.00		
13	3	0.60	10.50			
15	NR		< 100			
16	4	0.50			10.40	
18	3	-0.66	9.20			
23	0	2.81	12.80			
26	4	-0.25			9.62	
30	3	-0.85		9.00		
32	4	-0.11			9.77	
34	4	-0.27				9.60
36	1	-1.81	8.00			
42	2	1.08			11.00	
46	3	-0.75	9.10			
48	4	-0.08	9.80			
58	1	2.04	12.00			
59	4	0.12			10.00	
68	0	-2.87	6.90			
69	4	0.31	10.20			
70	NR		< 10			
73	0	3.01		13.00		
75	4	-0.42			9.44	
76	3	0.69			10.60	
80	3	0.69	10.60			
81	4	0.12	10.00			
86	4	0.07			9.95	
87	4	-0.46			9.40	
89	3	-0.85			9.00	
96	4	-0.17	9.70			
97	4	0.50	10.40			
102	0	-3.64		6.10		
105	4	0.40			10.30	
109	2	-1.33	8.50			
113	1	1.56	11.50			
118	3	0.69	10.60			
119	3	-0.85			9.00	
128	4	0.07			9.95	
133	0	-2.49	7.30			
134	3	-0.78			9.07	
138	4	-0.36			9.51	
141	2	-1.19	8.65			
142	4	-0.26			9.61	
144	4	-0.27	9.60			
145	0	9.08		19.30		
146	0	3.49		13.50		

MPV = 9.88  
 F-pseudosigma = 1.04  
 N = 66  
 Hu = 10.60  
 HI = 9.20

Lab	Rating	Z-value	3	4	6	11na
151	3	-0.57			9.29	
190	2	1.27	11.20			
191	2	1.27			11.20	
193	1	-1.81	8.00			
196	4	-0.03			9.85	
204	2	1.46	11.40			
212	4	-0.27			9.60	
213	2	-1.23	8.60			
215	2	1.08	11.00			
217	3	-0.71			9.14	
220	4	-0.37	9.50			
221	3	-0.55	9.31			
224	1	-1.72			8.10	
234	3	0.98	10.90			
236	0	31.62			42.70	
241	0	-2.39	7.40			
249	2	1.08	11.00			
255	4	0.12	10.00			
265	2	1.08			11.00	


Table 13. Statistical summary of reported data for standard reference water sample T-145 (trace constituents)—Continued  
 B (Boron)  $\mu\text{g/L}$



4. ICP			
6. ICP/MS			
22az. Color: azomethine			
N =	30	5	1
Minimum =	22.4	41.8	125.0
Maximum =	150.0	52.0	
Median =	45.6		
F-pseudosigma =	5.8		
Lab	Rating	Z-value	
1	4	0.02	45.7
3	3	-0.96	40.0
5	3	0.62	49.2
11	3	0.76	50.0
15	NR	< 50	
16	0	5.40	77.0
18	NR	< 50	
24	2	-1.50	36.9
26	2	-1.37	37.6
28	4	-0.02	45.5
42	4	0.24	47.0
46	0	-3.99	22.4
48	NR	< 100	
68	0	13.64	125.0
70	NR	< 50	
85	3	-0.55	42.4
86	4	0.02	45.7
119	3	-0.62	42.0
128	3	-0.81	40.9
129	0	13.64	125.0
132	0	-3.02	28.0
134	4	0.26	47.1
138	3	-0.65	41.8
141	2	-1.19	38.7
142	4	0.17	46.6
145	4	-0.15	44.7
158	4	0.38	47.8
180	3	0.55	48.8
191	4	-0.27	44.0
212	2	1.10	52.0
215	0	12.96	121.0
217	4	0.05	45.9
219	4	0.24	47.0
234	4	-0.26	44.1
235	3	-0.96	40.0
236	2	-1.41	37.4
240	1	-1.99	34.0
255	4	0.21	46.8
265	4	-0.27	44.0
273	0	17.94	150.0

MPV = 45.6  
 F-pseudosigma = 5.8  
 N = 36  
 Hu = 48.3  
 HI = 40.5

Table 13. Statistical summary of reported data for standard reference water sample T-145 (trace constituents)—Continued  
Ba (Barium)  $\mu\text{g/L}$




1. AA: direct air	4. ICP
2. AA: direct nitrous oxide	5. DCP
3. AA: graphite furnace	6. ICP/MS
N = 0	N = 0
Minimum = < 0.05	29.2
Maximum = 44.5	31.0
Median = 40.0	41.0
F-pseudosigma = 3.4	31.9
	42.0
	1
	2
	3
	4
	5
	6

MPV = 37.1  
F-pseudosigma = 1.9  
N = 63  
Hu = 38.5  
HI = 36.0

Lab	Rating	Z-value	1	2	3	4	5	6
1	4	-0.19						36.8
3	4	-0.32				36.5		
4	4	-0.05			37.0			
5	3	-0.59			36.0			
7	3	0.76			38.5			
11	3	-0.59		36.0				
13	2	1.46		39.8				
15	1	-1.56		34.2				
16	4	0.49				38.0		
18	4	-0.49		36.2				
24	2	-1.24		34.8				
25	3	0.65		38.3				
26	4	0.27		37.6				
28	3	0.81		38.6				
30	3	-0.59		36.0				
32	3	-0.92			35.4			
33	0	2.10			41.0			
36	0	-20.02 < 0.05						
40	3	-0.97			35.3			
46	4	0.27			37.6			
48	0	3.29		43.2				
59	4	-0.05			37.0			
68	3	0.76			38.5			
70	NR				< 50			
75	4	0.05			37.2			
81	2	-1.13			35.0			
83	3	-0.86			35.5			
85	4	0.32			37.7			
86	4	-0.16			36.8			
87	1	1.56		40.0				
89	NR				< 50			
96	NR				< 100			
97	0	-4.26		29.2				
102	0	2.64			42.0			
105	2	-1.24			34.8			
107	2	1.03		39.0				
113	4	-0.05			37.0			
119	2	1.03			39.0			
121	3	-0.59			36.0			
128	0	-2.81			31.9			
133	4	-0.38			36.4			
134	3	-0.81			35.6			
138	4	-0.38			36.4			
141	3	0.59			38.2			
142	4	0.05			37.2			
145	1	2.00			40.8			
146	3	0.76			38.5			
149	1	1.56		40.0				
151	4	0.38			37.8			
158	3	0.65			38.3			

Lab	Rating	Z-value	1	2	3	4	5	6
180	4	-0.49						36.2
191	4	0.49						38.0
196	4	-0.43						36.3
204	2	1.24						39.4
212	0	2.64						42.0
215	2	1.03						39.0
217	4	0.00						37.1
219	1	-1.67						34.0
224	0	-2.10						33.2
234	3	0.65						38.3
235	1	1.56						40.0
236	4	-0.43						36.3
240	0	-3.29						31.0
241	0	3.99					44.5	
255	4	-0.16						36.8
259	2	-1.13						35.0
265	4	0.49						38.0

Table 13. Statistical summary of reported data for standard reference water sample T-145 (trace constituents)—Continued  
 Be (Beryllium)  $\mu\text{g/L}$




1. AA: direct air	4. ICP
2. AA: direct nitrous oxide	6. ICP/MS
3. AA: graphite furnace	0. Other
N =	1      0      9      37      15
Minimum =	10.00      < 10      7.20      2.30      8.00
Maximum =	11.00      13.20      13.00
Median =	8.93      9.00      9.35
F-pseudosigma =	0.68      0.39      0.83

MPV = 9.04  
 F-pseudosigma = 0.70  
 N = 62  
 Hu = 9.64  
 HI = 8.70

Lab	Rating	Z-value	1	2	3	4	6	
1	4	-0.14				8.94		
3	4	0.24		9.20				
5	4	-0.28		8.84				
7	4	-0.48		8.70				
11	4	-0.05		9.00				
13	4	0.19		9.17				
15	4	-0.48		8.70				
16	3	-0.91			8.40			
18	4	0.24		9.20				
25	2	-1.34		8.10				
26	4	0.19		9.17				
30	0	-2.06		7.60				
32	3	0.88			9.65			
36	0	2.82		11.00				
40	0	-9.67			2.30			
42	0	5.69			13.00			
46	4	-0.11			8.96			
48	2	1.38			10.00			
59	2	1.10				9.80		
68	4	0.38			9.30			
69	1	-1.57		7.94				
70	4	0.19			9.17			
75	4	0.09			9.10			
76	1	1.67				10.20		
81	2	-1.49				8.00		
83	4	-0.34			8.80			
85	3	0.87			9.64			
86	4	-0.35			8.79			
89	2	-1.20			8.20			
96	2	1.38	10.00					
97	3	0.57		9.43				
102	4	-0.34			8.80			
105	2	1.38				10.00		
113	1	1.53			10.10			
114	NR		< 10					
119	4	-0.17		8.92				
121	2	1.38			10.00			
128	3	-0.91			8.40			
133	3	0.52			9.40			
134	4	0.14			9.13			
138	4	0.27			9.22			
141	4	0.05			9.07			
142	4	0.45			9.35			
144	0	-12.96	< 0.01		9.90			
145	2	1.24			9.90			
146	2	-1.21			8.19			
151	4	-0.09			8.97			
158	2	1.24			9.90			
180	4	-0.48			8.70			
191	2	1.38			10.00			

Table 13. Statistical summary of reported data for standard reference water sample T-145 (trace constituents)—Continued  
 Ca (Calcium) mg/L




0. Other		4. ICP					
1. AA: direct air		6. ICP/MS					
2. AA: direct nitrous oxide		20. Titrate: colorimetric					
N =	3	16	6	52	3	4	
Minimum =	30.3	12.4	24.0	12.1	30.1	0.0	
Maximum =	32.5	35.0	34.0	33.5	31.5	32.6	
Median =		30.0		31.0			
F-pseudosigma =		1.4		1.3			

Lab	Rating	Z-value	0	1	2	4	6	20
1	2	-1.32				29.0		
3	4	0.32				31.1		
4	4	-0.16				30.5		
5	1	-1.90				28.3		
7	3	0.88				31.8		
11	3	0.79				31.7		
13	1	1.98				33.2		
15	1	-1.67				28.6		
16	4	-0.48				30.1		
18	4	-0.08				30.6		
19	4	0.24				31.0		
23	0	-14.52		12.4				
24	3	-0.56				30.0		
25	4	0.32				31.1		
26	4	-0.08				30.6		
28	3	0.63				31.5		
30	2	1.03		32.0				
32	3	0.63				31.5		
33	4	-0.32	30.3					
36	3	-0.56		30.0				
42	0	-14.76				12.1		
43	4	0.24				31.0		
46	2	1.03				32.0		
48	3	0.71				31.6		
59	3	-0.56				30.0		
69	4	-0.40		30.2				
70	3	0.71				31.6		
75	3	-0.56		30.0				
81	4	-0.48				30.1		
83	3	-0.56				30.0		
85	4	0.24		31.0				
86	3	0.87				31.8		
87	2	-1.19		29.2				
89	3	-0.95		29.5				
97	2	-1.27		29.1				
102	1	1.83				33.0		
105	4	-0.32				30.3		
107	4	0.16		30.9				
109	1	-1.59		28.7				
113	1	1.98				33.2		
114	0	-5.32		24.0				
119	3	0.79				31.7		
121	3	-0.56				30.0		
128	2	1.03				32.0		
129	0	2.62		34.0				
132	4	0.08				30.8		
133	4	-0.24				30.4		
134	3	0.60				31.5		
138	3	0.63				31.5		
140	0	3.41		35.0				

MPV = 30.7  
 F-pseudosigma = 1.3  
 N = 84  
 Hu = 31.7  
 Hi = 30.0

Lab	Rating	Z-value	0	1	2	4	6	20
141	2	1.03				32.0		
142	4	-0.50				30.1		
145	1	1.67				32.8		
146	3	-0.87				29.6		
158	1	1.51				32.6		
180	4	0.24				31.0		
190	4	-0.32	30.3					30.6
191	4	-0.08				29.4		
193	2	-1.03						
203	2	-1.29	29.1					
204	3	0.79				31.7		
212	3	0.95				31.9		
215	3	0.79				31.7		
217	0	-3.02				26.9		
218	4	0.20				31.0		
219	3	-0.56				30.0		
220	4	0.24				31.0		
221	2	1.35				32.4		
224	1	-1.77				28.5		
234	4	0.24				31.0		
235	0	2.22				33.5		
236	4	-0.37				30.2		
240	1	-1.83				28.4		
241	1	-1.83				28.4		
255	4	-0.42				30.2		
257	0	2.62				34.0		
261	3	-0.51					30.1	
265	4	0.00				30.7		
268	2	-0.92				29.5		
270	2	1.43	32.5					
271	1	1.51				32.6		
272	0	-13.23				14.0		
273	2	1.35				32.4		
274	NR	-24.36				0.0		

Table 13. Statistical summary of reported data for standard reference water sample T-145 (trace constituents)—Continued  
 Cd (Cadmium)  $\mu\text{g/L}$



1. AA: direct air  
 3. AA: graphite furnace  
 4. ICP


6. ICP/MS

MPV = 9.33  
 F-pseudosigma = 0.82  
 N = 80  
 Hu = 10.00  
 HI = 8.90

	N =	4	29	34	13
Minimum =	7.60	4.60	5.90	9.00	
Maximum =	16.00	37.00	17.20	11.00	
Median =		9.35	9.23	9.39	
F-pseudosigma =		0.86	0.89	0.56	

Lab	Rating	Z-value	1	3	4	6
1	3	0.56	9.79			
3	3	0.94		10.10		
4	4	0.21		9.50		
5	3	-0.53	8.90			
7	0	4.50	13.00			
11	4	-0.40	9.00			
13	4	0.13	9.44			
15	0	9.65	17.20			
16	4	-0.28		9.10		
18	4	-0.16	9.20			
23	4	-0.21	9.16			
24	0	-2.12	7.60			
26	2	1.31	10.40			
28	1	-1.51	8.10			
30	4	0.33	9.60			
32	4	-0.18		9.18		
36	3	-0.90	8.60			
40	2	-1.02		8.50		
42	4	-0.40		9.00		
46	3	-0.60	8.84			
48	4	-0.16	9.20			
58	0	33.93	37.00			
59	4	-0.40		9.00		
68	3	0.82		10.00		
69	2	1.31	10.40			
70	3	-0.90	8.60			
73	4	-0.40		9.00		
75	4	-0.42	8.99			
80	NR	< 10				
81	3	0.82	10.00			
83	3	-0.65		8.80		
85	0	-2.12	7.60			
86	4	-0.16		9.20		
87	0	3.27	12.00			
89	4	0.16	9.46			
96	3	-0.53	8.90			
97	4	-0.16	9.20			
102	3	0.70		9.90		
105	3	0.65		9.86		
113	4	0.47	9.71			
114	NR	< 10				
118	1	1.56	10.60			
119	4	-0.16	9.20			
121	1	-1.63		8.00		
128	3	0.94		10.10		
132	2	-1.02	8.50			
133	1	1.56	10.60			
134	4	-0.09	9.26			
138	4	-0.31		9.08		
140	4	-0.40	9.00			

Table 13. Statistical summary of reported data for standard reference water sample T-145 (trace constituents)—Continued  
Co (Cobalt)




1. AA: direct air	6. ICP/MS
3. AA: graphite furnace	0. Other
4. ICP	0. Other
N =	1      7      32      13
Minimum =	19.0      9.0      4.8      8.0
Maximum =	10.6      14.6      14.0
Median =	10.0      10.2      9.7
F-pseudosigma =	0.4      1.2      0.6

MPV = 10.0  
F-pseudosigma = 0.9  
N = 53  
Hu = 10.6  
HI = 9.4

Lab	Rating	Z-value	1	3	4	6
1	4	-0.03			10.0	
3	0	-4.50		6.0		
4	3	0.67		10.6		
5	4	0.11		10.1		
7	1	1.57		11.4		
11	2	1.12		11.0		
13	0	-2.70		7.6		
15	NR		< 20			
16	4	0.45			10.4	
18	4	-0.11		9.9		
24	0	-5.85		4.8		
26	4	0.45		10.4		
30	4	0.22		10.2		
32	4	0.22			10.2	
40	0	-3.26		7.1		
42	0	4.50			14.0	
46	3	0.79		10.7		
48	NR		< 50			
68	1	1.69		11.5		
70	NR		< 50			
75	3	0.90		10.8		
81	0	-2.25			8.0	
86	3	0.56		10.5		
89	4	-0.03		10.0		
97	4	-0.36		9.7		
102	2	1.12		11.0		
105	4	-0.45			9.6	
119	3	-0.69			9.4	
121	4	0.00		10.0		
128	2	1.12		11.0		
132	3	-0.56		9.5		
134	4	0.26		10.2		
138	1	-1.53			8.6	
141	NR		< 10			
142	4	-0.33			9.7	
145	0	5.17		14.6		
146	3	-0.88			9.2	
158	3	0.67		10.6		
180	4	-0.37		9.7		
191	3	0.87		10.8		
196	3	-0.67			9.4	
212	4	0.00			10.0	
213	3	0.67		10.6		
215	2	-1.12		9.0		
217	3	-0.56		9.5		
219	3	-0.79		9.3		
221	4	0.00		10.0		
224	4	0.11			10.1	
234	3	0.67		10.6		
235	0	3.37			13.0	

Lab	Rating	Z-value	1	3	4	6
236	0	-4.95			5.6	
240	0	-4.72			5.8	
255	3	0.60			10.5	
257	0	10.12	19.0			
259	2	-1.12		9.0		
265	3	-0.56			9.5	
273	2	1.12			11.0	

Table 13. Statistical summary of reported data for standard reference water sample T-145 (trace constituents)—Continued




1. AA: direct air	4. ICP						
2. AA: direct nitrous oxide	6. ICP/MS						
3. AA: graphite furnace							
N =	1      3      30      30      14						
Minimum =	16.0      14.4      10.0      9.5      10.9						
Maximum =	30.0      20.1      19.3      18.0						
Median =	15.8      15.2      14.5						
F-pseudosigma =	2.1      1.2      1.3						
Lab	Rating	Z-value	1	2	3	4	6
1	4	0.14			15.5		
3	2	1.21				17.0	
4	4	-0.21				15.0	
5	4	0.14				15.5	
7	1	1.92				18.0	
10	4	0.50			16.0		
11	4	-0.21				15.0	
13	3	0.57				16.1	
15	3	0.57				16.1	
16	2	-1.07					13.8
18	3	-0.71				14.3	
23	0	-2.13			12.3		
26	4	0.07			15.4		
30	3	-0.71				14.3	
32	3	-0.57					14.5
36	0	-3.76			10.0		
40	0	-4.12				9.5	
42	1	1.92					18.0
46	3	0.78			16.4		
48	2	1.49			17.4		
58	2	1.21			17.0		
59	4	0.50					16.0
68	3	0.85					16.5
69	3	-0.78			14.2		
70	4	-0.07					15.2
73	4	0.50					16.0
75	3	0.57					16.1
76	4	-0.36					14.8
81	4	0.50			16.0		
83	4	-0.07					15.2
85	NR					< 10	
86	4	-0.43					14.7
87	3	-0.64		14.4			
89	1	1.56			17.5		
96	3	0.64			16.2		
97	3	0.64			16.2		
102	4	0.50				16.0	
105	4	0.07					15.4
113	4	0.21					15.6
114	0	3.34		20.0			
118	3	0.85			16.5		
119	1	-1.63			13.0		
128	1	-1.63					13.0
132	4	0.14				15.5	
133	4	0.00				15.3	
134	4	-0.49				14.6	
138	0	-3.12					10.9
140	4	0.50	16.0				
141	0	-2.34					12.0
142	3	-0.99					13.9

MPV = 15.3  
 F-pseudosigma = 1.4  
 N = 78  
 Hu = 16.2  
 HI = 14.3

Lab	Rating	Z-value	1	2	3	4	6
145	0	2.84				19.3	
146	3	-0.57				14.5	
149	2	1.21			17.0		
151	3	-0.57					14.5
158	1	-1.56			13.1		
180	3	-0.71				14.3	
190	0	2.70			19.1		
191	4	0.07					15.4
193	4	-0.21			15.0		
196	2	-1.21					13.6
204	0	3.20			19.8		
212	0	-2.34					12.0
213	4	0.07			15.4		
215	3	-0.92			14.0		
217	1	-1.85				12.7	
219	1	-1.63				13.0	
221	4	-0.21			15.0		
234	2	1.14			16.9		
235	4	-0.21			15.0		
236	4	-0.43				14.7	
241	3	-0.64			14.4		
245	1	-1.63			13.0		
249	1	1.63			17.6		
253	0	3.39			20.1		
255	3	-0.52				14.6	
257	0	10.44		30.0			
259	3	-0.92			14.0		
265	4	0.50					16.0
273	2	1.21				17.0	

Table 13. Statistical summary of reported data for standard reference water sample T-145 (trace constituents)—Continued  
 Cu (Copper)  $\mu\text{g/L}$




1. AA: direct air		6. ICP/MS			
3. AA: graphite furnace					
4. ICP					
N =	9	24	38	13	
Minimum =	7.7	2.6	4.6	9.3	
Maximum =	30.0	19.0	29.0	14.0	
Median =	11.0	11.0	11.1	11.3	
F-pseudosigma =	2.2	0.8	1.6	0.7	

MPV = 11.0  
 F-pseudosigma = 1.4  
 N = 84  
 Hu = 12.1  
 HI = 10.2

Lab	Rating	Z-value	1	3	4	6
1	4	0.04	11.1			
3	0	-2.14		8.0		
5	3	0.55		11.8		
7	1	2.04		13.9		
10	4	-0.16	10.8			
11	3	0.70		12.0		
13	4	0.20		11.3		
15	4	0.06		11.1		
16	4	0.27		11.4		
18	3	0.91	12.3			
19	0	-2.14	8.0			
23	4	0.34	11.5			
26	4	0.48		11.7		
28	0	-4.56		4.6		
30	4	-0.01		11.0		
32	4	0.20		11.3		
36	0	5.67	19.0			
40	3	-0.72		10.0		
42	2	1.41		13.0		
46	4	0.20	11.3			
48	4	0.20	11.3			
58	NR	< 50				
59	4	-0.16		10.8		
68	0	2.83		15.0		
69	3	-0.58	10.2			
70	3	0.55		11.8		
73	4	-0.01		11.0		
75	3	-0.58		10.2		
80	4	-0.23		10.7		
81	4	-0.01		11.0		
83	3	0.55		11.8		
85	2	1.12	12.6			
86	4	-0.09		10.9		
87	1	-1.93	8.3			
89	3	-0.58	10.2			
96	2	1.33		12.9		
97	3	-0.58		10.2		
102	4	-0.01		11.0		
105	4	0.27		11.4		
107	3	-0.58	10.2			
113	3	0.84		12.2		
114	NR	< 10				
118	4	-0.09		10.9		
119	3	-0.72		10.0		
121	4	-0.01		11.0		
128	2	-1.22		9.3		
129	0	13.48	30.0			
132	0	12.77		29.0		
133	1	-1.51		8.9		
134	4	0.01		11.0		

Table 13. Statistical summary of reported data for standard reference water sample T-145 (trace constituents)–Continued  
 Fe (Iron)  $\mu\text{g/L}$




1. AA: direct air			5. DCP		
3. AA: graphite furnace			6. ICP/MS		
4. ICP			22. Colorimetric		
N =	20	6	56	1	4
Minimum =	78	41	30	120	85
Maximum =	230	147	129	183	113
Median =	100		101		
F-pseudosigma =	15		6		
Lab	Rating	Z-value	1	3	4
1	3	-0.59		96	
3	0	2.41		120	
4	4	0.13		102	
5	4	-0.19		100	
7	4	0.32		104	
10	3	0.51	105		
11	2	-1.39		90	
13	1	2.03		117	
15	2	-1.13		92	
16	0	-3.00		77	
18	4	-0.13		100	
23	NR	< 500			
24	4	-0.35		98	
25	0	-2.28		83	
26	4	0.13		102	
30	0	16.34	230		
32	0	10.39		183	
33	0	2.41		120	
35	3	-0.51			97
36	0	4.94	140		
40	3	-0.82		95	
42	3	0.76		107	
43	2	1.14		110	
46	4	0.00		101	
48	0	-8.99		30	
58	4	-0.13	100		
59	4	-0.13		100	
68	3	0.51		105	
69	3	0.89	108		
70	4	0.13		102	
73	4	0.00		101	
75	4	-0.35		98	
80	4	-0.25	99		
81	1	-2.03			85
83	4	-0.13		100	
85	4	0.25		103	
86	4	0.25		103	
87	0	3.42	128		
89	0	3.17		126	
91	3	-0.63		96	
96	2	1.39	112		
97	4	0.13		102	
102	2	1.27		111	
105	4	0.25		103	
107	2	-1.39	90		
109	4	-0.46	97		
113	4	0.13		102	
114	4	-0.13	100		
119	4	0.38		104	
121	4	0.00		101	

MPV = 101  
 F-pseudosigma = 8  
 N = 89  
 Hu = 107  
 HI = 96

Lab	Rating	Z-value	1	3	4	5	6	22
128	2	-1.23						
129	0	-2.91	78					
132	0	3.55						
133	4	0.38						
134	4	-0.08						
138	4	0.38						
140	3	0.76	107					
141	1	-1.99						
142	3	0.63						
145	4	0.15						
146	4	-0.32						
149	4	-0.13	100					
151	2	-1.42						90
155	1	1.56						113
158	2	1.01					109	
180	4	-0.29						99
190	1	-1.65	88					
191	4	-0.38						98
203	3	-0.63	96					
204	2	-1.01						93
212	2	1.14						110
213	3	0.51					105	
215	3	0.63						106
217	3	0.63						106
219	2	-1.14						92
220	4	-0.38	98					
221	1	2.03						117
224	3	-0.82						95
234	4	0.13						102
235	2	1.14						110
236	4	-0.42						98
240	0	-2.28						83
241	0	2.53	121					
249	0	-7.60					41	
253	0	10.01	180					
255	4	-0.25						99
257	0	-2.53	81					
265	3	-0.51						97
273	3	-0.63						96
274	0	5.83						147

Table 13. Statistical summary of reported data for standard reference water sample T-145 (trace constituents)—Continued  
 K (Potassium) mg/L




1. AA: direct air	6. ICP/MS
4. ICP	7. Ion chromatography
5. DCP	12. Flame emission
N =	24      43      1      3      2      10
Minimum =	1.60      1.60      2.18      2.03      2.13      1.76
Maximum =	5.15      38.30      —      2.11      2.13      2.80
Median =	2.14      2.12      —      —      2.35
F-pseudosigma =	0.13      0.18      —      —      0.40

MPV = 2.13  
 F-pseudosigma = 0.16  
 N = 83  
 Hu = 2.28  
 HI = 2.06

Lab	Rating	Z-value	1	4	5	6	7	12	Lab	Rating	Z-value	1	4	5	6	7	12
1	4	0.25	2.17						142	1	-1.66	1.86					
3	0	14.53		4.50					145	3	0.55	2.22					
5	4	-0.06		2.12					146	0	2.94	2.61					
7	1	2.02		2.46					180	3	-0.80	2.00					
11	4	-0.06		2.12					190	4	0.00						2.13
13	4	-0.31		2.08					191	4	-0.12						2.11
15	0	221.79		38.30					193	3	-0.55	2.04					
16	0	-3.25	1.60						203	0	2.45	2.53					2.12
18	4	-0.18		2.10					204	4	-0.06						
19	4	-0.18		2.10					212	3	-0.80	2.00					
23	4	-0.25	2.09						215	4	0.43	2.20					
24	2	-1.17		1.94					217	1	-1.78	1.84					
25	2	1.10		2.31					218	4	-0.37	2.07					
26	4	0.00					2.13		219	4	0.43	2.20					
28	2	1.17		2.32					220	4	0.43	2.20					
32	4	-0.18		2.10					221	3	0.55	2.22					
33	4	0.31		2.18					224	2	-1.10	1.95					
36	0	-3.25	1.60						234	4	-0.06	2.12					
40	3	-0.86		1.99					235	0	8.40	3.50					
42	2	1.04		2.30					236	4	0.00	2.13					
43	4	0.43		2.20					241	4	-0.43	2.06					
46	3	0.80		2.26					249	2	1.41						2.36
48	4	0.00		2.13					255	4	-0.31	2.08					
51	4	0.18					2.16		257	0	2.88					2.60	
59	0	-2.64		1.70					259	2	1.04	2.30					
68	4	-0.18		2.10					261	0	-2.27						1.76
69	2	1.29					2.34		265	4	-0.25	2.09					
70	2	-1.04		1.96					268	0	2.63	2.55					
81	3	-0.61				2.03			270	0	3.25						2.66
83	4	0.43		2.20					271	0	4.11						2.80
85	0	2.27	2.50						272	3	-0.80						2.00
86	4	0.49		2.21					273	0	5.21	2.98					
87	0	18.52	5.15						274	0	3.37						2.68
89	3	-0.74		2.01													
97	4	0.06		2.14													
102	3	-0.80		2.00													
105	4	-0.31		2.08													
107	3	0.92	2.28														
109	4	0.43	2.20														
113	4	-0.06		2.12													
114	3	-0.80	2.00														
119	0	-3.25		1.60													
121	4	0.12	2.15														
128	3	-0.86		1.99													
129	4	-0.18	2.10														
132	3	0.98		2.29													
134	4	0.00	2.13														
138	4	-0.18		2.10													
140	3	0.74	2.25														
141	4	0.25		2.17													


Table 13. Statistical summary of reported data for standard reference water sample T-145 (trace constituents)—Continued



1. AA: direct air	6. ICP/MS					
3. AA: graphite furnace	0. Other					
4. ICP	0. Other					
N =	3      2      23      8					
Minimum =	26.2      21.0      2.8      22.8					
Maximum =	28.0      28.7      46.0      31.1					
Median =	27.3      27.4					
F-pseudosigma =	3.4      1.7					
Lab	Rating	Z-value	1	3	4	6
1	3	-0.72			25.5	
3	0	2.30			33.0	
4	0	3.50			36.0	
5	4	-0.44			26.2	
7	4	0.08			27.5	
11	0	6.85			44.3	
16	1	-1.81				22.8
24	4	0.12			27.6	
25	3	0.68			29.0	
26	3	-0.72			25.5	
30	2	-1.33			24.0	
32	4	0.00				27.3
40	3	-0.81			25.3	
42	0	-8.58			6.0	
68	0	-9.89			2.8	
69	3	0.56		28.7		
75	4	0.00			27.3	
76	4	0.12				27.6
85	4	-0.44	26.2			
105	1	-1.65			23.2	
109	4	-0.28	26.6			
134	4	0.17			27.7	
142	4	0.28			28.0	
145	1	1.61			31.3	
151	3	-0.72				25.5
191	1	1.53				31.1
196	4	-0.28				26.6
212	3	0.68				29.0
217	4	-0.40			26.3	
219	3	-0.93			25.0	
234	2	1.25				30.4
236	0	-2.70				20.6
257	4	0.28	28.0			
259	0	-2.54		21.0		
265	4	0.04				27.4
273	0	7.53			46.0	

MPV = 27.3  
 F-pseudosigma = 2.5  
 N = 36  
 Hu = 28.9  
 HI = 25.5

Table 13. Statistical summary of reported data for standard reference water sample T-145 (trace constituents)—Continued  
 Mg (Magnesium) mg/L




0. Other	4. ICP
1. AA: direct air	6. ICP/MS
2. AA: direct nitrous oxide	20. Titrate: colorimetric
N =	2      19      5      52      3      4
Minimum =	8.63      7.40      7.40      8.09      8.58      0.00
Maximum =	8.95      16.47      9.00      10.30      9.11      25.15
Median =	8.55      8.69
F-pseudosigma =	0.42      0.44

Lab	Rating	Z-value	0	1	2	4	6	20
1	2	-1.13				8.17		
3	4	-0.35				8.52		
4	4	0.06				8.70		
5	2	-1.28				8.10		
7	3	0.87				9.06		
11	2	1.17				9.20		
13	3	0.88				9.07		
15	2	-1.28				8.10		
16	4	-0.17				8.60		
18	4	-0.17				8.60		
19	4	0.28				8.80		
23	4	0.10		8.72				
24	3	-0.52				8.44		
25	3	-0.61				8.40		
26	3	0.95				9.10		
28	4	0.50				8.90		
30	3	0.72		9.00				
32	3	0.97				9.11		
33	3	0.61	8.95					
36	0	-2.40		7.60				
40	4	-0.43				8.48		
42	2	1.39				9.30		
43	4	0.28				8.80		
46	4	-0.01				8.67		
48	4	0.48				8.89		
51	0	17.38	16.47					
59	4	-0.39				8.50		
68	3	0.84				9.05		
69	3	-0.68	8.37					
70	4	0.12				8.73		
75	4	0.50	8.90					
81	4	-0.21				8.58		
83	4	-0.50				8.45		
85	3	0.55	8.92					
86	4	-0.01				8.67		
87	3	-0.70				8.36		
89	3	-0.86				8.29		
97	4	-0.28				8.55		
102	0	3.62				10.30		
105	4	0.01				8.68		
107	3	0.52				8.91		
109	4	-0.39				8.50		
113	4	0.37				8.84		
114	1	-1.51		8.00				
119	3	0.95				9.10		
121	3	-0.84				8.30		
128	3	-0.88				8.28		
129	3	0.72	9.00					
132	4	0.26				8.79		
133	3	-0.64				8.39		

MPV = 8.68  
 F-pseudosigma = 0.45  
 N = 85  
 Hu = 9.00  
 Hi = 8.40

Lab	Rating	Z-value	0	1	2	4	6	20
134	4	-0.23						8.57
138	4	0.46						8.88
140	1	1.84					9.50	
141	2	1.01						9.13
142	2	-1.10						8.18
145	2	1.35						9.28
146	2	-1.06						8.20
158	1	1.84						9.50
180	4	0.39						8.85
190	4	-0.10	8.63					
191	3	0.70						8.99
193	2	-1.06					8.20	
203	4	-0.30					8.54	
212	1	1.62						9.40
215	4	0.50						8.90
217	2	-1.30						8.09
218	4	-0.50					8.45	8.60
219	4	-0.17						
220	3	0.72				9.00		
221	3	-0.72				8.35		
224	3	-0.93						8.26
234	4	-0.37						8.51
235	0	2.06						9.60
236	3	0.72						9.00
240	4	-0.39						8.50
241	0	-2.84				7.40		
255	4	-0.12						8.62
257	0	-2.84					7.40	
261	0	4.74						10.80
265	4	0.06					8.70	
268	4	-0.07				8.65		
271	0	36.74						25.15
272	0	-5.82						6.06
273	0	2.06						9.60
274	0	-19.34						0.00

Table 13. Statistical summary of reported data for standard reference water sample T-145 (trace constituents)—Continued  
Mn (Manganese)  $\mu\text{g/L}$




1. AA: direct air	5. DCP
3. AA: graphite furnace	6. ICP/MS
4. ICP	22. Colorimetric
N =	14    9    47    1    13    1
Minimum =	10.0    8.0    16.0    20.0    18.7    30.7
Maximum =	27.0    25.2    25.1    23.0
Median =	20.4    22.4    20.8    20.9
F-pseudosigma =	2.7    1.8    1.0    1.8

MPV = 20.9  
F-pseudosigma = 1.5  
N = 85  
Hu = 22.0  
Hi = 20.0

Lab	Rating	Z-value	1	3	4	5	6	22
1	4	-0.44						
3	0	-2.63						
4	4	0.07						
5	4	0.07						
7	4	0.34						
10	0	2.77	25.0					
11	4	0.07						
13	3	-0.94						
15	0	2.09						
16	4	0.40						
18	4	-0.40						
23	3	0.61	21.8					
24	3	-0.74						
25	4	0.07						
26	4	0.40						
28	1	1.62						
30	4	0.07						
32	4	0.00						
33	3	-0.61						
36	3	0.74	22.0					
40	1	-1.89						
42	2	1.42						
43	3	-0.61						
46	4	-0.34						
48	3	-0.61						
58	NR	< 50						
59	4	0.07						
68	4	0.40						
69	NR	< 20						
70	4	-0.20						
73	4	0.07						
75	4	-0.07						
80	1	-1.62	18.5					
81	3	-0.61						
83	4	-0.40						
86	4	-0.40						
87	0	4.11	27.0					
89	4	-0.34	20.4					
91	4	-0.40						
96	0	4.11	27.0					
97	0	2.23	24.2					
102	2	1.42						
105	3	-0.88						
107	3	-0.61	20.0					
109	2	-1.08	19.3					
113	4	0.40						
114	4	0.07	21.0					
119	4	0.07						
121	3	0.74						
128	2	-1.48						
			18.7					

Table 13. Statistical summary of reported data for standard reference water sample T-145 (trace constituents)—Continued  
Mo (Molybdenum)  $\mu\text{g/L}$



2. AA: direct nitrous oxide

6. ICP/MS

3. AA: graphite furnace


4. ICP

	N =	1	10	20	12
Minimum =	72.0	8.2	5.0	8.2	
Maximum =		12.2	20.1	10.0	
Median =		9.8	9.0	9.0	
F-pseudosigma =		1.7	2.2	0.7	

MPV = 9.23  
F-pseudosigma = 1.29  
N = 43  
Hu = 10.00  
HI = 8.27

Lab	Rating	Z-value	2	3	4	6
1	4	-0.44		8.67		
3	0	-2.51			6.00	
5	NR				< 10	
7	NR				< 15	
11	3	0.60			10.00	
15	0	8.45		20.10		
16	4	-0.26			8.90	
18	NR				< 20	
23	NR		< 100			
26	3	0.52			9.90	
28	0	-2.20			6.40	
30	4	0.13			9.40	
32	3	-0.82			8.18	
40	3	0.52			9.90	
42	3	0.60			10.00	
48	4	0.05		9.30		
68	2	-1.50			7.30	
70	NR				< 50	
75	NR				< 10	
86	0	6.59			17.70	
87	2	1.38		11.00		
97	4	0.27		9.58		
105	4	0.12			9.38	
109	3	-0.80		8.20		
119	3	-0.70			8.33	
128	0	-2.90			5.50	
132	0	2.15			12.00	
134	4	-0.43			8.68	
138	3	-0.65			8.39	
141	NR				< 10	
142	4	0.06			9.31	
145	2	-1.27			7.60	
146	NR				< 10	
149	3	0.60		10.00		
151	4	-0.10			9.10	
180	3	0.99			10.50	
191	4	0.05			9.30	
196	3	-0.54			8.54	
212	3	-0.65			8.40	
215	3	-0.65			8.40	
217	1	-1.58			7.20	
219	0	-3.29			5.00	
221	3	-0.81		8.19		
224	0	2.93			13.00	
234	2	1.14		10.70		
235	3	-0.96			8.00	
236	NR				< 11	
241	1	1.69			11.40	
245	0	2.31			12.20	
255	4	0.00			9.23	

Table 13. Statistical summary of reported data for standard reference water sample T-145 (trace constituents)—Continued  
Na (Sodium) mg/L



1. AA: direct air			6. ICP/MS			7. Ion chromatography			12. Flame emission			MPV = 41.2					
4. ICP												F-pseudosigma = 1.9					
5. DCP												N = 84					
			N =	22	47	1	3	1	10			Hu =	42.6				
			Minimum =	34.0	35.5	42.2	40.0	41.4	22.4			Hi =	40.0				
			Maximum =	48.5	53.1		40.5		50.0								
			Median =	40.4	41.8				41.1								
			F-pseudosigma =	2.4	2.1				4.4								
Lab	Rating	Z-value	1	4	5	6	7	12	Lab	Rating	Z-value	1	4	5	6	7	12
1	3	-0.53			40.2				141	3	0.95	43.0					
3	2	1.27			43.6				142	4	0.30	41.8					
5	4	0.11			41.4				145	1	1.59	44.2					
7	2	1.13			43.3				146	3	0.63	42.4					
11	4	-0.26			40.7				180	4	0.26	41.7					
13	2	1.32			43.7				190	4	0.11		41.4				
15	1	-1.53			38.3				191	3	-0.58		40.1				
16	3	-0.53			40.2				193	2	-1.06	39.2					
18	3	-0.63			40.0				203	3	-0.95	39.4					
19	4	0.37			41.9				204	3	-0.95		39.4				
23	2	-1.38	38.6						212	1	1.69	44.4					
24	4	-0.37			40.5				215	3	0.63	42.4					
25	0	2.54			46.0				217	4	-0.16	40.9					
26	3	0.85			42.8				218	1	-1.69	38.0					
28	0	6.30			53.1				219	3	-0.63	40.0					
32	4	-0.37			40.5				220	1	-1.69	38.0					
33	3	0.53			42.2				221	4	0.00	41.2					
36	1	-1.69	38.0						224	0	-2.71	36.1					
40	3	-0.63			40.0				234	2	-1.06	39.2					
42	4	-0.11			41.0				235	0	3.60	48.0					
43	3	0.95			43.0				236	4	-0.46	40.3					
46	3	0.74			42.6				241	0	-2.22	37.0					
48	3	0.79			42.7				249	4	0.48		42.1				
51	4	-0.11					41.0		255	3	-0.56	40.2					
59	3	-0.63			40.0				257	3	-0.63		40.0				
68	3	0.95			43.0				259	4	0.42	42.0					
69	4	0.00					41.2		261	0	2.49		45.9				
70	4	0.32			41.8				265	4	-0.32	40.6					
75	3	0.69	42.5						268	4	-0.13	41.0					
81	3	-0.63			40.0				270	0	4.35		49.4				
83	4	-0.45			40.4				271	3	-0.63		40.0				
85	4	0.48	42.1						272	0	4.66		50.0				
86	3	0.85			42.8				273	1	1.96	44.9					
87	3	-0.85			39.6				274	0	-9.92		22.4				
89	3	-0.58			40.1												
97	4	0.32	41.8														
102	0	-2.22			37.0												
105	3	0.53			42.2												
107	3	-0.53			40.2												
109	4	0.00	41.2														
113	0	-3.02			35.5												
114	0	-3.81			34.0												
119	2	1.01			43.1												
121	2	-1.16			39.0												
128	4	-0.42			40.4												
129	4	0.42	42.0														
132	3	0.61			42.4												
134	4	0.00	41.2														
138	4	0.21			41.6												
140	0	3.86	48.5														

Table 13. Statistical summary of reported data for standard reference water sample T-145 (trace constituents)—Continued






Table 13. Statistical summary of reported data for standard reference water sample T-145 (trace constituents)—Continued



1. AA: direct air		6. ICP/MS					MPV = 12.7	
3. AA: graphite furnace							F-pseudosigma = 1.2	
4. ICP							N = 80	
N =	2	47	16	15			Hu =	13.7
Minimum =	10.0	1.3	5.6	6.0			HI =	12.0
Maximum =	49.0	42.0	115.0	18.0				
Median =		12.7	13.1	12.6				
F-pseudosigma =		1.2	6.2	0.3				
Lab	Rating	Z-value	1	3	4	6	Lab	Rating
1	4	-0.14	12.6				138	3
3	0	-5.51		6.0			140	0
4	4	0.21		13.0			141	3
5	3	0.62	13.5				142	4
7	NR		< 55				145	0
10	4	-0.04	12.7				146	0
11	4	0.21		13.0			149	3
13	3	0.62	13.5				151	4
15	1	2.01	15.2				158	0
16	3	0.54		13.4			180	NR
18	4	-0.04	12.7				190	2
19	4	0.37	13.2				191	4
23	2	-1.34	11.1				193	2
26	1	-1.59	10.8				196	4
28	0	-2.65		9.5			204	4
30	2	-1.02		11.5			212	0
32	4	-0.12		12.6			213	3
34	3	0.54	13.4				215	3
36	1	1.84	15.0				217	4
42	0	-5.51		6.0			220	3
46	3	0.78	13.7				221	4
48	4	0.21	13.0				224	2
58	0	23.92	42.0				234	2
59	4	-0.28		12.4			236	0
68	0	3.15	16.6				240	0
69	3	-0.61	12.0				241	0
70	4	0.04	12.8				245	1
73	0	2.66		16.0			249	3
75	4	-0.20	12.5				255	4
76	3	-0.53		12.1			257	0
80	3	-0.85	11.7				259	0
81	4	0.21	13.0				265	4
83	4	-0.28	12.4				273	0
85	NR		< 50				274	0
86	3	0.54	13.4					
87	2	-1.26		11.2				
89	3	-0.53	12.1					
96	4	0.29	13.1					
97	4	-0.04	12.7					
102	1	1.84		15.0				
105	3	0.86		13.8				
109	0	-2.24	10.0					
113	4	-0.45	12.2					
114	NR		< 10					
118	4	-0.20	12.5					
119	2	-1.02	11.5					
128	4	0.13		12.9				
132	0	7.98	22.5					
133	0	7.32	21.7					
134	4	-0.09	12.6					

Table 13. Statistical summary of reported data for standard reference water sample T-145 (trace constituents)—Continued  
 Sb (Antimony)  $\mu\text{g/L}$



3. AA: graphite furnace

11na. AA: Hydride  $\text{NaBH}_4$

4. ICP


6. ICP/MS

	N =	17	6	13	3
Minimum =	5.13	9.20	8.21	2.00	
Maximum =	14.00	26.10	11.30	8.58	
Median =	8.30		9.00		
F-pseudosigma =	1.07		0.34		

	MPV = 8.80
F-pseudosigma =	0.96
N = 39	
Hu = 9.50	
HI = 8.21	

Lab	Rating	Z-value	3	4	6	11na
1	4	-0.16	8.65			
3	0	3.33		12.00		
5	NR			< 20		
7	NR			< 30		
11	3	0.94			9.70	
13	2	-1.04	7.80			
15	4	0.47	9.25			
16	4	0.00			8.80	
18	1	-1.67	7.20			
26	NR			< 20		
30	4	0.42		9.20		
32	4	-0.39			8.43	
36	0	-2.92	6.00			
42	2	1.25			10.00	
46	3	0.52	9.30			
48	2	-1.25	7.60			
59	3	0.52			9.30	
68	3	-0.99	7.85			
69	3	-0.68	8.15			
70	0	2.71	11.40			
75	NR		< 50			
81	0	-2.92			< 6	
89	3	-0.52	8.30			
96	3	-0.63	8.20			
97	4	0.27	9.06			
102	0	-8.13		< 1		
105	4	0.24		9.03		
119	3	-0.73			8.10	
128	4	0.36			9.15	
134	4	-0.23			8.58	
138	4	0.00		8.80		
141	0	-3.82	5.13			
142	0	2.60			11.30	
146	NR		< 20			
151	4	0.25			9.04	
180	NR		< 31.4			
193	NR		< 10			
196	3	-0.61			8.21	
212	4	-0.21			8.60	
215	0	5.42	14.00			
217	4	-0.11			8.69	
234	1	1.77	10.50			
236	0	18.02		26.10		
240	0	5.42		14.00		
241	3	-0.52	8.30			
255	0	3.26		11.93		
257	0	-7.08			2.00	
265	4	0.21			9.00	

Table 13. Statistical summary of reported data for standard reference water sample T-145 (trace constituents)—Continued  
Se (Selenium)  $\mu\text{g/L}$



3. AA: graphite furnace      11na. AA: Hydride  $\text{NaBH}_4$

4. ICP

6. ICP/MS

	N =	27	7	14	12
Minimum =		5.0	10.5	9.6	8.7
Maximum =		12.1	128.3	13.0	19.2
Median =		9.5	13.0	11.1	10.0
F-pseudosigma =		1.2	0.0	0.7	0.9

MPV = 10.1

F-pseudosigma = 1.3


N = 60

Hu = 11.1

HI = 9.3


Lab	Rating	Z-value	3	4	6	11na
1	4	-0.06			10.0	
3	NR			< 10		
5	4	-0.05	10.0			
7	NR			< 50		
10	3	-0.81			9.0	
11	1	1.97		12.6		
13	2	-1.43	8.2			
15	1	1.58	12.1			
16	4	-0.04			10.0	
18	4	0.04	10.1			
23	3	-0.56			9.3	
26	3	-0.96			8.8	
30	3	0.73		11.0		
32	4	0.19			10.3	
34	4	0.12			10.2	
36	1	-1.58	8.0			
42	0	2.27			13.0	
46	4	-0.42	9.5			
48	2	-1.12	8.6			
58	NR		< 10			
59	3	0.73			11.0	
68	0	-2.43	6.9			
69	4	0.12	10.2			
70	3	0.73	11.0			
73	0	17.69		33.0		
75	4	0.04			10.1	
80	3	-0.58	9.3			
86	2	1.19			11.6	
87	0	7.05			19.2	
89	3	-0.58			9.3	
96	4	-0.04	10.0			
97	3	-0.98	8.8			
102	0	2.27		13.0		
105	3	0.96			11.3	
113	4	-0.42	9.5			
118	4	-0.35	9.6			
119	4	-0.04			10.0	
128	3	0.89			11.2	
133	3	0.58	10.8			
134	4	0.47			10.7	
138	4	-0.37			9.6	
141	3	-0.61	9.3			
142	3	0.58			10.8	
144	4	-0.04	10.0			
146	NR		< 10			
149	4	-0.04	10.0			
151	3	0.89			11.2	
180	NR		< 50.1			
190	3	0.73	11.0			
191	2	1.50			12.0	

Table 13. Statistical summary of reported data for standard reference water sample T-145 (trace constituents)—Continued  
 SiO<sub>2</sub> (Silica) mg/L



2. AA: direct nitrous oxide		6. ICP/MS						
4. ICP		22. Colorimetric						
5. DCP		22mb. Color: molybdate blue						
		N =	1	31	1	3	4	13
		Minimum =	11.8	5.7	10.9	11.1	11.0	10.7
		Maximum =		227.0		12.2	19.2	12.3
		Median =			11.3		11.4	
		F-pseudosigma =		0.9			0.7	
Lab	Rating	Z-value	2	4	5	6	22	22mb
1	2	-1.04		10.6				
3	0	-8.12		5.7				
4	4	0.15		11.4				
5	4	-0.29		11.1				
7	4	0.49		11.6				
11	0	-3.44		8.9				
13	4	-0.15		11.2				
15	2	1.45					12.3	
24	4	0.44		11.6				
25	1	2.05		12.7				
26	4	0.29		11.5				
32	4	0.15				11.4		
33	3	-0.58		10.9				
42	2	1.16		12.1				
43	4	-0.44		11.0				
70	3	-0.58				10.9		
76	2	1.31		12.2				
81	4	0.29				11.5		
83	2	-1.31		10.4				
87	3	0.73				11.8		
89	3	-0.87				10.7		
97	4	0.00				11.3		
104	3	-0.84				10.7		
105	3	-0.62		10.9				
107	3	0.87				11.9		
113	4	-0.44				11.0		
118	4	0.19				11.4		
119	2	1.02		12.0				
121	4	0.00		11.3				
128	3	0.73		11.8				
129	4	0.26				11.5		
134	4	-0.12		11.2				
138	3	-0.73				10.8		
140	4	-0.15				11.2		
142	0	2.18		12.8				
145	1	1.67		12.5				
158	1	1.60		12.4				
190	3	0.87				11.9		
191	4	-0.29		11.1				
203	3	-0.73				10.8		
204	4	0.15				11.4		
212	3	0.73		11.8				
215	4	0.00		11.3				
217	1	-1.99		9.9				
219	3	-0.73		10.8				
234	4	-0.44		11.0				
235	1	1.89		12.6				
236	0	-6.18		7.0				
240	2	-1.16		10.5				
241	3	0.73		11.8				


Table 13. Statistical summary of reported data for standard reference water sample T-145 (trace constituents)—Continued



0. Other	4. ICP							
1. AA: direct air	5. DCP							
2. AA: direct nitrous oxide	6. ICP/MS							
N =	2							
Minimum =	169							
Maximum =	390							
Median =	203							
F-pseudosigma =	7							
Lab	Rating	Z-value	0	1	2	4	5	6
1	3	-0.55				198		
3	4	0.47				207		
4	4	0.23				205		
5	4	0.23				205		
7	4	0.00				203		
11	4	-0.12				202		
16	0	-2.23				184		
18	4	-0.35				200		
24	4	0.12				204		
25	2	1.41				215		
28	4	0.47				207		
32	4	0.00					203	
33	4	-0.35					200	
40	0	-3.99				169		
42	4	-0.23				201		
59	4	-0.35				200		
68	3	0.82				210		
70	4	0.35				206		
81	0	-3.28					175	
85	2	1.29				214		
86	3	-0.82				196		
97	0	-3.99	169					
102	0	4.22				239		
105	2	-1.06				194		
109	0	-2.31	183					
113	4	0.00				203		
121	3	-0.94				195		
134	3	-0.82				196		
138	3	-0.59				198		
142	2	1.07				212		
145	1	1.55				216		
151	3	0.59					208	
190	0	21.94	390					
191	4	0.00				203		
196	2	-1.06				194		
212	3	0.82					210	
217	2	-1.17				193		
218	0	3.05	229					
219	3	0.82				210		
234	4	0.35				206		
235	3	0.59				208		
236	4	-0.43				199		
240	3	-0.70				197		
265	4	-0.35					200	

MPV =	203
F-pseudosigma =	9
N =	44
Hu =	208
Hi =	197

Table 13. Statistical summary of reported data for standard reference water sample T-145 (trace constituents)—Continued  
 Tl (Thallium)  $\mu\text{g/L}$



3. AA: graphite furnace

4. ICP


6. ICP/MS

	N =	17	4	16
Minimum =	7.9	10.0	9.0	
Maximum =	16.9	20.0	84.0	
Median =	13.3		15.6	
F-pseudosigma =	3.4		0.7	

	MPV =	15.3
F-pseudosigma =	2.7	
N =	37	
Hu =	16.0	
HI =	12.4	

Lab	Rating	Z-value	3	4	6
1	3	0.54	16.7		
3	NR			< 10	
11	1	1.76		20.0	
13	4	0.19	15.8		
15	2	-1.39	11.6		
16	4	0.26		16.0	
18	3	-0.56	13.8		
23	0	-3.81	< 5		
32	4	0.11		15.6	
36	0	-2.74	8.0		
42	0	-2.36			9.0
46	4	0.34	16.2		
48	2	-1.20	12.1		
59	4	0.07		15.5	
69	4	0.41	16.4		
70	3	-0.75	13.3		
76	1	-1.54		11.2	
81	0	25.74		84.0	
89	NR		< 10		
97	3	0.52	16.7		
102	2	1.39		19.0	
113	3	-0.82	13.1		
119	4	0.22		15.9	
128	4	0.00		15.3	
134	3	0.60	16.9		
138	4	-0.11		15.0	
141	NR		< 50		
142	3	0.52		16.7	
146	4	-0.49	14.0		
151	4	0.26		16.0	
180	NR		< 32.1		
191	4	0.19		15.8	
193	0	-2.36	9.0		
196	4	0.00		15.3	
212	4	0.26		16.0	
213	2	-1.09	12.4		
215	0	-3.07	< 7		
217	4	-0.04		15.2	
234	1	-1.65	10.9		
235	0	-2.79	7.9		
240	1	-1.99		10.0	
241	4	0.22	15.9		
265	4	-0.11		15.0	

Table 13. Statistical summary of reported data for standard reference water sample T-145 (trace constituents)—Continued  
 U (Uranium)  $\mu\text{g/L}$



4. ICP


6. ICP/MS

	N =	1	8
Minimum =	1.20	0.86	
Maximum =	1.20		
Median =	1.10		
F-pseudosigma =	0.08		

Lab	Rating	Z-value	4	6
1	4	-0.40	1.06	
7	NR	< 120		
16	3	1.00	1.20	
30	3	1.00	1.20	
75	NR	< 100		
119	4	-0.30	1.07	
142	0	-2.45	0.86	
196	4	0.20	1.12	
212	4	0.00	1.10	
217	4	0.00	1.10	
265	4	0.00	1.10	

MPV =	1.10
F-pseudosigma =	0.10
N =	9
Hu =	1.12
Hi =	1.07


Table 13. Statistical summary of reported data for standard reference water sample T-145 (trace constituents)—Continued  
 V (Vanadium)  $\mu\text{g/L}$



2. AA: direct nitrous oxide			6. ICP/MS		
3. AA: graphite furnace					
4. ICP					
	N =	1      4      30      11			
	Minimum =	39.0      9.5      6.0      9.0			
	Maximum =		14.2      18.8      15.0		
	Median =			11.5      11.8	
	F-pseudosigma =				1.6      1.1
Lab	Rating	Z-value	2	3	4
1	4	0.20		12.0	
3	0	-3.33		6.0	
5	4	0.42		12.4	
7	3	0.54		12.6	
11	4	-0.40		11.0	
13	NR		< 50		
15	NR		< 10		
16	1	1.94		15.0	
18	4	-0.28		11.2	
26	4	0.48		12.5	
28	1	1.71		14.6	
30	4	-0.17		11.4	
32	4	0.07		11.8	
40	3	-0.75		10.4	
42	1	1.94		15.0	
46	3	-0.75		10.4	
48	2	-1.28	9.5		
68	2	-1.08		9.9	
70	NR		< 50		
75	4	-0.11		11.5	
81	1	-1.57		9.0	
85	NR		< 20		
86	3	0.83		13.1	
89	2	1.48	14.2		
97	2	1.36	14.0		
102	3	-0.99		10.0	
105	4	-0.34		11.1	
119	3	-0.81		10.3	
128	3	-0.87		10.2	
134	3	-0.58		10.7	
138	4	-0.40		11.0	
141	3	0.60		12.7	
142	3	-0.58		10.7	
145	0	3.12		17.0	
146	4	-0.11		11.5	
158	4	0.48		12.5	
180	4	0.18		12.0	
191	4	0.18		12.0	
196	3	-0.87		10.2	
212	4	0.18		12.0	
217	1	-1.63		8.9	
219	3	-0.99		10.0	
224	0	4.17		18.8	
234	4	-0.40		11.0	
235	1	1.94		15.0	
236	4	0.24		12.1	
241	3	0.77	13.0		
255	4	-0.07		11.6	
257	0	16.02	39.0		
265	4	0.18		12.0	

MPV = 11.7  
 F-pseudosigma = 1.7  
 N = 46  
 Hu = 12.7  
 HI = 10.4

Table 13. Statistical summary of reported data for standard reference water sample T-145 (trace constituents)—Continued



1. AA: direct air	6. ICP/MS					
3. AA: graphite furnace						
4. ICP						
N =	13      7      36      14					
Minimum =	5.0      0.6      0.0      8.0					
Maximum =	66.0      13.8      20.0      14.5					
Median =	14.0      8.3      9.9      9.7					
F-pseudosigma =	3.7      3.7      1.7      1.5					
Lab	Rating	Z-value	1	3	4	6
1	3	0.59				11.4
3	2	-1.26			7.0	
4	1	1.69			14.0	
5	4	-0.13			9.7	
7	4	0.42			11.0	
10	0	2.53	16.0			
13	4	-0.04			9.9	
15	0	-2.11			< 5	
16	4	-0.25				9.4
18	NR				< 100	
23	NR		< 20			
24	4	0.08			10.2	
26	4	-0.34			9.2	
28	1	-1.69			6.0	
30	4	-0.17			9.6	
32	4	0.00				10.0
36	0	2.53	16.0			
42	4	0.00				10.0
48	NR				< 5	
58	NR		< 50			
59	4	0.00				10.0
68	4	0.21			10.5	
69	NR		< 50			
70	NR				< 20	
73	2	1.26			13.0	
75	4	0.42				11.0
80	0	-2.19			4.8	
81	3	-0.84				8.0
83	4	-0.08			9.8	
86	3	0.76			11.8	
87	0	2.66	16.3			
89	1	1.60			13.8	
96	4	0.42	11.0			
97	NR				< 4.6	
102	3	-0.51			8.8	
105	4	-0.42				9.0
113	3	-0.65			8.5	
114	NR		< 10			
118	0	2.11	15.0			
119	4	0.00			10.0	
121	3	0.84			12.0	
128	4	0.04			10.1	
132	4	-0.42			9.0	
133	4	0.04			10.1	
134	3	0.71			11.7	
138	3	-0.85				8.0
140	1	1.69	14.0			
141	4	-0.04			9.9	
142	3	-0.68				8.4
145	2	1.26			13.0	

MPV = 10.0  
 F-pseudosigma = 2.4  
 N = 70  
 Hu = 11.8  
 HI = 8.6

Lab	Rating	Z-value	1	3	4	6
146	NR				< 20	
151	4	-0.45				8.9
158	2	-1.18			7.2	
180	4	-0.21			9.5	
190	2	1.39	13.3			
191	3	-0.63				8.5
193	NR		< 50			
196	1	1.90				14.5
204	2	-1.18			7.2	
212	3	0.84				12.0
213	1	1.69	14.0			
215	0	-4.21			0.0	
217	3	-0.89			7.9	
219	4	0.42			11.0	
220	4	0.00	10.0			
221	3	0.55		11.3		
224	4	0.08			10.2	
234	2	-1.38			6.7	
235	3	-0.72		8.3		
236	3	-0.59			8.6	
240	4	0.42			11.0	
241	0	-2.11	5.0			
245	4	-0.30			9.3	
249	0	-3.96			0.6	
253	0	4.22	20.0			
255	4	-0.14				9.7
257	4	-0.42	9.0			
259	0	23.61	66.0			
265	4	0.21				10.5
273	0	4.22			20.0	
274	1	-1.76		5.8		

Table 14. Statistical summary of reported data for standard reference water sample M-140 (major constituents)

Definition of analytical methods, abbreviations, and symbols

Analytical methods


0 Other/Not reported	=	
1 AA: direct, air	=	atomic absorption: direct,air
2 AA: direct, N <sub>2</sub> O	=	atomic absorption: direct,nitrous oxide
3 AA: graphite furnace	=	atomic absorption: graphite furnace
4 ICP	=	inductively coupled plasma
5 DCP	=	direct current plasma
6 ICP/MS	=	mass spectrometry/inductively coupled plasma
7 IC	=	ion chromatography
12 Flame emission	=	flame emission
20 Titrate: color	=	titration: colorimetric [color reagent specified]
21 Titrate: electro	=	titration: electrometric
22 Color:	=	colorimetric [color reagent specified]
40 Ion electrode	=	ion selective electrode
41 Electro	=	electrometric: [type meter specified]
50 Gravimetric	=	gravimetric: [precipitate specified]
51 Turbidimetric	=	turbidimetric: [precipitate specified]

Abbreviations and symbols

N =	number of samples
MPV =	most probable value
F-pseudosigma =	nonparametric statistic deviation
Hu =	upper hinge value
Hi =	lower hinge value
µg/L =	micrograms per liter
mg/L =	milligrams per liter
µS/cm =	microsiemens per centimeter at 25° C
Lab =	laboratory code number
NR =	not rated, less than value reported
< =	less than

<u>Constituent</u>		<u>page</u>
Alk	Alkalinity as CaCO <sub>3</sub>	103
B	Boron	104
Ca	Calcium	105
Cl	Chloride	106
DSRD	Dissolved solids	107
F	Fluoride	108
K	Potassium	109
Mg	Magnesium	110
Na	Sodium	111
total P	Phosphorus	112
pH		113
SiO <sub>2</sub>	Silica	114
SO <sub>4</sub>	Sulfate	115
Sp Con	Specific Conductance	116
Sr	Strontium	117
V	Vanadium	118

Table 14. Statistical summary of reported data for standard reference water sample M-140 (major constituents)—Continued  
 Alkalinity (as CaCO<sub>3</sub>) mg/L




0. Other	21. Titrate: colorimetric
20. Titrate: colorimetric	21h <sub>2</sub> SO <sub>4</sub> , Titrate: sulfuric acid
20br. Titrate: brom cresol	22. Colorimetric
N =	5      15      4      62      12      5
Minimum =	107      107      108      4      104      99
Maximum =	136      310      118      140      120      118
Median =	117      114      114
F-pseudosigma =	8      3      3

MPV = 114  
 F-pseudosigma = 3  
 N = 103  
 Hu = 116  
 HI = 112

Lab	Rating	Z-value	0	20	20br	21	21h <sub>2</sub> SO <sub>4</sub>	22
1	4	0.24				115		
2	0	-2.80				106		
3	3	0.58				116		
5	4	-0.44				113		
7	3	0.81				117		
10	3	0.58	116					
11	3	0.58				116		
13	0	-2.12				108		
15	0	-6.17				96		
16	3	-0.54	113					
18	0	-2.12				108		
19	4	0.24				115		
23	2	-1.11				111		
24	4	0.24				115		
25	3	0.58				116		
26	4	-0.10				114		
32	3	0.58				116		
33	3	-0.57				113		
36	4	0.24				115		
38	4	0.41				116		
39	0	-4.82				100		
40	4	-0.44				113		
42	0	8.67				140		
43	4	0.24				115		
46	4	-0.37				113		
48	3	-0.77				112		
50	2	1.25				118		
51	3	0.58				116		
55	0	-2.12				108		
56	2	1.43				119		
57	1	1.93	120					
59	3	0.85	117					
68	3	0.58				116		
69	2	1.25				118		
70	4	0.24				115		
75	4	-0.10				114		
76	0	-2.12				108		
80	3	-0.77	112					
81	4	0.24				115		
83	3	-0.57	113					
85	2	-1.11				111		
87	1	1.93				120		
89	4	-0.10				114		
90	0	-3.92				103		
92	4	-0.47				113		
96	4	0.24				115		
97	3	0.98	117					
105	2	-1.04				111		
107	4	-0.10				114		
109	0	2.70				122		
113	4	0.24				115		
114	4	0.24				115		
118	4	0.24				115		
119	2	-1.45				110		
127	3	0.58				116		

Lab	Rating	Z-value	0	20	20br	21	21h <sub>2</sub> SO <sub>4</sub>	22
128	3	0.58					116	
129	2	1.25					118	
132	4	-0.04						114
133	4	0.24						115
134	0	7.39	136					
138	4	-0.10					114	
141	3	0.58					116	
142	3	0.58					116	
143	4	-0.44					113	
145	0	-5.16						99
146	4	-0.10					114	
149	4	-0.10				114		
151	4	-0.10						114
155	3	0.75				117		
158	2	-1.11					111	
180	2	-1.11					111	
190	4	0.24					115	
191	2	1.25					118	
193	4	-0.10					114	
203	0	-3.23						105
204	2	-1.11						111
212	3	-0.77						112
213	3	-0.77						112
215	3	-0.77						112
217	3	0.91						117
218	0	-2.80						106
220	3	-0.64						112
224	0	-3.47						104
234	3	-0.77				112		
236	4	0.04						114
240	3	-0.81						112
241	4	0.41						116
244	4	0.24						115
247	0	-37.34						4
249	0	-2.46	107					
255	4	-0.44						113
256	0	-2.46				107		
257	4	0.24						115
258	0	3.41				124		
259	4	-0.10						114
261	1	1.93				120		
262	0	-2.09						108
265	0	7.32				136		
268	2	1.25				310		118
272	0	66.00				310		
273	0	3.27						124
274	0	38.56				229		
276	2	1.02				117		


Table 14. Statistical summary of reported data for standard reference water sample M-140 (major constituents)–Continued



4. ICP		22az. Color: azomethine					
6. ICP/MS		22cu. Color: curcumin					
22. Colorimetric							
	N =	31					
	Minimum =	20.0					
	Maximum =	145.0					
	Median =	40.9					
	F-pseudosigma =	4.4					
Lab	Rating	Z-value	4	6	22	22az	22cu
1	4	0.38	43.6				
3	4	0.26	43.0				
5	3	0.55	44.5				
10	0	3.56					60.0
15	NR		< 50				
16	0	3.87	61.6				
18	NR		< 50				
24	3	-0.67	38.2				
26	2	-1.23	35.3				
28	2	-1.29	35.0				
42	0	-4.01		21.0			
46	0	-3.91	21.5				
48	0	-4.20	20.0				
50	1	1.82					51.0
57	NR		< 100				
68	0	20.06	145.0				
70	NR		< 50				
75	4	-0.14	40.9				
85	4	-0.20	40.6				
86	4	0.48	44.1				
119	4	0.26	43.0				
127	4	-0.44	39.4				
128	1	-1.66	33.1				
129	0	2.59					55.0
132	4	0.17	42.5				
134	4	0.13	42.3				
138	3	-0.79		37.6			
141	3	-0.80	37.5				
142	4	-0.13	41.0				
145	2	1.29	48.3				
180	3	0.55	44.5				
212	3	0.65		45.0			
215	4	-0.32	40.0				
217	4	-0.20	40.6				
219	4	0.46	44.0				
234	3	-0.51	39.0				
235	3	-0.71	38.0				
236	4	-0.24	40.4				
240	0	3.31	58.7				
255	0	-3.74	22.4				
256	NR		< 20				
259	0	8.42					85.0
262	0	-3.81		22.0			
265	4	0.26		43.0			
273	0	5.50	70.0				

MPV = 41.6  
 F-pseudosigma = 5.2  
 N = 40  
 Hu = 44.8  
 HI = 37.8

Table 14. Statistical summary of reported data for standard reference water sample M-140 (major constituents)--Continued  
 Ca (Calcium) mg/L




0. Other		4. ICP					
1. AA: direct air		6. ICP/MS					
2. AA: direct nitrous oxide		20. Titrate: colorimetric					
		N =	4	16	6	56	3
		Minimum =	53.1	57.9	55.7	2.9	59.1
		Maximum =	63.0	66.4	68.0	68.0	61.3
		Median =	61.8			60.8	63.1
		F-pseudosigma =	3.1			2.6	3.9
Lab	Rating	Z-value	0	1	2	4	6
1	2	-1.40				56.3	
3	0	2.19				67.5	
5	4	-0.27				59.8	
10	4	-0.47	59.2				
11	0	-18.50				2.9	
13	1	1.52				65.4	
15	0	-2.39				53.2	
16	3	-0.66				58.6	
18	4	-0.18				60.1	
19	4	-0.11				60.3	
23	2	1.49	65.3				
24	4	-0.27				59.8	
25	2	1.20				64.4	
26	4	0.14				61.1	
28	4	0.27				61.5	
30	3	0.75	63.0				
32	4	0.21				61.3	
33	4	-0.47	59.2				
36	2	-1.49				56.0	
38	3	-0.82				58.1	
40	0	-2.26				53.6	
42	3	0.85				63.3	
43	4	0.11				61.0	
46	3	-0.59				58.8	
48	4	0.49				62.2	
50	3	0.75				63.0	
55	4	-0.47				59.2	
56	4	0.36	61.8				
57	3	0.75				63.0	
59	0	2.35				68.0	
68	3	0.75				63.0	
69	3	-0.59	58.8				
70	4	0.24				61.4	
75	3	-0.79				58.2	
80	1	1.71				66.0	
81	4	-0.24				59.9	
83	3	-0.72				58.4	
85	4	0.46	62.1				
86	3	0.62				62.6	
87	2	-1.49	56.0				
89	3	-0.76				58.3	
90	3	-0.72					58.4
97	3	-0.88				57.9	
102	4	-0.08				60.4	
105	4	0.11				61.0	
107	3	-0.85	58.0				
109	3	-0.63				58.7	
113	1	1.71				66.0	
119	4	0.14				61.1	
121	4	-0.21				60.0	

MPV = 60.7  
 F-pseudosigma = 3.1  
 N = 96  
 Hu = 63.0  
 HI = 58.8

Lab	Rating	Z-value	0	1	2	4	6	20
128	4	0.43				62.0		
129	4	0.43				62.0		
132	3	-0.60					58.8	
133	4	-0.31					59.7	
134	4	0.13					61.1	
138	4	0.43					62.0	
140	3	0.59				62.5		
141	3	0.85					63.3	
142	3	-0.53					59.0	
145	3	0.67					62.8	
146	2	-1.20					56.9	
151	4	0.17					61.2	
155	4	0.00						60.7
180	4	-0.08					60.4	
190	3	-0.66	58.6					
191	4	-0.50						59.1
204	4	-0.27					59.8	
212	3	0.72					62.9	
215	4	0.43					62.0	
217	0	-3.22					50.6	
218	1	-1.58					55.7	
219	3	-0.85					58.0	
220	4	-0.21				60.0		
221	1	1.84				66.4		
224	1	1.68					65.9	
234	4	-0.02					60.6	
235	1	1.87					66.5	
236	4	-0.33					59.6	
240	4	-0.40					59.4	
241	3	0.75	63.0					
249	2	1.46						65.2
255	4	-0.01					60.6	
256	3	0.82						63.2
257	0	2.35					68.0	
258	1	2.05						67.0
259	4	-0.34						59.6
261	2	1.27						64.6
262	3	0.75	63.0					
265	4	0.11					61.0	
268	0	-2.21	53.85					
270	0	-2.41	53.1					
271	3	0.79						63.1
272	0	-10.70						27.3
273	3	0.75					63.0	
274	0	-6.50						40.4
276	2	1.04						63.9

Table 14. Statistical summary of reported data for standard reference water sample M-140 (major constituents)--Continued  
 Cl (Chloride) mg/L




0. Other	21. Titrate: electrometric
7. Ion chromatography	22. Colorimetric
20. Titrate: colorimetric	40. Ion selective electrode
N =	3      41      20      9      27      4
Minimum =	27.0      23.2      22.0      23.9      23.5      19.6
Maximum =	27.6      35.8      42.5      27.2      29.0      26.2
Median =	25.5      27.0      25.2      26.2
F-pseudosigma =	1.2      3.9      0.8      0.6

Lab	Rating	Z-value	0	7	20	21	22	40
1	3	-0.78		24.7				
2	2	1.35			27.7			
3	3	0.64				26.7		
4	4	0.16		26.0				
5	3	-0.71		24.8				
7	2	-1.21		24.1				
10	4	0.36			26.3			
11	3	0.64			26.7			
13	3	-0.71		24.8				
15	4	0.43		26.4				
16	0	2.18	28.9					
18	3	0.85			27.0			
19	3	-0.57			25.0			
23	4	0.28				26.2		
24	4	-0.14			25.6			
25	3	-0.85		24.6				
26	4	0.28		26.2				
30	4	-0.43		25.2				
32	3	0.78		26.9				
33	2	-1.07		24.3				
36	4	0.14			26.0			
39	0	2.99			30.0			
40	3	0.78				26.9		
42	0	7.12		35.8				
43	4	0.14				26.0		
46	3	0.57			26.6			
48	2	-1.28			24.0			
50	3	-0.57			25.0			
51	2	1.42		27.8				
55	4	0.00			25.8			
56	1	-1.61				23.5		
57	3	-0.57		25.0				
59	2	-1.28		24.0				
64	4	0.36		26.3				
68	4	0.50			26.5			
69	4	0.28			26.2			
70	4	0.00		25.8				
75	4	0.21				26.1		
76	2	1.49		27.9				
80	3	-0.57		25.0				
81	2	-1.35			23.9			
83	4	-0.43			25.2			
85	3	-0.57		25.0				
87	0	2.28				29.0		
89	3	-0.57		25.0				
96	3	-0.64			24.9			
97	4	-0.07			25.7			
102	4	-0.21				25.5		
105	4	-0.43		25.2				
107	3	1.00			27.2			
109	1	-1.78		23.3				
113	3	-0.64		24.9				
114	3	-0.57			25.0			
119	3	-0.78		24.7				
127	4	0.43		26.4				

MPV = 25.8  
 F-pseudosigma = 1.4  
 N = 103  
 Hu = 26.9  
 Hi = 25.0

Lab	Rating	Z-value	0	7	20	21	22	40
128	4	-0.28		25.4				
129	3	-0.57		25.0				
133	0	-4.41						19.6
134	3	0.70			26.8			
138	4	0.07		25.9				
140	3	0.68					26.8	
141	4	-0.14					25.6	
142	4	-0.46					25.2	
143	4	0.43					26.4	
145	3	-0.63		24.9				
146	3	0.85					27.0	
149	3	0.85			27.0			
151	3	-0.57			25.0			
158	3	-0.64					24.9	
180	3	-0.85					24.6	
183	1	1.72			28.2			
190	0	2.06			28.7			
191	4	0.21			26.1			
196	1	-1.85			23.2			
203	2	-1.21					24.1	
204	4	-0.36					25.3	
212	4	0.28			26.2			
213	3	0.78				26.9		
215	2	-1.28				24.0		
217	4	0.07			25.9			
220	3	0.71					26.8	
221	4	-0.28					25.4	
224	2	1.26			27.6			
234	2	-1.28				24.0		
236	0	2.30				29.0		
240	2	-1.35			23.9			
241	0	-2.71				22.0		
247	4	-0.07			25.7			
249	0	3.42				30.6		
253	2	-1.14					24.2	
255	4	0.36					26.3	
256	1	1.72				28.2		
257	4	0.14					26.0	
258	0	4.56				32.2		
259	3	-0.57				25.0		
261	1	-1.64			23.5			
262	4	-0.07				25.7		
265	4	-0.21				25.5		
268	4	-0.14			25.6			
271	3	0.85			27.0			
272	0	11.92				42.5		
273	3	0.85			27.0			
274	0	11.00				41.3		
276	0	6.91				35.5		

Table 14. Statistical summary of reported data for standard reference water sample M-140 (major constituents)—Continued  
 DSRD (Dissolved solids) mg/L




50. Gravimetric

	N =	67
	Minimum =	0
	Maximum =	424
	Median =	382
	F-pseudosigma =	0
Lab	Rating	Z-value
1	3	-0.75
3	3	-0.75
5	4	0.25
10	4	0.25
11	1	-1.51
13	4	0.38
15	3	-0.94
16	3	-0.63
18	3	0.94
19	4	-0.50
23	3	-0.69
25	3	-0.88
26	4	-0.25
32	3	-0.63
36	1	1.76
38	4	-0.25
40	4	-0.25
43	4	0.25
50	4	0.00
55	4	-0.13
57	4	0.50
59	3	-0.69
69	2	1.19
70	4	0.19
75	3	1.00
76	3	1.00
80	1	1.82
81	0	2.64
85	4	-0.25
87	0	-2.13
89	3	0.75
90	1	1.76
92	4	-0.50
96	2	1.25
97	1	1.51
105	2	-1.07
109	4	0.31
113	3	0.88
114	4	-0.44
118	3	1.00
119	3	-0.75
127	4	0.31
129	4	-0.19
134	3	0.94
138	3	-0.88
140	4	-0.31
141	4	-0.38
142	3	0.69
143	3	1.00
146	4	-0.50

MPV = 382  
 F-pseudosigma = 16  
 N = 67  
 Hu = 396  
 Hi = 374

Lab	Rating	Z-value	50
151	4	-0.25	378
155	3	0.95	397
158	4	0.38	388
190	0	-23.94	0
212	1	-1.76	354
215	4	-0.44	375
217	4	-0.06	381
221	3	0.88	396
224	4	-0.11	380
234	4	0.00	382
236	4	0.19	385
240	4	-0.13	380
241	0	-6.15	284
253	2	1.38	404
255	4	0.13	384
257	3	0.82	395
259	4	0.13	384

Table 14. Statistical summary of reported data for standard reference water sample M-140 (major constituents)—Continued  
 F (Fluoride) mg/L




7. Ion chromatography			22. Colorimetric			
10. AA: extraction			40. Ion selective electrode			
21. Titrate: electrometric						
Lab	Rating	Z-value	7	10	21	40
1	3	0.81				0.560
2	1	1.62				0.590
3	1	1.86				0.599
7	0	4.05	0.680			
10	4	-0.27				0.520
11	3	-0.81				0.500
13	3	-0.54				0.510
15	4	-0.24				0.521
16	4	0.38				0.544
18	3	-0.54				0.510
23	3	0.70				0.556
24	3	0.81				0.560
25	3	-0.54				0.510
26	2	1.08	0.570			
32	0	-2.67				0.431
36	1	-1.70				0.467
39	4	0.27				0.540
40	3	-0.65				0.506
46	4	-0.19				0.523
48	0	4.32				0.690
50	4	0.00				0.530
55	4	-0.27				0.520
57	4	-0.27				0.520
59	3	0.54				0.550
69	2	1.35				0.580
70	2	1.35				0.580
76	4	0.32				0.542
80	3	-0.54				0.510
81	4	0.22				0.538
83	0	3.24				0.650
89	0	-12.03				0.084
96	4	-0.13				0.525
97	3	-0.94				0.495
105	3	0.54	0.550			
107	4	-0.08				0.527
109	1	1.62				0.590
113	4	-0.13				0.525
114	4	0.00				0.530
119	4	0.00				0.530
127	2	-1.46	0.476			
128	4	-0.27	0.520			
129	0	3.10	0.645			
134	1	1.62				0.590
138	4	0.03				0.531
140	0	-2.40				0.441
141	2	1.08				0.570
142	1	1.97				0.603
145	1	-1.89	0.460			
146	3	0.73				0.557
149	4	0.00	0.530			

MPV = 0.530  
 F-pseudosigma = 0.037  
 N = 71  
 Hu = 0.560  
 HI = 0.510

Lab	Rating	Z-value	7	10	21	22	40
151	3	-0.54					0.517
158	4	0.27					0.547
190	4	-0.24					0.521
196	4	-0.13					0.525
212	3	-0.54					0.517
215	3	-0.81					0.507
217	0	-2.97					0.427
224	0	10.52	0.920				
234	2	-1.21	0.485				
240	4	0.13					0.535
241	1	1.62					0.597
247	3	-0.54	0.510				
255	3	0.70					0.556
257	3	0.81					0.567
258	3	-0.54					0.510
259	3	-0.81					0.507
262	3	-0.76					0.502
265	4	-0.27					0.527
272	0	39.66					2.007
273	0	-2.70					0.430
274	0	6.21					0.760

Table 14. Statistical summary of reported data for standard reference water sample M-140 (major constituents)—Continued  
K (Potassium) mg/L




1. AA: direct air	6. ICP/MS
4. ICP	7. Ion chromatography
5. DCP	12. Flame emission
N =	26    43    1    3    2    12
Minimum =	2.00    2.20    2.62    2.53    2.58    2.35
Maximum =	3.20    4.20    —    2.68    2.68    3.91
Median =	2.57    2.58    —    2.83
F-pseudosigma =	0.09    0.16    —    0.39

MPV = 2.58  
F-pseudosigma = 0.14  
N = 87  
Hu = 2.69  
Hi = 2.50

Lab	Rating	Z-value	1	4	5	6	7	12
1	4	-0.07	2.57					
2	0	-3.41	2.10					
3	2	-1.07		2.43				
5	2	-1.49		2.37				
10	4	0.00	2.58					
13	1	-1.99		2.30				
15	1	-1.56		2.36				
16	3	-0.57	2.50					
18	0	-2.70		2.20				
19	4	0.14	2.60					
23	3	0.71	2.68					
24	3	-0.92	2.45					
26	3	0.71		2.68				
28	1	-1.85		2.32				
32	4	-0.36			2.53			
33	4	0.28		2.62				
36	0	-3.41	2.10					
38	4	0.28	2.62					
40	4	0.43		2.64				
42	3	0.85		2.70				
43	4	0.14		2.60				
46	0	-2.34		2.25				
48	4	0.00		2.58				
50	3	-0.57		2.50				
51	4	-0.14			2.56			
56	0	3.12	3.02					
57	0	4.40	3.20					
59	NR		< 5					
64	3	0.71	2.68					
68	0	2.27	2.90					
69	1	1.70		2.82				
70	3	-0.71	2.48					
80	0	-4.12	2.00					
81	4	-0.36	2.53					
83	4	0.14	2.60					
85	0	2.63	2.95					
86	4	0.43	2.64					
87	3	-0.78	2.47					
89	2	-1.42	2.38					
97	4	0.14	2.60					
102	1	-1.99	2.30					
105	4	0.21	2.61					
107	2	1.21	2.75					
109	4	-0.14	2.56					
113	2	1.35	2.77					
119	3	-0.57	2.50					
121	4	-0.14	2.56					
127	4	-0.36	2.53					
128	4	0.00	2.58					
129	3	-0.57	2.50					

Lab	Rating	Z-value	1	4	5	6	7	12
132	3	0.57	2.66					
134	4	-0.28	2.54					
138	4	0.07	2.59					
140	4	0.00	2.58					
141	4	-0.07	2.57					
142	1	-1.99	2.30					
145	4	-0.28	2.54					
146	0	4.54	3.22					
151	3	0.71		2.68				
180	2	-1.14	2.42					
190	4	0.00		2.58				
191	4	-0.14		2.56				
212	3	-0.57	2.50					
215	1	1.56	2.80					
217	0	11.50	4.20					
218	2	-1.21	2.41					
219	1	1.56	2.80					
220	4	0.14	2.60					
221	4	0.21	2.61					
224	0	3.69	3.10					
234	4	0.43	2.64					
235	0	11.15	4.15					
236	1	-1.92	2.31					
241	4	-0.14	2.56					
249	3	0.71		2.68				
255	4	0.28	2.62					
256	0	9.44		3.91				
257	0	5.11		3.30				
258	2	1.21		2.75				
259	3	-0.57	2.50					
261	1	-1.63		2.35				
262	1	1.78		2.83				
265	4	-0.43	2.52					
268	1	1.93	2.85					
270	0	4.47		3.21				
271	3	-0.57		2.50				
272	0	2.98		3.00				
273	1	1.56	2.80					
274	0	3.55		3.08				

Table 14. Statistical summary of reported data for standard reference water sample M-140 (major constituents)—Continued  
 Mg (Magnesium) mg/L




0. Other	4. ICP
1. AA: direct air	6. ICP/MS
2. AA: direct nitrous oxide	20. Titrate: colorimetric
N = 6	
Minimum = 16.60	10.95
Maximum = 21.40	19.27
Median = 18.00	18.18
F-pseudosigma = 0.85	1.11

Lab	Rating	Z-value	0	1	2	4	6	20
1	3	-0.63				17.4		
2	2	-1.04		17.0				
3	0	2.71			20.6			
5	3	-0.94				17.1		
10	4	0.42		18.4				
11	0	-18.23			0.5			
13	2	1.04			19.0			
15	0	-2.08			16.0			
16	4	-0.21			17.8			
18	4	-0.02			18.0			
19	4	0.42			18.4			
23	4	0.00	18.0					
24	4	0.10			18.1			
25	2	1.25			19.2			
26	2	1.25			19.2			
28	4	0.21			18.2			
30	2	-1.04		17.0				
32	0	2.06			20.0			
33	4	-0.21	17.8					
36	2	-1.04		17.0				
38	4	0.14		18.1				
40	2	-1.25			16.8			
42	2	1.35			19.3			
43	2	1.04			19.0			
46	3	-0.82			17.2			
48	4	0.31			18.3			
50	4	0.00			18.0			
51	0	-7.34	11.0					
55	2	1.46			19.4			
56	2	1.32	19.3					
57	0	2.08			20.0			
59	0	2.60			20.5			
68	4	0.00			18.0			
69	3	-0.52	17.5					
70	4	0.10			18.1			
75	4	0.42	18.4					
80	2	-1.04	17.0					
81	4	0.21			18.2			
83	3	-0.83			17.2			
85	4	0.31	18.3					
86	4	0.10			18.1			
87	3	-0.83	17.2					
89	4	0.21	18.2					
97	4	0.00	18.0					
102	1	1.98			19.9			
105	3	0.83			18.8			
107	4	-0.31	17.7					
109	4	-0.42			17.6			
113	1	1.98			19.9			
119	3	0.63			18.6			


Lab	Rating	Z-value	0	1	2	4	6	20
121	4	-0.31				17.7		
127	3	0.63				18.6		
128	3	-0.83				17.2		
129	4	0.00			18.0			
132	4	-0.21				17.8		
133	2	-1.15				16.9		
134	4	-0.13				17.9		
138	3	0.52				18.5		
140	3	0.52			18.5			
141	3	0.83				18.8		
142	4	-0.17				17.8		
145	4	-0.20				17.8		
146	4	-0.42				17.6		
151	4	0.10				18.1		
155	4	0.32				18.3		
180	4	0.42				18.4		
190	4	-0.21	17.8					
191	4	0.42				18.4		
204	0	-17.64				1.1		
212	1	1.77				19.7		
215	3	0.63				18.6		
217	0	-2.50				15.6		
218	4	-0.32			17.7			
219	2	1.04				19.0		
220	4	0.00			18.0			
221	4	0.10			18.1			
224	0	4.96				22.8		
234	4	-0.42				17.6		
235	2	1.46				19.4		
236	2	1.24				19.2		
240	4	-0.42				17.6		
241	0	-4.17			14.0			
255	4	0.17				18.2		
256	2	-1.46	16.6					
257	0	-8.54			9.8			
258	2	-1.46	16.6					
259	2	-1.04	17.0					
261	3	-0.63				17.4		
262	0	3.54	21.4					
265	4	-0.21			17.8			
268	3	-0.52			17.5			
271	0	-7.95				10.4		
272	0	-4.85				13.3		
273	0	2.81				20.7		
274	0	13.32				30.8		
276	0	-2.08				16.0		

Table 14. Statistical summary of reported data for standard reference water sample M-140 (major constituents)—Continued  
Na (Sodium) mg/L



1. AA: direct air		6. ICP/MS		7. Ion chromatography		12. Flame emission				MPV =	39.0
		N =	24	51	1	3	1	12	F-pseudosigma =	1.9	
4.	ICP	Minimum =	34.4	1.5	38.2	38.5	39.3	20.6			
5.	DCP	Maximum =	50.8	48.3		42.6		50.0			
		Median =	38.1	39.6				38.7			
		F-pseudosigma =	2.2	1.8				3.9			
Lab	Rating	Z-value	1	4	5	6	7	12	Lab	Rating	Z-value
1	4	-0.42		38.2					127	4	0.00
2	0	-2.08	35.0						128	4	-0.47
3	0	-2.33		34.5					129	2	-1.04
5	4	0.31		39.6					132	4	0.13
10	4	-0.10	38.8						134	4	0.14
11	0	-19.47		1.5					138	4	0.36
13	2	1.14		41.2					140	0	5.19
15	0	-2.33		34.5					141	3	0.83
16	4	-0.26		38.5					142	4	0.30
18	4	-0.01		39.0					145	2	1.08
19	2	1.14		41.2					146	1	1.66
23	0	-2.18	34.8						151	1	1.87
24	4	-0.10		38.8					180	4	0.47
25	1	1.92		42.7					190	4	0.16
26	3	0.99		40.9					191	4	-0.26
28	1	1.92		42.7					212	1	1.76
32	4	0.26			39.5				215	2	1.30
33	4	-0.42			38.2				217	0	-2.18
36	0	-2.08	35.0						218	1	-1.56
38	0	-2.39	34.4						219	4	0.00
40	3	-0.99		37.1					220	1	-1.56
42	4	0.10		39.2					221	4	0.42
43	3	0.52		40.0					224	2	1.08
46	4	0.31		39.6					234	4	0.10
48	4	0.42		39.8					235	0	4.83
50	3	0.52		40.0					236	4	-0.43
51	4	-0.10			38.8				241	2	-1.04
56	0	6.12	50.8						249	4	0.05
57	1	-1.56		36.0					255	4	-0.17
59	0	2.85		44.5					256	2	-1.02
64	3	-0.57	37.9						257	4	0.00
68	3	0.78		40.5					258	3	-0.78
69	4	-0.21			38.6				259	4	0.00
70	4	0.21		39.4					261	0	3.58
75	4	-0.31	38.4						262	3	-0.78
80	2	-1.04	37.0						265	4	-0.47
81	4	0.05		39.1					268	3	0.91
83	4	-0.47		38.1					270	0	4.09
85	4	0.42	39.8						271	0	-2.59
86	2	1.14		41.2					272	0	5.71
87	3	-0.88	37.3						273	0	2.70
89	4	-0.42	38.2						274	0	-9.55
97	3	0.57	40.1								
102	0	-3.01		33.2							
105	3	0.93		40.8							
107	4	0.31	39.6								
109	4	-0.23		38.6							
113	1	-1.61		35.9							
119	4	0.47		39.9							
121	4	-0.05		38.9							


Table 14. Statistical summary of reported data for standard reference water sample M-140 (major constituents)—Continued  
total P as P (total Phosphorus as Phosphorus) mg/L



0. Other		22. Colorimetric					
4. ICP		22m. Color: phosphomolybdate					
6. ICP/MS		22sncl. Color: stanous chloride					
N =		2	2	1	8	43	1
Minimum =		0.05	0.04	0.03	0.02	0.01	0.04
Maximum =		0.05	0.05	0.03	0.71	0.10	0.04
Median =		0.05	0.05	0.03	0.04	0.03	0.04
F-pseudosigma =		0.00	0.00	0.00	0.00	0.00	0.00
Lab	Rating	Z-value	0	4	6	22	22m
1	4	-0.18				0.030	
2	1	1.80				0.052	
3	4	-0.18				0.030	
11	2	1.17				0.045	
15	NR					< 0.05	
16	1	1.71	0.051				
18	4	-0.45				0.027	
22	4	0.00				0.032	
23	4	0.18				0.034	
32	1	1.58	0.050				
38	4	0.00				0.032	
39	4	0.18				0.034	
48	0	3.42				0.070	
55	3	-0.72				0.024	
57	3	0.72				0.040	
59	3	0.72				0.040	
64	4	-0.18				0.030	
68	1	1.80				0.052	
70	NR					< 0.1	
81	4	-0.36				0.028	
83	NR		< 100				
87	4	-0.36				0.028	
89	4	-0.27				0.029	
92	4	-0.45				0.027	
97	0	-2.70				< 0.002	
102	0	-2.25				0.007	
105	2	-1.17				0.019	
107	4	-0.36				0.028	
113	4	-0.36				0.028	
114	4	-0.09				0.031	
118	2	-1.08				0.020	
119	4	-0.18				0.030	
127	NR					< 0.01	
129	4	-0.45				0.027	
132	4	-0.18				0.030	
133	4	-0.36				0.028	
134	4	-0.18				0.030	
138	4	0.00				0.032	
140	NR					< 0.02	
141	NR					< 0.05	
142	2	-1.08				0.020	
143	4	0.09				0.033	
145	4	0.27				0.035	
146	NR					< 0.1	
155	4	-0.24				0.029	
158	3	0.81				0.041	
180	1	1.80				0.052	
190	1	1.62				0.050	
191	4	0.18				0.034	
212	3	0.99				0.043	

MPV =	0.032
F-pseudosigma =	0.011
N =	57
Hu =	0.043
HI =	0.028
Lab	Rating
213	2
215	3
224	4
234	4
235	1
236	3
240	3
241	4
243	4
249	1
257	0
258	0
259	3
273	0
274	0

Table 14. Statistical summary of reported data for standard reference water sample M-140 (major constituents)—Continued  
pH



—○— 41

41. Direct reading


N =	108
Minimum =	7.42
Maximum =	8.56
Median =	8.28
F-pseudosigma =	0.17

Lab	Rating	Z-value	41
1	4	-0.42	8.21
2	2	-1.44	8.04
3	0	-2.70	7.83
5	3	-0.66	8.17
7	4	-0.48	8.20
10	3	0.84	8.42
11	4	-0.12	8.26
13	4	0.48	8.36
15	2	-1.02	8.11
16	2	-1.08	8.10
18	4	-0.18	8.25
19	1	1.62	8.55
23	4	0.00	8.28
24	4	-0.48	8.20
25	4	0.48	8.36
26	4	0.30	8.33
30	3	0.72	8.40
32	3	0.72	8.40
33	4	0.00	8.28
36	4	0.30	8.33
38	3	0.72	8.40
39	3	0.72	8.40
40	4	0.42	8.35
42	0	-2.28	7.90
43	4	-0.42	8.21
46	3	0.60	8.38
48	1	-1.68	8.00
50	4	0.18	8.31
51	4	0.06	8.29
55	1	-1.68	8.00
56	4	0.42	8.35
57	4	-0.48	8.20
59	3	0.72	8.40
64	3	0.60	8.38
68	3	0.60	8.38
69	3	0.72	8.40
70	4	0.00	8.28
75	4	-0.06	8.27
76	4	-0.12	8.26
80	3	-0.66	8.17
81	3	0.54	8.37
85	3	0.54	8.37
86	4	0.42	8.35
87	1	-1.98	7.95
89	3	0.66	8.39
92	3	-0.84	8.14
96	4	0.36	8.34
97	3	0.54	8.37
105	4	0.36	8.34
107	2	1.26	8.49
109	3	0.84	8.42
113	4	0.24	8.32
114	2	-1.26	8.07
118	0	-2.28	7.90
119	2	1.32	8.50

MPV =	8.28
F-pseudosigma =	0.17
N =	108
Hu =	8.37
Hi =	8.15

Lab	Rating	Z-value	41
127	4	0.18	8.31
128	3	0.90	8.43
129	1	-1.74	7.99
132	2	-1.38	8.05
133	3	-0.78	8.15
134	3	0.96	8.44
138	4	0.06	8.29
140	4	0.30	8.33
141	4	0.42	8.35
142	4	-0.12	8.26
143	3	0.60	8.38
145	4	0.12	8.30
146	3	-0.78	8.15
149	3	0.72	8.40
151	3	0.66	8.39
155	4	-0.24	8.24
158	4	-0.12	8.26
180	3	0.72	8.40
183	4	0.30	8.33
190	2	-1.08	8.10
191	3	0.66	8.39
203	4	0.18	8.31
204	3	-0.72	8.16
212	4	0.12	8.30
213	4	0.42	8.35
215	4	-0.18	8.25
217	3	0.72	8.40
218	1	-1.56	8.02
221	2	-1.20	8.08
224	3	-0.66	8.17
234	3	-0.90	8.13
236	4	-0.24	8.24
240	4	-0.24	8.24
241	4	0.12	8.30
243	0	-2.22	7.91
244	4	0.36	8.34
247	1	1.68	8.56
249	1	-1.80	7.98
253	4	-0.48	8.20
255	2	1.26	8.49
256	2	-1.38	8.05
257	4	-0.42	8.21
258	4	-0.48	8.20
259	2	1.32	8.50
261	1	-1.98	7.95
262	0	-5.16	7.42
265	3	-0.96	8.12
268	2	-1.08	8.10
271	0	-2.70	7.83
272	4	0.00	8.28
273	4	-0.48	8.20
274	0	-2.28	7.90
276	0	-3.48	7.70

Table 14. Statistical summary of reported data for standard reference water sample M-140 (major constituents)—Continued  
 SiO<sub>2</sub> (Silica) mg/L




2. AA: direct nitrous oxide	6. ICP/MS
4. ICP	22. Colorimetric
5. DCP	22mb. Color: molybdate blue
	N = 2 30 1 2 6 25
Minimum =	5.50 4.66 7.10 7.50 6.65 5.88
Maximum =	7.68 11.04 8.71 14.93 7.95
Median =	7.43 7.29
F-pseudosigma =	0.61 0.00

Lab	Rating	Z-value	2	4	5	6	22	22mb
1	3	-1.00		6.89				
3	0	2.35		8.43				
5	3	-0.59		7.08				
10	4	-0.11				7.30		
13	4	0.04		7.37				
15	2	1.35				7.97		
18	4	0.20				7.44		
23	3	0.87				7.75		
24	3	0.89		7.76				
25	0	8.03		11.04				
26	4	0.39		7.53				
32	4	0.33				7.50		
33	3	-0.54				7.10		
38	4	-0.13					7.29	
40	4	-0.37		7.18				
42	2	1.41		8.00				
43	4	0.33		7.50				
46	3	-0.67				7.04		
50	4	-0.11				7.30		
55	4	0.02		7.36				
57	4	0.11		7.40				
59	3	0.76				7.70		
68	3	-0.65				7.05		
70	3	-0.54				7.10		
80	0	-4.03		5.50				
81	4	-0.24				7.24		
83	2	-1.28		6.76				
85	4	-0.07				7.32		
87	4	-0.37				7.18		
89	2	-1.20				6.80		
97	4	0.04				7.37		
102	2	-1.50				6.66		
105	4	0.22		7.45				
107	4	-0.30				7.21		
113	4	-0.02				7.34		
118	4	0.26				7.47		
119	3	-0.76		7.00				
121	4	-0.50		7.12				
127	4	-0.30		7.21				
128	3	0.96		7.79				
129	2	1.31				7.95		
134	4	0.25		7.47				
138	4	-0.46				7.14		
140	4	0.11				7.40		
142	0	2.31		8.41				
145	2	1.31		7.95				
151	0	-3.20				5.88		
155	3	-0.58				7.08		
190	3	0.76				7.70		
191	0	2.96				8.71		

MPV = 7.35  
 F-pseudosigma = 0.46  
 N = 66  
 Hu = 7.70  
 HI = 7.08


Lab	Rating	Z-value	2	4	5	6	22	22mb
203	3	-0.98					6.90	
204	4	0.15					7.42	
212	2	1.20				7.90		
215	1	1.59				8.08		
217	0	-2.76				6.08		
219	4	0.33				7.50		
234	4	-0.33				7.20		
235	1	1.98				8.26		
236	0	-5.85				4.66		
240	2	-1.02				6.88		
241	3	0.72	7.68					
255	4	0.07					7.38	
256	1	-1.52					6.65	
259	3	-0.76					7.00	
265	2	-1.09				6.85		
274	0	16.49						14.93

Table 14. Statistical summary of reported data for standard reference water sample M-140 (major constituents)—Continued  
 $\text{SO}_4$  (Sulfate) mg/L



0. Other						22mtb. Color: methyl thymol blue					
7. Ion chromatography						50. Gravimetric					
22. Colorimetric						51. Turbidimetric					
N =	6	51	2	17	4	18					
Minimum =	16	5	150	120	143	17					
Maximum =	174	197	158	168	152	200					
Median =	150		149		149						
F-pseudosigma =	5		7		11						
Lab	Rating	Z-value	0	7	22	22mtb	50	51			
1	4	0.03		150					114	4	0.30
2	4	0.30				152			119	2	-1.05
3	2	-1.20		142					127	2	1.35
4	4	0.34		152					128	3	0.75
5	4	0.15		151					129	4	0.00
7	3	0.67		155					134	4	0.30
10	4	0.15				151			138	4	-0.15
11	0	-21.75		5					140	3	0.75
13	4	0.00		150					141	3	0.90
15	2	-1.50		140					142	1	1.95
16	2	-1.03				143			145	4	-0.18
18	4	-0.30				148			146	3	0.75
19	4	-0.15				149			149	2	1.50
23	2	1.20			158				151	4	0.15
24	4	0.00				150			158	4	0.30
25	4	-0.30		148					180	2	-1.50
26	4	0.00		150					190	0	4.05
30	4	0.00		150					191	4	-0.15
32	3	-0.60		146					193	4	-0.15
33	4	-0.42		147					196	3	-0.75
36	2	1.50				160			203	0	2.70
39	4	0.00		150					204	3	0.75
40	0	-4.65		119					212	3	0.90
42	0	7.04		197					215	3	0.75
43	4	0.00				150			217	2	1.20
46	0	5.25		185					219	0	-16.49
48	0	-6.89				104			220	4	-0.34
50	4	-0.15				149			221	3	-0.60
51	3	0.82		156					224	4	0.32
55	0	-2.55				133			234	4	0.45
56	4	-0.13				149			235	0	3.60
57	2	-1.50				140			236	4	0.34
59	4	-0.30		148					240	4	-0.45
64	4	0.10		151					241	4	0.30
69	2	1.20			158				247	2	-1.08
70	3	-0.60		146					249	2	-1.39
75	4	0.45				153			253	0	-4.95
80	0	-3.00		130					255	0	-4.50
81	3	-0.75				145			256	3	0.65
83	4	0.15		151					257	0	6.75
85	4	-0.30		148					258	0	7.46
87	0	-3.45				127			259	4	-0.30
89	4	-0.30		148					261	0	-20.13
92	4	-0.39				147			262	4	0.00
96	3	-0.75				145			265	3	0.75
97	4	0.45				153			268	3	0.55
102	2	-1.05				143			271	0	-20.01
105	2	-1.50		140					274	4	-0.09
109	3	0.78		155							17
113	3	0.60		154							149

Table 14. Statistical summary of reported data for standard reference water sample M-140 (major constituents)—Continued  
 Sp Cond (Specific Conductance)  $\mu\text{S}/\text{cm}$



0. Other  
 41. Direct reading


N =	1	99
Minimum =	620	70
Maximum =	719	
Median =	600	
F-pseudosigma =	19	

Lab	Rating	Z-value	0	41
1	4	0.02	600	
2	3	-0.76	586	
3	4	0.43	608	
5	3	0.65	612	
7	1	-1.56	571	
10	4	0.32	606	
11	4	-0.16	597	
13	4	0.43	608	
15	4	-0.05	599	
16	4	-0.12	598	
18	1	-2.00	563	
19	0	3.56	666	
23	1	-1.73	568	
24	4	0.00	600	
25	4	0.38	607	
26	4	0.32	606	
32	4	0.05	601	
33	1	-1.56	571	
36	0	-28.60	70	
38	3	0.65	612	
39	3	-0.86	584	
40	4	0.43	608	
42	4	0.22	604	
43	4	0.00	600	
46	4	-0.05	599	
48	4	0.43	608	
50	2	1.19	622	
51	2	-1.08	580	
55	2	1.35	625	
56	4	-0.14	597	
57	0	3.24	660	
59	3	-0.66	588	
64	3	0.70	613	
68	4	0.43	608	
70	3	-0.70	587	
75	3	0.54	610	
80	4	0.00	600	
81	4	0.00	600	
86	2	1.30	624	
87	0	-15.11	320	
89	3	-0.59	589	
90	0	-2.59	552	
92	4	-0.17	597	
96	3	0.97	618	
97	2	1.35	625	
102	0	6.42	719	
105	3	0.81	615	
107	4	-0.49	591	
109	4	-0.49	591	
113	3	-0.54	590	

MPV = 600  
 F-pseudosigma = 19  
 N = 100  
 Hu = 612  
 Hi = 587

Lab	Rating	Z-value	0	41
114	4	0.43	608	
118	3	-0.54	590	
119	4	-0.16	597	
127	4	-0.27	595	
128	3	0.81	615	
129	3	-0.97	582	
132	0	-7.72	457	
134	4	0.32	606	
138	4	-0.16	597	
140	0	5.83	708	
141	4	0.38	607	
142	4	0.22	604	
145	3	0.81	615	
146	4	-0.27	595	
149	3	0.65	612	
151	4	0.11	602	
155	1	1.67	631	
158	4	0.49	609	
180	2	1.08	620	
183	0	-3.51	535	
190	4	0.22	604	
193	1	-1.73	568	
203	4	-0.11	598	
204	4	-0.32	594	
212	1	2.00	637	
215	2	-1.40	574	
217	2	1.35	625	
218	0	-2.37	556	
224	0	-2.16	560	
234	3	0.59	611	
236	3	-0.59	589	
240	2	-1.30	576	
241	0	-10.03	414	
243	4	-0.27	595	
244	4	0.00	600	
247	4	-0.27	595	
249	3	0.92	617	
253	2	1.35	625	
255	4	-0.27	595	
257	4	0.00	600	
258	0	-3.29	539	
259	4	0.49	609	
261	0	-4.21	522	
262	3	-0.97	582	
268	0	-2.16	560	
271	0	-2.97	545	
272	0	2.43	645	
273	2	1.08	620	
274	1	1.83	634	
276	0	-2.70	550	


Table 14. Statistical summary of reported data for standard reference water sample M-140 (major constituents)—Continued  
Sr (Strontium)  $\mu\text{g/L}$



0. Other			4. ICP					
1. AA: direct air			5. DCP					
2. AA: direct nitrous oxide			6. ICP/MS					
N =	2	2	1	33	1	4		
Minimum =	578	680	673	17	685	1		
Maximum =	958	693		784		710		
Median =				668				
F-pseudosigma =				32				
Lab	Rating	Z-value	0	1	2	4	5	6
1	4	-0.15				666		
3	0	3.67				784		
5	4	-0.03				670		
11	0	-21.27				17		
16	2	-1.40				628		
18	3	-0.68				650		
23	3	0.72	693					
24	4	0.16				676		
25	1	1.92				730		
28	4	0.00				671		
32	4	0.26				679		
33	4	0.46				685		
40	3	-0.68				650		
42	4	-0.49				656		
55	3	-0.88				644		
68	4	0.13				675		
70	4	0.07				673		
81	4	-0.29				662		
85	1	1.53				718		
86	4	-0.16				666		
97	0	-3.02	578					
102	1	1.98				732		
105	0	-3.80				554		
113	4	0.29				680		
121	4	0.23				678		
127	2	-1.30				631		
134	3	-0.57				654		
138	4	-0.42				658		
142	3	0.73				694		
145	3	0.91				699		
151	2	1.27				710		
190	0	9.33	958					
191	0	-21.79				1		
212	4	-0.36				660		
217	0	-3.02				578		
218	4	0.07	673					
219	3	0.94				700		
234	4	-0.10				668		
235	2	1.24				709		
236	3	-0.73				649		
259	4	0.29	680					
265	4	-0.42				658		
273	0	3.22				770		

MPV = 671  
F-pseudosigma = 31  
N = 43  
Hu = 693  
Hi = 652

Table 14. Statistical summary of reported data for standard reference water sample M-140 (major constituents)—Continued



2. AA: direct nitrous oxide	6. ICP/MS						
3. AA: graphite furnace	22. Colorimetric						
4. ICP	0. Other						
N =	1      2      11      4      1						
Minimum =	39.00      1.00      1.00      1.20      1.40						
Maximum =	1.40      50.00      5.19						
Median =	4.70						
F-pseudosigma =	3.34						
Lab	Rating	Z-value	2	3	4	6	22
1	3	-0.67					1.40
3	4	-0.14			3.00		
5	NR				< 4		
13	NR				< 50		
15	NR				< 10		
16	3	0.59				5.19	
18	NR				< 5		
23	NR			< 5			
26	NR				< 4		
28	0	2.25			10.20		
42	NR					< 5	
48	NR				< 200		
57	NR				< 100		
68	0	15.48			50.00		
70	NR				< 50		
81	NR				< 3		
85	NR				< 20		
86	3	0.77			5.73		
89	NR			< 10			
97	NR			< 0			
105	NR				< 13		
127	3	-0.67		1.40			
134	NR				< 1		
138	NR					< 2	
141	NR				< 10		
142	4	0.00				3.42	
145	4	0.43			4.70		
146	NR				< 10		
180	3	0.51			4.95		
212	4	-0.44				2.10	
219	3	-0.80			1.00		
224	1	1.62			8.30		
234	4	-0.47			2.02		
235	NR				< 5		
236	4	0.13			3.80		
241	3	-0.80		1.00			
255	3	-0.67			1.40		
257	0	11.82	39.00				
265	3	-0.74				1.20	

MPV =	3.42
F-pseudosigma =	3.01
N =	19
Hu =	5.46
HI =	1.40

Table 15. Statistical summary of reported data for standard reference water sample N-51 (nutrients)

Definition of analytical methods, abbreviations, and symbols		
<u>Analytical methods</u>		
0. Other/Not reported	=	
4. ICP	= inductively coupled plasma	
5. DCP	= direct coupled plasma	
7. IC	= ion chromatography	
20. Titrate: color	= titration: colorimetric (color reagent specified)	
21. Titrate: electro	= titration: electrometric	
22. Color:	= colorimetric [color reagent specified]	
40. Ion electrode	= ion selective electrode	
<u>Abbreviations and symbols</u>		
N =	number of samples	
MPV =	most probable value	
F-pseudosigma =	nonparametric statistic deviation	
Hu =	upper hinge value	
Hi =	lower hinge value	
mg/L =	milligrams per liter	
Lab =	laboratory code number	
NR =	not rated, less than value reported	
< =	less than	
<u>Constituent</u>		
NH <sub>3</sub> as N	Ammonia as nitrogen	page
NH <sub>3</sub> +Org N as N	Ammonia plus organic nitrogen	120
NO <sub>3</sub> +NO <sub>2</sub> as N	Nitrate plus nitrite as nitrogen	121
Total P as P	Total Phosphorus as phosphorus	122
PO <sub>4</sub> as P	Orthophosphate as phosphorus	123
		124

Table 15. Statistical summary of reported data for standard reference water sample N-51 (nutrient constituents)--Continued  
 NH<sub>3</sub> as N (Ammonia as nitrogen) mg/L






Table 15. Statistical summary of reported data for standard reference water sample N-51 (nutrient constituents)—Continued  
 $\text{NH}_3 + \text{Org. N as N}$  (Ammonia + Organic nitrogen as nitrogen) mg/L



0. Other		22n. Color: Nesslerization						
22. Colorimetric		22p. Color: phenate						
22nitro. Color: nitroprusside		40. Ion selective electrode						
N =		3	7	1	3	31	5	
Minimum =		0.11	0.08	0.29	0.25	0.12	0.24	
Maximum =		0.48	0.35		0.36	0.53	0.50	
Median =					0.23	0.28		
F-pseudosigma =		0.09			0.09			
Lab	Rating	Z-value	0	22	22nitro	22n	22p	40
1	3	-0.72				0.22		
2	0	2.07					0.50	
10	4	-0.29				0.26		
11	3	0.69				0.36		
15	0	2.35				0.53		
16	1	-1.73	0.11					
18	2	-1.15				0.17		
21	4	0.00				0.29		
23	0	2.07				0.50		
25	0	-2.06	0.08					
38	4	-0.39			0.25			
48	3	0.99				0.39		
55	4	0.09		0.30				
56	3	-0.64		0.23				
58	1	1.88				0.48		
59	3	-0.88				0.20		
68	4	0.40		0.33				
70	4	0.00				0.29		
81	2	1.08				0.40		
85	3	-0.88				0.20		
87	1	-1.67				0.12		
89	2	-1.39				0.15		
91	4	0.30				0.32		
96	3	-0.61				0.23		
97	4	0.41				0.33		
102	4	0.10				0.30		
104	4	-0.24				0.27		
105	NR			< 0.2				
118	4	-0.49				0.24		
119	3	0.69				0.36		
127	4	0.18				0.31		
128	3	-0.88				0.20		
129	4	0.07			0.30			
133	4	-0.49				0.24		
134	4	0.08				0.30		
138	4	-0.10				0.28		
140	3	-0.78	0.21					
141	NR			< 1				
142	3	-0.86				0.20		
143	4	0.00				0.29		
145	3	-0.78				0.21		
155	3	-0.75				0.21		
180	3	-0.76				0.21		
190	2	1.09				0.40		
203	3	0.59		0.35				
204	3	-1.00		0.19				
212	4	0.40				0.33		
213	NR			< 1				
215	0	2.17				0.51		
221	1	1.85	0.48					

MPV =	0.29
F-pseudosigma =	0.10
N =	51
Hu =	0.35
HI =	0.21
Lab	Rating
224	1
240	1
241	4
253	4
	Z-value
	0
	22
	22nitro
	22n
	22p
	40

Table 15. Statistical summary of reported data for standard reference water sample N-51 (nutrient constituents)—Continued  
 $\text{NO}_3 + \text{NO}_2$  as N (Nitrate + nitrite as nitrogen) mg/L




0. Other		22cd. Cd diazotization						
7. Ion chromatography		22h. Color: hydrazine diazotization						
22. Colorimetric		22sulf. Color: sulfanilamide						
		N =	2	3	6	11	4	
		Minimum =	0.07	0.00	0.00	0.00	0.01	
		Maximum =	0.40	0.12	0.02	1.16	0.11	
		Median =				0.01		
		F-pseudosigma =				0.02		
Lab	Rating	Z-value	0	7	22	22cd	22h	22sulf
1	NR					< 0.005		
2	NR				< 0.02			
5	NR					< 0.03		
7	NR						< 0.05	
10	NR					< 0.01		
13	NR				< 0.06			
15	NR					< 0.02		
16	1	1.69	0.07					
18	NR					< 0.01		
21	4	-0.03				0.01		
23	NR					< 0.05		
25	4	-0.17			0.01			
36	NR					< 0.05		
38	4	-0.25				0.00		
48	NR						< 0.08	
53	0	10.79	0.40					
56	NR				< 0.02			
59	NR					< 0.04		
64	NR					< 0.02		
68	4	-0.17			0.01			
69	NR					< 0.05		
70	NR					< 0.1		
75	NR					< 0.1		
80	NR				< 0.01			
81	4	-0.08				0.01		
83	NR				< 0.02			
85	3	0.67				0.04		
87	NR					< 0.01		
88	0	32.29				1.16		
89	NR					< 0.05		
91	NR						< 0.02	
92	4	-0.03				0.01		
96	NR						< 0.05	
97	NR						< 0.003	
104	4	-0.20				0.00		
105	NR				< 0.04			
113	NR					< 0.015		
114	NR					< 0.04		
118	1	1.94					0.08	
119	4	0.00			0.01			
127	NR					< 0.01		
128	NR					< 0.01		
129	4	-0.25				0.00		
132	0	2.78					0.11	
133	0	4.64				0.18		
134	4	-0.25					0.00	
138	NR						< 0.005	
140	4	-0.25				0.00		
141	NR						< 0.05	
142	4	0.20				0.02		

MPV = 0.01  
F-pseudosigma = 0.04  
N = 28  
Hu = 0.05  
HI = 0.01


Lab	Rating	Z-value	0	7	22	22cd	22h	22sulf
143	4	0.42					0.03	
145	NR						< 0.02	
146	NR						< 0.05	
149	0	3.06			0.12			
151	NR						< 0.05	
155	4	-0.23					0.00	
158	NR							< 0.02
180	NR						< 0.01	
190	4	0.03					0.01	
191	NR					< 0.01		
193	NR						< 0.02	
196	NR					< 0.05		
197	4	0.25				0.02		
203	NR							< 0.02
204	NR							< 0.02
212	NR							< 0.1
215	NR						< 0.01	
220	4	-0.14						0.01
221	4	0.14				0.02		
224	2	-1.23						0.06
234	NR						< 0.01	
240	NR						< 0.1	
241	4	-0.03						0.01
243	NR							< 0.01
247	NR						< 0.01	

Table 15. Statistical summary of reported data for standard reference water sample N-51 (nutrient constituents)—Continued  
 total P as P (total Phosphorus as phosphorus) mg/L



0. Other				22m. Color:phosphomolybdate								
4. ICP												
22. Colorimetric												
N =	1	0	4	56				MPV =	0.04			
Minimum =	0.07	< 0.075	0.02	0.00				F-pseudosigma =	0.01			
Maximum =			0.08	0.14				N =	61			
Median =				0.04				Hu =	0.05			
F-pseudosigma =				0.01				HI =	0.03			
Lab	Rating	Z-value		0	4	22	22m	Lab	Rating	Z-value	0	
1	3	-0.70				0.03		149	4	0.28		0.04
2	4	0.49				0.05		151	3	0.63		0.05
10	3	0.63				0.05		155	3	-0.66		0.03
11	4	-0.21				0.04		158	3	-0.63		0.03
13	NR					< 0.05		180	4	0.42		0.05
15	0	5.48				0.12		183	3	0.91		0.05
16	1	2.04	0.07					190	0	6.89		0.14
18	3	-0.70				0.03		193	3	-0.70		0.03
21	4	-0.21				0.04		203	4	0.35		0.05
22	2	1.26				0.06		204	2	-1.05		0.03
23	NR					< 0.1		212	3	0.98		0.05
36	NR					< 0.05		213	4	0.00		0.04
38	2	1.05				0.06		215	4	0.00		0.04
39	0	-2.65				0.00		221	3	-0.84	0.03	
46	3	-0.91				0.03		224	3	-0.84		0.03
48	4	0.00				0.04		234	3	-0.70		0.03
55	4	0.07				0.04		240	1	1.76		0.07
56	3	-0.53				0.03		241	4	-0.42		0.03
58	0	6.32				0.13		243	3	0.70		0.05
59	3	-0.70				0.03						
68	0	3.02				0.08						
70	NR					< 0.1						
81	4	-0.42				0.03						
83	NR					< 0.075						
85	4	-0.07				0.04						
87	3	0.91				0.05						
89	4	0.28				0.04						
91	3	0.70				0.05						
92	4	-0.49				0.03						
96	3	-0.63				0.03						
97	0	-2.11				0.01						
102	2	-1.05				0.03						
104	4	0.42				0.05						
105	2	-1.41				0.02						
113	4	-0.42				0.03						
114	4	0.14				0.04						
118	0	-2.11				0.01						
119	2	1.41				0.06						
127	NR					< 0.01						
129	4	-0.07				0.04						
132	3	-0.70				0.03						
133	4	0.00				0.04						
134	4	-0.35				0.04						
138	3	0.70				0.05						
140	NR					< 0.02						
141	2	1.41				0.06						
142	3	0.77				0.05						
143	4	0.42				0.05						
145	4	0.00				0.04						
146	NR					< 0.1						

Table 15. Statistical summary of reported data for standard reference water sample N-51 (nutrient constituents)—Continued  
 $\text{PO}_4$  as P (orthophosphate as phosphorus) mg/L



22m. Color: phosphomolybdate			
0. Other			
7. Ion chromatography			
22. Colorimetric			
N =	1	5	4
Minimum =	0.02	0.01	0.01
Maximum =		0.03	0.02
Median =			0.02
F-pseudosigma =			0.01
Lab	Rating	Z-value	0 7 22 22m
1	NR		< 0.001
2	2	-1.08	0.00
5	0	2.43	0.04
10	4	0.45	0.02
13	NR		< 0.05
15	NR		< 0.02
16	4	0.09	0.02
18	4	0.00	0.02
21	4	-0.09	0.02
23	NR		< 0.1
25	4	-0.09	0.02
28	0	2.16	0.04
33	NR		< 0.01
36	NR		< 0.05
38	3	-0.54	0.01
39	3	-0.72	0.01
46	3	-0.99	0.01
48	4	0.09	0.02
53	0	9.08	0.12
55	0	5.49	0.08
56	4	0.18	0.02
58	0	7.55	0.10
59	NR		< 0.01
64	3	-0.54	0.01
70	NR		< 0.1
80	NR		< 0.05
81	NR		< 0.005
83	3	-0.56	0.01
85	4	0.00	0.02
87	3	-0.72	0.01
88	0	3.06	0.05
89	3	-0.72	0.01
92	3	0.72	0.02
96	NR		< 0.01
97	4	-0.36	0.01
102	2	-1.08	0.00
104	3	-0.99	0.01
105	4	0.27	0.02
113	NR		< 0.004
118	NR		< 0.01
119	NR	-1.44	0.00
127	NR		< 0.05
128	NR		< 0.01
129	3	0.81	0.03
132	3	-0.54	0.01
133	3	-0.54	0.01
134	2	-1.26	0.00
138	2	-1.01	0.00
140	NR		< 0.01
141	0	3.06	0.05

MPV = 0.02  
F-pseudosigma = 0.01  
N = 52  
Hu = 0.02  
HI = 0.01

Lab	Rating	Z-value	0	7	22	22m
142	0	6.03				0.08
143	4	-0.36				0.01
145	NR					< 0.01
146	NR					< 0.05
149	4	0.36			0.02	
151	3	0.81			0.03	
155	3	-0.87				0.01
158	3	0.54				0.02
180	NR					< 0.01
183	1	1.53				0.03
190	0	11.33				0.14
191	4	0.36			0.02	
196	NR				< 0.05	
203	3	0.54				0.02
204	4	-0.36				0.01
212	3	-0.54				0.01
213	NR					< 0.02
215	NR					< 0.01
220	2	-1.35				0.00
221	3	-0.72			0.01	
224	3	0.72				0.02
234	3	-0.99			0.01	
240	NR				< 0.1	
241	4	0.18				0.02
247	4	0.18			0.02	

Table 16. Statistical summary of reported data for standard reference water sample N-52 (nutrients)

<u>Definition of analytical methods, abbreviations, and symbols</u>		
<u>Analytical methods</u>		
0. Other/Not reported		
4. ICP	= inductively coupled plasma	
5. DCP	= direct coupled plasma	
7. IC	= ion chromatography	
20. Titrate: color	= titration: colorimetric (color reagent specified)	
21. Titrate: electro	= titration: electrometric	
22. Color:	= colorimetric [color reagent specified]	
40. Ion electrode	= ion selective electrode	
<u>Abbreviations and symbols</u>		
N =	number of samples	
MPV =	most probable value	
F-pseudosigma =	nonparametric statistic deviation	
Hu =	upper hinge value	
Hi =	lower hinge value	
mg/L =	milligrams per liter	
Lab =	laboratory code number	
NR =	not rated, less than value reported	
< =	less than	
<u>Constituent</u>		
NH <sub>3</sub> as N	Ammonia as nitrogen	<u>page</u>
NH <sub>3</sub> +Org N as N	Ammonia plus organic nitrogen	126
NO <sub>3</sub> +NO <sub>2</sub> as N	Nitrate plus nitrite as nitrogen	127
Total P as P	Total Phosphorus as phosphorus	128
PO <sub>4</sub> as P	Orthophosphate as phosphorus	129
		130

Table 16. Statistical summary of reported data for standard reference water sample N-52 (nutrient constituents)—Continued  
 NH<sub>3</sub> as N (Ammonia as nitrogen) mg/L





Table 16. Statistical summary of reported data for standard reference water sample N-52 (nutrient constituents)—Continued  
 $\text{NH}_3 + \text{Org N as N}$  (Ammonia + Organic nitrogen as nitrogen) mg/L



0. Other			22n. Color: Nesslerization					
20. Titrate: colorimetric			22p. Color: phenate					
22. Colorimetric			40. Ion selective electrode					
	N =		2	4	10	2	34	6
	Minimum =		1.94	2.30	1.44	2.18	1.89	1.87
	Maximum =		2.64	4.00	3.62	2.23	3.70	2.57
	Median =				2.39		2.39	
	F-pseudosigma =				0.32		0.20	
Lab	Rating	Z-value	0	20	22	22N	22p	40
1	3	-0.92					2.16	
3	3	-0.74			2.20			
10	3	0.52				2.48		
11	3	0.97			2.58			
15	3	0.97				2.58		
16	1	-1.91	1.94					
18	0	-2.14				1.89		
25	2	1.24	2.64					
36	3	-0.97			2.15			
38	3	-0.61				2.23		
46	3	-0.56				2.24		
48	4	-0.29				2.30		
55	4	0.16				2.40		
56	1	1.69			2.74			
57	0	5.55	3.60					
58	3	0.56				2.49		
59	4	-0.29				2.30		
68	4	0.25			2.42			
70	4	-0.11				2.34		
81	3	-0.97				2.15		
85	4	0.29				2.43		
87	2	-1.19				2.10		
89	3	-0.65				2.22		
91	4	0.29				2.43		
96	4	0.11				2.39		
97	3	-0.61				2.23		
102	3	-0.61				2.23		
104	4	0.29				2.43		
105	0	-3.89			1.50			
108	4	-0.11			2.34			
113	2	-1.01				2.14		
118	4	0.47				2.47		
119	0	-2.23				1.87		
127	2	1.33				2.66		
128	3	0.61				2.50		
129	3	-0.84			2.18			
133	2	-1.42				2.05		
134	4	-0.29				2.30		
138	4	0.07				2.38		
140	4	-0.07			2.35			
141	3	0.88				2.56		
142	4	-0.16				2.33		
143	3	0.74				2.53		
145	4	-0.11				2.34		
155	2	-1.09				2.12		
180	4	0.47				2.47		
190	1	1.51				2.70		
203	0	4.15			1.44			
204	3	0.52				2.48		
212	3	0.61				2.50		

MPV = 2.37  
F-pseudosigma = 0.22  
N = 58  
Hu = 2.50  
HI = 2.20

Lab	Rating	Z-value	0	20	22	22N	22p	40
213	0	7.35		4.00				
215	3	0.65					2.51	
221	4	-0.29			2.30			
224	0	6.02					3.70	
240	4	0.25						2.42
241	3	-0.98						2.15
249	3	0.92						2.57
253	0	5.64					3.62	
255	NR						< 5	

Table 16. Statistical summary of reported data for standard reference water sample N-52 (nutrient constituents)—Continued  
 $\text{NO}_3 + \text{NO}_2$  as N (Nitrate + nitrite as nitrogen) mg/L






Table 16. Statistical summary of reported data for standard reference water sample N-52 (nutrient constituents)—Continued  
total P as P (total Phosphorus as phosphorus) mg/L



0. Other				22m. Color:phosphomolybdate							
4. ICP											
22. Colorimetric											
N =	1	4	6	65				MPV =	1.60		
Minimum =	1.97	1.50	1.10	0.82				F-pseudosigma =	0.06		
Maximum =					1.69	1.76	4.03	N =	76		
Median =							1.60	Hu =	1.64		
F-pseudosigma =							0.05	Hi =	1.56		
Lab	Rating	Z-value		0	4	22	22m	Lab	Rating	Z-value	
1	3	-0.84				1.55		141	4	-0.17	1.59
3	3	-0.84			1.55			142	3	0.84	1.65
7	3	0.67				1.64		143	4	0.00	1.60
10	4	0.34				1.62		145	2	1.01	1.66
11	4	-0.34				1.58		146	0	40.98	4.03
13	4	0.17				1.61		155	2	-1.37	1.52
15	2	1.35				1.68		158	4	0.00	1.60
16	0	6.24	1.97					180	2	1.35	1.68
18	4	-0.17				1.59		183	4	-0.34	1.58
19	3	0.67				1.64		190	0	-6.41	1.22
22	4	-0.17				1.59		193	2	1.18	1.67
23	0	-2.87				1.43		203	4	0.44	1.63
25	2	1.35	1.68					204	4	-0.17	1.59
36	0	-3.37				1.40		212	4	0.00	1.60
38	4	-0.44				1.57		213	4	-0.34	1.58
42	1	-1.69	1.50					215	0	-13.15	0.82
46	3	-0.84				1.55		221	0	2.70	1.76
48	1	-1.69				1.50		224	3	-0.84	1.55
55	3	-0.67				1.56		234	4	0.34	1.62
56	0	-8.43	1.10					240	0	7.25	2.03
57	4	0.00				1.60		241	3	-0.67	1.56
58	2	-1.01				1.54		243	2	1.18	1.67
59	4	0.00				1.60		248	3	-0.67	1.56
68	3	0.67	1.64					249	0	2.53	1.75
70	4	0.34				1.62		253	0	-2.87	1.43
81	4	-0.17				1.59		255	4	-0.34	1.58
83	4	0.00	1.60								
85	4	0.17				1.61					
86	1	1.52	1.69								
87	3	0.51				1.63					
89	4	0.34				1.62					
91	3	0.51				1.63					
92	4	0.00				1.60					
96	4	-0.34				1.58					
97	2	-1.38				1.52					
102	3	-0.51				1.57					
104	4	0.17				1.61					
105	3	0.84	1.65								
107	2	1.01				1.66					
113	4	0.00				1.60					
114	4	0.00				1.60					
118	0	5.06				1.90					
119	4	0.34				1.62					
127	2	-1.01				1.54					
129	1	-1.62				1.50					
132	3	0.51				1.63					
133	3	0.51				1.63					
134	3	0.84				1.65					
138	2	-1.35				1.52					
140	2	1.01			1.66						

Table 16. Statistical summary of reported data for standard reference water sample N-52 (nutrient constituents)—Continued  
 PO<sub>4</sub> as P (orthophosphate as phosphorus) mg/L



22m. Color:phosphomolybdate				
0. Other				
7. Ion chromatography				
22. Colorimetric				
N =	1	10	8	62
Minimum =	1.45	1.05	0.58	0.12
Maximum =		1.24	1.25	3.49
Median =		1.18	1.11	1.16
F-pseudosigma =		0.06	0.26	0.05

Lab	Rating	Z-value	0	7	22	22m
1	3	0.67			1.20	
3	3	-0.67			1.12	
5	3	-0.67			1.12	
7	1	-1.85			1.05	
10	3	0.67			1.20	
11	3	-0.51			1.13	
13	3	0.67			1.20	
15	4	-0.17			1.15	
16	0	4.89	1.45			
18	3	0.51			1.19	
19	4	-0.34			1.14	
23	4	-0.34			1.14	
25	4	0.17			1.17	
26	2	1.35			1.24	
30	4	0.00			1.16	
33	3	-0.67	1.12			
36	4	-0.17			1.15	
38	3	-0.57			1.13	
42	0	-9.80			0.58	
46	0	-17.54			0.12	
48	1	-1.85			1.05	
53	0	39.32			3.49	
55	3	0.84			1.21	
56	0	-8.38			0.66	
57	0	2.36			1.30	
58	0	-4.05			0.92	
59	4	0.00			1.16	
64	4	0.10			1.17	
70	4	0.00			1.16	
80	2	1.35			1.24	
81	1	-1.69			1.06	
83	3	0.67	1.20			
85	4	-0.17			1.15	
87	3	0.51			1.19	
88	3	-0.84			1.11	
89	4	0.00			1.16	
92	4	0.34			1.18	
96	4	-0.34			1.14	
97	4	0.00			1.16	
102	3	-0.84			1.11	
104	4	0.13			1.17	
105	4	0.17			1.17	
107	3	-0.67			1.12	
108	1	2.02			1.28	
113	4	-0.17			1.15	
118	4	0.34			1.18	
119	4	0.34			1.18	
127	4	0.00			1.16	
128	3	0.51			1.19	
129	0	-2.09			1.04	

MPV = 1.16  
 F-pseudosigma = 0.06  
 N = 81  
 Hu = 1.20  
 Hi = 1.12

Lab	Rating	Z-value	0	7	22	22m
132	1	1.52			1.25	
133	3	-0.67			1.12	
134	3	0.51			1.19	
138	4	-0.17			1.15	
140	0	-2.53			1.01	
141	3	0.84			1.21	
142	4	-0.17			1.15	
143	4	-0.17			1.15	
145	4	0.17			1.17	
146	0	39.12			3.48	
155	2	-1.32			1.08	
158	3	-0.94			1.10	
180	3	0.51			1.19	
183	2	-1.01			1.10	
190	4	-0.34			1.14	
191	3	0.84			1.21	
203	2	1.08			1.22	
212	3	0.67			1.20	
213	1	-1.69			1.06	
215	0	-10.45			0.54	
220	0	2.70			1.32	
221	2	-1.18			1.09	
224	0	2.53			1.31	
234	3	-0.84			1.11	
240	3	0.84			1.21	
241	4	0.00			1.16	
247	3	0.57			1.19	
248	2	1.18			1.23	
249	4	0.34			1.18	
253	3	0.67			1.20	
255	3	-0.51			1.13	

Table 17. Statistical summary of reported data for standard reference water sample P-27 (low ionic strength constituents)

Definition of analytical methods, abbreviations, and symbols	
<u>Analytical methods</u>	
0 Other/Not reported	=
1 AA: direct, air	= atomic absorption: direct,air
2 AA: direct, N <sub>2</sub> O	= atomic absorption: direct,nitrous oxide
3 AA: graphite furnace	= atomic absorption: graphite furnace
4 ICP	= inductively coupled plasma
5 DCP	= direct current plasma
6 ICP/MS	= mass spectrometry/inductively coupled plasma
7 IC	= ion chromatography
12 Flame emission	= flame emission
20 Titrate: color	= titration: colorimetric [color reagent specified]
21 Titrate: electro	= titration: electrometric
22 Color:	= colorimetric [color reagent specified]
40 Ion electrode	= ion selective electrode
41 Electro	= electrometric: [type meter specified]
50 Gravimetric	= gravimetric: [precipitate specified]
51 Turbidimetric	= turbidimetric: [precipitate specified]


  

<u>Abbreviations and symbols</u>	
N =	number of samples
MPV =	most probable value
F-pseudosigma =	nonparametric statistic deviation
Hu =	upper hinge value
HI =	lower hinge value
mg/L =	milligrams per liter
µS/cm =	microsiemens per centimeter at 25° C
Lab =	laboratory code number
NR =	not rated, less than value reported
< =	less than

<u>Constituent</u>	<u>page</u>
Acid	Acidity as CaCO <sub>3</sub>
Ca	Calcium
Cl	Chloride
F	Fluoride
K	Potassium
Mg	Magnesium
Na	Sodium
pH	
PO <sub>4</sub> as P	Orthophosphate as Phosphorus
SO <sub>4</sub>	Sulfate
Sp Cond	Specific Conductance

Table 17. Statistical summary of reported data for standard reference water sample P-27 (low ionic strength constituents)—Continued  
 Acidity (as CaCO<sub>3</sub>) mg/L




0. Other
20. Titrate: colorimetric
21. Titrate: electrometric
N = 2 3 17
Minimum = 5.08 0.42 0.20
Maximum = 5.10 15.00 27.00
Median = 4.01
F-pseudosigma = 3.19

MPV = 4.74  
 F-pseudosigma = 3.19  
 N = 22  
 Hu = 6.57  
 HI = 2.26

Lab	Rating	Z-value	0	20	21
1	3	-0.70		2.50	
3	NR			< 10	
5	3	0.84		7.42	
7	4	0.14		5.20	
15	NR			< 2	
25	0	6.97		27.00	
38	3	0.66		6.85	
81	2	-1.33		0.50	
83	4	0.11	5.10		
89	3	-0.78		2.26	
105	2	-1.35		0.42	
109	3	0.57		6.57	
132	4	0.11	5.08		
141	3	-0.73		2.40	
146	NR			< 10	
190	4	0.39		6.00	
215	4	-0.36		3.60	
220	3	0.77		7.21	
224	4	-0.23		4.01	
240	3	-0.90		1.88	
247	2	-1.42		0.20	
257	2	-1.09		1.25	
272	0	3.21	15.00		
273	4	0.14		5.20	
276	4	-0.11	4.40		

Table 17. Statistical summary of reported data for standard reference water sample P-27 (low ionic strength constituents)—Continued  
 Ca (Calcium) mg/L




0. Other		4. ICP					
1. AA: direct air		7. Ion chromatography					
2. AA: direct nitrous oxide		20. Titrate: colorimetric					
N =		3	13	3	29	4	7
Minimum =		1.06	1.42	2.40	2.40	2.19	1.60
Maximum =		2.37	2.60	2.80	3.30	3.16	4.28
Median =		2.39			2.60		2.50
F-pseudosigma =		0.16			0.19		0.66
Lab	Rating	Z-value	0	1	2	4	7
1	4	-0.12			2.50		
2	2	1.09			2.80		
3	3	-0.53			2.40		
5	4	-0.49			2.41		
15	2	1.19			2.82		
23	4	0.29	2.60				
25	4	-0.12			2.50		
26	1	1.77			2.96		
28	2	1.31			2.85		
33	3	-0.66	2.37				
36	3	-0.53			2.40		
38	4	-0.25			2.47		
42	0	3.16			3.30		
46	4	-0.37			2.44		
48	4	0.00			2.53		
58	0	-3.41	1.70				
59	NR				< 5		
64	4	-0.12			2.50		
81	4	-0.45			2.42		
83	4	-0.12			2.50		
89	3	-0.99	2.29				
105	4	0.21			2.58		
107	3	-0.57	2.39				
109	4	0.21			2.58		
110	0	-4.43	1.45				
113	2	1.35			2.86		
119	4	0.16			2.57		
132	3	0.74			2.71		
134	4	0.30			2.60		
138	4	0.29			2.60		
140	3	-0.57	2.39				
141	3	0.86			2.74		
145	3	0.57			2.67		
146	4	-0.41			2.43		
155	3	0.79			2.72		
158	0	2.59			3.16		
180	4	0.37			2.62		
190	2	-1.40			2.19		
191	3	-0.86	2.32				
204	3	0.90			2.75		
215	4	0.29			2.60		
220	4	-0.12	2.50				
221	4	-0.04	2.52				
224	4	0.26			2.59		
235	3	0.90			2.75		
240	4	0.49			2.65		
241	3	-0.94	2.30				
255	4	0.16			2.57		
256	4	-0.12			2.50		
257	2	1.11			2.80		

MPV = 2.53  
 F-pseudosigma = 0.24  
 N = 59  
 Hu = 2.73  
 Hi = 2.40

Lab	Rating	Z-value	0	1	2	4	7	20
258	0	-3.82						1.60
261	0	3.00						3.26
262	4	-0.12			2.50			
265	1	1.52						2.90
268	0	-4.56			1.42			
270	0	-6.03	1.06					
271	0	7.18						4.28
272	3	-0.51						2.41
273	2	1.11						2.80
276	0	-3.00						1.80

Table 17. Statistical summary of reported data for standard reference water sample P-27 (low ionic strength constituents)—Continued  
 Cl (Chloride) mg/L



7. Ion chromatography

8. AA: cold vapor

20. Titrate: colorimetric

21. Titrate: electrometric

22. Colorimetric

40. Ion selective electrode

MPV = 1.20

F-pseudosigma = 0.49

N = 57


Hu = 1.78

HI = 1.13

	N =	29	1	12	4	10	1
Minimum =	0.27	13.00	0.00	1.01	0.87	1.27	
Maximum =	2.39		8.40	1.40	2.00		
Median =	1.14		2.00		1.30		
F-pseudosigma =	0.10		1.50		0.47		

Lab	Rating	Z-value	7	10	20	21	22	40
1	4	0.00	1.20					
2	4	-0.15	1.13					
3	2	1.19				1.78		
5	0	2.45	2.39					
7	4	-0.14	1.13					
15	3	0.89	1.63					
23	4	0.14			1.27			
25	4	0.00	1.20					
26	4	-0.14	1.13					
33	4	0.08	1.24					
36	NR			< 5				
39	1	1.65		2.00				
42	4	-0.21	1.10					
46	3	0.92			1.65			
48	1	1.65			2.00			
58	3	0.60		1.49				
59	4	-0.21	1.10					
64	4	0.00	1.20					
81	4	-0.10			1.15			
83	NR			< 5				
89	4	0.19		1.29				
105	4	-0.12	1.14					
107	NR				< 0.6			
109	4	-0.39			1.01			
110	4	-0.15	1.13					
113	1	1.61	1.98					
119	4	-0.04	1.18					
134	4	-0.05	1.17					
138	4	-0.27	1.07					
140	3	-0.68			0.87			
141	4	0.41			1.40			
143	4	0.00			1.20			
145	4	-0.33	1.04					
146	3	0.68			1.53			
158	1	-1.92	0.27					
180	4	-0.41			1.00			
183	4	-0.39			1.01			
190	4	-0.23	1.09					
191	2	1.30	1.83					
196	4	-0.02	1.19					
197	3	-0.54	0.94					
203	NR			< 2				
204	4	0.41			1.40			
215	1	1.65		2.00				
220	4	-0.04			1.18			
221	3	0.56		1.47				
224	4	0.33	1.36					
240	4	-0.14	1.13					
241	0	4.74		3.50				
247	2	1.24	1.80					

Table 17. Statistical summary of reported data for standard reference water sample P-27 (low ionic strength constituents)—Continued  
F (Fluoride) mg/L



7. Ion chromatography		40. Ion selective electrode		50. Gravimetric	
8. AA: cold vapor		21. Titrate: electrometric			
N =	4	1	1	24	1
Minimum =	0.08	0.40	0.07	0.01	0.07
Maximum =	0.45			0.54	
Median =				0.10	
F-pseudosigma =				0.03	
Lab	Rating	Z-value	7	10	21
1	NR				< 0.1
3	2	-1.13			0.063
7	NR		< 0.5		
15	4	0.25			0.108
23	NR				< 0.1
25	4	0.00			0.100
26	3	0.61	0.120		
36	NR				< 0.1
39	0	-2.88			0.006
48	0	4.29			0.240
58	4	-0.25			0.092
59	NR				< 0.2
81	3	-0.95			0.069
83	0	2.15			0.170
89	0	13.40			0.537
105	NR		< 0.2		
107	3	-0.71			0.077
109	4	-0.31			0.090
113	4	-0.30			0.090
119	4	0.00			0.100
134	3	0.61			0.120
138	4	0.03			0.101
140	2	-1.44			0.053
141	3	0.92			0.130
145	3	-0.61	0.080		
146	NR				< 0.2
190	3	-0.83		0.073	
196	4	0.18			0.106
215	4	0.00			0.100
224	0	10.73	0.450		
240	3	-0.74			0.076
241	4	0.49			0.116
247	4	0.00	0.100		
255	NR				< 0.2
257	4	-0.46			0.085
258	0	2.15			0.170
262	3	-0.55			0.082
265	NR				< 0.1
272	0	9.35			0.405
273	0	9.20		0.400	

$$\begin{aligned} \text{MPV} &= 0.100 \\ \text{F-pseudosigma} &= 0.033 \\ N &= 31 \\ Hu &= 0.125 \\ HI &= 0.081 \end{aligned}$$

Table 17. Statistical summary of reported data for standard reference water sample P-27 (low ionic strength constituents)—Continued  
K (Potassium) mg/L





Table 17. Statistical summary of reported data for standard reference water sample P-27 (low ionic strength constituents)—Continued  
Mg (Magnesium) mg/L



0. Other		4. ICP					
1. AA: direct air		7. Ion chromatography					
2. AA: direct nitrous oxide		20. Titrate: colorimetric					
		N =	4	13	2	27	4
		Minimum =	0.440	0.300	0.630	0.387	0.427
		Maximum =	1.940	0.490	0.720	3.580	0.550
		Median =	0.450		0.460		10.370
		F-pseudosigma =	0.018		0.045		
Lab	Rating	Z-value	0	1	2	4	7
1	4	-0.21				0.450	
2	1	1.70					0.545
3	3	-0.82			0.420		
5	3	-0.53			0.434		
15	0	3.59			0.639		
23	NR		< 0.5				
25	3	-0.82			0.420		
26	4	0.19				0.470	
28	4	-0.21			0.450		
33	4	-0.41	0.440				
36	NR		< 0.5				
38	4	-0.47	0.437				
42	3	0.80			0.500		
46	4	-0.37			0.442		
48	3	0.80			0.500		
58	0	-2.83	0.320				
59	NR				< 5		
64	4	-0.21	0.450				
81	3	-0.55			0.433		
83	3	-0.53			0.434		
89	4	-0.15	0.453				
105	2	-1.48			0.387		
107	4	-0.21	0.450				
109	3	0.59	0.490				
110	4	-0.21	0.450				
113	3	0.88			0.504		
119	2	-1.02			0.410		
132	4	-0.21			0.450		
134	4	-0.27			0.447		
138	4	-0.01			0.460		
140	4	0.13	0.467				
141	3	0.51			0.486		
145	3	0.59			0.490		
146	NR				< 0.5		
158	1	1.80			0.550		
180	4	0.17			0.469		
190	3	-0.67				0.427	
191	4	0.19	0.470				
204	0	62.81			3.580		
215	4	-0.01			0.460		
220	2	-1.22	0.400				
221	4	0.01	0.461				
224	2	1.18			0.519		
235	3	0.59			0.490		
240	3	-0.61			0.430		
241	0	-3.23	0.300				
255	4	0.19			0.470		
256	0	20.93	1.500				
257	0	3.41			0.630		
258	0	29.79	1.940				

MPV = 0.461  
F-pseudosigma = 0.050  
N = 54  
Hu = 0.504  
HI = 0.437

Lab	Rating	Z-value	0	1	2	4	7	20
261	0	29.99						1.950
262	4	-0.41				0.440		
265	0	5.22					0.720	
268	4	0.39				0.480		
271	0	199.52						10.370
272	0	-9.27						0.000
273	2	1.22					0.521	
276	0	18.92						1.400

Table 17. Statistical summary of reported data for standard reference water sample P-27 (low ionic strength constituents)—Continued  
Na (Sodium) mg/L





Table 17. Statistical summary of reported data for standard reference water sample P-27 (low ionic strength constituents)—Continued  
pH



8. AA: cold vapor

41. Direct reading

N =	1	61
Minimum =	7.03	4.69
Maximum =		8.23
Median =		6.90
F-pseudosigma =		0.32


MPV =	6.92
F-pseudosigma =	0.32
N =	62
Hu =	7.03
Hi =	6.60

Lab	Rating	Z-value	10	41
1	4	0.23	6.99	
2	1	-1.83	6.33	
3	0	-2.18	6.22	
5	0	-2.21	6.21	
7	4	0.20	6.98	
15	0	-2.14	6.23	
23	4	0.05	6.93	
25	3	0.67	7.13	
26	4	0.11	6.95	
33	3	-0.55	6.74	
36	4	0.14	6.96	
38	3	0.58	7.10	
39	4	-0.05	6.90	
46	0	2.08	7.58	
48	1	-1.61	6.40	
58	0	3.62	8.07	
59	3	-0.70	6.69	
64	4	0.30	7.01	
81	4	-0.11	6.88	
89	4	0.30	7.01	
92	4	0.36	7.03	
105	3	-0.83	6.65	
107	3	0.74	7.15	
109	0	-2.33	6.17	
110	4	0.36	7.03	
113	4	0.42	7.05	
119	2	-1.33	6.49	
132	3	0.74	7.15	
134	4	0.05	6.93	
138	4	-0.42	6.78	
140	1	-1.55	6.42	
141	4	0.14	6.96	
143	3	-0.70	6.69	
145	2	1.21	7.30	
146	4	-0.05	6.90	
155	2	-1.21	6.53	
158	0	-6.96	4.69	
180	3	0.58	7.10	
183	1	1.71	7.46	
190	3	-0.99	6.60	
203	2	-1.30	6.50	
204	3	-0.55	6.74	
215	4	0.42	7.05	
221	0	-2.58	6.09	
224	4	-0.49	6.76	
240	3	-0.55	6.74	
241	4	-0.27	6.83	
243	0	4.12	8.23	
244	4	0.20	6.98	
247	4	0.30	7.01	

Lab	Rating	Z-value	10	41
255	3	0.89	7.20	
256	4	0.39	7.04	
257	4	0.36	7.03	
258	1	-1.92	6.30	
261	2	-1.49	6.44	
262	0	-5.43	5.18	
265	4	-0.49	6.76	
268	4	-0.49	6.76	
271	4	0.20	6.98	
272	3	0.93	7.21	
273	4	0.36	7.03	
276	4	-0.08	6.89	

Table 17. Statistical summary of reported data for standard reference water sample P-27 (low ionic strength constituents)—Continued  
 $\text{PO}_4$  as P (orthophosphate as phosphorus)

mg/L




0. Other	22. Colorimetric
4. ICP	22m. Color: phosphomolybdate
7. Ion chromatography	22sncl. Color: stannous fluoride
N =	1      0      1      5      13      0
Minimum =	0.000      0.400      0.002      0.000
Maximum =	0.150      0.150      0.030
Median =	0.002
F-pseudosigma =	0.007

MPV = 0.002  
F-pseudosigma = 0.014  
N = 20  
Hu = 0.021  
HI = 0.001

Lab	Rating	Z-value	0	4	7	22	22m	22sncl
1	NR					< 0.001		
3	NR					< 0.01		
7	NR					< 0.16		
15	NR						< 0.02	
23	NR						< 0.01	
26	NR					< 0.5		
28	4	0.00					0.002	
33	NR					< 0.01		
36	NR						< 0.025	
38	4	-0.07					0.001	
39	NR						< 0.005	
42	4	0.00					0.002	
48	NR						< 0.005	
58	3	0.55					0.010	
64	4	-0.07					0.001	
81	NR						< 0.005	
83	3	0.51					0.009	
89	NR						< 0.002	
92	NR						< 0.005	
105	4	0.00					0.002	
107	NR						< 0.002	
113	NR						< 0.004	
119	NR	-0.14					0.000	
132	NR						< 0.01	
134	NR	-0.14					0.000	
138	NR						< 0.004	
140	NR						< 0.01	
141	NR						< 0.05	
143	4	0.00					0.002	
145	NR						< 0.01	
146	NR						< 0.05	
155	4	-0.06					0.001	
180	NR						< 0.01	
190	2	1.31					0.021	
191	0	27.53					0.400	
196	NR						< 0.05	
204	NR						< 0.002	
215	1	1.94					0.030	
224	4	-0.07					0.001	
235	NR						< 0.5	
240	NR						< 0.1	
241	4	0.07					0.003	
247	NR						< 0.001	
257	2	1.25					0.020	
258	0	2.50					0.038	
261	NR						< 0.05	
271	0	10.24					0.150	
273	NR						-0.14 0.000	


Table 17. Statistical summary of reported data for standard reference water sample P-27 (low ionic strength constituents)—Continued  
 $\text{SO}_4$  (Sulfate) mg/L



0. Other		22. Colorimetric					
4. ICP		50. Gravimetric					
7. Ion chromatography		51. Turbidimetric					
N		2	3	30	5	3	7
Minimum		2.00	2.52	2.19	2.03	0.41	0.00
Maximum		17.04	2.89	30.50	6.38	3.70	60.00
Median					2.40		2.00
F-pseudosigma					0.21		3.04
Lab	Rating	Z-value	0	4	7	22	50
1	4	0.04			2.43		
2	4	-0.27			2.32		
3	0	2.66			3.40		
5	0	5.27			4.37		
7	4	-0.04			2.40		
15	4	0.23			2.50		
23	NR				< 2.5		
26	4	0.12			2.46		
33	4	-0.20			2.34		
36	NR				< 5		
39	4	-0.31			2.30		
42	NR				< 2.5		
46	4	-0.05			2.40		
48	0	155.36					60.00
58	2	-1.12				2.00	
59	4	-0.31			2.30		
64	4	0.07			2.44		
81	3	0.63					2.65
83	4	0.28			2.52		
89	4	0.18			2.48		
92	2	-1.44					1.88
105	4	-0.36			2.28		
109	0	3.47					3.70
110	4	-0.22			2.34		
113	NR				< 1		
119	4	-0.09			2.38		
134	4	-0.19			2.35		
138	4	-0.28			2.31		
140	2	-1.12	2.00				
141	4	0.42			2.57		
145	3	-0.61			2.18		
146	NR				< 5		
158	0	75.77			30.50		
180	4	0.50					2.60
190	2	1.04			2.80		
191	4	-0.09			2.38		
196	4	-0.04			2.40		
197	4	-0.15			2.36		
203	2	1.20			2.86		
204	2	-1.04			2.03		
215	1	2.01					3.16
220	0	17.50					8.90
221	0	-5.40					0.41
224	3	0.55			2.62		
235	2	1.28	2.89				
240	3	0.71			2.68		
241	2	-1.12					2.00
247	2	1.04			2.80		
256	0	39.46	17.04				
257	2	1.50			2.97		

MPV =	2.42							
F-pseudosigma =	0.37							
N =	50							
Hu =	2.80							
HI =	2.30							
Lab	Rating	Z-value	0	4	7	22	50	51
258	2	-1.15						
261	0	-5.18	< 0.5					
262	0	10.70						
265	4	-0.07					2.39	
268	4	-0.45					2.25	
271	0	-6.52						0.00

Table 17. Statistical summary of reported data for standard reference water sample P-27 (low ionic strength constituents)—Continued



0. Other				
41. Direct reading				
	N =	1	56	
	Minimum =	50.7	20.0	
	Maximum =		263.0	
	Median =		26.1	
	F-pseudosigma =		1.07	
Lab	Rating	Z-value	0	41
1	3	0.63	26.8	
2	0	-4.33	21.3	
3	4	0.45	26.6	
5	4	0.18	26.3	
7	3	-0.72	25.3	
15	4	0.00	26.1	
23	3	0.81	27.0	
25	3	0.81	27.0	
26	3	0.99	27.2	
33	4	-0.27	25.8	
36	1	-1.98	23.9	
38	4	0.00	26.1	
39	2	1.26	27.5	
48	2	-1.44	24.5	
58	4	0.00	26.1	
59	3	-0.63	25.4	
64	3	-0.54	25.5	
81	4	0.18	26.3	
89	4	-0.27	25.8	
105	3	0.81	27.0	
107	4	-0.09	26.0	
109	3	0.81	27.0	
113	4	-0.09	26.0	
119	4	-0.09	26.0	
132	0	-5.04	20.5	
134	4	0.45	26.6	
138	3	0.54	26.7	
140	0	5.76	32.5	
141	3	0.63	26.8	
143	4	-0.18	25.9	
145	1	1.71	28.0	
146	0	4.05	30.6	
155	1	1.53	27.8	
158	4	0.00	26.1	
180	4	-0.09	26.0	
183	0	-2.61	23.2	
190	2	1.35	27.6	
193	3	0.63	26.8	
203	0	213.05	263.0	
204	4	0.18	26.3	
215	3	0.54	26.7	
224	3	-0.99	25.0	
240	2	-1.26	24.7	
241	0	-5.53	20.0	
243	4	-0.36	25.7	
244	4	-0.18	25.9	
247	3	0.94	27.1	
255	4	0.09	26.2	
257	3	-0.54	25.5	
258	0	2.97	29.4	

MPV =	26.1
F-pseudosigma =	1.1
N =	57
Hu =	27.0
Hi =	25.5

Table 18. *Statistical summary of reported data for standard reference water sample Hg-23 (mercury)*

---

Definition of analytical methods, abbreviations, and symbols

---

Analytical methods

- 0 Other/Not reported  
6 ICP/MS = mass spectrometry/inductively coupled plasma  
8 AA: cold vapor = atomic absorption: cold vapor
- 

Abbreviations and symbols


N =	number of samples
MPV =	most probable value
F-pseudosigma =	nonparametric statistic deviation
Hu =	upper hinge value
Hi =	lower hinge value
µg/L =	micrograms per liter
Lab =	laboratory code number
NR =	not rated, less than value reported
< =	less than

---

<u>Constituent</u>		<u>page</u>
Hg	Mercury	144

---

Table 18. Statistical summary of reported data for standard reference water sample Hg-23 (mercury)--Continued  
 Hg (Mercury)  $\mu\text{g/L}$



0. Other  
 6. ICP/MS  
 8. AA: cold vapor

N =	1	2	55
Minimum =	0.33	0.20	0.25
Maximum =		0.29	4.40
Median =			0.35
F-pseudosigma =			0.07

MPV =	0.34
F-pseudosigma =	0.07
N =	58
Hu =	0.40
Hi =	0.30

Lab	Rating	Z-value	0	6	8
1	3	-0.59		0.30	
3	0	56.43		4.40	
7	4	-0.03		0.34	
10	4	-0.24		0.33	
11	3	-0.87		0.28	
13	NR		< 0.4		
15	3	-0.73		0.29	
16	4	0.10		0.35	
18	3	0.52		0.38	
26	4	0.38		0.37	
28	0	28.61		2.40	
32	3	-0.73		0.29	
34	4	-0.37		0.32	
36	3	-0.59		0.30	
39	3	-0.59		0.30	
42	4	-0.31		0.32	
46	4	-0.41		0.31	
48	4	0.38		0.37	
50	0	2.47		0.52	
51	3	0.66		0.39	
55	3	-0.65		0.30	
58	NR		< 0.5		
59	4	-0.31		0.32	
68	2	-1.23		0.25	
69	4	-0.17		0.33	
70	3	0.76		0.40	
76	3	-0.77	0.29		
81	3	-0.59		0.30	
86	3	-0.51		0.31	
87	0	2.19		0.50	
89	3	0.66		0.39	
96	4	0.35		0.37	
97	NR		< 0.42		
105	4	0.34		0.37	
108	0	2.61	0.53		
113	3	0.66		0.39	
119	3	0.66		0.39	
127	4	-0.38		0.32	
133	4	-0.17	0.33		
134	3	-0.87		0.28	
138	4	-0.31		0.32	
141	2	1.19		0.43	
142	2	1.36		0.44	
144	4	0.03		0.35	
145	2	1.22		0.43	
146	4	0.23		0.36	
149	4	0.10		0.35	
193	2	1.22		0.43	
212	3	-0.87		0.28	
213	0	5.67		0.75	

Table 19. Most probable values for constituents and properties in standard reference samples distributed in September 1996

[MPV, most probable value;  $\mu\text{g/L}$ , microgram per liter;  $\text{mg/L}$ , milligram per liter;  $\mu\text{S/cm}$ , microsiemen per centimeter at 25 degrees Celsius]

**T-143 (trace constituents)**

Analyte	MPV	F-pseudosigma	N	Analyte	MPV	F-pseudosigma	N
Ag	19.8 $\mu\text{g/L}$	1.4	65	Mg	10.4 $\text{mg/L}$	0.5	88
Al	22.1 $\mu\text{g/L}$	8.3	50	Mn	18.2 $\mu\text{g/L}$	1.9	86
As	15.2 $\mu\text{g/L}$	1.2	66	Mo	36.1 $\mu\text{g/L}$	4.3	54
B	35.0 $\mu\text{g/L}$	5.2	35	Na	34.0 $\text{mg/L}$	1.6	88
Ba	81.9 $\mu\text{g/L}$	4.5	68	Ni	71.0 $\mu\text{g/L}$	5.0	81
Be	8.50 $\mu\text{g/L}$	0.66	61	Pb	83.4 $\mu\text{g/L}$	7.1	84
Ca	53.7 $\text{mg/L}$	2.2	86	Sb	16.6 $\mu\text{g/L}$	1.5	46
Cd	19.1 $\mu\text{g/L}$	1.5	81	Se	9.63 $\mu\text{g/L}$	1.64	60
Co	17.0 $\mu\text{g/L}$	1.2	53	$\text{SiO}_2$	23.4 $\text{mg/L}$	1.7	56
Cr	37.0 $\mu\text{g/L}$	2.6	79	Sr	306 $\mu\text{g/L}$	15	45
Cu	22.3 $\mu\text{g/L}$	1.9	90	Tl	10.0 $\mu\text{g/L}$	1.0	33
Fe	222 $\mu\text{g/L}$	14	93	U	12.0 $\mu\text{g/L}$	0.9	9
K	2.50 $\text{mg/L}$	0.21	84	V	30.0 $\mu\text{g/L}$	3.0	54
Li	18.0 $\mu\text{g/L}$	2.1	34	Zn	20.0 $\mu\text{g/L}$	2.2	75

**T-145 (trace constituents)**

Analyte	MPV	F-pseudosigma	N	Analyte	MPV	F-pseudosigma	N
Ag	7.55 $\mu\text{g/L}$	0.92	61	Mg	8.68 $\text{mg/L}$	0.45	85
Al	67.6 $\mu\text{g/L}$	11.0	64	Mn	20.9 $\mu\text{g/L}$	1.5	85
As	9.88 $\mu\text{g/L}$	1.04	66	Mo	9.23 $\mu\text{g/L}$	1.29	43
B	45.6 $\mu\text{g/L}$	5.8	36	Na	41.2 $\text{mg/L}$	1.9	84
Ba	37.1 $\mu\text{g/L}$	1.9	63	Ni	11.0 $\mu\text{g/L}$	1.3	66
Be	9.04 $\mu\text{g/L}$	0.70	62	Pb	12.7 $\mu\text{g/L}$	1.2	80
Ca	30.7 $\text{mg/L}$	1.3	84	Sb	8.80 $\mu\text{g/L}$	0.96	39
Cd	9.33 $\mu\text{g/L}$	0.82	80	Se	10.1 $\mu\text{g/L}$	1.3	60
Co	10.0 $\mu\text{g/L}$	0.9	53	$\text{SiO}_2$	11.3 $\text{mg/L}$	0.7	53
Cr	15.3 $\mu\text{g/L}$	1.4	78	Sr	203 $\mu\text{g/L}$	9	44
Cu	11.0 $\mu\text{g/L}$	1.4	84	Tl	15.3 $\mu\text{g/L}$	2.7	37
Fe	101 $\mu\text{g/L}$	8	89	U	1.10 $\mu\text{g/L}$	0.10	9
K	2.13 $\text{mg/L}$	0.16	83	V	11.7 $\mu\text{g/L}$	1.7	46
Li	27.3 $\mu\text{g/L}$	2.5	36	Zn	10.0 $\mu\text{g/L}$	2.4	70

**M-140 (major constituents)**

Analyte	MPV	F-pseudosigma	N	Analyte	MPV	F-pseudosigma	N
Alkalinity	114 $\text{mg/L}$	3	103	Na	39.0 $\text{mg/L}$	1.9	92
B	41.6 $\text{mg/L}$	5.2	40	total P	0.032 $\text{mg/L}$	0.011	57
Ca	60.7 $\text{mg/L}$	3.1	96	pH	8.28 units	0.17	108
Cl	25.8 $\text{mg/L}$	1.4	104	$\text{SiO}_2$	7.35 $\text{mg/L}$	0.46	66
DSRD	382 $\text{mg/L}$	16	67	$\text{SO}_4$	150 $\text{mg/L}$	7	98
F	0.530 $\text{mg/L}$	0.037	71	Sp Cond	600 $\mu\text{S/cm}$	19	100
K	2.58 $\text{mg/L}$	0.14	87	Sr	671 $\mu\text{g/L}$	31	43
Mg	18.0 $\text{mg/L}$	1.0	96	V	3.42 $\mu\text{g/L}$	3.01	19

**N-51 (nutrients)**

Analyte	MPV	F-pseudosigma	N
$\text{NH}_3 \text{ as N}$	0.07 $\text{mg/L}$	0.05	59
$\text{NH}_3\text{-OrgN as N}$	0.29 $\text{mg/L}$	0.10	51
$\text{NO}_3\text{-NO}_2 \text{ as N}$	0.01 $\text{mg/L}$	0.04	28
Total P as P	0.04 $\text{mg/L}$	0.01	61
$\text{PO}_4 \text{ as P}$	0.02 $\text{mg/L}$	0.01	52

**N-52 (nutrients)**

Analyte	MPV	F-pseudosigma	N
$\text{NH}_3 \text{ as N}$	1.33 $\text{mg/L}$	0.09	84
$\text{NH}_3\text{-OrgN as N}$	2.37 $\text{mg/L}$	0.22	58
$\text{NO}_3\text{-NO}_2 \text{ as N}$	1.72 $\text{mg/L}$	0.10	86
total P as P	1.60 $\text{mg/L}$	0.06	76
$\text{PO}_4 \text{ as P}$	1.16 $\text{mg/L}$	0.06	81

**P-27 (low ionic strength constituents)**

Analyte	MPV	F-pseudosigma	N	Analyte	MPV	F-pseudosigma	N
Acidity	4.74 $\text{mg/L}$	3.19	22	Na	1.34 $\text{mg/L}$	0.10	55
Ca	2.53 $\text{mg/L}$	0.24	59	pH	6.92 units	0.32	62
Cl	1.20 $\text{mg/L}$	0.49	57	$\text{PO}_4 \text{ as P}$	0.002 $\text{mg/L}$	0.014	20
F	0.100 $\text{mg/L}$	0.033	31	$\text{SO}_4$	2.42 $\text{mg/L}$	0.37	50
K	0.336 $\text{mg/L}$	0.038	45	Sp Cond	26.1 $\mu\text{S/cm}$	1.1	57
Mg	0.461 $\text{mg/L}$	0.050	54				

**Hg-23 (mercury)**

Analyte	MPV	F-pseudosigma	N
Hg	0.34 $\mu\text{g/L}$	0.07	58